

Fluor Hanford
WSCF Analytical Lab
P.O. Box 1000
Richland, WA 99352
Telephone 373-7495
Telefax 372-0456

FLUOR

Memorandum

M4W41-SLF-07-655

To: H. Hampt E6-35

Date: September 18, 2007

From: S. L. Fitzgerald, Manager
WSCF Analytical Lab



cc: w/Attachments

T. F. Dale S3-30
H. K. Meznarich S3-30
P. D. Mix S3-30
J. E. Trechter S3-30

S. J. Trent E6-35
D. D. Wright S3-30
File/LB

Subject: FINAL RESULTS FOR SAMPLE DELIVERY GROUP WSCF20071413 – SAF NUMBER F07-057

Reference: (1) Groundwater Protection Program-Letter of Instruction, FH-EIS-2003-MEM-001, October 31, 2002
(2) HNF-SD-CD-QAPP-017, Rev. 8, Waste Sampling & Characterization Facility Quality Assurance Plan

This letter contains the following information for sample delivery group WSCF20071413:

- Cover Sheet (Attachment 1)
- Narrative (Attachment 2)
- Analytical Results (Attachment 3)
- Sample Receipt Information (Attachment 4)

SLF/grf

Attachments 4

M4W41-SLF-07-655

ATTACHMENT 1

COVER SHEET

Consisting of 2 pages
Including cover page

WSCF SAF NUMBER CROSS REFERENCE

Group#: WSCF20071413
Data Deliverable Date: 20-sep-2007
Data Deliverable: Cover Sheet

SAF#	Sample ID	WSCF#	Matrix
F07-057	B1NRR5	W07GR02090	WATER

M4W41-SLF-07-655

ATTACHMENT 2

NARRATIVE

Consisting of 3 pages
Including cover page

Sample Delivery Group	WSCF20071413
Sample Matrix	Water
Data Deliverable	Summary Report

Introduction

One groundwater sample was received at the WSCF Laboratory on August 9, 2007. Sample was analyzed for the analytes indicated on the attached copy of the chain of custody (COC) form in accordance with the *Groundwater Remediation Program – Letter of Instruction*, referenced in the cover letter.

The narrative (Attachment 2) will address sample characteristics, analyses requested and general information in performance of the analytical methods. A Data Summary Report (Attachment 3) includes analytical results, a comment report detailing method abnormalities, tentatively identified peaks if applicable, method references, and Laboratory QC information as applicable. Copies of the chain of custody and sample receipt documentation are included as Attachment 4.

It should be noted that the attached chain of custody was stamped “iced” and initialed by the WSCF Laboratory Sample Custodian during sample receiving, indicating the presence of ice in the sample container.

Analytical Methodology for Requested Analyses

Refer to *WSCF Method References Report*, page 10, for a complete listing of approved analytical methods.


Organic Comments


VOA – The hold time requirement for this analysis was met. A Matrix Spike, Matrix Spike Duplicate, Blank and Laboratory Control Sample, were analyzed with this delivery group per the GRP Letter of Instruction. See pages 13 through 15 for QC details. Analytical Note(s):

- Matrix Spike and Matrix Spike Duplicate were analyzed on sample# B1P0W2 (SDG# 20071455, SAF# F07-052).

All QC controls are within the established limits.

I certify that this data package is in compliance with the LOI, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the Laboratory Analytical Manager and Client Services as verified by the following signatures.


Scot L. Fitzgerald
WSCF Analytical Laboratory Manager


Pauline D. Mix
WSCF Client Services

Abbreviations

Hg – mercury

IC – ion chromatography

ICP – inductively coupled plasma

ICP/AES – ICP/atomic emission spectroscopy

ICP/MS – ICP/mass spectrometry

Total U – total uranium

AT/TB – total alpha/total beta

AEA – Alpha Energy Analysis

WTPH-G – Total Hydrocarbons-Gasoline

Am – americium

Cm – curium

Pu – plutonium

Np – neptunium

GEA – gamma energy analysis

H3 – Tritium

Sr – Strontium 89, 90

WTPH-D – Total Hydrocarbons-Diesel

TSS – Total Suspended Solids

M4W41-SLF-07-655

ATTACHMENT 3

ANALYTICAL RESULTS


Consisting of 9 pages
Including cover page

WSCF
ANALYTICAL RESULTS REPORT

for
Groundwater Remediation Program

Richland, WA 99354

Attention: Steve Trent

Analytical:  S. Fitzgerald 9/18/07
Client Services:  P. D. Mix 9/18/2007

All results are reported on an "as received" basis unless otherwise noted in the comment section.

This information is intended for the use of the addressee only. If the reader of this report is not the intended recipient or is not authorized by the recipient to receive the report, you are hereby notified that any dissemination, distribution or copying of this report is strictly prohibited. If you have received this report in error, please notify WSCF Laboratory immediately by telephone at (509) 373-7020 or (509) 531-8004. Information designation of this report is the responsibility of the customer.

Contract#: FH-EIS-2003-MEM-001
Report#: WSCF20071413
Report Date: 18-sep-2007
Report WGPP/ver. 5.2
Groundwater Remediation Program

Department: Organic

W13q Worklist/Batch/QC Report for Group# WSCF20071413

WL#	S#	Batch	QC#	Tray Type	Sample#	Test
			37821	BLANK		VOA Ground Water Protection
			37821	LCS		VOA Ground Water Protection
			37821	SAMPLE	W07GR02090	VOA Ground Water Protection
			37821	SURR	W07GR02090	VOA Ground Water Protection
			37821	MS	W07GR02158	VOA Ground Water Protection
			37821	MSD	W07GR02158	VOA Ground Water Protection
			37821	SPK-RPD	W07GR02158	VOA Ground Water Protection

WSCF METHOD REFERENCES REPORT

Department: Organic

The results provided in this report were generated using the following WSCF Laboratory procedures. For your convenience, this table provides a listing of the regulatory or industry methods that are referenced by each of these WSCF procedures. Please note that the most recent version of the regulatory or industry method is listed here even though the WSCF procedure may reference an older version of the method. Also, a reference to a regulatory or industry method here does not necessarily indicate a verbatim implementation of that method.

LA-523-455	LA-523-455: VOLATILE SAMPLE ANALYSIS BY SW-846
EPA SW-846 8000B	DETERMINATIVE CHROMATOGRAPHIC SEPARATIONS
EPA SW-846 8260B	VOLATILE ORGANIC COMPOUNDS BY GAS CHROMATOGRAPHY/MASS SPECTROMETRY (GC/MS)
HEIS 8260_VOA_GCMS	Volatile Organic Compounds by Gas Chromatography/Mass Spectrometry (GC/MS)

Note: A complete list of WSCF analytical procedures and referenced regulatory or industry methods is available online at <http://www2.ri.gov/phmc/as-dol>.

Report Date: 18-sep-2007
Report #: WSCF20071413
Report WGPPM/5.2

WSCF ANALYTICAL RESULTS REPORT

Attention: Steve Trent
SAF Number: F07-057
Sample #: W07GR02090
Client ID: BINRR5

Matrix: WATER
Group #: WSCF20071413
Department: Organic
Sampled: 08/09/07
Received: 08/09/07

Test Performed	CAS #	Method	RQ	Result	Unit	TP Err	Unit	DF	MDL	PQL	Analysis Date
VOA Ground Water Protection											
1,1-Dichloroethene	75-35-4	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
Trichloroethene	79-01-6	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
Benzene	71-43-2	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
Toluene	108-88-3	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
Chlorobenzene	108-90-7	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
1,1-Dichloroethane	75-34-3	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
Ethylbenzene	100-41-4	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
Styrene	100-42-5	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
cis-1,3-Dichloropropene	10061-01-5	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
trans-1,3-Dichloropropene	10061-02-6	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
1,2-Dichloroethane	107-06-2	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
4-Methyl-2-Pentanone	108-10-1	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
Dibromochloromethane	124-48-1	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
Tetrachloroethene	127-18-4	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
Xylenes (total)	1330-20-7	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
1,2-Dichloroethene(Total)	540-59-0	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
Carbon tetrachloride	56-23-5	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
2-Hexanone	591-78-6	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
Acetone	67-64-1	LA-523-455	U	7.50	ug/L			1.00	1.0		08/16/07
Chloroform	67-66-3	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
1,1,1-Trichloroethane	71-55-6	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
Bromomethane	74-83-9	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
1,1,2,2-Tetrachloroethane	79-34-5	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
Chloromethane	74-87-3	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07

MDL = Minimum Detection Limit U - Analyzed for but not detected above limiting criteria.(org)

RQ = Result Qualifier

TP Err = Total Propagated Error

DF = Dilution Factor

* - Indicates results that have NOT been validated; + - Indicates more than six qualifier symbols

Report *WGPP/ver. 5.2*

Groundwater Remediation Program

WSCF ANALYTICAL RESULTS REPORT

Attention: Steve Trent
SAF Number: F07-057
Sample # W07GR02090
Client ID: B1NRR5
TRENT WSCF
Matrix: WATER
Group #: WSCF20071413
Department: Organic
Sampled: 08/09/07
Received: 08/09/07

Test Performed	CAS #	Method	RQ	Result	Unit	TP Err	Unit	DF	MDL	PQL	Analysis Date
Chloroethane	75-00-3	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
Vinyl chloride	75-01-4	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
Methylenechloride	75-09-2	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
Carbon disulfide	75-15-0	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
Bromoform	75-25-2	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
Bromodichloromethane	75-27-4	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
1,2-Dichloropropane	78-87-5	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
2-Butanone	78-93-3	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
1,1,2-Trichloroethane	79-00-5	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
1-Butanol	71-36-3	LA-523-455	U	< 100	ug/L			1.00	1.0e+02		08/16/07
n-Butylbenzene	104-51-8	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
trans-1,2-Dichloroethylene	156-60-5	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07
cis-1,2-Dichloroethylene	156-59-2	LA-523-455	U	< 1.00	ug/L			1.00	1.0		08/16/07

MDL = Minimum Detection Limit U - Analyzed for but not detected above limiting criteria. (org)

RQ = Result Qualifier

TP Err = Total Propagated Error

DF = Dilution Factor

* - Indicates results that have NOT been validated; + - Indicates more than six qualifier symbols
 Report WGPP/ver. 5.2

Groundwater Remediation Program

WSCF ANALYTICAL LABORATORY QC REPORT

Department: **Organic**

SDG Number: WSCF20071413
 Matrix: WATER
 Test: VOA Ground Water Protection

Sample Date: 08/09/07
 Receive Date: 08/09/07

QC Type	Analyte	CAS #	QC Found	QC Yield	Units	Lower Limit	Upper Limit	RPD(%)	RPD Limit	RQ	Analysis Date
Lab ID: W07GR02090											
BATCH QC ASSOCIATED WITH SAMPLE											
SURR	4-Bromofluorobenzene(Surr)	460-00-4	47.650	95.300	% Recov	84.000	116.000				08/16/07
SURR	1,2-Dichloroethane-d4(Surr)	17060-07-0	42.640	85.300	% Recov	82.000	136.000				08/16/07
SURR	Toluene-d8(Surr)	2037-26-5	52.390	105.000	% Recov	89.000	119.000				08/16/07
Lab ID: W07GR02158											
BATCH QC ASSOCIATED WITH SAMPLE											
MS	1,1-Dichloroethene	75-35-4	26.930	108.000	% Recov	63.000	117.000				08/16/07
MS	Benzene	71-43-2	24.870	99.500	% Recov	75.000	129.000				08/16/07
MS	4-Bromofluorobenzene(Surr)	460-00-4	47.330	94.700	% Recov	84.000	116.000				08/16/07
MS	Chlorobenzene	108-90-7	21.460	85.800	% Recov	79.000	119.000				08/16/07
MS	1,2-Dichloroethane-d4(Surr)	17060-07-0	42.080	84.200	% Recov	82.000	136.000				08/16/07
MS	Toluene-d8(Surr)	2037-26-5	49.370	98.700	% Recov	89.000	119.000				08/16/07
MS	Toluene	108-88-3	25.250	101.000	% Recov	76.000	120.000				08/16/07
MS	Trichloroethene	79-01-6	21.570	86.300	% Recov	73.000	123.000				08/16/07
MSD	1,1-Dichloroethene	75-35-4	26.070	104.000	% Recov	63.000	117.000				08/16/07
MSD	Benzene	71-43-2	25.000	100.000	% Recov	75.000	129.000				08/16/07
MSD	4-Bromofluorobenzene(Surr)	460-00-4	47.810	95.600	% Recov	84.000	116.000				08/16/07
MSD	Chlorobenzene	108-90-7	21.170	84.700	% Recov	79.000	119.000				08/16/07
MSD	1,2-Dichloroethane-d4(Surr)	17060-07-0	43.020	86.000	% Recov	82.000	136.000				08/16/07
MSD	Toluene-d8(Surr)	2037-26-5	51.890	104.000	% Recov	89.000	119.000				08/16/07
MSD	Toluene	108-88-3	25.280	101.000	% Recov	76.000	120.000				08/16/07
MSD	Trichloroethene	79-01-6	20.540	82.200	% Recov	73.000	123.000				08/16/07
SPK-RPD	1,1-Dichloroethene	75-35-4	104.000		RPD			3.774	20.000		08/16/07
SPK-RPD	Benzene	71-43-2	100.000		RPD			0.501	20.000		08/16/07
SPK-RPD	Chlorobenzene	108-90-7	84.700		RPD			1.290	20.000		08/16/07
SPK-RPD	1,2-Dichloroethane-d4(Surr)	17060-07-0	86.000		RPD			2.115	20.000		08/16/07

WSCF ANALYTICAL LABORATORY QC REPORT

Department: **Organic**

SDG Number: WSCF20071413

Matrix: WATER

Test: VOA Ground Water Protection

Sample Date: 08/15/07
Receive Date: 08/15/07

QC Type	Analyte	CAS #	QC Found	QC Yield	Units	Lower Limit	Upper Limit	RPD(%)	RPD Limit	RQ	Analysis Date
SPK-RPD	Toluene-d8(Surr)	2037-26-5	104.000		RPD			5.229	20.000		08/16/07
SPK-RPD	Toluene	108-88-3	101.000		RPD			0.000	20.000		08/16/07
SPK-RPD	Trichloroethane	79-01-6	82.200		RPD			4.866	20.000		08/16/07
BATCH QC											
BLANK	1,1-Dichloroethane	75-34-3	< 1.0	n/a	ug/L					U	08/16/07
BLANK	1,1,1-Trichloroethane	71-55-6	< 1.0	n/a	ug/L					U	08/16/07
BLANK	1,1,2-Trichloroethane	79-00-5	< 1.0	n/a	ug/L					U	08/16/07
BLANK	1,1,2,2-Tetrachloroethane	79-34-5	< 1.0	n/a	ug/L	0.000	5.000			U	08/16/07
BLANK	1,1-Dichloroethene	75-35-4	< 1.0	n/a	ug/L					U	08/16/07
BLANK	1,2-Dichloroethane	107-06-2	< 1.0	n/a	ug/L					U	08/16/07
BLANK	1,2-Dichloroethene(Total)	540-59-0	< 1.0	n/a	ug/L					U	08/16/07
BLANK	1-Butanol	71-36-3	< 100	n/a	ug/L					U	08/16/07
BLANK	2-Hexanone	591-78-6	< 1.0	n/a	ug/L					U	08/16/07
BLANK	4-Methyl-2-Pentanone	108-10-1	< 1.0	n/a	ug/L					U	08/16/07
BLANK	Acetone	67-64-1	< 1.0	n/a	ug/L					U	08/16/07
BLANK	Bromodichloromethane	75-27-4	< 1.0	n/a	ug/L					U	08/16/07
BLANK	Benzene	71-43-2	< 1.0	n/a	ug/L					U	08/16/07
BLANK	4-Bromofluorobenzene(Surr)	460-00-4	47.720	95.400	% Recov	84.000	116.000			U	08/16/07
BLANK	Bromoform	75-25-2	< 1.0	n/a	ug/L					U	08/16/07
BLANK	n-Butylbenzene	104-51-8	< 1.0	n/a	ug/L					U	08/16/07
BLANK	Carbon disulfide	75-15-0	< 1.0	n/a	ug/L					U	08/16/07
BLANK	Carbon tetrachloride	56-23-5	< 1.0	n/a	ug/L					U	08/16/07
BLANK	Dibromochloromethane	124-48-1	< 1.0	n/a	ug/L					U	08/16/07
BLANK	Chloroform	67-66-3	< 1.0	n/a	ug/L					U	08/16/07
BLANK	Chlorobenzene	108-90-7	< 1.0	n/a	ug/L					U	08/16/07
BLANK	cis-1,2-Dichloroethylene	156-59-2	< 1.0	n/a	ug/L					U	08/16/07
BLANK	cis-1,3-Dichloropropene	10061-01-5	< 1.0	n/a	ug/L					U	08/16/07
BLANK	Chloroethane	75-00-3	< 1.0	n/a	ug/L					U	08/16/07
BLANK	1,2-Dichloroethane-d4(Surr)	17060-07-0	42.650	85.300	% Recov	82.000	136.000			U	08/16/07

WSCF ANALYTICAL LABORATORY QC REPORT

Department: **Organic**

SDG Number: WSCF20071413

Matrix: WATER

Test: VOA Ground Water Protection

Sample Date:
Receive Date:

QC Type	Analyte	CAS #	QC Found	QC Yield	Units	Lower Limit	Upper Limit	RPD(%)	RPD Limit	RQ	Analysis Date
BLANK	trans-1,2-Dichloroethylene	156-60-5	< 1.0	n/a	ug/L					U	08/16/07
BLANK	1,2-Dichloropropane	78-87-5	< 1.0	n/a	ug/L					U	08/16/07
BLANK	Ethylbenzene	100-41-4	< 1.0	n/a	ug/L					U	08/16/07
BLANK	Bromomethane	74-83-9	< 1.0	n/a	ug/L					U	08/16/07
BLANK	Chloromethane	74-87-3	< 1.0	n/a	ug/L					U	08/16/07
BLANK	2-Butanone	78-93-3	< 1.0	n/a	ug/L					U	08/16/07
BLANK	Methylenechloride	75-09-2	< 1.0	n/a	ug/L					U	08/16/07
BLANK	Tetrachloroethene	127-18-4	< 1.0	n/a	ug/L					U	08/16/07
BLANK	Styrene	100-42-5	< 1.0	n/a	ug/L					U	08/16/07
BLANK	Xylenes (total)	1330-20-7	< 1.0	n/a	ug/L	0.000	5.000			U	08/16/07
BLANK	Toluene-d8(Surr)	2037-26-5	48.620	97.200	% Recov	89.000	119.000			U	08/16/07
BLANK	Toluene	108-88-3	< 1.0	n/a	ug/L					U	08/16/07
BLANK	trans-1,3-Dichloropropene	10061-02-6	< 1.0	n/a	ug/L					U	08/16/07
BLANK	Trichloroethene	79-01-6	< 1.0	n/a	ug/L					U	08/16/07
BLANK	Vinyl chloride	75-01-4	< 1.0	n/a	ug/L					U	08/16/07
LCS	1,1-Dichloroethene	75-35-4	26.850	107.000	% Recov	70.000	130.000			U	08/16/07
LCS	Benzene	71-43-2	25.720	103.000	% Recov	70.000	130.000			U	08/16/07
LCS	4-Bromofluorobenzene(Surr)	460-00-4	47.370	94.700	% Recov	84.000	116.000			U	08/16/07
LCS	Chlorobenzene	108-90-7	22.120	88.500	% Recov	70.000	130.000			U	08/16/07
LCS	1,2-Dichloroethane-d4(Surr)	17060-07-0	42.220	84.400	% Recov	82.000	136.000			U	08/16/07
LCS	Toluene-d8(Surr)	2037-26-5	49.570	99.100	% Recov	89.000	119.000			U	08/16/07
LCS	Toluene	108-88-3	25.600	102.000	% Recov	70.000	130.000			U	08/16/07
LCS	Trichloroethene	79-01-6	21.350	85.400	% Recov	70.000	130.000			U	08/16/07

M4W41-SLF-07-655

ATTACHMENT 4

SAMPLE RECEIPT INFORMATION

Consisting of 3 pages
Including cover page

Waste Sampling and Characterization Facility
P.O. BOX 1970 S3-30, Richland, WA 99352
PHONE: (509) 373-7004/FAX: (509) 373-7134

9/20/07

ACKNOWLEDGMENT OF SAMPLES RECEIVED

File KB

Groundwater Remediation Program

Richland, WA 99354
Attn: Steve Trent

Customer Code: GPP
PO#: 122868/ES3
Group#: 20071413
Project#: F07-057
Proj Mgr: Steve Trent E6-35
Phone: 373-5869

The following samples were received from you on 08/09/07. They have been scheduled for the tests listed beside each sample. If this information is incorrect, please contact your service representative. Thank you for using Waste Sampling and Characterization Facility.

Sample#	Sample Id	Tests Scheduled	Matrix	Sample Date
W07GR02090	B1NRR5	TRENT @VOA-GPP	Water	08/09/07

Test Acronym Description

Test Acronym	Description
@VOA-GPP	VOA Ground Water Protection

FLUOR HANFORD INC.
COLLECTOR: Popel/Mokler *CONNELLY*
COMPANY CONTACT: Trent, SJ
TELEPHONE NO.: 373-5869
PROJECT COORDINATOR: TRENT, SJ
PRICE CODE: 8N
DATA TURNAROUND: 45 Days / 45 Days

CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST
PROJECT DESIGNATION: 216-A-2 and 216-AE-21 Characterization Sampling and Analysis - QC
SAF NO.: F07-057
METHOD OF SHIPMENT: GOVERNMENT VEHICLE
FIELD LOGBOOK NO.: COA 122868 ES3
BILL OF LADING/AIR BILL NO.: N/A

SHIPPED TO: Waste Sampling & Characterization
OFFSITE PROPERTY NO.: N/A
PRESERVATION: HCl or H2SO4 to pH <2/Cool AC
TYPE OF CONTAINER: ags*
NO. OF CONTAINER(S): 4
VOLUME: 40ml
SAMPLE ANALYSIS: SEE ITEM (1) IN SPECIAL INSTRUCTIONS

SAMPLE NO.	MATRIX*	SAMPLE DATE	SAMPLE TIME
B1NRR5	WATER	8-7-7	0830
W076202090			

CHAIN OF POSSESSION	SIGN/PRINT NAMES	DATE/TIME
RELINQUISHED BY/REMOVED FROM	RECEIVED BY/STORED IN	DATE/TIME
<i>R. P. Foster</i>	AZ SITZ FRIG	8/7/07 1020
RELINQUISHED BY/REMOVED FROM	RECEIVED BY/STORED IN	DATE/TIME
AZ SITZ FRIG	<i>R. P. Foster</i>	8/9/07 0950
RELINQUISHED BY/REMOVED FROM	RECEIVED BY/STORED IN	DATE/TIME
<i>R. P. Foster</i>	<i>R. P. Foster</i>	8/9/07 1015
RELINQUISHED BY/REMOVED FROM	RECEIVED BY/STORED IN	DATE/TIME
RELINQUISHED BY/REMOVED FROM	RECEIVED BY/STORED IN	DATE/TIME
RELINQUISHED BY/REMOVED FROM	RECEIVED BY/STORED IN	DATE/TIME

SPECIAL INSTRUCTIONS

(1)VOA - 8260B (TCL); VOA - 8260B (Add-On) {1-Butanol, cis-1,2-Dichloroethylene, n-Butylbenzene, trans-1,2-Dichloroethylene}

ICED Initial Date
[Signature]

LABORATORY SECTION	RECEIVED BY	DATE/TIME
FINAL SAMPLE DISPOSITION	DISPOSAL METHOD	DATE/TIME