

8/3/2016



August 02, 2016

Mr. Scot Fitzgerald
CH2MHill Plateau Remediation Company
MSIN R3-50 CHPRC
PO Box 1600
Richland, Washington 99352

Re: CHPRC SAF W16-007
Work Order: 401205
SDG: GEL401205

Dear Mr. Fitzgerald:

GEL Laboratories, LLC (GEL) appreciates the opportunity to provide the enclosed analytical results for the sample(s) we received on July 09, 2016. This original data report has been prepared and reviewed in accordance with GEL's standard operating procedures.

Our policy is to provide high quality, personalized analytical services to enable you to meet your analytical needs on time every time. We trust that you will find everything in order and to your satisfaction. If you have any questions, please do not hesitate to call me at (843) 556-8171, ext. 4505.

Sincerely,

B Luthman
Brielle Luthman for
Heather Shaffer
Project Manager

Purchase Order: 300072 - 7H
Chain of Custody: W16-007-082
Enclosures




Table of Contents

Case Narrative.....	1
Chain of Custody and Supporting Documentation.....	5
Data Review Qualifier Definitions.....	8
Laboratory Certifications.....	10
General Chem Analysis.....	12
Case Narrative.....	13
Sample Data Summary.....	17
Quality Control Summary.....	22

Case Narrative

**General Narrative
for
CH2MHill Plateau Remediation Company
CHPRC SAF W16-007
SDG: GEL401205**

August 02, 2016

Laboratory Identification:

GEL Laboratories LLC
2040 Savage Road
Charleston, South Carolina 29407
(843) 556-8171

Summary

Sample receipt

The sample(s) arrived at GEL Laboratories, LLC, Charleston, South Carolina on July 09, 2016, for analysis. The samples were delivered with proper chain of custody documentation and signatures. All sample containers arrived without any visible signs of tampering or breakage. There are no additional comments concerning sample receipt.

Items of Note All efforts were made by the lab to meet any short hold times. Samples that were analyzed outside of the initial hold time but still within 2X hold time will be noted in the lab case narrative and DER.

Sample Identification

The laboratory received the following samples:

<u>Laboratory Identification</u>	<u>Sample Description</u>
401205001	B35P81
401205002	B35P85
401205003	B35NC2
401205004	B35P83


Case Narrative

Sample analyses were conducted using methodology as outlined in GEL Laboratories, LLC (GEL) Standard Operating Procedures. Any technical or administrative problems during analysis, data review, and reduction are contained in the analytical case narratives in the enclosed data package.

Data Package

The enclosed data package contains the following sections: General Narrative, Chain of Custody and Supporting Documentation, and data from the following fractions: General Chemistry.

We certify that this package is in compliance with the SOW, both technically and for completeness, including a full description of, explanation of, and corrective actions for, any and all deviations, from either the analyses requested or the case narrative requested. Release of the data contained in this hard copy data package has been authorized by the Laboratory Analytical Manager (or designee) and the laboratory's client services representative as verified by their signatures on this report.


Brielle Luthman for
Heather Shaffer
Project Manager

**General Chemistry
Technical Case Narrative
CH2MHill Plateau Remediation Company (CPRC)
SDG #: GEL401205
Work Order #: 401205**

Carbon, Total Organic

All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable, with the following exceptions.

Technical Information

Sample Re-analysis

Samples 1203581758 (MB), 1203581759 (LCS), 1203582071 (B35P83DUP), 1203582072 (B35P83PS), 401205001 (B35P81), 401205002 (B35P85), 401205003 (B35NC2) and 401205004 (B35P83) were re-analyzed due to CCV failure. The reanalysis data with passing instrument QC was reported.

Total Organic Halogens (TOX)

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

Certification Statement

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless otherwise noted in the analytical case narrative.

Chain of Custody and Supporting Documentation

CH2MHill Plateau Remediation Company
 401205
CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST
 C.O.C. # **W16-007-082**
 Page 1 of 1

Collector: Dan Woehle CHPRC
SAF No.: W16-007
Contact/Requester: Karen Waters-Husted
Telephone No.: 509-376-4650
Project Title: RCRA, JULY 2016
Sampling Origin: Hanford Site
Purchase Order/Charge Code: 300071
Shipped To (Lab): GEL Laboratories, LLC
Logbook No.: HNF-N-506 80144
Ice Chest No.: 6005-458
Method of Shipment: Commercial Carrier
Bill of Lading/Air Bill No.: 776700158709
Priority: 30 Days
Offsite Property No.: 6810

Protocol: RCRA
SPECIAL INSTRUCTIONS: Hold Time
Total Activity Exemption: Yes No
Special Handling: N/A

POSSIBLE SAMPLE HAZARDS/REMARKS
 *** Contains Radioactive Material at concentrations that are not regulated for transportation per 49 CFR / IATA Dangerous Goods Regulations but are not releasable per DOE Order 458.1

Sample No.	Filter	* Date	Time	No./Type Container	9020_TOX	Sample Analysis	Holding Time	Preservative
B35P81	N	JUL 08 2016	1105	1x1-L aGs*	9020_TOX: COMMON		28 Days	H2SO4 to pH <2/Cool <=6C
B35P81	N			1x250-mL aG	9060_TOX: COMMON		28 Days	HCl or H2SO4 to pH <2/Cool <=6C
B35P85	N			1x1-L aGs*	9020_TOX: COMMON		28 Days	H2SO4 to pH <2/Cool <=6C
B35P85	N			1x250-mL aG	9060_TOX: COMMON		28 Days	HCl or H2SO4 to pH <2/Cool <=6C
B35NC2	N			1x1-L aGs*	9020_TOX: COMMON		28 Days	H2SO4 to pH <2/Cool <=6C
B35NC2	N			1x250-mL aG	9060_TOX: COMMON		28 Days	HCl or H2SO4 to pH <2/Cool <=6C
B35P83	N			1x1-L aGs*	9020_TOX: COMMON		28 Days	H2SO4 to pH <2/Cool <=6C
B35P83	N	JUL 08 2016	1105	1x250-mL aG	9060_TOX: COMMON		28 Days	HCl or H2SO4 to pH <2/Cool <=6C

Relinquished By	Print	Sign	Date/Time	Received By	Print	Sign	Date/Time	Matrix *
Dan Woehle CHPRC			JUL 08 2016 1210	Lesly Wall CHPRC			JUL 08 2016 1210	S = Soil, SE = Sediment, SO = Solid, SL = Sludge, W = Water, O = Oil, A = Air, DS = Drum Solids, DL = Drum Liquids, T = Tissue, WI = Wipe, L = Liquid, V = Vegetation, X = Other
Relinquished By			JUL 08 2016 1400	Received By	FEDEX			
Relinquished By				Received By			7-9-16 0955	
Relinquished By				Received By				

FINAL SAMPLE DISPOSITION
 Disposal Method (e.g., Return to customer, per lab procedure, used in process)
 Disposed By
 Date/Time
 Disposal Method (e.g., Return to customer, per lab procedure, used in process)
 Disposed By
 Date/Time
 Disposal Method (e.g., Return to customer, per lab procedure, used in process)
 Disposed By
 Date/Time

SAMPLE RECEIPT & REVIEW FORM

Client: <u>C/PC</u>		SDG/AR/COC/Work Order:	
Received By: <u>MK</u>		Date Received: <u>7-9-16</u>	
Suspected Hazard Information		Yes	No
COC/Samples marked as radioactive?		<input checked="" type="checkbox"/>	<input type="checkbox"/>
Classified Radioactive II or III by RSO?		<input checked="" type="checkbox"/>	<input type="checkbox"/>
COC/Samples marked containing PCBs?		<input checked="" type="checkbox"/>	<input type="checkbox"/>
Package, COC, and/or Samples marked as beryllium or asbestos containing?		<input checked="" type="checkbox"/>	<input type="checkbox"/>
Shipped as a DOT Hazardous?		<input checked="" type="checkbox"/>	<input type="checkbox"/>
Samples identified as Foreign Soil?		<input checked="" type="checkbox"/>	<input type="checkbox"/>

*If Net Counts > 100cpm on samples not marked "radioactive", contact the Radiation Safety Group for further investigation.

Maximum Net Counts Observed* (Observed Counts - Area Background Counts): 0

If yes, Were swipes taken of sample containers < action levels?

If yes, samples are to be segregated as Safety Controlled Samples, and opened by the GEL Safety Group.

Hazard Class Shipped: UN#:

Sample Receipt Criteria		Yes	NA	No	Comments/Qualifiers (Required for Non-Conforming Items)
1	Shipping containers received intact and sealed?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Circle Applicable: Seals broken Damaged container Leaking container Other (describe)
2	Samples requiring cold preservation within (0 ≤ 6 deg. C)?*	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Preservation Method: <u>Ice bags</u> Blue ice Dry ice None Other (describe) *all temperatures are recorded in Celsius
2a	Daily check performed and passed on IR temperature gun?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Temperature Device Serial #: <u>130462961</u> Secondary Temperature Device Serial # (If Applicable):
3	Chain of custody documents included with shipment?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
4	Sample containers intact and sealed?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Circle Applicable: Seals broken Damaged container Leaking container Other (describe)
5	Samples requiring chemical preservation at proper pH?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Sample ID's, containers affected and observed pft: If Preservation added, Lot#:
6	Do Low Level Perchlorate samples have headspace as required?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Sample ID's and containers affected:
7	VOA vials contain acid preservation?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	(If unknown, select No)
8	VOA vials free of headspace (defined as < 6mm bubble)?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Sample ID's and containers affected:
9	Are Encore containers present?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	(If yes, immediately deliver to Volatiles laboratory)
10	Samples received within holding time?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	ID's and tests affected:
11	Sample ID's on COC match ID's on bottles?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Sample ID's and containers affected:
12	Date & time on COC match date & time on bottles?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Sample ID's affected:
13	Number of containers received match number indicated on COC?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Sample ID's affected:
14	Are sample containers identifiable as GEL provided?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
15	COC form is properly signed in relinquished/received sections?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
16	Carrier and tracking number.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Circle Applicable: FedEx Air <u>FedEx Ground</u> UPS Field Services Courier Other <u>7767 0615 8709</u>

Comments (Use Continuation Form if needed):

PM (or PMA) review: Initials EM Date 7/11/16 Page 1 of 1

Data Review Qualifier Definitions

Project Specific Qualifier Definitions for GEL Client Code: CPRC

Qualifier	Qualifier Definition	Department	Fraction
U	Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.		
J	The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate). Value is estimated	Organics	
P	Aroclor target analyte with greater than 25% difference between column analyses.	Organics	
C	Analyte has been confirmed by GC/MS analysis	Organics	Pesticide
B	The analyte was detected in both the associated QC blank and in the sample.	Organics	
E	Concentration exceeds the calibration range of the instrument	Organics	
A	The TIC is a suspected aldol-condensation product	Organics	Semi-Volatile
X	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier		
N	Spike Sample recovery is outside control limits.		
*	Duplicate analysis not within control limits	Inorganics	
>	Result greater than quantifiable range or greater than upper limit of the analysis range	General Chemistry	
Z	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier		
B	The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate).	Inorganics	Metals
D	Results are reported from a diluted aliquot of sample.		
E	Reported value is estimated due to interferences. See comment in narrative.	Inorganics	Metals
M	Duplicate precision not met.	Inorganics	Metals
o	Analyte failed to recover within LCS limits (Organics only)	Organics	
S	Reported value determined by the Method of Standard Additions (MSA)	Inorganics	
T	Spike and/or spike duplicate sample recovery is outside control limits.	Organics	
W	Post-digestion spike recovery for GFAA out of control limit. Sample absorbency < 50% of spike absorbency.	Inorganics	
B	The associated QC sample blank has a result $\geq 2X$ the MDA and, after corrections, result is \geq MDA for this sample	Radiological	
Y	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier		
+	Correlation coefficient for Method of Standard Additions (MSA) is < 0.995	Inorganics	
B	The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate).	General Chemistry	
C	Target analyte was detected in the sample and the associated blank. The associated blank concentration is \geq EQL or is > 5% of the measured concentration and/or decision level for associated samples.	Inorganics	Metals
C	Target analyte was detected in the sample and the associated blank. The associated blank concentration is \geq EQL or is > 5% of the measured concentration and/or decision level for associated samples.	General Chemistry	
<	Sample is below the EPA guidance level for Reactive Releasable Cyanide and/or Reactive Releasable Sulfide	General Chemistry	
UX	Gamma Spectroscopy--Uncertain identification	Radiological	

Laboratory Certifications

List of current GEL Certifications as of 02 August 2016

State	Certification
Alaska	UST-0110
Arkansas	88-0651
CLIA	42D0904046
California	2940
Colorado	SC00012
Connecticut	PH-0169
Delaware	SC00012
DoD ELAP/ ISO17025 A2LA	2567.01
Florida NELAP	E87156
Foreign Soils Permit	P330-15-00283, P330-15-00253
Georgia	SC00012
Georgia SDWA	967
Hawaii	SC00012
Idaho Chemistry	SC00012
Idaho Radiochemistry	SC00012
Illinois NELAP	200029
Indiana	C-SC-01
Kansas NELAP	E-10332
Kentucky SDWA	90129
Kentucky Wastewater	90129
Louisiana NELAP	03046 (AI33904)
Louisiana SDWA	LA160006
Maryland	270
Massachusetts	M-SC012
Michigan	9976
Mississippi	SC00012
Nebraska	NE-OS-26-13
Nevada	SC000122016-1
New Hampshire NELAP	205415
New Jersey NELAP	SC002
New Mexico	SC00012
New York NELAP	11501
North Carolina	233
North Carolina SDWA	45709
North Dakota	R-158
Oklahoma	9904
Pennsylvania NELAP	68-00485
S.Carolina Radchem	10120002
South Carolina Chemistry	10120001
Tennessee	TN 02934
Texas NELAP	T104704235-16-11
Utah NELAP	SC000122016-20
Vermont	VT87156
Virginia NELAP	460202
Washington	C780
West Virginia	997404

General Chem Analysis

Case Narrative

**General Chemistry
 Technical Case Narrative
 CH2MHill Plateau Remediation Company (CPRC)
 SDG #: GEL401205
 Work Order #: 401205**

Product: Carbon, Total Organic

Analytical Method: SW846 9060A

Analytical Procedure: GL-GC-E-093 REV# 14

Analytical Batch: 1580470

The following samples were analyzed using the above methods and analytical procedure(s).

<u>GEL Sample ID#</u>	<u>Client Sample Identification</u>
401205001	B35P81
401205002	B35P85
401205003	B35NC2
401205004	B35P83
1203581758	Method Blank (MB)
1203581759	Laboratory Control Sample (LCS)
1203582071	401205004(B35P83) Sample Duplicate (DUP)
1203582072	401205004(B35P83) Post Spike (PS)

The samples in this SDG were analyzed on an "as received" basis.

Data Summary:

All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable, with the following exceptions.

Technical Information

Sample Re-analysis

Samples 1203581758 (MB), 1203581759 (LCS), 1203582071 (B35P83DUP), 1203582072 (B35P83PS), 401205001 (B35P81), 401205002 (B35P85), 401205003 (B35NC2) and 401205004 (B35P83) were re-analyzed due to CCV failure. The reanalysis data with passing instrument QC was reported.

Product: Total Organic Halogens (TOX)

Analytical Method: 9020_TOX

Analytical Procedure: GL-GC-E-007 REV# 14

Analytical Batch: 1580815

The following samples were analyzed using the above methods and analytical procedure(s).

<u>GEL Sample ID#</u>	<u>Client Sample Identification</u>
401205001	B35P81
401205002	B35P85
401205003	B35NC2
401205004	B35P83
1203582531	Method Blank (MB)
1203582532	Laboratory Control Sample (LCS)
1203582533	401205001(B35P81) Sample Duplicate (DUP)
1203582534	401205001(B35P81) Post Spike (PS)

The samples in this SDG were analyzed on an "as received" basis.

Data Summary:

All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable, with the following exceptions.

Miscellaneous Information

Additional Comments

A pair of nitrate wash blanks is analyzed at the start of the batch. Although they are designated as ICB, they are performed for calculating purposes only. The value of the nitrate wash blanks are averaged and subtracted from all samples. Neither of these values should exceed 0.6 ug Cl. The PQL limit typically applied to ICB results does not apply in this application, since the results are used only to determine background concentrations and are subtracted from all calculated results.

Breakthrough effect

Breakthrough effect: If the value for a sample is greater than the reporting limit (10 ug/L), the result for the second slug should not be greater than 25% of the combined value of the first and second slug. Results which do not meet these criteria are designated with a "Fail" comment in the Breakthrough effect column on the Logbook page; however, the "fail" designation is not applicable for samples with a result of less than 10 ug/L.

Certification Statement

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless otherwise noted in the analytical case narrative.

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

**Qualifier Definition Report
for**

CPRC001 CH2MHill Plateau Remediation Company

Client SDG: GEL401205 GEL Work Order: 401205

The Qualifiers in this report are defined as follows:


B The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate).

U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.

Review/Validation

GEL requires all analytical data to be verified by a qualified data reviewer. In addition, all CLP-like deliverables receive a third level review of the fractional data package.

The following data validator verified the information presented in this data report:

Signature: 

Name: Aubrey Kingsbury

Date: 02 AUG 2016

Title: Analyst I

Sample Data Summary

8/3/2016

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis

Report Date: August 2, 2016

Company : CH2MHill Plateau Remediation Company
 Address : MSIN R3-50 CHPRC
 PO Box 1600
 Richland, Washington 99352
 Contact: Mr. Scot Fitzgerald
 Project: CHPRC SAF W16-007

Client Sample ID:	B35NC2	Project:	CPRCOW16007
Sample ID:	401205003	Client ID:	CPRC001
Matrix:	WATER		
Collect Date:	08-JUL-16 11:05		
Receive Date:	09-JUL-16		
Collector:	Client		

Parameter	Qualifier	Result	DL	RL	Units	PF	DF	Analyst	Date	Time	Batch	Method
Carbon Analysis												
9060_TOX: COMMON "As Received"												
Total Organic Carbon #1	U	330	330	1000	ug/L		1	TSM	07/13/16	0353	1580470	1
Total Organic Carbon #2	U	330	330	1000	ug/L		1					
Total Organic Carbon #3	U	330	330	1000	ug/L		1					
Total Organic Carbon #4	B	331	330	1000	ug/L		1					
Total Organic Carbon Average	U	330	330	1000	ug/L		1					

Halogen Analysis												
9020_TOX: COMMON "As Received"												
Total Organic Halogens	U	3.33	3.33	10.0	ug/L		1	RMJ	07/19/16	2151	1580815	2

The following Analytical Methods were performed:

Method	Description	Analyst Comments
1	SW846 9060A	
2	9020_TOX	

Notes:

Column headers are defined as follows:

DF: Dilution Factor	Lc/LC: Critical Level
DL: Detection Limit	PF: Prep Factor
MDA: Minimum Detectable Activity	RL: Reporting Limit
MDC: Minimum Detectable Concentration	SQL: Sample Quantitation Limit

Quality Control Summary

GEL LABORATORIES LLC

2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Report Date: August 2, 2016

Page 1 of 2

CH2M Hill Plateau Remediation Company

MSIN R3-50 CHPRC

PO Box 1600

Richland, Washington

Contact: Mr. Scot Fitzgerald

Workorder: 401205

Parmname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date	Time
Carbon Analysis											
Batch	1580470										
QC1203582071	401205004	DUP									
Total Organic Carbon Average		U	330	U	330	ug/L	N/A		TSM	07/13/16	05:13
QC1203581759	LCS										
Total Organic Carbon Average	10000				9930	ug/L	99.3	(80%-120%)		07/12/16	15:57
QC1203581758	MB										
Total Organic Carbon Average			U		330	ug/L				07/12/16	15:47
QC1203582072	401205004	PS									
Total Organic Carbon Average	10.0	U	0.246		9.78	mg/L	95.3	(75%-125%)		07/13/16	05:54
Halogen Analysis											
Batch	1580815										
QC1203582533	401205001	DUP									
Total Organic Halogens		U	3.33	U	3.33	ug/L	N/A		RMJ	07/19/16	19:07
QC1203582532	LCS										
Total Organic Halogens	100				112	ug/L	112	(80%-120%)		07/19/16	18:36
QC1203582531	MB										
Total Organic Halogens			U		3.33	ug/L				07/19/16	18:12
QC1203582534	401205001	PS									
Total Organic Halogens	100	U	2.80		110	ug/L	107	(75%-125%)		07/19/16	20:05

Notes:

The Qualifiers in this report are defined as follows:

- < Sample is below the EPA guidance level for Reactive Releasable Cyanide and/or Reactive Releasable Sulfide
- > Result greater than quantifiable range or greater than upper limit of the analysis range
- B The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate).
- C Target analyte was detected in the sample and the associated blank. The associated blank concentration is \geq EQL or is $>$ 5% of the measured concentration and/or decision level for associated samples.
- D Results are reported from a diluted aliquot of sample.
- N Spike Sample recovery is outside control limits.
- U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
- X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
- Y Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier

GEL LABORATORIES LLC

2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 401205

Page 2 of 2

Parmname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date	Time
Z	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier										

N/A indicates that spike recovery limits do not apply when sample concentration exceeds spike conc. by a factor of 4 or more or %RPD not applicable.

^ The Relative Percent Difference (RPD) obtained from the sample duplicate (DUP) is evaluated against the acceptance criteria when the sample is greater than five times (5X) the contract required detection limit (RL). In cases where either the sample or duplicate value is less than 5X the RL, a control limit of +/- the RL is used to evaluate the DUP result.

* Indicates that a Quality Control parameter was not within specifications.

For PS, PSD, and SDILT results, the values listed are the measured amounts, not final concentrations.

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless qualified on the QC Summary.