



July 29, 2016  
ALS1607275

Ft. Collins, Colorado

LIMS Version: 6.820

Page 1 of 1

Friday, July 29, 2016

Karen Waters-Husted  
CH2M HILL Plateau Remediation Company  
2420 Stevens Center  
Richland, WA 99352

Re: ALS Workorder: 1607275  
Project Name: SURV, APRIL 2016  
Project Number: S16-004

Dear Ms. Waters-Husted:

One water sample was received from CH2M HILL Plateau Remediation Company, on 7/15/2016. The sample was scheduled for the following analysis:

Metals

The results for these analyses are contained in the enclosed reports.

The data contained in the following report have been reviewed and approved by the personnel listed below. In addition, ALS certifies that the analyses reported herein are true, complete and correct within the limits of the methods employed.

Thank you for your confidence in ALS Environmental. Should you have any questions, please call.

Sincerely,

ALS Environmental  
Julie Ellingson  
Project Manager

ADDRESS 225 Commerce Drive, Fort Collins, Colorado, USA 80524 | PHONE +1 970 490 1511 | FAX +1 970 490 1522  
ALS GROUP USA, CORP. Part of the ALS Laboratory Group An ALS Limited Company

Environmental

[www.alsglobal.com](http://www.alsglobal.com)

RIGHT SOLUTIONS RIGHT PARTNER

1 of 13

ALS Environmental – Fort Collins is accredited by the following accreditation bodies for various testing scopes in accordance with requirements of each accreditation body. All testing is performed under the laboratory management system, which is maintained to meet these requirement and regulations. Please contact the laboratory or accreditation body for the current scope testing parameters.

ALS Environmental – Fort Collins	
Accreditation Body	License or Certification Number
AIHA	214884
Alaska (AK)	UST-086
Alaska (AK)	CO01099
Arizona (AZ)	AZ0742
California (CA)	06251CA
Colorado (CO)	CO01099
Connecticut (CT)	PH-0232
Florida (FL)	E87914
Idaho (ID)	CO01099
Kansas (KS)	E-10381
Kentucky (KY)	90137
L-A-B (DoD ELAP/ISO 170250)	L2257
Louisiana (LA)	05057
Maryland (MD)	285
Missouri (MO)	175
Nebraska(NE)	NE-OS-24-13
Nevada (NV)	CO000782008A
New York (NY)	12036
North Dakota (ND)	R-057
Oklahoma (OK)	1301
Pennsylvania (PA)	68-03116
Tennessee (TN)	2976
Texas (TX)	T104704241
Utah (UT)	CO01099
Washington (WA)	C1280

July 29, 2016

ALS1607275

# ALS Environmental -- FC

## Sample Number(s) Cross-Reference Table

---

**OrderNum:** 1607275

**Client Name:** CH2M HILL Plateau Remediation Company

**Client Project Name:** SURV, APRIL 2016

**Client Project Number:** S16-004

**Client PO Number:** BOA 54854

---

Client Sample Number	Lab Sample Number	COC Number	Matrix	Date Collected	Time Collected
B34LP3	1607275-1		WATER	14-Jul-16	10:03

01 / 02

<b>CH2MHill Plateau Remediation Company</b>		<b>CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST</b> 1607275		C.O.C. # <b>S16-004-001</b>
Collector Kevin Patterson CHPRC	Contact/Requester Karen Waters-Husted	Telephone No. 509-376-4650	Page 1 of 1	
SAF No. S16-004	Sampling Origin Hanford Site	Purchase Order/Charge Code 300071		
Project Title SURV, APRIL 2016	Logbook No. HNF-N-506 05 / 94	Ice Chest No. GWS-555		
Shipped To (Lab) ALS Environmental Ft. Collins	Method of Shipment Commercial Carrier	Bill of Lading/Air Bill No. 77767 4908 8892		
Protocol CERCLA	Priority: 30 Days	Offsite Property No. 6028	Total Activity Exemption: Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>	
<b>POSSIBLE SAMPLE HAZARDS/REMARKS</b> *** Contains Radioactive Material at concentrations that are not regulated for transportation per 49 CFR / IATA Dangerous Goods Regulations but are not releasable per DOE Order 458.1		<b>SPECIAL INSTRUCTIONS</b> N/A		
Sample No. B34LP3 ①	Filter N	Date WJUL 14 2016	Time 1003	No/Type Container 1x500-mL G/P
Sample Analysis 6020_METALS_JCPMS: Uranium (1)		Holding Time 6 Months	Preservative HNO3 to pH <2	

1607275

Relinquished By Kevin Patterson CHPRC	Print 	Date JUL 14 2016	Time 150	Received By J.C. Fulton CHPRC	Sign 	Date/Time JUL 14 2016 1150	Matrix * S = Soil SE = Sediment SO = Solid SL = Sludge W = Water O = Oil A = Air
Relinquished By J.C. Fulton CHPRC	Print 	Date JUL 14 2016	Time 1400	Received By FEDEX	Sign FEDEX	Date/Time JUL 14 2016 1150	Matrix * DS = Drum Solids DL = Drum Liquids T = Tissue WI = Wipe L = Liquid V = Vegetation X = Other
Relinquished By FED EX	Print 	Date JUL 14 2016	Time 1400	Received By CTImbteC	Sign CTImbteC	Date/Time 7-15-14 0930	
Relinquished By 4	Print 	Date JUL 14 2016	Time 1400	Received By	Sign	Date/Time	
<b>FINAL SAMPLE DISPOSITION</b>		Disposal Method (e.g., Return to customer, per lab procedure, used in process)		Disposed By		Date/Time	



**ALS Environmental - Fort Collins**  
**CONDITION OF SAMPLE UPON RECEIPT FORM**

Client: CHPRC  
Project Manager: JK

Workorder No: 1607275  
Initials: CDI Date: 7-15-16

1. Does this project require any special handling in addition to standard ALS procedures?		YES	<input checked="" type="radio"/> NO
2. Are custody seals on shipping containers intact?	NONE	<input checked="" type="radio"/> YES	NO
3. Are Custody seals on sample containers intact?	NONE	<input checked="" type="radio"/> YES	NO
4. Is there a COC (Chain-of-Custody) present or other representative documents?		<input checked="" type="radio"/> YES	NO
5. Are the COC and bottle labels complete and legible?		<input checked="" type="radio"/> YES	NO
6. Is the COC in agreement with samples received? (IDs, dates, times, no. of samples, no. of containers, matrix, requested analyses, etc.)		<input checked="" type="radio"/> YES	NO
7. Were airbills / shipping documents present and/or removable?	DROP OFF	<input checked="" type="radio"/> YES	NO
8. Are all aqueous samples requiring preservation preserved correctly? (excluding volatiles)	N/A	<input checked="" type="radio"/> YES	NO
9. Are all aqueous non-preserved samples pH 4-9?	<input checked="" type="radio"/> N/A	YES	NO
10. Is there sufficient sample for the requested analyses?		<input checked="" type="radio"/> YES	NO
11. Were all samples placed in the proper containers for the requested analyses?		<input checked="" type="radio"/> YES	NO
12. Are all samples within holding times for the requested analyses?		<input checked="" type="radio"/> YES	NO
13. Were all sample containers received intact? (not broken or leaking, etc.)		<input checked="" type="radio"/> YES	NO
14. Are all samples requiring no headspace (VOC, GRO, RSK/MEE, Rx CN/S, radon) headspace free? Size of bubble: ___ < green pea ___ > green pea	<input checked="" type="radio"/> N/A	YES	NO
15. Do any water samples contain sediment? Amount Amount of sediment: ___ dusting ___ moderate ___ heavy	N/A	YES	<input checked="" type="radio"/> NO
16. Were the samples shipped on ice?		YES	<input checked="" type="radio"/> NO
17. Were cooler temperatures measured at 0.1-6.0°C? IR gun used*: #2 #4		YES	<input checked="" type="radio"/> NO
Cooler #: <u>1</u>			
Temperature (°C): <u>42.5</u>			
No. of custody seals on cooler: <u>2</u>			
DOT Survey/ Acceptance Information	External µR/hr reading: <u>13</u>		
	Background µR/hr reading: <u>11</u>		
Were external µR/hr readings ≤ two times background and within DOT acceptance criteria? <input checked="" type="radio"/> YES / <input type="radio"/> NO / <input type="radio"/> NA (If no, see Form 008.)			

**Additional Information:** PROVIDE DETAILS BELOW FOR A NO RESPONSE TO ANY QUESTION ABOVE, EXCEPT #1 AND #16.

If applicable, was the client contacted? YES / NO /  NA Contact: \_\_\_\_\_ Date/Time: \_\_\_\_\_

Project Manager Signature / Date: [Signature] 7/15/16

ORIGIN ID: PSCA (509) 528-9426  
LESLEY WALL  
CH2M  
6267 LATAH ST.  
8268 LATAH ST.  
RICHLAND, WA 99354  
UNITED STATES US

SHIP DATE: 14JUL16  
ACTWGHT: 27.00 LB  
CAD: 107066051/NET3730  
BILL THIRD PARTY

TO JULIE ELLINGSON  
ALS GLOBAL  
225 COMMERCE DRIVE

13-2

540115C8D727F

FORT COLLINS CO 80524  
REF: PTR# 6828

(970) 490-1511  
NW  
PO: DEPT:



FRI - 15 JUL 10:30A  
PRIORITY OVERNIGHT

TRK# 7767 4908 8892

DSR  
80524  
CO-US DEN

XH FTCA



After printing this label:

1. Use the 'Print' button on this page to print your label to your laser or inkjet printer.
2. Fold the printed page along the horizontal line.
3. Place label in shipping pouch and affix it to your shipment so that the barcode portion of the label can be read and scanned.

Warning: Use only the printed original label for shipping. Using a photocopy of this label for shipping purposes is fraudulent and could result in additional billing charges, along with the cancellation of your FedEx account number.

Use of this system constitutes your agreement to the service conditions in the current FedEx Service Guide, available on fedex.com. FedEx will not be responsible for any claim in excess of \$100 per package, whether the result of loss, damage, delay, non-delivery, misdelivery, or misinformation, unless you declare a higher value, pay an additional charge, document your actual loss and file a timely claim. Limitations found in the current FedEx Service Guide apply. Your right to recover from FedEx for any loss, including intrinsic value of the package, loss of sales, income interest, profit, attorney's fees, costs, and other forms of damage whether direct, incidental, consequential, or special is limited to the greater of \$100 or the authorized declared value. Recovery cannot exceed actual documented loss. Maximum for items of extraordinary value is \$1,000, e.g. jewelry, precious metals, negotiable instruments and other items listed in our ServiceGuide. Written claims must be filed within strict time limits, see current FedEx Service Guide.

ALS1607275



## Metals Case Narrative

---

### **CH2M HILL Plateau Remediation Company**

SURV, APRIL 2016 -- S16-004

Work Order Number: 1607275

1. This report consists of 1 water sample.
2. The sample was received intact at ambient temperature by ALS on 07/15/16.
3. The sample had a pH less than 2 upon receipt.
4. The sample was prepared and analyzed based on SW-846, 3<sup>rd</sup> Edition procedures.

For analysis by ICP-MS, the sample was digested following method 3005A and the current revision of SOP 806.

5. Analysis by ICP-MS followed method 6020A and the current revision of SOP 827.
6. All standards and solutions are NIST traceable and were used within their recommended shelf life.
7. The sample was prepared and analyzed within the established hold time.

All in house quality control procedures were followed, as described below.

8. General quality control procedures.
  - A preparation (method) blank and laboratory control sample were digested and analyzed with the samples in this digestion batch.
  - The preparation (method) blank associated with this digestion batch was below the reporting limit for the requested analyte. The sample result has been compared to the blank result.
  - All laboratory control sample criteria were met.



- All initial and continuing calibration blanks were below the reporting limit for the requested analyte.
- All initial and continuing calibration verifications were within the acceptance criteria for the requested analyte.
- The interference check samples associated with Method 6020A were analyzed.

9. Matrix specific quality control procedures.

Sample 1607299-4 was designated as the quality control sample for this analysis. Results for the shared quality control samples are included at the client's request.

Similarity of matrix and therefore relevance of the QC results should not be automatically inferred for any sample other than the native sample selected for QC.

- A matrix spike and matrix spike duplicate were digested and analyzed with this batch. All acceptance criteria for accuracy were met.
- A matrix spike duplicate was digested and analyzed with this batch. All acceptance criteria for precision were met.
- A serial dilution was analyzed with this ICP batch. All acceptance criteria were met.

10. It is a standard practice that samples for ICP-MS are analyzed at a 10X dilution. This dilution factor can be considered an artifact of the prep, and does not indicate a secondary dilution. Therefore, it is not flagged as a dilution.

The data contained in the following report have been reviewed and approved by the personnel listed below. In addition, ALS certifies that the analyses reported herein are true, complete and correct within the limits of the methods employed.

  
\_\_\_\_\_  
Emily Knopel  
Inorganics Primary Data Reviewer

7/28/16  
Date

  
\_\_\_\_\_  
April E. Llyza  
Inorganics Final Data Reviewer

7/28/16  
Date



### Inorganic Data Reporting Qualifiers

The following qualifiers are used as needed by the laboratory when reporting results of inorganic analyses.

- Result qualifier -- A "B" is entered if the reported value was obtained from a reading that was less than the Reporting Limit but greater than or equal to the Method Detection Limit (MDL). If the analyte was analyzed for but not detected a "U" is entered. For samples, negative values are reported as non-detects ("U" flagged). For blanks, if the absolute value of the negative value is above the MDL and below the reporting limit, then the result is "B" flagged.
- QC qualifier -- Specified entries and their meanings are as follows:
  - E - The reported value is estimated because of the presence of interference. An explanatory note may be included in the narrative.
  - M - Duplicate injection precision was not met.
  - N - Spiked sample recovery not within control limits. A post spike is analyzed for all ICP analyses when the matrix spike and or spike duplicate fail and the native sample concentration is less than four times the spike added concentration.
  - Z - Spiked recovery not within control limits. An explanatory note may be included in the narrative.
  - \* - Duplicate analysis (relative percent difference) not within control limits.
  - S - SAR value is estimated as one or more analytes used in the calculation were not detected above the detection limit.
  - C - The analyte was detected in both the sample and the associated QC blank, and the sample concentration was  $\leq 20X$  the blank concentration.
  - D - Analyte was reported at a secondary dilution factor, typically  $DF > 1$  (i.e., the primary preparation required dilution to either bring the analyte within the calibration range or to minimize interference). Required for organics/wetchem if the sample was diluted.

# Total Recoverable ICPMS Metals

Method SW6020A

Sample Results

Lab Name: ALS -- Fort Collins

Work Order Number: 1607275

Client Name: CH2M HILL Plateau Remediation Company

ClientProject ID: SURV, APRIL 2016 S16-004

Field ID: B34LP3	Sample Matrix: WATER	Prep Batch: IP160722-1	Analyst: Brent A. Stanfield
Lab ID: 1607275-1	% Moisture: N/A	QCBatchID: IP160722-1-2	Sample Aliquot: 50 ml
	Date Collected: 14-Jul-16	Run ID: IM160725-12A3	Final Volume: 50 ml
	Date Extracted: 22-Jul-16	Cleanup: NONE	Result Units: UG/L
Analysis ReqCode: 6020_METALS_I	Date Analyzed: 25-Jul-16	Basis: As Received	Clean DF: 1
	Prep Method: SW3005 Rev A	File Name: 066SMPL_	

CASNO	Target Analyte	Dilution Factor	Result	RptLimit/ LOQ/LOD	MDL/DL	Result Qualifier	EPA Qualifier
7440-61-1	URANIUM	10	1.8	0.1	0.027		

Data Package ID: IM1607275-1

July 29, 2016

ALS1607275

# ICPMS Metals

Method SW6020A

Method Blank

Lab Name: ALS -- Fort Collins

Work Order Number: 1607275

Client Name: CH2M HILL Plateau Remediation Company

ClientProject ID: SURV, APRIL 2016 S16-004

Lab ID: IP160722-1MB

Sample Matrix: WATER

% Moisture: N/A

Date Collected: N/A

Date Extracted: 22-Jul-16

Date Analyzed: 25-Jul-16

Prep Batch: IP160722-1

QCBatchID: IP160722-1-2

Run ID: IM160725-12A3

Cleanup: NONE

Basis: N/A

File Name: 063SMPL\_

Sample Aliquot: 50 ml

Final Volume: 50 ml

Result Units: UG/L

Clean DF: 1

CASNO	Target Analyte	DF	Result	RptLimit/ LOQ/LOD	MDL/DL	Result Qualifier	EPA Qualifier
7440-61-1	URANIUM	10	0.027	0.1	0.027	U	

Data Package ID: IM1607275-1

Date Printed: Thursday, July 28, 2016

ALS -- Fort Collins

Page 1 of 1

LIMS Version: 6.820

July 29, 2016

ALS1607275

# ICPMS Metals

Method SW6020A

## Laboratory Control Sample

Lab Name: ALS -- Fort Collins

Work Order Number: 1607275

Client Name: CH2M HILL Plateau Remediation Company

ClientProject ID: SURV, APRIL 2016 S16-004

Lab ID: IM160722-1LCS

Sample Matrix: WATER

% Moisture: N/A

Date Collected: N/A

Date Extracted: 07/22/2016

Date Analyzed: 07/25/2016

Prep Method: SW3005A

Prep Batch: IP160722-1

QCBatchID: IP160722-1-2

Run ID: IM160725-12A3

Cleanup: NONE

Basis: N/A

File Name: 064SMPL\_

Sample Aliquot: 50 ml

Final Volume: 50 ml

Result Units: UG/L

Clean DF: 1

CASNO	Target Analyte	Spike Added	LCS Result	Reporting Limit	Result Qualifier	LCS % Rec.	Control Limits
7440-61-1	URANIUM	10	10.8	0.1		108	80 - 120%

Data Package ID: IM1607275-1

July 29, 2016

ALS1607275

# ICPMS Metals

Method SW6020A

## Matrix Spike And Matrix Spike Duplicate

Lab Name: ALS -- Fort Collins

Work Order Number: 1607275

Client Name: CH2M HILL Plateau Remediation Company

ClientProject ID: SURV, APRIL 2016 S16-004

Field ID: SHARED QC	Sample Matrix: WATER	Prep Batch: IP160722-1	Sample Aliquot: 50 ml
LabID: 1607299-4MS	% Moisture: N/A	QCBatchID: IP160722-1-2	Final Volume: 50 ml
	Date Collected: 15-Jul-16	Run ID: IM160725-12A3	Result Units: UG/L
	Date Extracted: 22-Jul-16	Cleanup: NONE	File Name: 075SMPL_
	Date Analyzed: 25-Jul-16	Basis: As Received	
	Prep Method: SW3005 Rev A		

CASNO	Target Analyte	Sample Result	Samp Qual	MS Result	MS Qual	Reporting Limit	Spike Added	MS % Rec.	Control Limits
7440-61-1	URANIUM	12		22.4		0.1	10	104	75 - 125%

Field ID: SHARED QC	Sample Matrix: WATER	Prep Batch: IP160722-1	Sample Aliquot: 50 ml
LabID: 1607299-4MSD	% Moisture: N/A	QCBatchID: IP160722-1-2	Final Volume: 50 ml
	Date Collected: 15-Jul-16	Run ID: IM160725-12A3	Result Units: UG/L
	Date Extracted: 22-Jul-16	Cleanup: NONE	File Name: 076SMPL_
	Date Analyzed: 25-Jul-16	Basis: As Received	
	Prep Method: SW3005 Rev A		

CASNO	Target Analyte	MSD Result	MSD Qual	Spike Added	MSD % Rec.	Reporting Limit	RPD Limit	RPD
7440-61-1	URANIUM	22.4		10	103	0.1	20	0

Data Package ID: IM1607275-1