

July 13, 2016

TestAmerica

THE LEADER IN ENVIRONMENTAL TESTING

ANALYTICAL REPORT

TestAmerica Laboratories, Inc.

TestAmerica Richland

2800 George Washington Way

Richland, WA 99352

Tel: (509)375-3131

TestAmerica Job ID: 300-3405-1

TestAmerica Sample Delivery Group: WC1228

Client Project/Site: X16-043

For:


CH2M Hill Plateau Remediation Company

PO BOX 1600

Mail Stop R3-50

Richland, Washington 99352

Attn: CPP Sample Management



Authorized for release by:

7/13/2016 12:08:26 PM

Steven Campbell, Project Mgmt. Assistant

steven.campbell@testamericainc.com

Designee for

Whitney Ritari, Project Manager I

(509)375-3131 ext164

whitney.ritari@testamericainc.com

LINKS

Review your project
results through

TotalAccess

Have a Question?



Visit us at:

www.testamericainc.com

I certify that this Certificate of Analysis is in compliance with the SOW, both technically and for completeness, for other than the conditions detailed in this package. Release of the data contained in this hard copy data package has been authorized by the Laboratory Manager, or a designee as verified by the attached signature.

The laboratory is accredited to EPA 7196A in Solid and Chemical Materials, per Washington State Department of Ecology this is approved for non-discharge water samples. The sample/ sample duplicate RPD agreement is not applicable if one or both results are less than five times the MDL.



Table of Contents

Cover Page	1
Table of Contents	2
Definitions/Glossary	3
Case Narrative	4
Chain of Custody	5
Detection Summary	10
Client Sample Results	11
QC Sample Results	13
QC Association Summary	14
Lab Chronicle	15
Method Summary	18
Sample Summary	19
Receipt Checklists	20

Definitions/Glossary

Client: CH2M Hill Plateau Remediation Company
 Project/Site: X16-043

TestAmerica Job ID: 300-3405-1
 SDG: WC1228

Qualifiers

General Chemistry

Qualifier	Qualifier Description
U	Analyzed for but not detected.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
□	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CNF	Contains no Free Liquid
DER	Duplicate error ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision level concentration
MDA	Minimum detectable activity
EDL	Estimated Detection Limit
MDC	Minimum detectable concentration
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
NC	Not Calculated
ND	Not detected at the reporting limit (or MDL or EDL if shown)
PQL	Practical Quantitation Limit
QC	Quality Control
RER	Relative error ratio
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)



July 13, 2016

Case Narrative

Client: CH2M Hill Plateau Remediation Company
Project/Site: X16-043

TestAmerica Job ID: 300-3405-1
SDG: WC1228

Job ID: 300-3405-1

Laboratory: TestAmerica Richland

Narrative

Job Narrative
300-3405-1

Comments

No additional comments.

Receipt

The samples were received on 7/1/2016 2:40 PM; the samples arrived in good condition, properly preserved and, where required, on ice. The temperature of the cooler at receipt was 8.6° C.

General Chemistry

No analytical or quality issues were noted, other than those described in the Definitions/Glossary page.

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

CH2MHill Plateau Remediation Company

CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST

C.O.C. #

X16-043-037

Page 1 of 1

Collector	Kevin Patterson CHPRC	Contact/Requester	WATERS-HUSTED, K	Telephone No.	376-4650
SAF No.	X16-043	Sampling Origin	Hanford Site	Purchase Order/Charge Code	304027
Project Title	100-KW Rebound Study, June 27, 2016	Logbook No.	HNF-N-506 <i>Se/4D</i>	Ice Chest No.	N/A
Shipped To (Lab)	TestAmerica Incorporated, Richland	Method of Shipment	GOVERNMENT VEHICLE	Bill of Lading/Air Bill No.	N/A
Protocol	CERCLA	Priority:	30 Days	Offsite Property No.	N/A
POSSIBLE SAMPLE HAZARDS/REMARKS			SPECIAL INSTRUCTIONS		
*Contains Radioactive Material at concentrations that are not regulated for transportation per 49 CFR/ATA Dangerous Goods Regulations but are not releasable per DOE Order 458.1.			Hold Time		
			Total Activity Exemption: Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>		

Sample No.	Filter	* Date	Time	No./Type Container	Sample Analysis	Holding Time	Preservative
B360H3	N	<i>W JUL 11 7:05</i>	<i>0805</i>	1x60-mL aG	7196_CR6: COMMON	24 Hours	Cool <=6C
B360H2	Y	<i>W JUL 11 7:05</i>	<i>0805</i>	1x60-mL aG	7196_CR6: COMMON	24 Hours	Cool <=6C

*A 3405
WCBAS*

300-3405 COC



Relinquished By	Kevin Patterson CHPRC	Print	Sign	Date/Time	Received By	Frank Hest CHPRC	Print	Sign	Date/Time
Relinquished By	<i>Frank Hest CHPRC</i>			JUL 01 2016	Received By	<i>J. Bock, TART</i>			JUL 11 2016
Relinquished By				JUL 01 2016	Received By				JUL 01 2016
Relinquished By					Received By				

Matrix *	
S	= Soil
SE	= Sediment
SO	= Solid
SL	= Sludge
W	= Water
O	= Oil
A	= Air
DS	= Drum Solids
DL	= Drum Liquids
T	= Tissue
WL	= Wipe
L	= Liquid
V	= Vegetation
X	= Other

FINAL SAMPLE DISPOSITION	Disposed Method (e.g., Return to customer, per lab procedure, used in process)	Disposed By	Date/Time
---------------------------------	--	-------------	-----------



CH2MHill Plateau Remediation Company

CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST

C.O.C. #
X16-043-020
Page 1 of 1

Collector	Kevin Patterson CHPRC	Contact/Requester	WATERS-HUSTED, K	Telephone No.	376-4650
SAF No.	X16-043	Sampling Origin	Hanford Site	Purchase Order/Charge Code	304027
Project Title	100-KW Rebound Study, June 27, 2016	Logbook No.	HNF-N-506 <i>Site / 41</i>	Ice Chest No.	N/A
Shipped To (Lab)	TestAmerica Incorporated, Richland	Method of Shipment	GOVERNMENT VEHICLE	Bill of Lading/Air Bill No.	N/A
Protocol	CERCLA	Priority:	30 Days	Offsite Property No.	N/A

POSSIBLE SAMPLE HAZARDS/REMARKS
 *Contains Radioactive Material at concentrations that are not be regulated for transportation per 49 CFR/ATA Dangerous Goods Regulations but are not releasable per DOE Order 458.1.

Sample No.	Filter	* Date	Time	No./Type Container	Sample Analysis	Holding Time	Preservative
B360C9	N	WILL 11 7/11/16	1:50	1x60-ml aG	7196_CR6: COMMON	24 Hours	Cool <=6C
B360C8	Y	W	Y	1x60-ml aG	7196_CR6: COMMON	24 Hours	Cool <=6C

*3x05
WCI222*

Relinquished By	Kevin Patterson CHPRC	Print	Sign	Date/Time	1/2/15	Received By	Print	Sign	Date/Time	7/2/16
Relinquished By	<i>Frank Hild</i> CHPRC	<i>[Signature]</i>	<i>[Signature]</i>	Date/Time	JUL 01 2016	Received By	<i>[Signature]</i>	<i>[Signature]</i>	Date/Time	JUL 01 2016
Relinquished By				Date/Time	JUL 01 2016	Received By	J. Bock, TARL		Date/Time	JUL 01 2016
Relinquished By				Date/Time		Received By			Date/Time	

S	= Soil	DS	= Drum Solids
SE	= Sediment	T	= Tissue
SO	= Solid	WI	= Wipe
SL	= Sludge	L	= Liquid
W	= Water	V	= Vegetation
O	= Oil	X	= Other
A	= Air		

FINAL SAMPLE DISPOSITION Disposal Method (e.g., Return to customer, per lab procedure, used in process) Disposed By Date/Time

CH2M/Hill Plateau Remediation Company

CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST

C.O.C. #
X16-043-017
Page 1 of 1

Collector	Kevin Paterson CHPRC	Contact/Requester	WATERS-HUSTED, K	Telephone No.	376-4650
SAF No.	X16-043	Sampling Origin	Hanford Site	Purchase Order/Charge Code	304027
Project Title	100-KW Rebound Study, June 27, 2016	Logbook No.	HNF-N-506 <i>Site 40</i>	Ice Chest No.	N/A
Shipped To (Lab)	TestAmerica Incorporated, Richland	Method of Shipment	GOVERNMENT VEHICLE	Bill of Lading/Air Bill No.	N/A
Protocol	CERCLA	Priority:	30 Days	Offsite Property No.	N/A
POSSIBLE SAMPLE HAZARDS/REMARKS			SPECIAL INSTRUCTIONS		
*Contains Radioactive Material at concentrations that are not regulated for transportation per 49 CFR/LATA Dangerous Goods Regulations but are not releasable per DOE Order 458.1.			Hold Time		
			Total Activity Exemption: Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>		

Sample No.	Filter	* Date	Time	No/Type Container	Sample Analysis	Holding Time	Preservative
B360C4	N	JUL 01 2016	1007	1x60-ml aG	7196_Cr6: COMMON	24 Hours	Cool <=6C
B360C3	Y	W		1x60-ml aG	7196_Cr6: COMMON	24 Hours	Cool <=6C

Waters

Relinquished By	Kevin Paterson CHPRC	Print	Sign	Date/Time	Received By	Frank Hagg CHPRC	Print	Sign	Date/Time	Matrix *
Relinquished By	Frank Hagg CHPRC	Print	Sign	Date/Time	Received By	J. Bock, TARL	Print	Sign	Date/Time	Matrix *
Relinquished By		Print	Sign	Date/Time	Received By		Print	Sign	Date/Time	Matrix *

FINAL SAMPLE DISPOSITION	Disposal Method (e.g., Return to customer, per lab procedure, used in process)	Disposed By	Date/Time
--------------------------	--	-------------	-----------

CH2MHill Plateau Remediation Company

CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST

C.O.C. #
X16-043-023

Page 1 of 1

Collector	Kevin Patterson CHPRC	Contact/Requester	WATERS-HUSTED, K	Telephone No.	376-4650
SAF No.	X16-043	Sampling Origin	Hanford Site	Purchase Order/Charge Code	304027
Project Title	100-KW Rebound Study, June 27, 2016	Logbook No.	HNF-N-506	Ice Chest No.	N/A
Shipped To (Lab)	TestAmerica Incorporated, Richland	Method of Shipment	GOVERNMENT VEHICLE	Bill of Lading/Air Bill No.	N/A
Protocol	CERCLA	Priority:	30 Days	Offsite Property No.	N/A
POSSIBLE SAMPLE HAZARDS/REMARKS			SPECIAL INSTRUCTIONS		
*Contains Radioactive Material at concentrations that are not be regulated for transportation per 49 CFR/IATA Dangerous Goods Regulations but are not releasable per DOE Order 458.1.			Hold Time		
			Total Activity Exemption: Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>		

Sample No.	Filter	*	Date	Time	No./Type Container	Sample Analysis	Holding Time	Preservative
B360D3	Y		JUL 01 2016	0920	1x60-ml aG	7196_CR6: COMMON	24 Hours	Cool <=6C
B360D4	N		W		1x60-ml aG	7196_CR6: COMMON	24 Hours	Cool <=6C

3405
W01228

Relinquished By	Kevin Patterson CHPRC	Print	Sign	Date/Time	10/25	Received By	Frank Hill CHPRC	Print	Sign	Date/Time	10/25
Relinquished By	Frank Hill CHPRC	Print	Sign	Date/Time	JUL 01 2016	Received By	J. Bock, TARL	Print	Sign	Date/Time	JUL 01 2016
Relinquished By		Print	Sign	Date/Time		Received By		Print	Sign	Date/Time	

- Matrix *
- S = Soil
 - SE = Sediment
 - SO = Solid
 - SL = Sludge
 - W = Water
 - O = Oil
 - A = Air
 - DS = Drum Solids
 - DL = Drum Liquids
 - T = Tissue
 - WI = Wipe
 - L = Liquid
 - V = Vegetation
 - X = Other

FINAL SAMPLE DISPOSITION	Disposal Method (e.g., Return to customer, per lab procedure, used in process)	Disposed By	Date/Time
---------------------------------	--	-------------	-----------

Detection Summary

Client: CH2M Hill Plateau Remediation Company
 Project/Site: X16-043

TestAmerica Job ID: 300-3405-1
 SDG: WC1228

Client Sample ID: B360H3

Lab Sample ID: 300-3405-1

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Cr (VI)	0.022		0.0040	0.0015	mg/L	1		7196A	Total/NA

Client Sample ID: B360H2

Lab Sample ID: 300-3405-2

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Cr (VI)	0.023		0.0040	0.0015	mg/L	1		7196A	Total/NA

Client Sample ID: B360H0

Lab Sample ID: 300-3405-3

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Cr (VI)	0.021		0.0040	0.0015	mg/L	1		7196A	Total/NA

Client Sample ID: B360H1

Lab Sample ID: 300-3405-4

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Cr (VI)	0.021		0.0040	0.0015	mg/L	1		7196A	Total/NA

Client Sample ID: B360C9

Lab Sample ID: 300-3405-5

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Cr (VI)	0.0097		0.0040	0.0015	mg/L	1		7196A	Total/NA

Client Sample ID: B360C8

Lab Sample ID: 300-3405-6

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Cr (VI)	0.0094		0.0040	0.0015	mg/L	1		7196A	Total/NA

Client Sample ID: B360C4

Lab Sample ID: 300-3405-7

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Cr (VI)	0.015		0.0040	0.0015	mg/L	1		7196A	Total/NA

Client Sample ID: B360C3

Lab Sample ID: 300-3405-8

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Cr (VI)	0.015		0.0040	0.0015	mg/L	1		7196A	Total/NA

Client Sample ID: B360D3

Lab Sample ID: 300-3405-9

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Cr (VI)	0.035		0.0040	0.0015	mg/L	1		7196A	Total/NA

Client Sample ID: B360D4

Lab Sample ID: 300-3405-10

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Cr (VI)	0.035		0.0040	0.0015	mg/L	1		7196A	Total/NA

This Detection Summary does not include radiochemical test results.

TestAmerica Richland

Client Sample Results

Client: CH2M Hill Plateau Remediation Company
 Project/Site: X16-043

TestAmerica Job ID: 300-3405-1
 SDG: WC1228

Client Sample ID: B360H3

Lab Sample ID: 300-3405-1

Date Collected: 07/01/16 08:05

Matrix: Water

Date Received: 07/01/16 14:40

General Chemistry

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Cr (VI)	0.022		0.0040	0.0015	mg/L			07/01/16 16:16	1

Client Sample ID: B360H2

Lab Sample ID: 300-3405-2

Date Collected: 07/01/16 08:05

Matrix: Water

Date Received: 07/01/16 14:40

General Chemistry

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Cr (VI)	0.023		0.0040	0.0015	mg/L			07/01/16 16:17	1

Client Sample ID: B360H0

Lab Sample ID: 300-3405-3

Date Collected: 07/01/16 08:05

Matrix: Water

Date Received: 07/01/16 14:40

General Chemistry

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Cr (VI)	0.021		0.0040	0.0015	mg/L			07/01/16 16:17	1

Client Sample ID: B360H1

Lab Sample ID: 300-3405-4

Date Collected: 07/01/16 08:05

Matrix: Water

Date Received: 07/01/16 14:40

General Chemistry

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Cr (VI)	0.021		0.0040	0.0015	mg/L			07/01/16 16:17	1

Client Sample ID: B360C9

Lab Sample ID: 300-3405-5

Date Collected: 07/01/16 10:50

Matrix: Water

Date Received: 07/01/16 14:40

General Chemistry

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Cr (VI)	0.0097		0.0040	0.0015	mg/L			07/01/16 16:18	1

Client Sample ID: B360C8

Lab Sample ID: 300-3405-6

Date Collected: 07/01/16 10:50

Matrix: Water

Date Received: 07/01/16 14:40

General Chemistry

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Cr (VI)	0.0094		0.0040	0.0015	mg/L			07/01/16 16:18	1

Client Sample ID: B360C4

Lab Sample ID: 300-3405-7

Date Collected: 07/01/16 10:07

Matrix: Water

Date Received: 07/01/16 14:40

General Chemistry

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Cr (VI)	0.015		0.0040	0.0015	mg/L			07/01/16 16:21	1

Client Sample Results

Client: CH2M Hill Plateau Remediation Company
 Project/Site: X16-043

TestAmerica Job ID: 300-3405-1
 SDG: WC1228

Client Sample ID: B360C3

Date Collected: 07/01/16 10:07

Date Received: 07/01/16 14:40

Lab Sample ID: 300-3405-8

Matrix: Water

General Chemistry

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Cr (VI)	0.015		0.0040	0.0015	mg/L			07/01/16 16:22	1

Client Sample ID: B360D3

Date Collected: 07/01/16 09:20

Date Received: 07/01/16 14:40

Lab Sample ID: 300-3405-9

Matrix: Water

General Chemistry

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Cr (VI)	0.035		0.0040	0.0015	mg/L			07/01/16 16:22	1

Client Sample ID: B360D4

Date Collected: 07/01/16 09:20

Date Received: 07/01/16 14:40

Lab Sample ID: 300-3405-10

Matrix: Water

General Chemistry

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Cr (VI)	0.035		0.0040	0.0015	mg/L			07/01/16 16:22	1

QC Sample Results

Client: CH2M Hill Plateau Remediation Company
 Project/Site: X16-043

TestAmerica Job ID: 300-3405-1
 SDG: WC1228

Method: 7196A - Chromium, Hexavalent

Lab Sample ID: MB 300-4775/3
 Matrix: Water
 Analysis Batch: 4775

Client Sample ID: Method Blank
 Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Cr (VI)	0.0015	U	0.0040	0.0015	mg/L			07/01/16 16:13	1

Lab Sample ID: LCS 300-4775/4
 Matrix: Water
 Analysis Batch: 4775

Client Sample ID: Lab Control Sample
 Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
Cr (VI)	0.250	0.251		mg/L		101	80 - 120

Lab Sample ID: 300-3405-1 MS
 Matrix: Water
 Analysis Batch: 4775

Client Sample ID: B360H3
 Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec. Limits
Cr (VI)	0.022		0.0499	0.0721		mg/L		101	75 - 125

Lab Sample ID: 300-3405-1 DU
 Matrix: Water
 Analysis Batch: 4775

Client Sample ID: B360H3
 Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	DU Result	DU Qualifier	Unit	D	RPD	RPD Limit
Cr (VI)	0.022		0.0219		mg/L		1	20

QC Association Summary

Client: CH2M Hill Plateau Remediation Company
 Project/Site: X16-043

TestAmerica Job ID: 300-3405-1
 SDG: WC1228

General Chemistry

Analysis Batch: 4775

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
300-3405-1	B360H3	Total/NA	Water	7196A	
300-3405-1 DU	B360H3	Total/NA	Water	7196A	
300-3405-1 MS	B360H3	Total/NA	Water	7196A	
300-3405-2	B360H2	Total/NA	Water	7196A	
300-3405-3	B360H0	Total/NA	Water	7196A	
300-3405-4	B360H1	Total/NA	Water	7196A	
300-3405-5	B360C9	Total/NA	Water	7196A	
300-3405-6	B360C8	Total/NA	Water	7196A	
300-3405-7	B360C4	Total/NA	Water	7196A	
300-3405-8	B360C3	Total/NA	Water	7196A	
300-3405-9	B360D3	Total/NA	Water	7196A	
300-3405-10	B360D4	Total/NA	Water	7196A	
LCS 300-4775/4	Lab Control Sample	Total/NA	Water	7196A	
MB 300-4775/3	Method Blank	Total/NA	Water	7196A	

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

Lab Chronicle

Client: CH2M Hill Plateau Remediation Company
 Project/Site: X16-043

TestAmerica Job ID: 300-3405-1
 SDG: WC1228

Client Sample ID: B360H3

Date Collected: 07/01/16 08:05

Date Received: 07/01/16 14:40

Lab Sample ID: 300-3405-1

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	7196A		1			4775	07/01/16 16:16	RLC	TAL RCH

Client Sample ID: B360H2

Date Collected: 07/01/16 08:05

Date Received: 07/01/16 14:40

Lab Sample ID: 300-3405-2

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	7196A		1			4775	07/01/16 16:17	RLC	TAL RCH

Client Sample ID: B360H0

Date Collected: 07/01/16 08:05

Date Received: 07/01/16 14:40

Lab Sample ID: 300-3405-3

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	7196A		1			4775	07/01/16 16:17	RLC	TAL RCH

Client Sample ID: B360H1

Date Collected: 07/01/16 08:05

Date Received: 07/01/16 14:40

Lab Sample ID: 300-3405-4

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	7196A		1			4775	07/01/16 16:17	RLC	TAL RCH

Client Sample ID: B360C9

Date Collected: 07/01/16 10:50

Date Received: 07/01/16 14:40

Lab Sample ID: 300-3405-5

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	7196A		1			4775	07/01/16 16:18	RLC	TAL RCH

Client Sample ID: B360C8

Date Collected: 07/01/16 10:50

Date Received: 07/01/16 14:40

Lab Sample ID: 300-3405-6

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	7196A		1			4775	07/01/16 16:18	RLC	TAL RCH

Lab Chronicle

Client: CH2M Hill Plateau Remediation Company
 Project/Site: X16-043

TestAmerica Job ID: 300-3405-1
 SDG: WC1228

Client Sample ID: B360C4

Date Collected: 07/01/16 10:07
 Date Received: 07/01/16 14:40

Lab Sample ID: 300-3405-7

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	7196A		1			4775	07/01/16 16:21	RLC	TAL RCH

Client Sample ID: B360C3

Date Collected: 07/01/16 10:07
 Date Received: 07/01/16 14:40

Lab Sample ID: 300-3405-8

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	7196A		1			4775	07/01/16 16:22	RLC	TAL RCH

Client Sample ID: B360D3

Date Collected: 07/01/16 09:20
 Date Received: 07/01/16 14:40

Lab Sample ID: 300-3405-9

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	7196A		1			4775	07/01/16 16:22	RLC	TAL RCH

Client Sample ID: B360D4

Date Collected: 07/01/16 09:20
 Date Received: 07/01/16 14:40

Lab Sample ID: 300-3405-10

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	7196A		1			4775	07/01/16 16:22	RLC	TAL RCH

Client Sample ID: B360H3

Date Collected: 07/01/16 08:05
 Date Received: 07/01/16 14:40

Lab Sample ID: 300-3405-1 DU

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	7196A		1			4775	07/01/16 16:17	RLC	TAL RCH

Client Sample ID: Lab Control Sample

Date Collected: N/A
 Date Received: N/A

Lab Sample ID: LCS 300-4775/4

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	7196A		1			4775	07/01/16 16:13	RLC	TAL RCH

Lab Chronicle

Client: CH2M Hill Plateau Remediation Company
 Project/Site: X16-043

TestAmerica Job ID: 300-3405-1
 SDG: WC1228

Client Sample ID: Method Blank

Lab Sample ID: MB 300-4775/3

Date Collected: N/A

Matrix: Water

Date Received: N/A

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	7196A		1			4775	07/01/16 16:13	RLC	TAL RCH

Client Sample ID: B360H3

Lab Sample ID: 300-3405-1 MS

Date Collected: 07/01/16 08:05

Matrix: Water

Date Received: 07/01/16 14:40

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	7196A		1		3.0 mL	4775	07/01/16 16:14	RLC	TAL RCH

Laboratory References:

TAL RCH = TestAmerica Richland, 2800 George Washington Way, Richland, WA 99352, TEL (509)375-3131

Method Summary

Client: CH2M Hill Plateau Remediation Company
Project/Site: X16-043

TestAmerica Job ID: 300-3405-1
SDG: WC1228

Method	Method Description	Protocol	Laboratory
7196A	Chromium, Hexavalent	SW846	TAL RCH

Protocol References:

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

Laboratory References:

TAL RCH = TestAmerica Richland, 2800 George Washington Way, Richland, WA 99352, TEL (509)375-3131



July 13, 2016

Sample Summary

Client: CH2M Hill Plateau Remediation Company
Project/Site: X16-043

TestAmerica Job ID: 300-3405-1
SDG: WC1228

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
300-3405-1	B360H3	Water	07/01/16 08:05	07/01/16 14:40
300-3405-2	B360H2	Water	07/01/16 08:05	07/01/16 14:40
300-3405-3	B360H0	Water	07/01/16 08:05	07/01/16 14:40
300-3405-4	B360H1	Water	07/01/16 08:05	07/01/16 14:40
300-3405-5	B360C9	Water	07/01/16 10:50	07/01/16 14:40
300-3405-6	B360C8	Water	07/01/16 10:50	07/01/16 14:40
300-3405-7	B360C4	Water	07/01/16 10:07	07/01/16 14:40
300-3405-8	B360C3	Water	07/01/16 10:07	07/01/16 14:40
300-3405-9	B360D3	Water	07/01/16 09:20	07/01/16 14:40
300-3405-10	B360D4	Water	07/01/16 09:20	07/01/16 14:40

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

Login Sample Receipt Checklist

Client: CH2M Hill Plateau Remediation Company

Job Number: 300-3405-1

SDG Number: WC1228

Login Number: 3405

List Number: 1

Creator: Bock, Julie A

List Source: TestAmerica Richland

Question	Answer	Comment
Radioactivity wasn't checked or is <=/ background as measured by a survey meter.	True	
The cooler's custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	N/A	
Multiphasic samples are not present.	N/A	
Samples do not require splitting or compositing.	True	
Residual Chlorine Checked.	N/A	

