

# TestAmerica

THE LEADER IN ENVIRONMENTAL TESTING

## ANALYTICAL REPORT

TestAmerica Laboratories, Inc.  
TestAmerica Richland  
2800 George Washington Way  
Richland, WA 99352  
Tel: (509)375-3131

TestAmerica Job ID: 300-1771-1  
TestAmerica Sample Delivery Group: WC0676  
Client Project/Site: X16-001

For:  
CH2M Hill Plateau Remediation Company  
PO BOX 1600  
Mail Stop R3-50  
Richland, Washington 99352

Attn: CPP Sample Management



Authorized for release by:  
12/10/2015 11:23:02 AM  
Steven Campbell, Quality Assurance Assistant  
[steven.campbell@testamericainc.com](mailto:steven.campbell@testamericainc.com)

Designee for  
Whitney Ritari, Project Manager I  
(509)375-3131 ext164  
[whitney.ritari@testamericainc.com](mailto:whitney.ritari@testamericainc.com)

*I certify that this Certificate of Analysis is in compliance with the SOW, both technically and for completeness, for other than the conditions detailed in this package. Release of the data contained in this hard copy data package has been authorized by the Laboratory Manager, or a designee as verified by the attached signature.*

*The laboratory is accredited to EPA 7196A in Solid and Chemical Materials, per Washington State Department of Ecology this is approved for non-discharge water samples. The sample/ sample duplicate RPD agreement is not applicable if one or both results are elss than five times the MDL.*

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13



### LINKS

Review your project results through  
**TotalAccess**

Have a Question?



Visit us at:  
[www.testamericainc.com](http://www.testamericainc.com)



# Table of Contents

Cover Page . . . . .	1
Table of Contents . . . . .	2
Definitions/Glossary . . . . .	3
Case Narrative . . . . .	4
Chain of Custody . . . . .	5
Detection Summary . . . . .	6
Client Sample Results . . . . .	7
QC Sample Results . . . . .	8
QC Association Summary . . . . .	10
Lab Chronicle . . . . .	11
Method Summary . . . . .	13
Sample Summary . . . . .	14
Receipt Checklists . . . . .	15

## Definitions/Glossary

Client: CH2M Hill Plateau Remediation Company  
 Project/Site: X16-001

TestAmerica Job ID: 300-1771-1  
 SDG: WC0676



### Qualifiers

#### HPLC/IC

Qualifier	Qualifier Description
D	The reported value is from a dilution.
U	Analyzed for but not detected.

### Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
▫	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CNF	Contains no Free Liquid
DER	Duplicate error ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision level concentration
MDA	Minimum detectable activity
EDL	Estimated Detection Limit
MDC	Minimum detectable concentration
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
NC	Not Calculated
ND	Not detected at the reporting limit (or MDL or EDL if shown)
PQL	Practical Quantitation Limit
QC	Quality Control
RER	Relative error ratio
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)

### Case Narrative

Client: CH2M Hill Plateau Remediation Company  
Project/Site: X16-001

TestAmerica Job ID: 300-1771-1  
SDG: WC0676



**Job ID: 300-1771-1**

**Laboratory: TestAmerica Richland**

**Narrative**

**Job Narrative  
300-1771-1**

**Comments**

No additional comments.

**Receipt**

The sample was received on 12/7/2015 2:45 PM; the sample arrived in good condition, properly preserved and, where required, on ice. The temperature of the cooler at receipt was 6.4° C.

**HPLC/IC**

No analytical or quality issues were noted, other than those described in the Definitions/Glossary page.



CH2MHill Plateau Remediation Company

CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST

C.O.C. #

X16-001-091

Page 1 of 1

Collector	KC. Paterson/CHPRC	Contact/Requester	Karen Waters-Husted	Telephone No.	509-376-4650
SAF No.	X16-001	Sampling Origin	Hanford Site	Purchase Order/Charge Code	303271
Project Title	100-BG-5 RI, OCTOBER 2015	Logbook No.	HNF-N-506 <u>8051</u>	Ice Chest No.	N/A
Shipped To (Lab)	TestAmerica Incorporated, Richland	Method of Shipment	GOVERNMENT VEHICLE	Bill of Lading/Air Bill No.	N/A
Protocol	CERCLA	Priority:	30 Days	Offsite Property No.	N/A

POSSIBLE SAMPLE HAZARDS/REMARKS

\*\*\* Contains Radioactive Material at concentrations that are not regulated for transportation per 49 CFR / IATA Dangerous Goods Regulations but are not releasable per DOE Order 458.1

SPECIAL INSTRUCTIONS

Submit deliverables & invoices to CCP Sample Management.

Hold Time

Total Activity Exemption: Yes  No

Sample No.	Filter	*	Date	Time	No/Type Container	Sample Analysis	Holding Time	Preservative
B321B0	N		WDEC 07 2015	1047	1x500-ml P	300.0_ANIONS_IC: COMMON	48 Hours	Cool <=6C



300-1771 COC

# 1771

W0676

Relinquished By	Print	Sign	Date/Time	Received By	Print	Sign	Date/Time	Matrix *
KC. Paterson/CHPRC			DEC 07 2015 1205	F.M. Hall/CHPRC			DEC 07 2015 1205	S = Soil SE = Sediment SO = Solid SL = Sludge W = Water O = Oil A = Air
Relinquished By			DEC 07 2015 1445	J. Friesz, TARL			DEC 07 2015 1445	DS = Drums Solids DL = Tissue T = Tissue WI = Wipe L = Liquid V = Vegetation X = Other
Relinquished By				Received By				

FINAL SAMPLE DISPOSITION

Disposal Method (e.g., Return to customer, per lab procedure, used in process)

Disposed By

Date/Time

### Detection Summary

Client: CH2M Hill Plateau Remediation Company  
Project/Site: X16-001

TestAmerica Job ID: 300-1771-1  
SDG: WC0676

**Client Sample ID: B32TB0**

**Lab Sample ID: 300-1771-1**

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Chloride	10	D	0.40	0.20	mg/L	2		300.0	Total/NA
Nitrate as N	2.3	D	0.056	0.028	mg/L	2		300.0	Total/NA
Fluoride	0.16	D	0.10	0.050	mg/L	2		300.0	Total/NA
Sulfate	43	D	0.50	0.25	mg/L	2		300.0	Total/NA



This Detection Summary does not include radiochemical test results.

TestAmerica Richland

### Client Sample Results

Client: CH2M Hill Plateau Remediation Company  
Project/Site: X16-001

TestAmerica Job ID: 300-1771-1  
SDG: WC0676

**Client Sample ID: B32TB0**

**Lab Sample ID: 300-1771-1**

**Date Collected: 12/07/15 10:47**

**Matrix: Water**

**Date Received: 12/07/15 14:45**

**Method: 300.0 - Anions, Ion Chromatography**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Chloride	10	D	0.40	0.20	mg/L			12/08/15 15:29	2
Nitrate as N	2.3	D	0.056	0.028	mg/L			12/08/15 15:29	2
Fluoride	0.16	D	0.10	0.050	mg/L			12/08/15 15:29	2
Nitrite as N	0.038	U	0.076	0.038	mg/L			12/08/15 15:29	2
Sulfate	43	D	0.50	0.25	mg/L			12/08/15 15:29	2



QC Sample Results

Client: CH2M Hill Plateau Remediation Company  
Project/Site: X16-001

TestAmerica Job ID: 300-1771-1  
SDG: WC0676

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

Method: 300.0 - Anions, Ion Chromatography

Lab Sample ID: MB 300-2283/5  
Matrix: Water  
Analysis Batch: 2283

Client Sample ID: Method Blank  
Prep Type: Total/NA

Analyte	MB	MB	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Nitrate as N	0.014	U	0.028	0.014	mg/L			12/08/15 12:13	1
Nitrite as N	0.019	U	0.038	0.019	mg/L			12/08/15 12:13	1

Lab Sample ID: LCS 300-2283/6  
Matrix: Water  
Analysis Batch: 2283

Client Sample ID: Lab Control Sample  
Prep Type: Total/NA

Analyte	Spike	LCS	LCS	Unit	D	%Rec	%Rec.	Limits
Nitrate as N	2.26	2.36		mg/L		104	80 - 120	
Nitrite as N	3.04	3.19		mg/L		105	80 - 120	

Lab Sample ID: 300-1760-A-1 MS  
Matrix: Water  
Analysis Batch: 2283

Client Sample ID: Matrix Spike  
Prep Type: Total/NA

Analyte	Sample	Sample	Spike	MS	MS	Unit	D	%Rec	%Rec.	Limits
Nitrite as N	0.038	U	0.609	0.568	D	mg/L		93	75 - 125	

Lab Sample ID: 300-1760-A-1 MS  
Matrix: Water  
Analysis Batch: 2283

Client Sample ID: Matrix Spike  
Prep Type: Total/NA

Analyte	Sample	Sample	Spike	MS	MS	Unit	D	%Rec	%Rec.	Limits
Nitrate as N	12	D	1.13	13.2	D	mg/L		80	75 - 125	

Lab Sample ID: 300-1760-A-1 DU  
Matrix: Water  
Analysis Batch: 2283

Client Sample ID: Duplicate  
Prep Type: Total/NA

Analyte	Sample	Sample	DU	DU	Unit	D	RPD	RPD	Limit
Nitrite as N	0.038	U	0.038	U	mg/L		NC	20	

Lab Sample ID: 300-1760-A-1 DU  
Matrix: Water  
Analysis Batch: 2283

Client Sample ID: Duplicate  
Prep Type: Total/NA

Analyte	Sample	Sample	DU	DU	Unit	D	RPD	RPD	Limit
Nitrate as N	12	D	12.2	D	mg/L		0.7	20	

Lab Sample ID: MB 300-2284/5  
Matrix: Water  
Analysis Batch: 2284

Client Sample ID: Method Blank  
Prep Type: Total/NA

Analyte	MB	MB	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Chloride	0.10	U	0.20	0.10	mg/L			12/08/15 12:13	1
Fluoride	0.025	U	0.050	0.025	mg/L			12/08/15 12:13	1
Sulfate	0.13	U	0.25	0.13	mg/L			12/08/15 12:13	1

QC Sample Results

Client: CH2M Hill Plateau Remediation Company  
Project/Site: X16-001

TestAmerica Job ID: 300-1771-1  
SDG: WC0676

Method: 300.0 - Anions, Ion Chromatography (Continued)

Lab Sample ID: LCS 300-2284/6  
Matrix: Water  
Analysis Batch: 2284

Client Sample ID: Lab Control Sample  
Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
Chloride	16.0	16.5		mg/L		103	80 - 120
Fluoride	4.00	4.14		mg/L		103	80 - 120
Sulfate	20.0	20.7		mg/L		104	80 - 120

Lab Sample ID: 300-1760-A-1 MS  
Matrix: Water  
Analysis Batch: 2284

Client Sample ID: Matrix Spike  
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec. Limits
Chloride	14	D	3.20	17.3	D	mg/L		92	75 - 125
Fluoride	0.33	D	0.800	1.12	D	mg/L		100	75 - 125
Sulfate	39	D	4.00	41.9	D	mg/L		79	75 - 125

Lab Sample ID: 300-1760-A-1 DU  
Matrix: Water  
Analysis Batch: 2284

Client Sample ID: Duplicate  
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	DU Result	DU Qualifier	Unit	D	RPD	RPD Limit
Chloride	14	D	14.2	D	mg/L		1	20
Fluoride	0.33	D	0.321	D	mg/L		1	20
Sulfate	39	D	38.3	D	mg/L		1	20

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

**QC Association Summary**

Client: CH2M Hill Plateau Remediation Company  
 Project/Site: X16-001

TestAmerica Job ID: 300-1771-1  
 SDG: WC0676

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

**HPLC/IC**

**Analysis Batch: 2283**

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
300-1760-A-1 DU	Duplicate	Total/NA	Water	300.0	
300-1760-A-1 DU	Duplicate	Total/NA	Water	300.0	
300-1760-A-1 MS	Matrix Spike	Total/NA	Water	300.0	
300-1760-A-1 MS	Matrix Spike	Total/NA	Water	300.0	
300-1771-1	B32TB0	Total/NA	Water	300.0	
LCS 300-2283/6	Lab Control Sample	Total/NA	Water	300.0	
MB 300-2283/5	Method Blank	Total/NA	Water	300.0	

**Analysis Batch: 2284**

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
300-1760-A-1 DU	Duplicate	Total/NA	Water	300.0	
300-1760-A-1 MS	Matrix Spike	Total/NA	Water	300.0	
300-1771-1	B32TB0	Total/NA	Water	300.0	
LCS 300-2284/6	Lab Control Sample	Total/NA	Water	300.0	
MB 300-2284/5	Method Blank	Total/NA	Water	300.0	

Lab Chronicle

Client: CH2M Hill Plateau Remediation Company  
Project/Site: X16-001

TestAmerica Job ID: 300-1771-1  
SDG: WC0676

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

**Client Sample ID: B32TB0**

**Lab Sample ID: 300-1771-1**

Date Collected: 12/07/15 10:47

Matrix: Water

Date Received: 12/07/15 14:45

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	300.0		2	10 mL		2283	12/08/15 15:29	CPM	TAL RCH
Total/NA	Analysis	300.0		2	10 mL		2284	12/08/15 15:29	CPM	TAL RCH

**Client Sample ID: Duplicate**

**Lab Sample ID: 300-1760-A-1 DU**

Date Collected: 12/06/15 14:24

Matrix: Water

Date Received: 12/07/15 09:25

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	300.0		2	10 mL		2283	12/08/15 13:13	CPM	TAL RCH
Total/NA	Analysis	300.0		2	10 mL		2284	12/08/15 13:13	CPM	TAL RCH
Total/NA	Analysis	300.0		5	10 mL		2283	12/08/15 16:14	CPM	TAL RCH

**Client Sample ID: Lab Control Sample**

**Lab Sample ID: LCS 300-2283/6**

Date Collected: N/A

Matrix: Water

Date Received: N/A

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	300.0		1	10 mL		2283	12/08/15 12:28	CPM	TAL RCH

**Client Sample ID: Lab Control Sample**

**Lab Sample ID: LCS 300-2284/6**

Date Collected: N/A

Matrix: Water

Date Received: N/A

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	300.0		1	10 mL		2284	12/08/15 12:28	CPM	TAL RCH

**Client Sample ID: Method Blank**

**Lab Sample ID: MB 300-2283/5**

Date Collected: N/A

Matrix: Water

Date Received: N/A

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	300.0		1	10 mL		2283	12/08/15 12:13	CPM	TAL RCH

**Client Sample ID: Method Blank**

**Lab Sample ID: MB 300-2284/5**

Date Collected: N/A

Matrix: Water

Date Received: N/A

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	300.0		1	10 mL		2284	12/08/15 12:13	CPM	TAL RCH

### Lab Chronicle

Client: CH2M Hill Plateau Remediation Company  
Project/Site: X16-001

TestAmerica Job ID: 300-1771-1  
SDG: WC0676

**Client Sample ID: Matrix Spike**

**Lab Sample ID: 300-1760-A-1 MS**

**Date Collected: 12/06/15 14:24**

**Matrix: Water**

**Date Received: 12/07/15 09:25**

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	300.0		2	10 mL		2283	12/08/15 12:58	CPM	TAL RCH
Total/NA	Analysis	300.0		2	10 mL		2284	12/08/15 12:58	CPM	TAL RCH
Total/NA	Analysis	300.0		5	10 mL		2283	12/08/15 15:59	CPM	TAL RCH

**Laboratory References:**

TAL RCH = TestAmerica Richland, 2800 George Washington Way, Richland, WA 99352, TEL (509)375-3131



### Method Summary

Client: CH2M Hill Plateau Remediation Company  
Project/Site: X16-001

TestAmerica Job ID: 300-1771-1  
SDG: WC0676



Method	Method Description	Protocol	Laboratory
300.0	Anions, Ion Chromatography	MCAWW	TAL RCH

**Protocol References:**

MCAWW = "Methods For Chemical Analysis Of Water And Wastes", EPA-600/4-79-020, March 1983 And Subsequent Revisions.

**Laboratory References:**

TAL RCH = TestAmerica Richland, 2800 George Washington Way, Richland, WA 99352, TEL (509)375-3131

### Sample Summary

Client: CH2M Hill Plateau Remediation Company  
Project/Site: X16-001

TestAmerica Job ID: 300-1771-1  
SDG: WC0676

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

---

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
300-1771-1	B32TB0	Water	12/07/15 10:47	12/07/15 14:45

**Login Sample Receipt Checklist**

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

Client: CH2M Hill Plateau Remediation Company

Job Number: 300-1771-1  
SDG Number: WC0676

**Login Number: 1771**  
**List Number: 1**  
**Creator: Friesz, Jordan D**

**List Source: TestAmerica Richland**

Question	Answer	Comment
Radioactivity wasn't checked or is <math>\leq</math> background as measured by a survey meter.	True	
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time.	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <math><6\text{mm}</math> (1/4").	N/A	
Multiphasic samples are not present.	N/A	
Samples do not require splitting or compositing.	N/A	
Residual Chlorine Checked.	N/A	