

October 26, 2015

TestAmerica

THE LEADER IN ENVIRONMENTAL TESTING

ANALYTICAL REPORT

TestAmerica Laboratories, Inc.
TestAmerica Richland
2800 George Washington Way
Richland, WA 99352
Tel: (509)375-3131

TestAmerica Job ID: 300-1106-1
TestAmerica Sample Delivery Group: WC0452
Client Project/Site: CH2M Hill Chemical

For:
CH2M Hill Plateau Remediation Company
PO BOX 1600
Mail Stop R3-50
Richland, Washington 99352

Attn: CPP Sample Management



Authorized for release by:
10/26/2015 12:29:21 PM
Steven Campbell, Quality Assurance Assistant
steven.campbell@testamericainc.com

Designee for
Whitney Ritari, Project Manager I
(509)375-3131 ext164
whitney.ritari@testamericainc.com

I certify that this Certificate of Analysis is in compliance with the SOW, both technically and for completeness, for other than the conditions detailed in this package. Release of the data contained in this hard copy data package has been authorized by the Laboratory Manager, or a designee as verified by the attached signature.

The laboratory is accredited to EPA 7196A in Solid and Chemical Materials, per Washington State Department of Ecology this is approved for non-discharge water samples. The sample/ sample duplicate RPD agreement is not applicable if one or both results are elss than five times the MDL.



LINKS

Review your project
results through
TotalAccess

Have a Question?



Visit us at:
www.testamericainc.com



Table of Contents

Cover Page	1
Table of Contents	2
Definitions/Glossary	3
Case Narrative	4
Chain of Custody	5
Detection Summary	6
Client Sample Results	7
QC Sample Results	8
QC Association Summary	10
Lab Chronicle	11
Method Summary	13
Sample Summary	14
Receipt Checklists	15

Qualifiers

HPLC/IC

Qualifier	Qualifier Description
U	Analyzed for but not detected.
D	The reported value is from a dilution.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CNF	Contains no Free Liquid
DER	Duplicate error ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision level concentration
MDA	Minimum detectable activity
EDL	Estimated Detection Limit
MDC	Minimum detectable concentration
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
NC	Not Calculated
ND	Not detected at the reporting limit (or MDL or EDL if shown)
PQL	Practical Quantitation Limit
QC	Quality Control
RER	Relative error ratio
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)

Client: CH2M Hill Plateau Remediation Company
Project/Site: CH2M Hill Chemical

TestAmerica Job ID: 300-1106-1
SDG: WC0452

Job ID: 300-1106-1

Laboratory: TestAmerica Richland

Narrative

**Job Narrative
300-1106-1**

Comments

No additional comments.

Receipt

The sample was received on 10/15/2015 10:00 AM; the sample arrived in good condition, properly preserved and, where required, on ice. The temperature of the cooler at receipt was 8.0° C.

HPLC/IC

No analytical or quality issues were noted, other than those described in the Definitions/Glossary page.

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

CH2M Hill Plateau Remediation Company

CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST

C.O.C. # **X16-001-173**
Page 1 of 1

Collector	S.W. King/CHPRC	Contact/Requester	Karen Waters-Husted	Telephone No.	509-376-4650
SAF No.	X16-001	Sampling Origin	Hanford Site	Purchase Order/Charge Code	303269
Project Title	100-BC-5 RI, OCTOBER 2015	Logbook No.	HNF-N-50678/55	Ice Chest No.	N/A
Shipped To (Lab)	TestAmerica Incorporated, Richland	Method of Shipment	GOVERNMENT VEHICLE	Bill of Lading/Air Bill No.	N/A
Protocol	CERCLA	Priority:	30 Days	Offsite Property No.	N/A
POSSIBLE SAMPLE HAZARDS/REMARKS		SPECIAL INSTRUCTIONS		Total Activity Exemption: Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>	
*** Contains Radioactive Material at concentrations that are not regulated for transportation per 49 CFR / IATA Dangerous Goods Regulations but are not releasable per DOE Order 458.1		Submitt deliverables & invoices to CFP Sample Management			
Sample No.	B32V95	Filter	N	Date	OCT 14 2015 1200
		* Date		Time	
		No/Type Container	1x60-mL P	Sample Analysis	300.0 ANIONS_IC: COMMON
		Holding Time	48 Hours	Preservative	Cool <=6C

October 26, 2015



300-1106 COC

W 0845Z
1106

Relinquished By	Print	Signature	Date/Time	Received By	Print	Signature	Date/Time	Matrix *
Relinquished By			OCT 14 2015 1505	Received By			OCT 14 2015 1505	S = Soil SE = Sediment SO = Solid SL = Sludge W = Water O = Oil A = Air
Relinquished By			OCT 15 2015 0730	Received By			OCT 15 2015 0730	DS = Dura Solids DL = Dura Liquids T = Tissue WI = Wipe L = Liquid V = Vegetation X = Other
Relinquished By			OCT 15 2015 1000	Received By			OCT 15 2015 1000	
FINAL SAMPLE DISPOSITION	Disposal Method (e.g., Return to customer, per lab procedure, used in process)			Disposed By			Date/Time	

October 26, 2015
Detection Summary

Client: CH2M Hill Plateau Remediation Company
Project/Site: CH2M Hill Chemical

TestAmerica Job ID: 300-1106-1
SDG: WC0452

Client Sample ID: B32V95

Lab Sample ID: 300-1106-1

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Chloride	8.1	D	0.40	0.20	mg/L	2		300.0	Total/NA
Nitrate as N	2.3	D	0.056	0.028	mg/L	2		300.0	Total/NA
Fluoride	0.14	D	0.10	0.050	mg/L	2		300.0	Total/NA
Sulfate	40	D	0.50	0.25	mg/L	2		300.0	Total/NA

This Detection Summary does not include radiochemical test results.

TestAmerica Richland

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

October 26, 2015
Client Sample Results

Client: CH2M Hill Plateau Remediation Company
Project/Site: CH2M Hill Chemical

TestAmerica Job ID: 300-1106-1
SDG: WC0452

Client Sample ID: B32V95
Date Collected: 10/14/15 12:00
Date Received: 10/15/15 10:00

Lab Sample ID: 300-1106-1
Matrix: Water

Method: 300.0 - Anions, Ion Chromatography

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Chloride	8.1	D	0.40	0.20	mg/L			10/15/15 18:56	2
Nitrate as N	2.3	D	0.056	0.028	mg/L			10/15/15 18:56	2
Fluoride	0.14	D	0.10	0.050	mg/L			10/15/15 18:56	2
Nitrite as N	0.038	U	0.076	0.038	mg/L			10/15/15 18:56	2
Sulfate	40	D	0.50	0.25	mg/L			10/15/15 18:56	2



~~October 26, 2015~~
QC Sample Results

Client: CH2M Hill Plateau Remediation Company
Project/Site: CH2M Hill Chemical

TestAmerica Job ID: 300-1106-1
SDG: WC0452

Method: 300.0 - Anions, Ion Chromatography

Lab Sample ID: MB 300-1495/20
Matrix: Water
Analysis Batch: 1495

Client Sample ID: Method Blank
Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Nitrate as N	0.014	U	0.028	0.014	mg/L			10/15/15 15:41	1
Nitrite as N	0.019	U	0.038	0.019	mg/L			10/15/15 15:41	1

Lab Sample ID: LCS 300-1495/21
Matrix: Water
Analysis Batch: 1495

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
Nitrate as N	2.26	2.32		mg/L		103	80 - 120
Nitrite as N	3.04	3.13		mg/L		103	80 - 120

Lab Sample ID: 300-1101-A-2 MS
Matrix: Water
Analysis Batch: 1495

Client Sample ID: Matrix Spike
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec. Limits
Nitrite as N	0.038	U	0.609	0.531	D	mg/L		87	75 - 125

Lab Sample ID: 300-1101-A-2 MS
Matrix: Water
Analysis Batch: 1495

Client Sample ID: Matrix Spike
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec. Limits
Nitrate as N	32	D	4.52	36.9	D	mg/L		100	75 - 125

Lab Sample ID: 300-1101-A-2 DU
Matrix: Water
Analysis Batch: 1495

Client Sample ID: Duplicate
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	DU Result	DU Qualifier	Unit	D	RPD	RPD Limit
Nitrite as N	0.038	U	0.038	U	mg/L		NC	20

Lab Sample ID: 300-1101-A-2 DU
Matrix: Water
Analysis Batch: 1495

Client Sample ID: Duplicate
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	DU Result	DU Qualifier	Unit	D	RPD	RPD Limit
Nitrate as N	32	D	32.3	D	mg/L		0.4	20

Lab Sample ID: MB 300-1496/20
Matrix: Water
Analysis Batch: 1496

Client Sample ID: Method Blank
Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Chloride	0.10	U	0.20	0.10	mg/L			10/15/15 15:41	1
Fluoride	0.025	U	0.050	0.025	mg/L			10/15/15 15:41	1
Sulfate	0.13	U	0.25	0.13	mg/L			10/15/15 15:41	1

TestAmerica Richland

Client: CH2M Hill Plateau Remediation Company
Project/Site: CH2M Hill Chemical

TestAmerica Job ID: 300-1106-1
SDG: WC0452

Method: 300.0 - Anions, Ion Chromatography (Continued)

Lab Sample ID: LCS 300-1496/21
Matrix: Water
Analysis Batch: 1496

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
Chloride	16.0	16.2		mg/L		101	80 - 120
Fluoride	4.00	4.09		mg/L		102	80 - 120
Sulfate	20.0	20.4		mg/L		102	80 - 120

Lab Sample ID: 300-1101-A-2 MS
Matrix: Water
Analysis Batch: 1496

Client Sample ID: Matrix Spike
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec. Limits
Chloride	24	D	3.20	26.8	D	mg/L		96	75 - 125
Fluoride	0.36	D	0.800	1.15	D	mg/L		98	75 - 125
Sulfate	55	D	4.00	58.5	D	mg/L		99	75 - 125

Lab Sample ID: 300-1101-A-2 DU
Matrix: Water
Analysis Batch: 1496

Client Sample ID: Duplicate
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	DU Result	DU Qualifier	Unit	D	RPD	RPD Limit
Chloride	24	D	23.7	D	mg/L		0.3	20
Fluoride	0.36	D	0.362	D	mg/L		0.1	20
Sulfate	55	D	54.4	D	mg/L		0.3	20

October 26, 2015
QC Association Summary

Client: CH2M Hill Plateau Remediation Company
Project/Site: CH2M Hill Chemical

TestAmerica Job ID: 300-1106-1
SDG: WC0452

HPLC/IC

Analysis Batch: 1495

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
300-1101-A-2 DU	Duplicate	Total/NA	Water	300.0	
300-1101-A-2 DU	Duplicate	Total/NA	Water	300.0	
300-1101-A-2 MS	Matrix Spike	Total/NA	Water	300.0	
300-1101-A-2 MS	Matrix Spike	Total/NA	Water	300.0	
300-1106-1	B32V95	Total/NA	Water	300.0	
LCS 300-1495/21	Lab Control Sample	Total/NA	Water	300.0	
MB 300-1495/20	Method Blank	Total/NA	Water	300.0	

Analysis Batch: 1496

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
300-1101-A-2 DU	Duplicate	Total/NA	Water	300.0	
300-1101-A-2 MS	Matrix Spike	Total/NA	Water	300.0	
300-1106-1	B32V95	Total/NA	Water	300.0	
LCS 300-1496/21	Lab Control Sample	Total/NA	Water	300.0	
MB 300-1496/20	Method Blank	Total/NA	Water	300.0	

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

Client: CH2M Hill Plateau Remediation Company
 Project/Site: CH2M Hill Chemical

TestAmerica Job ID: 300-1106-1
 SDG: WC0452

Client Sample ID: B32V95
Date Collected: 10/14/15 12:00
Date Received: 10/15/15 10:00

Lab Sample ID: 300-1106-1
Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	300.0		2	10 mL		1495	10/15/15 18:56	CPM	TAL RCH
Total/NA	Analysis	300.0		2	10 mL		1496	10/15/15 18:56	CPM	TAL RCH

Client Sample ID: Duplicate
Date Collected: 10/14/15 14:00
Date Received: 10/15/15 10:00

Lab Sample ID: 300-1101-A-2 DU
Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	300.0		2	10 mL		1495	10/15/15 16:41	CPM	TAL RCH
Total/NA	Analysis	300.0		2	10 mL		1496	10/15/15 16:41	CPM	TAL RCH
Total/NA	Analysis	300.0		20	10 mL		1495	10/15/15 20:12	CPM	TAL RCH

Client Sample ID: Lab Control Sample
Date Collected: N/A
Date Received: N/A

Lab Sample ID: LCS 300-1495/21
Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	300.0		1	10 mL		1495	10/15/15 15:56	CPM	TAL RCH

Client Sample ID: Lab Control Sample
Date Collected: N/A
Date Received: N/A

Lab Sample ID: LCS 300-1496/21
Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	300.0		1	10 mL		1496	10/15/15 15:56	CPM	TAL RCH

Client Sample ID: Method Blank
Date Collected: N/A
Date Received: N/A

Lab Sample ID: MB 300-1495/20
Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	300.0		1	10 mL		1495	10/15/15 15:41	CPM	TAL RCH

Client Sample ID: Method Blank
Date Collected: N/A
Date Received: N/A

Lab Sample ID: MB 300-1496/20
Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	300.0		1	10 mL		1496	10/15/15 15:41	CPM	TAL RCH

Client: CH2M Hill Plateau Remediation Company
 Project/Site: CH2M Hill Chemical

TestAmerica Job ID: 300-1106-1
 SDG: WC0452

Client Sample ID: Matrix Spike

Lab Sample ID: 300-1101-A-2 MS

Date Collected: 10/14/15 14:00

Matrix: Water

Date Received: 10/15/15 10:00

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	300.0		2	10 mL		1495	10/15/15 16:26	CPM	TAL RCH
Total/NA	Analysis	300.0		2	10 mL		1496	10/15/15 16:26	CPM	TAL RCH
Total/NA	Analysis	300.0		20	10 mL		1495	10/15/15 19:57	CPM	TAL RCH

Laboratory References:

TAL RCH = TestAmerica Richland, 2800 George Washington Way, Richland, WA 99352, TEL (509)375-3131

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

Client: CH2M Hill Plateau Remediation Company
Project/Site: CH2M Hill Chemical

TestAmerica Job ID: 300-1106-1
SDG: WC0452

Method	Method Description	Protocol	Laboratory
300.0	Anions, Ion Chromatography	MCAWW	TAL RCH

Protocol References:

MCAWW = "Methods For Chemical Analysis Of Water And Wastes", EPA-600/4-79-020, March 1983 And Subsequent Revisions.

Laboratory References:

TAL RCH = TestAmerica Richland, 2800 George Washington Way, Richland, WA 99352, TEL (509)375-3131



Client: CH2M Hill Plateau Remediation Company
Project/Site: CH2M Hill Chemical

TestAmerica Job ID: 300-1106-1
SDG: WC0452

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
300-1106-1	B32V95	Water	10/14/15 12:00	10/15/15 10:00

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

Login Sample Receipt Checklist

Client: CH2M Hill Plateau Remediation Company

Job Number: 300-1106-1
SDG Number: WC0452

Login Number: 1106
List Number: 1
Creator: Friesz, Jordan D

List Source: TestAmerica Richland

Question	Answer	Comment
Radioactivity wasn't checked or is <=/ background as measured by a survey meter.	True	
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time.	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	N/A	
Multiphasic samples are not present.	N/A	
Samples do not require splitting or compositing.	N/A	
Residual Chlorine Checked.	N/A	

