0078789

SAF-RC-052
Remaining Sites Confirmation Sampling - Water
FINAL DATA PACKAGE

COMPLETE COPY OF DATA PACKAGE TO:

Kathy Wendt H4-21 KW 9/2208
INITIAL/DATE

COMMENTS:

SDG J00196 **SAF-RC-052**

Rad only X Chem only Rad & Chem

X Complete Partial

Waste Site: 100-D-78

RECEIVED
SEP 29 2008
EDMC

RECEIVED
SEP 2008

Analytical Data Package Prepared For

Washington Closure Hanford

Radiochemical Analysis By

TestAmerica

2800 G.W. Way, Richland Wa, 99354, (509)-375-3131.

Assigned Laboratory Code: TARL

Data Package Contains 21 Pages

Report No.: 39876

Results in this report relate only to the sample(s) analyzed.

SDG No.	Order No.	Client Sample ID (List Order)	Lot-Sa No.	Work Order	Report DB ID	Batch No.
J00196	RC-052	J17J49	J8I180139-1	KW4MM1AA	9KW4MM10	8262284

TestAmerica

THE LEADER IN ENVIRONMENTAL TESTING

Certificate of Analysis

Washington Hanford Closure
2620 Fermi Avenue
Richland, WA 99354

September 19, 2008

Attention: Joan Kessner

SAF Number	:	RC-052
Date SDG Closed	:	September 17, 2008
Number of Samples	:	One (1)
Sample Type	:	Water
SDG Number	:	J00196
Data Deliverable	:	15-Day / Summary

CASE NARRATIVE

I. Introduction

On September 17, 2008 one water sample was received at TestAmerica for chemistry analysis. Upon receipt, the sample was assigned the following laboratory ID number to correspond with the Washington Closure Hanford (WCH) specific ID:

WCH ID#	TALR ID#	MATRIX	DATE OF RECEIPT
J17J49	KW4MM	WATER	9/17/08

II. Sample Receipt

The sample was received in good condition and no anomalies were noted during check-in.

III. Analytical Results/Methodology

The analytical results for this report are presented by laboratory sample ID. Each set of data includes sample identification information, analytical results and the appropriate associated statistical errors. The requested analyses were:

Chemical Analysis
Hexavalent Chromium by EPA method 7196A

IV. Quality Control

The analytical results for each analysis performed includes a minimum of one laboratory control sample (LCS), one method (reagent) blank, and one duplicate sample analysis. Any exceptions have been noted in the "Comments" section.

QC and sample results are reported in the same units.

V. Comments

Chemical Analysis

Hexavalent Chromium by EPA method 7196A:

The matrix spike and matrix spike duplicate recovered low. When colorizing the MS and MSD, there were no changes in the color; which suggests that the sample is acting as a reducing media for the Cr+6. Data will be accepted. Except as noted, the LCS, batch blank, sample, sample duplicate (J17J49), sample matrix spike (J17J49) and sample matrix spike duplicate (J17J49) results are within contractual requirements.

I certify that this Certificate of Analysis is in compliance with the SOW, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the Laboratory Manager, or a designee as verified by the following signature.

Reviewed and approved:

for Sandra Seger
Project Manager

Drinking Water Method Cross References

	DRINKING WATER ASTM METHOD CROSS REFERENCES	
Referenced Method	**Isotope(s)**	**TestAmerica Richland's SOP No.**
EPA 901.1	Cs-134, I-131	RICH-RC-5017
EPA 900.0	Alpha & Beta	RICH-RC-5014
EPA 00-02	Gross Alpha (Coprecipitation)	RICH-RC-5021
EPA 903.0	Total Alpha Radium (Ra-226)	RICH-RC-5027
EPA 903.1	Ra-226	RICH-RC-5005
EPA 904.0	Ra-228	RICH-RC-5005
EPA 905.0	Sr-89/90	RICH-RC-5006
ASTM D5174	Uranium	RICH-RC-5058
EPA 906.0	Tritium	RICH-RC-5007

Results in this report relate only to the sample(s) analyzed.

Uncertainty Estimation

TestAmerica Richland has adopted the internationally accepted approach to estimating uncertainties described in "NIST Technical Note 1297, 1994 Edition". The approach, "Law of Propagation of Errors", involves the identification of all variables in an analytical method which are used to derive a result. These variables are related to the analytical result (R) by some functional relationship, R = constants * f(x,y,z,...). The components (x,y,z) are evaluated to determine their contribution to the overall method uncertainty. The individual component uncertainties (u_i) are then combined using a statistical model that provides the most probable overall uncertainty value. All component uncertainties are categorized as type A, evaluated by statistical methods, or type B, evaluated by other means. Uncertainties not included in the components, such as sample homogeneity, are combined with the component uncertainty as the square root of the sum-of-the-squares of the individual uncertainties. The uncertainty associated with the derived result is the combined uncertainty (u_c) multiplied by the coverage factor (1,2, or 3).

When three or more sample replicates are used to derive the analytical result, the type A uncertainty is the standard deviation of the mean value (S/?n), where S is the standard deviation of the derived results. The type B uncertainties are all other random or non-random components that are not included in the standard deviation.

The derivation of the general "Law of Propagation of Errors" equations and specific example are available on request.

Report Definitions

Action Lev	An agreed upon activity level used to trigger some action when the final result is greater than or equal to the Action Level. Often the Action Level is related to the Decision Limit.
Batch	The QC preparation batch number that relates laboratory samples to QC samples that were prepared and analyzed together.
Bias	Defined by the equation (Result/Expected)-1 as defined by ANSI N13.30.
COC No	Chain of Custody Number assigned by the Client or TestAmerica.
Count Error (#s)	Poisson counting statistics of the gross sample count and background. The uncertainty is absolute and in the same units as the result. For Liquid Scintillation Counting (LSC) the batch blank count is the background.
Total Uncert (#s) ***u_c - Combined Uncertainty.***	All known uncertainties associated with the preparation and analysis of the sample are propagated to give a measure of the uncertainty associated with the result, ***u_c the combined uncertainty.*** The uncertainty is absolute and in the same units as the result.
(#s), Coverage Factor	The coverage factor defines the width of the confidence interval, 1, 2 or 3 standard deviations.
CRDL (RL)	Contractual Required Detection Limit as defined in the Client's Statement Of Work or TestAmerica "default" nominal detection limit. Often referred to the reporting level (RL)
Lc	Decision Level based on instrument background or blank, adjusted by the Efficiency, Chemical Yield, and Volume associated with the sample. The Type I error probability is approximately 5%. Lc=(1.645 * Sqrt(2*(BkgrndCnt/BkgrndCntMin)/SCntMin)) * (ConvFct/(Eff*Yld*Abn*Vol) * IngrFct). For LSC methods the batch blank is used as a measure of the background variability. Lc cannot be calculated when the background count is zero.
Lot-Sample No	The number assigned by the LIMS software to track samples received on the same day for a given client. The sample number is a sequential number assigned to each sample in the Lot.
MDC|MDA	Detection Level based on instrument background or blank, adjusted by the Efficiency, Chemical Yield, and Volume with a Type I and II error probability of approximately 5%. MDC = (4.65 * Sqrt((BkgrndCnt/BkgrndCntMin)/SCntMin) + 2.71/SCntMin) * (ConvFct/(Eff * Yld * Abn * Vol) * IngrFct). For LSC methods the batch blank is used as a measure of the background variability.
Primary Detector	The instrument identifier associated with the analysis of the sample aliquot.
Ratio U-234/U-238	The U-234 result divided by the U-238 result. The U-234/U-238 ratio for natural uranium in NIST SRM 4321C is 1.038.
Rst/MDC	Ratio of the Result to the MDC. A value greater than 1 may indicate activity above background at a high level of confidence. Caution should be used when applying this factor and it should be used in concert with the qualifiers associated with the result.
Rst/TotUcert	Ratio of the Result to the Total Uncertainty. If the uncertainty has a coverage factor of 2 a value greater than 1 may indicate activity above background at approximately the 95% level of confidence assuming a two-sided confidence interval. Caution should be used when applying this factor and it should be used in concert with the qualifiers associated with the result.
Report DB No	Sample Identifier used by the report system. The number is based upon the first five digits of the **Work Order** Number.
RER	The equation Replicate Error Ratio = (S-D)/[sqrt($TPUs^2$ + $TPUd^2$)] as defined by ICPT BOA where S is the original sample result, D is the result of the duplicate, TPUs is the total uncertainty of the original sample and TPUd is the total uncertainty of the duplicate sample.
SDG	Sample Delivery Group Number assigned by the Client or assigned by TestAmerica upon sample receipt.
Sum Rpt Alpha Spec Rst(s)	The sum of the reported alpha spec results for tests derived from the same sample excluding duplicate result where the results are in the same units.
Work Order	The LIMS software assign test specific identifier.
Yield	The recovery of the tracer added to the sample such as Pu-242 used to trace a Pu-239/40 method.

Sample Results Summary

Date: 19-Sep-08

TestAmerica TARL

Ordered by Method, Batch No., Client Sample ID.

Report No. : 39876

SDG No: J00196

Batch	Client Id Work Order	Parameter	Result +- Uncertainty (2s)	Qual	Units	Tracer Yield	MDC or MDA	CRDL	RPD
8262284	7196_CR6								
	J17J49								
	KW4MM1AA	HEXCHROME	2.70E-02 +- 0.00E+00		mg/L	N/A	2.00E-03	3.50E-01	
	KW4MM1AE	HEXCHROME	2.50E-02 +- 0.00E+00		mg/L	N/A	2.00E-03	3.50E-01	7.7

No. of Results: 2

TestAmerica **RPD** - Relative Percent Difference.

rptSTLRchSaSum
mary2 V5.1.7
A2002

QC Results Summary

Date: 19-Sep-08

TestAmerica TARL

Ordered by Method, Batch No, QC Type,.

Report No. : 39876

SDG No.: J00196

Batch Work Order	Parameter	Result +- Uncertainty (2s)	Qual	Units	Tracer Yield	LCS Recovery	Bias	MDC|MDA
7196_CR6								
8262284 MATRIX SPIKE, J17J49								
KW4MM1AC	HEXCHROME	1.20E-02 +- 0.00E+00		mg/L	N/A	5%	-1.0	2.00E-03
KW4MM1AD	HEXCHROME	4.70E-02 +- 0.00E+00		mg/L	N/A	18%	-0.8	2.00E-03
8262284 LCS,								
KW41X1AC	HEXCHROME	5.03E-01 +- 0.00E+00		mg/L	N/A	101%	0.0	2.00E-03
8262284 BLANK QC,								
KW41X1AA	HEXCHROME	2.00E-03 +- 0.00E+00	U	mg/L	N/A			2.00E-03

No. of Results: 4

TestAmerica

rptSTLRchQcSummary V5.1.7 A2002

Bias - (Result/Expected)-1 as defined by ANSI N13.30.

U Qual - Analyzed for but not detected above limiting criteria. Limit criteria is less than the Mdc/Mda or Total Uncert or not identified by gamma scan software.

FORM I

SAMPLE RESULTS

Date: 19-Sep-08

Lab Name:	TestAmerica	SDG:	J00196	Collection Date:	9/17/2008 9:00:00 AM
Lot-Sample No.:	J8I180139-1	Report No. :	39876	Received Date:	9/17/2008 12:35:00 PM
Client Sample ID:	J17J49	COC No. :	RC-052-009	Matrix:	WATER

Ordered by Client Sample ID, Batch No.

Parameter	Result	Qual	Count Error (2 s)	Total Uncert(2 s)	MDC|MDA, Action Lev	Rpt Unit, Lc	Yield CRDL(RL)	Rst/MDC, Rst/TotUcert	Analysis, Prep Date	Total Sa Size	Aliquot Size	Primary Detector
Batch: 8262284	7196_CR6			Work Order:	KW4MM1AA		Report DB ID:	9KW4MM10				
HEXCHROME	*2.70E-02*			0.0E+00	2.00E-03	mg/L	N/A	*(13.5)*	9/17/08		100.0	
							3.50E-01	*N/A*			ML	

No. of Results: 1 Comments:

TestAmerica
rptSTLRchSample
V5.1.7 A2002

MDC|MDA,Lc - Detection, Decision Level based on instrument background or blank, adjusted by the sample Efficiency, Yield, and Volume.
U Qual - Analyzed for but not detected above limiting criteria. Limit criteria is less than the Mdc/Mda or Total Uncert or not identified by gamma scan software.

FORM II

Date: 19-Sep-08

DUPLICATE RESULTS

Lab Name:	TestAmerica	**SDG:**	J00196	**Collection Date:**	9/17/2008 9:00:00 AM
Lot-Sample No.:	J8I180139-1	**Report No. :**	39876	**Received Date:**	9/17/2008 12:35:00 PM
Client Sample ID:	J17J49	**COC No. :**	RC-052-009	**Matrix:**	WATER

Parameter	Result, Orig Rst	Qual	Count Error (2 s)	Total Uncert(2 s)	MDC|MDA, Action Lev	Rpt Unit, CRDL	Yield	Rst/MDC, Rst/TotUcert	Analysis, Prep Date	Total Sa Size	Aliquot Size	Primary Detector
Batch: 8262284	7196_CR6			**Work Order:** KW4MM1AE		**Report DB ID:** KW4MM1ER			**Orig Sa DB ID:** 9KW4MM10			
HEXCHROME	2.50E-02			0.0E+00	2.00E-03	mg/L	N/A	(12.5)	9/17/08		100.0	
	2.70E-02		RPD 7.7			3.50E-01		N/A			ML	

No. of Results: 1 **Comments:**

TestAmerica RPD - Relative Percent Difference.

rptSTLRchDupV5.1 .7 A2002 MDC|MDA,Lc - Detection, Decision Level based on instrument background or blank, adjusted by the sample Efficiency, Yield, and Volume.

FORM II

BLANK RESULTS

Date: 19-Sep-08

Lab Name: TestAmerica

Matrix: WATER

SDG: J00196

Report No. : 39876

Parameter	Result	Qual	Count Error (2 s)	Total Uncert(2 s)	MDC|MDA ,	Rpt Unit, CRDL	Yield	Rst/MDC, Rst/TotUcert	Analysis, Prep Date	Total Sa Size	Aliquot Size	Primary Detector
Batch: 8262284	7196_CR6			Work Order:	KW41X1AA		Report DB ID:	KW41X1AB				
HEXCHROME	2.00E-03	U		0.0E+00	2.00E-03	mg/L	N/A	1.	9/17/08		100.0	
						3.50E-01		N/A			ML	

No. of Results: 1 Comments:

TestAmerica
rptSTLRchBlank
V5.1.7 A2002

MDC|MDA,Lc - Detection, Decision Level based on instrument background or blank, adjusted by the sample Efficiency, Yield, and Volume.
U Qual - Analyzed for but not detected above limiting criteria. Limit criteria is less than the Mdc/Mda or Total Uncert or not identified by gamma scan software.

FORM II

LCS RESULTS

Date: 19-Sep-08

Lab Name: TestAmerica

Matrix: WATER

SDG: J00196

Report No. : 39876

Parameter	Result	Qual	Count Error (2 s)	Total Uncert(2 s)	MDC|MDA	Report Unit	Yield	Expected	Expected Uncert	Recovery, Bias	Analysis, Prep Date	Aliquot Size	Primary Detector
Batch: 8262284	7196_CR6			**Work Order:** KW41X1AC			**Report DB ID:** KW41X1AS						
HEXCHROME	5.03E-01			0.0E+00	2.00E-03	mg/L	N/A	5.00E-01		101%	9/17/08	100.0	
							Rec Limits:	70	130	0.0		ML	

No. of Results: 1 **Comments:**

TestAmerica **Bias** - (Result/Expected)-1 as defined by ANSI N13.30.

rptSTLRchLcs
V5.1.7 A2002

FORM II

MATRIX SPIKE RESULTS

Date: 19-Sep-08

Lab Name: TestAmerica **SDG:** J00196

Lot-Sample No.: J8I180139-1, J17J49 **Report No. :** 39876 **Matrix:** WATER

Parameter	SpikeResult, Orig Rst	Qual	Count Error (2 s)	Total Uncert(2 s)	MDC|MDA	Rpt Unit, CRDL	Yield	Rec-overy	Expected, Uncert	Analysis, Prep Date	Aliquot Size	Analy Method, Primary Detector
Batch: 8262284	Work Order: KW4MM1AC			Report DB ID: KW4MM1CW			Orig Sa DB ID: 9KW4MM10					
HEXCHROME	1.20E-02			0.0E+00	2.00E-03	mg/L	N/A	4.56%	2.63E-01	9/17/08	100.0 ML	7196_CR6
	2.70E-02											
Batch: 8262284	Work Order: KW4MM1AD			Report DB ID: KW4MM1DW			Orig Sa DB ID: KW4MM1CW					
HEXCHROME	4.70E-02			0.0E+00	2.00E-03	mg/L	N/A	17.87%	2.63E-01	9/17/08	100.0 ML	7196_CR6
	1.20E-02											

Number of Results: 2

Comments:

RER - Replicate Error Ratio = (S-D)/[sqrt(sq(TPUs)+sq(TPUd))] as defined by ICPT BOA.

Bias - (Result/Expected)-1 as defined by ANSI N13.30.

FORM II

Date: 19-Sep-08

MATRIX SPIKE DUPLICATE RESULTS

Lab Name: TestAmerica **SDG:** J00196

Lot-Sample No.: J8I180139-1, J17J49 **Report No. :** 39876 **Matrix:** WATER

Parameter	SpikeResult, Orig Rst	Qual	Count Error (2 s)	Total Uncert(2 s)	MDC|MDA	Rpt Unit, CRDL	Yield	Rec-overy	Expected, Uncert	Analysis, Prep Date	Aliquot Size	Primary Detector
Batch: 8262284	7196_CR6			Work Order:	KW4MM1AC		Report DB ID:	KW4MM1CW		Orig Sa DB ID:	KW4MM1DW	
HEXCHROME	1.20E-02			0.0E+00	2.00E-03	mg/L	N/A	4.56%	2.63E-01	9/17/08	100.0	
	4.70E-02	*RPD*	118.6								ML	
Batch: 8262284	7196_CR6			Work Order:	KW4MM1AD		Report DB ID:	KW4MM1DW		Orig Sa DB ID:	KW4MM1CW	
HEXCHROME	4.70E-02			0.0E+00	2.00E-03	mg/L	N/A	17.87%	2.63E-01	9/17/08	100.0	
	1.20E-02	*RPD*	118.6								ML	

No. of Results: 2 **Comments:**

TestAmerica RER - Replicate Error Ratio = (S-D)/[sqrt(sq(TPUs)+sq(TPUd))] as defined by ICPT BOA.
rptSTLRchMsDup2 Bias - (Result/Expected)-1 as defined by ANSI N13.30.
V5.1.7 A2002

TestAmerica
THE LEADER IN ENVIRONMENTAL TESTING

Richland Laboratory
Data Review Check List
Hexavalent Chromium

Batch Number(s): 8262284

Lab Sample Numbers or J00196 Due 10/2

Method/Test/Parameter: Cr+6 in Water / RL-WC-003

Review Item	Yes (✓)	No (✓)	N/A (✓)	2nd Level Review (✓)
A. Initial Calibration				
1. Performed at required frequency with required number of levels?	✓			✓
2. Correlation coefficient within QC limits?	✓			✓
3. Initial calibration verification (ICV) analyzed immediately after calibration and results within QC limits?	✓			✓
4. Initial calibration blank (ICB) analyzed immediately after ICV and concentrations of all parameters ≤ reporting limit?	✓			✓
B. Continuing Calibration				
1. CCV analyzed at required frequency and all parameters within QC limits?	✓			✓
2. CCB analyzed at required frequency and all results ≤ reporting limit?	✓			✓
C. Sample Analysis				
1. Were any samples with concentrations above the linear range for any parameter diluted and reanalyzed?	✓			✓
2. Were all sample holding times met?	✓			✓
D. QC Samples				
1. All results for the preparation blank below limits?	✓			✓
2. MS or MS/MSD recoveries within QC limits and %RPD (for MSD) acceptable?		✓		✓
3. LCS percent recovery within QC limits and %RPD (for LCSD) acceptable?	✓			✓
4. Analytical spikes within QC limits where applicable?	✓			✓
5. ICP only: One serial dilution performed per SDG?			✓	✓
6. ICP only: CRDL standard (CRI or CRA) analyzed at required frequency?			✓	✓
7. ICP only: Interference check samples (ICSA, ICSAB) and HICAL analyzed at the required frequencies and within QC limits?			✓	✓

Review Item	Yes (✓)	No (✓)	N/A (✓)	2nd Level Review (✓)
E. Other	✓			
1. Are all nonconformances included and noted?				✓
2. Is the correct date and time of analysis shown?	✓			✓
3. Did the analyst sign and date the front page of the analytical run?	✓			✓
4. Correct methodology used?	✓			✓
5. Transcriptions checked?	✓			✓
6. Calculations checked at minimum frequency?	✓			✓
7. Units checked?	✓			✓

Comments on any "No" response:
The matrix spike and matrix spike duplicate recovered very low. Strongly suspect there was too much reduction occurring. –See NCM

Analyst: [signature] Date: 9/18/08
Second-Level Review: JSeger Date: 9/18/08

Clouseau
Nonconformance Memo

TestAmerica
THE PARER IN ENVIRONMENTAL TESTING

NCM #: **10-13003**
NCM Initiated By: LIEM DINH
Date Opened: 09/18/2008
Date Closed:

Classification: **Anomaly**
Status: **GLREVIEW**
Production Area: Classical Chemistry
Tests: None
Lot #'s (Sample #'s): ,
QC Batches: None.,

Nonconformance: Other (describe in detail)
Subcategory: Other (explanation required)

Problem Description / Root Cause

Name	Date	Description
LIEM DINH	09/18/2008	Batch Number: 8262284 SDG: J00196 The matrix spike and matrix spike duplicate recovered very low. When colorizing the MS and MSD, there were no changes in the color. Strongly suspect there was too much reduction occurring.
Liem Dinhl	09/18/2008	Batch Number: 8262284 SDG: J00196 The matrix spike and matrix spike duplicate recovered very low. When colorizing the MS and MSD, there were no changes in the color. Strongly suspect there was too much reduction occurring.

Corrective Action

Name	Date	Corrective Action
LIEM DINH	09/18/2008	Report Data
Liem Dinhl	09/18/2008	Report Data

Client Notification Summary

Client	Project Manager	Notified	Response	How Notified	Note	Response	Response Note

Quality Assurance Verification

Verified By	Due Date	Status	Notes
		This section not yet completed by QA.	

Approval History

Date Approved	Approved By	Position

Date Printed: 9/18/2008

Washington Closure Hanford	CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST		RC-052-009	Page 1 of 1
Collector Rios	Company Contact Bill Hudson	Telephone No. 372-9695	Project Coordinator KESSNER, JH	Price Code 7G Data Turnaround 15 Days
Project Designation Remaining Sites Confirmation Sampling - Water	Sampling Location 100-D-78		SAF No. RC-052	
Ice Chest No.	Field Logbook No. EL-1601-2	COA 010D78A000	Method of Shipment	
Shipped To TestAmerica Incorporated, Richland	Offsite Property No.		Bill of Lading/Air Bill No.	

POSSIBLE SAMPLE HAZARDS/REMARKS

J8I180139 ~~J6I180128~~ N8
J00196 ~~W05516~~

Special Handling and/or Storage

Due: ~~10/03/08~~ ~~10/03/08~~ 10/2/08

Preservation	Cool 4C											
Type of Container	G/P											
No. of Container(s)	1											
Volume	~~500mL~~ 1000mL	BH 9/17/08										
SAMPLE ANALYSIS	Chromium Hex - 7196											

Sample No.	Matrix *	Sample Date	Sample Time										
J17J49	WATER	9/17/08	0900	x				KW4mm					

CHAIN OF POSSESSION		Sign/Print Names		SPECIAL INSTRUCTIONS	Matrix *
Relinquished By/Removed From	Date/Time 1132 9-17-08	Received By/Stored In BILL HUDSON Bill Hudson	Date/Time ~~1120~~ 1132 9/17/08	BH 9/17/08	S=Soil SE=Sediment SO=Solid SI=Sludge W = Water O=Oil A=Air DS=Drum Solids DL=Drum Liquids T=Tissue WI=Wipe L=Liquid V=Vegetation X=Other
Relinquished By/Removed From BILL HUDSON B Hudson	Date/Time 1235 9/17/08	Received By/Stored In X [illegible] WLANE TAL	Date/Time 1235 9/17/08		
Relinquished By/Removed From	Date/Time	Received By/Stored In	Date/Time		
Relinquished By/Removed From	Date/Time	Received By/Stored In	Date/Time		
Relinquished By/Removed From	Date/Time	Received By/Stored In	Date/Time		
Relinquished By/Removed From	Date/Time	Received By/Stored In	Date/Time		

LABORATORY SECTION	Received By	Title	Date/Time
FINAL SAMPLE DISPOSITION	Disposal Method	Disposed By	Date/Time

WCH-EE-011

TestAmerica
THE LEADER IN ENVIRONMENTAL TESTING

Sample Check-in List

Date/Time Received: 91708 1235 GM Screen Result 0.1K

Client: WCH SDG #: ~~W0551L~~ J00196 NA[ ] SAF #: RC-052 NA [ ]

Work Order Number: J8I180128-139 18 Chain of Custody # RC-052-009

Shipping Container ID: N/A Air Bill # N/A

1. Custody Seals on shipping container intact? NA [ ] Yes [✓] No [ ]
2. Custody Seals dated and signed? NA [ ] Yes [✓] No [ ]
3. Chain of Custody record present? NA [ ] Yes [✓] No [ ]
4. Cooler Temperature: _____ NA [✓] 5. Vermiculite/packing materials is NA [✓] Wet [ ] Dry [ ]
6. Number of samples in shipping container: 1
7. Sample holding times exceeded? NA [✓] Yes [ ] No [ ]
8. Samples have:
   - ____ Tape
   - ✓ Custody Seals
   - ____ Hazard Lables
   - ✓ Appropriate Sample Lables
9. Samples are:
   - ✓ In Good Condition
   - ____ Broken
   - ____ Leaking
   - ____ Have Air Bubbles
   - (Only for samples requiring no head space.)
10. Sample pH taken? NA [ ] pH<2 [ ] pH>2 [✓] pH>9 [ ] Amount HNO_3 Added _____
11. Sample Location, Sample Collector Listed? *
    *For documentation only. No corrective action needed.
12. Were any anomalies identified in sample receipt? Yes [ ] No [✓]
13. Description of anomalies (include sample numbers): _____

Sample Custodian: [signature] Date: 91708

Client Sample ID	Analysis Requested	Condition	Comments/Action

Client Informed on _____ by _____ Person Contacted _____

[ ] No action necessary; process as is.

Project Manager _____ Date _____

9/18/2008 11:02:29 AM

Sample Preparation/Analysis

127642, Washington Closure Hanford LLC
Bechtel Hanford, Inc.

AnalyDueDate: 10/02/2008

88 NO SAMPLE PREPARATION PERFORMED / DIRECT INJECTION
EA Chromium, Hexavalent (7196A)
5I CLIENT: HANFORD

Batch: 8262284 **WATER** **mg/L**
SEQ Batch, Test: None All Tests: 8262284 88EA,

PM, Quote: SS , 27023

Balance Id:
Pipet #:
Sep1 DT/Tm Tech:
Sep2 DT/Tm Tech:
Prep Tech:

Work Order, Lot, Sample DateTime	Total Amt/Unit	Initial Aliquot Amt/Unit	QC Tracer Prep Date	Count Time Min	Detector Id	Count On | Off (24hr) Circle	CR Analyst, Init/Date	Comments:
1 KW4MM-1-AA								
J8I180139-1-SAMP								
09/17/2008 09:00		AmtRec: LP	#Containers: 1			Scr:	Alpha:	Beta:
2 KW4MM-1-AC-S								
J8I180139-1-MS								
09/17/2008 09:00		AmtRec: LP	#Containers: 1			Scr:	Alpha:	Beta:
3 KW4MM-1-AD-D								
J8I180139-1-MSD								
09/17/2008 09:00		AmtRec: LP	#Containers: 1			Scr:	Alpha:	Beta:
4 KW4MM-1-AE-X								
J8I180139-1-DUP								
09/17/2008 09:00		AmtRec: LP	#Containers: 1			Scr:	Alpha:	Beta:
5 KW41X-1-AA-B								
J8I180000-284-BLK								
09/17/2008 09:00		AmtRec:	#Containers: 1			Scr:	Alpha:	Beta:
6 KW41X-1-AC-C								
J8I180000-284-LCS								
09/17/2008 09:00		AmtRec:	#Containers: 1			Scr:	Alpha:	Beta:

TestAmerica Laboratories, Inc.
19

9/18/2008 11:02:30 AM

Sample Preparation/Analysis

88 NO SAMPLE PREPARATION PERFORMED / DIRECT INJECTION
EA Chromium, Hexavalent (7196A)
5I CLIENT: HANFORD

Balance Id:
Pipet #:
Sep1 DT/Tm Tech:
Sep2 DT/Tm Tech:
Prep Tech:

AnalyDueDate: 10/02/2008

Batch: 8262284 mg/L
SEQ Batch, Test: None

Work Order, Lot, Sample DateTime	Total Amt/Unit	Initial Aliquot Amt/Unit	QC Tracer Prep Date	Count Time Min	Detector Id	Count On | Off (24hr) Circle	CR Analyst, Init/Date	Comments:

Comments:

All Clients for Batch:
127642, Washington Closure Hanford LLC Bechtel Hanford, Inc. , SS , 27023

KW4MM1AA-SAMP Constituent List:

KW4MM1AC-MS Constituent List:

KW4MM1AD-MSD:

KW41X1AA-BLK:

KW41X1AC-LCS:

KW4MM1AA-SAMP Calc Info:
Uncert Level (#s).: 2 Decay to SaDt: Y Blk Subt.: N Sci.Not.: Y ODRs: B
KW4MM1AC-MS Calc Info:
Uncert Level (#s).: 2 Decay to SaDt: Y Blk Subt.: N Sci.Not.: Y ODRs: B
KW4MM1AD-MSD:
Uncert Level (#s).: 2 Decay to SaDt: Y Blk Subt.: N Sci.Not.: Y ODRs: B
KW41X1AA-BLK:
Uncert Level (#s).: 2 Decay to SaDt: Y Blk Subt.: N Sci.Not.: Y ODRs: B
KW41X1AC-LCS:
Uncert Level (#s).: 2 Decay to SaDt: Y Blk Subt.: N Sci.Not.: Y ODRs: B

Approved By _____ Date: _____

TestAmerica Laboratories, Inc.

TestAmerica

THE LEADER IN ENVIRONMENTAL TESTING

TA Richland
Hexavalent Chromium - Water

Analyst:	L. Dinh
Start Date:	9/17/2008
Start Time:	
End Date:	9/17/2008
End Time	
Analyst Signature:	
Date:	9/18/2008

Calibration Curve Information			
	Amount	Conc.(mg/L)	ABS.
Blank	0.000	0.000	0.000
Std. 1	0.100	0.050	0.100
Std. 2	0.500	0.250	0.479
Std. 3	0.750	0.375	0.726
Std. 4	1.500	0.750	1.407
Std 5	2.000	1.000	1.850
Standard Volume (mL):			100.000
Date of Curve:			9/17/2008

SOP Information	
RL-WC-003	
Revision 0	
MDL (mg/L)	0.002

BATCH #	8262284
SDG #	J00196
Matrix	Water

Instrument Information	
Instrument:	Hach DR2010
Wavelength:	540
R Squared	0.99971
Slope:	1.85077
Intercept:	0.01232

	Calibration Information:	ICV Information:		LCS Information:		Matrix Spike Information:	
Dilution ID #	Cr-08-00179	Cr-08-00180		Cr-08-00179		Cr-08-00179	
Prep Date:	09/17/08	09/17/08		09/17/08		09/17/08	
Concentration (mg/L)	50	50		50		50	
Expiration Date:	09/18/08	09/18/08		09/18/08		09/18/08	
Pipettor(s)	70,190	190		190		190	
Volume Used (Expected Value		1.000	0.50000	1.00	0.50000	0.50	0.26316

Expected values are only amounts added in mg and not final concentrations

Sample ID	Client ID	Type	Sample Volume (mL)	Sample ABS.	Blank ABS.	Corrected ABS.	Dilution Factor	Curve Conc. (mg/L)	Final Conc. (mg/L)	% Rec.
n/a	n/a	ICV	100.000	0.946	0.000	0.946	1	0.5045	0.504	100.90%
n/a	n/a	ICB	100.000	0.000	0.000	0.000	1	<MDL	<MDL	
KW41X1AA	n/a	Prep Blank	100.000	0.001		0.001	1	-0.0061	<MDL	
KW41X1AC	n/a	LCS	100.000	0.943		0.943	1	0.5029	0.503	100.57%
KW4MM1AA	J17J49'	Sample	100.000	0.063		0.063	1	0.0274	0.027	
KW4MM1AC-S	J17J49'-MS	MS	100.000	0.084		0.084	1	0.0387	~~0.039~~ 0.038	4.31%
KW4MM1AD-D	J17J49'-MSD	MSD	100.000	0.149		0.149	1	0.0739	~~0.074~~ 0.07	17.66%
KW4MM1AE-X	J17J49'-DUP	Duplicate	100.000	0.058		0.058	1	0.0247	0.025	SKS 9/18/08
			100.000				1			
			100.000				1			
			100.000				1			
			100.000				1			
n/a	n/a	CCV	100.000	0.946		0.946	1	0.5045	0.504	100.90%
n/a	n/a	CCB	100.000	-0.001		-0.001	1	<MDL	<MDL	
			100.000				1			
			100.000				1			
			100.000				1			
			100.000				1			
			100.000				1			