



A TETRA TECH COMPANY
August 7, 2003

Steve Trent
Sample Management Project Coordinator
Fluor Hanford - Central Plateau Project
825 Jadwin Avenue
Richland, WA 99352

H2120

0060795

Dear Mr. Trent:

We submitted two moisture content results for B16HB7. The test was originally performed on 4/1/03 and was performed again on 5/6/03. The result from 5/6/03 is a duplicate that should not have been reported. The revised 5/06/03 datasheet is enclosed. Sorry for the confusion.

Sincerely,

David E. Germeroth

RECEIVED
NOV 24 2003

EDMC



Moisture Content, ASTM D2216

MAXIM Technologies, Inc.
1908 Innerbelt Business Center Drive
St. Louis, MO 63114-5700
314-426-0880

Client: ThermoNutech/Eberline

Date: 5/6/03

Project Number: 9905093-100

File: Eberline-Moisture-5-6-03

Performed By: Ron Swaller

Maxim Sample Number	Client Sample Name	Container Number	Container Weight	Wet Sample + Container	Dry Sample + Container	Moisture Content (%)
030523001-007	B16T00	5X	11.88	39.08	38.35	2.8
030523001-003	B16HB9	Q102	11.55	37.05	36	4.3
030523001-005	B16RY8	25	11.82	36.48	35.54	4.0
030523001-006	B16RY9	36	12.08	37.26	36.85	1.7
030523001-004	B16HCO	43	12.06	35.55	34.59	4.3
030523001-004	B16HCO-dup	X-18	11.55	37.74	36.71	4.1

Remarks: weights in grams