



EBERLINE
SERVICES

RECEIVED APRIL 19, 2010

0101703
REVISION 1

EBERLINE ANALYTICAL CORPORATION
2030 Wright Avenue
Richmond, California 94804-3048
Phone (510) 235-2633 Fax (510) 235-0438
Toll Free (800) 841-5487
www.eberlineservices.com

February 12, 2010

Mr. Michael Neely
CH2M Hill Plateau Remediation Company
P.O. Box 1600
Mail Stop – B6-06
Richland, WA 99352

Reference: **P.O. #33677**
Eberline Analytical S0-02-032-7764, SDG H4154

Dear Mr. Neely:

Enclosed is a data report for three water samples designated under SAF No. F10-034 received at Eberline Analytical on February 5, 2009. The samples were analyzed according to the accompanying chain-of-custody documents.


Please call if you have any questions concerning this report.

Sincerely,

N. Joseph Verville
Client Services Manager

NJV/ljb

Enclosure: Data Package



Eberline Analytical
W.O. No. S0-02-032-7764

CH2M Hill Plateau Remediation Company
SDG H4154

Case Narrative

Page 1 of 1

1.0 GENERAL

CH2M Hill Plateau Remediation Company (CHPRC) Sample Delivery Group H4154 was composed of three water samples designated under SAF No. F10-034 with a Project Designation of: AARA 200-CW-3 Sampling and Analysis – Verification - QC.

The samples were received as stated on the chain-of-custody documents. Any discrepancies are noted on the Eberline Analytical Sample Receipt Checklist.

The analysis requested on the Chain of Custody for sample B23WM0 was tritium by Ion Exchange. Client was contacted by phone and informed that Eberline Analytical does not perform this analysis and the client requested that the analysis be changed to tritium by distillation.

2.0 ANALYSIS NOTES

2.1 Tritium Analysis


The tritium QC MS analysis data sheet denotes an "X" qualifier, which indicates that some data was manually entered and may need to be double checked; in this case the "added amount" was manually entered, and subsequently double checked. The results for both the original and duplicate analyses were less than their respective MDA's, therefore no RPD is calculated, and there is no associated control limit. No problems were encountered during the course of the analyses.

2.2 Nickel-63 Analysis

The results for both the original and duplicate analyses were less than their respective MDA's, therefore no RPD is calculated, and there is no associated control limit. No problems were encountered during the course of the analyses.

3.0 Case Narrative Certification Statement

"I certify that this data package is in compliance with the SOW, both technically and for completeness, for other than the conditions detailed above. Release of the data obtained in this hard copy data package has been authorized by the Laboratory Manager or a designee, as verified by the following signature."



N. Joseph Verville
Client Services Manager

2/12/10
Date

Problem and Discrepancy Report

Eberline

SDG H4154

1. The data package has the following issues:

a) Sample Receipt Checklist, #12, nothing is checked.

Resolution: *Provide correction.*

Lab Response: **Provide correction.**

Please correct the issues and resubmit the hard copy data package.

SDG 7764
 Contact N. Joseph Verville

Client CHPRC
 Contract No. 33677
 Case no SDG_H4154

S U M M A R Y D A T A S E C T I O N

T A B L E O F C O N T E N T S	
About this section	1
Sample Summaries	3
Prep Batch Summary	5
Work Summary	6
Method Blanks	8
Lab Control Samples	9
Duplicates	10
Matrix Spikes	12
Data Sheets	13
Method Summaries	16
Report Guides	18
End of Section	32

UB

Prepared by

W. Quill

 Reviewed by

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-TOC
 Version 3.06
 Report date 02/12/10

SDG 7764
 Contact N. Joseph Verville

REPORT GUIDE

Client CHPRC
 Contract No. 33677
 Case no SDG H4154

ABOUT THE DATA SUMMARY SECTION

The Data Summary Section of a Data Package has all data, in several useful orders, necessary for first level, routine review of the data package for a Sample Delivery Group (SDG). This section follows the Data Package Narrative, which has an overview of the data package and a discussion of special problems. It is followed by the Raw Data Section, which has full details.

The Data Summary Section has several groups of reports:

SAMPLE SUMMARIES

The Sample and QC Summary Reports show all samples, including QC samples, reported in one SDG. These reports cross-reference client and lab sample identifiers.

PREPARATION BATCH SUMMARY

The Preparation Batch Summary Report shows all preparation batches (lab groupings reflecting how work was organized) relevant to the reported SDG with information necessary to check the completeness and consistency of the SDG.

WORK SUMMARY

The Work Summary Report shows all samples and work done on them relevant to the reported SDG.

METHOD BLANKS

The Method Blank Reports, one for each Method Blank relevant to the SDG, show all results and primary supporting information for the blanks.

LAB CONTROL SAMPLES

The Lab Control Sample Reports, one for each Lab Control Sample relevant to the SDG, show all results, recoveries and primary supporting information for these QC samples.

REPORT GUIDES

Page 1

SUMMARY DATA SECTION

Page 1

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-RG
 Version 3.06
 Report date 02/12/10

SDG 7764
Contact N. Joseph Verville

GUIDE, cont.

Client CHPRC
Contract No. 33677
Case no SDG H4154

ABOUT THE DATA SUMMARY SECTION

DUPLICATES

The Duplicate Reports, one for each Duplicate and Original sample pair relevant to the SDG, show all results, differences and primary supporting information for these QC samples.

MATRIX SPIKES

The Matrix Spike Reports, one for each Spiked and Original sample pair relevant to the SDG, show all results, recoveries and primary supporting information for these QC samples.

DATA SHEETS

The Data Sheet Reports, one for each client sample in the SDG, show all results and primary supporting information for these samples.

METHOD SUMMARIES

The Method Summary Reports, one for each test used in the SDG, show all results, QC and method performance data for one analyte on one or two pages. (A test is a short code for the method used to do certain work to the client's specification.)

REPORT GUIDES

The Report Guides, one for each of the above groups of reports, have documentation on how to read the associated reports.

REPORT GUIDES

Page 2

SUMMARY DATA SECTION

Page 2

Lab id EBRLNE
Protocol CHPRC
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 02/12/10

EBERLINE ANALYTICAL/RICHMOND

SAMPLE DELIVERY GROUP H4154

REVISION 1

SDG 7764
 Contact N. Joseph Verville

LAB SAMPLE SUMMARY

Client CHPRC
 Contract No. 33677
 Case no SDG H4154

LAB	CLIENT SAMPLE ID	LOCATION	MATRIX	LEVEL	SAF NO	CHAIN OF CUSTODY	COLLECTED
S002032-01	B23WL8	216-N-1 COVERED POND;FTB	WATER		F10-034	F10-034-008	01/29/10 07:10
S002032-02	B23WL9	216-N-1 COVERED POND;EB	WATER		F10-034	F10-034-010	01/29/10 07:55
S002032-03	B23WM0	216-N-1 COVERED POND;FTB	WATER		F10-034	F10-034-011	01/29/10 07:10
S002032-04	Lab Control Sample		WATER		F10-034		
S002032-05	Method Blank		WATER		F10-034		
S002032-06	Duplicate (S002032-01)	216-N-1 COVERED POND;FTB	WATER		F10-034		01/29/10 07:10
S002032-07	Duplicate (S002032-03)	216-N-1 COVERED POND;FTB	WATER		F10-034		01/29/10 07:10
S002032-08	Spike (S002032-03)	216-N-1 COVERED POND;FTB	WATER		F10-034		01/29/10 07:10

LAB SUMMARY

Page 1

SUMMARY DATA SECTION

Page 3

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-LS
 Version 3.06
 Report date 02/12/10

SDG 7764
 Contact N. Joseph Verville

QC SUMMARY

Client CHPRC
 Contract No. 33677
 Case no SDG H4154

QC BATCH	CHAIN OF CUSTODY	CLIENT SAMPLE ID	MATRIX	SOLIDS	SAMPLE AMOUNT	BASIS AMOUNT	DAYS SINCE RECEIVED	LAB COLL SAMPLE ID	DEPARTMENT SAMPLE ID
7764	F10-034-008	B23WL8	WATER		1.0 L		02/05/10 7	S002032-01	7764-001
	F10-034-010	B23WL9	WATER		1.0 L		02/05/10 7	S002032-02	7764-002
	F10-034-011	B23WM0	WATER		1.0 L		02/05/10 7	S002032-03	7764-003
		Method Blank	WATER					S002032-05	7764-005
		Lab Control Sample	WATER					S002032-04	7764-004
		Duplicate (S002032-01)	WATER		1.0 L		02/05/10 7	S002032-06	7764-006
		Duplicate (S002032-03)	WATER		1.0 L		02/05/10 7	S002032-07	7764-007
		Spike (S002032-03)	WATER		1.0 L		02/05/10 7	S002032-08	7764-008

QC SUMMARY

Page 1

SUMMARY DATA SECTION

Page 4

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-QS
 Version 3.06
 Report date 02/12/10

EBERLINE ANALYTICAL/RICHMOND

REVISION 1

SAMPLE DELIVERY GROUP H4154

SDG 7764
 Contact N. Joseph Verville

PREP BATCH SUMMARY

Client CHPRC
 Contract No. 33677
 Case no SDG H4154

TEST	MATRIX	METHOD	PREPARATION ERROR		PLANCHETS ANALYZED				QUALI-PIERS		
			BATCH	2σ %	CLIENT	MORE	RE BLANK	LCS		DUP/ORIG MS/ORIG	
Liquid Scintillation Counting											
H	WATER	Tritium in Water	7243-091	10.0	1		1	1	1/1	1/1	X
NI_L	WATER	Nickel-63 in Liquid	7243-091	11.2	2		1	1	1/1		

Duplicates and Matrix Spikes are those with original (Client) sample in this Sample Delivery Group.
 Blank and LCS planchets are those in the same preparation batch as some Client, Duplicate or Spike sample.

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-PBS
 Version 3.06
 Report date 02/12/10

EBERLINE ANALYTICAL/RICHMOND

REVISION 1

SAMPLE DELIVERY GROUP H4154

SDG 7764
Contact N. Joseph Verville

Client CHPRC
Contract No. 33677
Case no SDG H4154

LAB WORK SUMMARY

LAB SAMPLE COLLECTED RECEIVED	CLIENT SAMPLE ID LOCATION CUSTODY	MATRIX	PLANCHET	TEST	SUP- FIX ANALYZED	REVIEWED	BY	METHOD
S002032-01 01/29/10 02/05/10	B23WL8 216-N-1 COVERED POND;FTB WATER F10-034-008 F10-034		7764-001	NI_L	02/10/10	02/11/10	BW	Nickel-63 in Liquid
S002032-02 01/29/10 02/05/10	B23WL9 216-N-1 COVERED POND;EB WATER F10-034-010 F10-034		7764-002	NI_L	02/10/10	02/11/10	BW	Nickel-63 in Liquid
S002032-03 01/29/10 02/05/10	B23MM0 216-N-1 COVERED POND;FTB WATER F10-034-011 F10-034		7764-003	H	02/10/10	02/11/10	BW	Tritium in Water
S002032-04	Lab Control Sample F10-034	WATER	7764-004	H	02/10/10	02/11/10	BW	Tritium in Water
			7764-004	NI_L	02/10/10	02/11/10	BW	Nickel-63 in Liquid
S002032-05	Method Blank F10-034	WATER	7764-005	H	02/10/10	02/11/10	BW	Tritium in Water
			7764-005	NI_L	02/10/10	02/11/10	BW	Nickel-63 in Liquid
S002032-06 01/29/10 02/05/10	Duplicate (S002032-01) 216-N-1 COVERED POND;FTB WATER F10-034		7764-006	NI_L	02/10/10	02/11/10	BW	Nickel-63 in Liquid
S002032-07 01/29/10 02/05/10	Duplicate (S002032-03) 216-N-1 COVERED POND;FTB WATER F10-034		7764-007	H	02/10/10	02/11/10	BW	Tritium in Water
S002032-08 01/29/10 02/05/10	Spike (S002032-03) 216-N-1 COVERED POND;FTB WATER F10-034		7764-008	H	02/10/10	02/11/10	BW	Tritium in Water

WORK SUMMARY

Page 1

SUMMARY DATA SECTION

Page 6

Lab id EBRLINE
Protocol CHPRC
Version Ver 1.0
Form DVD-LWS
Version 3.06
Report date 02/12/10

EBERLINE ANALYTICAL/RICHMOND

SAMPLE DELIVERY GROUP H4154

REVISION 1

SDG 7764
 Contact N. Joseph Verville

WORK SUMMARY, cont.

Client CHPRC
 Contract No. 33677
 Case no SDG H4154

COUNTS OF TESTS BY SAMPLE TYPE

TEST	SAP No	METHOD	REFERENCE	CLIENT	MORE	RE	BLANK	LCS	DUP SPIKE	TOTAL
H	F10-034	Tritium in Water	906.0_H3_LSC	1			1	1	1	5
NI_L	F10-034	Nickel-63 in Liquid	NI63_LSC	2			1	1	1	5
TOTALS				3			2	2	2	10

WORK SUMMARY

Page 2

SUMMARY DATA SECTION

Page 7

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-LWS
 Version 3.06
 Report date 02/12/10

EBERLINE ANALYTICAL/RICHMOND
 SAMPLE DELIVERY GROUP H4154

REVISION 1

7764-005

Method Blank

METHOD BLANK

SDG <u>7764</u>	Client/Case no <u>CHPRC</u>	SDG <u>H4154</u>
Contact <u>N. Joseph Verville</u>	Contract <u>No. 33677</u>	
Lab sample id <u>S002032-05</u>	Client sample id <u>Method Blank</u>	
Dept sample id <u>7764-005</u>	Material/Matrix <u>WATER</u>	
	SAF No <u>F10-034</u>	

ANALYTE	CAS NO	RESULT pCi/L	2σ ERR (COUNT)	MDA pCi/L	RDL pCi/L	QUALI- FIERS	TEST
Tritium	10028-17-8	6.22	97	165	400	U	H
Nickel 63	13981-37-8	0.390	1.7	2.89	15.0	U	NI_L

QC-BLANK #72287

METHOD BLANKS

Page 1

SUMMARY DATA SECTION

Page 8

Lab id <u>EBRINE</u>
Protocol <u>CHPRC</u>
Version <u>Ver 1.0</u>
Form <u>DVD-DS</u>
Version <u>3.06</u>
Report date <u>02/12/10</u>

EBERLINE ANALYTICAL/RICHMOND

REVISION 1

SAMPLE DELIVERY GROUP H4154

7764-004

Lab Control Sample

LAB CONTROL SAMPLE

SDG <u>7764</u>	Client/Case no <u>CHPRC</u>	<u>SDG H4154</u>
Contact <u>N. Joseph Verville</u>	Contract No. <u>33677</u>	
Lab sample id <u>S002032-04</u>	Client sample id <u>Lab Control Sample</u>	
Dept sample id <u>7764-004</u>	Material/Matrix <u>WATER</u>	
	SAF No <u>F10-034</u>	

ANALYTE	RESULT pCi/L	2σ ERR (COUNT)	MDA pCi/L	RDL pCi/L	QUALI- FIERS	TEST	ADDED pCi/L	2σ ERR pCi/L	REC %	3σ LMTS (TOTAL)	PROTOCOL LIMITS
Tritium	2480	170	164	400		H	2680	110	93	82-118	80-120
Nickel 63	230	6.2	2.90	15.0		NI_L	262	10	88	84-116	80-120

QC-LCS #72286

Lab id <u>EBRLNE</u>
Protocol <u>CHPRC</u>
Version <u>Ver 1.0</u>
Form <u>DVD-LCS</u>
Version <u>3.05</u>
Report date <u>02/12/10</u>

EBERLINE ANALYTICAL/RICHMOND

REVISION 1

SAMPLE DELIVERY GROUP H4154

7764-006

B23WL8

DUPLICATE

SDG <u>7764</u>	Client/Case no <u>CHPRC</u>	<u>SDG H4154</u>
Contact <u>N. Joseph Verville</u>	Contract No. <u>33677</u>	
<u>DUPLICATE</u>	<u>ORIGINAL</u>	
Lab sample id <u>S002032-06</u>	Lab sample id <u>S002032-01</u>	Client sample id <u>B23WL8</u>
Dept sample id <u>7764-006</u>	Dept sample id <u>7764-001</u>	Location/Matrix <u>216-N-1 COVERED POND;FTB WATER</u>
	Received <u>02/05/10</u>	Collected/Volume <u>01/29/10 07:10</u> <u>1.0 L</u>
		Custody/SAF No <u>F10-034-008</u> <u>F10-034</u>

ANALYTE	DUPLICATE pCi/L	2σ ERR (COUNT)	MDA pCi/L	RDL pCi/L	QUALI- FIERS	TEST	ORIGINAL pCi/L	2σ ERR (COUNT)	MDA pCi/L	QUALI- FIERS	RPD %	3σ TOT	DER σ
Nickel 63	1.42	1.7	2.84	15.0	U	NI_L	0.469	1.7	2.90	U	-		0.8

QC-DUP#1 72288

AARA 200-CW-3 Sampling and Analysis-Verification-QC

Lab id <u>EBRLNE</u>
Protocol <u>CHPRC</u>
Version <u>Ver 1.0</u>
Form <u>DVD-DUP</u>
Version <u>3.06</u>
Report date <u>02/12/10</u>

EBERLINE ANALYTICAL/RICHMOND

REVISION 1

SAMPLE DELIVERY GROUP H4154

7764-007

B23WM0

DUPLICATE

SDG <u>7764</u>		Client/Case no <u>CHPRC</u>		SDG <u>H4154</u>	
Contact <u>N. Joseph Verville</u>		Contract No. <u>33677</u>			
DUPLICATE		ORIGINAL			
Lab sample id <u>S002032-07</u>	Lab sample id <u>S002032-03</u>	Client sample id <u>B23WM0</u>			
Dept sample id <u>7764-007</u>	Dept sample id <u>7764-003</u>	Location/Matrix <u>216-N-1 COVERED POND;PTB WATER</u>			
	Received <u>02/05/10</u>	Collected/Volume <u>01/29/10 07:30 1.0 L</u>			
		Custody/SAF No <u>F10-034-011 F10-034</u>			

ANALYTE	DUPLICATE pCi/L	2σ ERR (COUNT)	MDA pCi/L	RDL pCi/L	QUALI- FIERS	TEST	ORIGINAL pCi/L	2σ ERR (COUNT)	MDA pCi/L	QUALI- FIERS	RPD &	3σ TOT	DER σ
Tritium	87.2	100	165	400	U	H	116	100	165	U	-	0.4	

QC-DUP#3 72289

AARA 200-CN-3 Sampling and Analysis-Verification-QC

Lab id <u>EBRLNE</u>
Protocol <u>CHPRC</u>
Version <u>Ver 1.0</u>
Form <u>DVD-DUP</u>
Version <u>3.06</u>
Report date <u>02/12/10</u>

MATRIX SPIKE

SDG <u>7764</u>	Client/Case no <u>CHPRC</u>	SDG <u>H4154</u>
Contact <u>N. Joseph Verville</u>	Contract No. <u>33677</u>	
MATRIX SPIKE	ORIGINAL	
Lab sample id <u>S002032-08</u>	Lab sample id <u>S002032-03</u>	Client sample id <u>B23WM0</u>
Dept sample id <u>7764-008</u>	Dept sample id <u>7764-003</u>	Location/Matrix <u>216-N-1 COVERED POND;FTB WATER</u>
	Received <u>02/05/10</u>	Collected/Volume <u>01/29/10 07:10 1.0 L</u>
		Custody/SAP No <u>F10-034-011 F10-034</u>

ANALYTE	SPIKE pCi/L	2σ ERR (COUNT)	MDA pCi/L	RDL pCi/L	QUALI- FIERS TEST	ADDED pCi/L	2σ ERR pCi/L	ORIGINAL pCi/L	2σ ERR (COUNT)	REC 3σ % (TOTAL)	IAMS LIMITS	PROTOCOL LIMITS
Tritium	12800	340	165	400	X H	14300	570	116	100	89	85-115	60-140

QC-MS#3 72300

AARA 200-CW-3 Sampling and Analysis-Verification-QC

Lab id <u>EBRLNE</u>
Protocol <u>CHPRC</u>
Version <u>Ver 1.0</u>
Form <u>DVD-MS</u>
Version <u>3.06</u>
Report date <u>02/12/10</u>

7764-001

B23WL8

DATA SHEET

SDG <u>7764</u>	Client/Case no <u>CHPRC</u>	<u>SDG H4154</u>
Contact <u>N. Joseph Verville</u>	Contract <u>No. 33677</u>	
Lab sample id <u>S002032-01</u>	Client sample id <u>B23WL8</u>	
Dept sample id <u>7764-001</u>	Location/Matrix <u>216-N-1 COVERED POND; FTB WATER</u>	
Received <u>02/05/10</u>	Collected/Volume <u>01/29/10 07:10</u>	<u>1.0 L</u>
	Custody/SAF No <u>F10-034-008</u>	<u>F10-034</u>

ANALYTE	CAS NO	RESULT pCi/L	2σ ERR (COUNT)	MDA pCi/L	RDL pCi/L	QUALI- FIERS	TEST
Nickel 63	13981-37-8	0.469	1.7	2.90	15.0	U	NI_L

AARA 200-CW-3 Sampling and Analysis-Verification-QC

Lab id <u>EBRINE</u>
Protocol <u>CHPRC</u>
Version <u>Ver 1.0</u>
Form <u>DVD-DS</u>
Version <u>3.06</u>
Report date <u>02/12/10</u>

EBERLINE ANALYTICAL/RICHMOND
 SAMPLE DELIVERY GROUP H4154

REVISION 1

7764-002

B23WL9

DATA SHEET

SDG <u>7764</u>	Client/Case no <u>CHPRC</u>	<u>SDG H4154</u>
Contact <u>N. Joseph Verville</u>	Contract No. <u>33677</u>	
Lab sample id <u>S002032-02</u>	Client sample id <u>B23WL9</u>	
Dept sample id <u>7764-002</u>	Location/Matrix <u>216-N-1 COVERED POND;EB WATER</u>	
Received <u>02/05/10</u>	Collected/Volume <u>01/29/10 07:55</u> <u>1.0 L</u>	
	Custody/SAF No <u>F10-034-010</u> <u>F10-034</u>	

ANALYTE	CAS NO	RESULT pCi/L	2 σ ERR (COUNT)	MDA pCi/L	RDL pCi/L	QUALI- FIERS	TEST
Nickel 63	13981-37-8	1.83	1.7	2.83	15.0	U	NI_L

AARA 200-CW-3 Sampling and Analysis-Verification-QC

Lab id <u>EBRLNE</u>
Protocol <u>CHPRC</u>
Version <u>Ver 1.0</u>
Form <u>DVD-DS</u>
Version <u>3.06</u>
Report date <u>02/12/10</u>

7764-003

B23WM0

DATA SHEET

SDG <u>7764</u>	Client/Case no <u>CHPRC</u>	SDG <u>H4154</u>
Contact <u>N. Joseph Verville</u>	Contract <u>No. 33677</u>	
Lab sample id <u>S002032-03</u>	Client sample id <u>B23WM0</u>	
Dept sample id <u>7764-003</u>	Location/Matrix <u>216-N-1 COVERED POND; FTB WATER</u>	
Received <u>02/05/10</u>	Collected/Volume <u>01/29/10 07:10</u> <u>1.0 L</u>	
	Custody/SAF No <u>F10-034-011</u> <u>F10-034</u>	

ANALYTE	CAS NO	RESULT pCi/L	2σ ERR (COUNT)	MDA pCi/L	RDL pCi/L	QUALI- FIERS	TEST
Tritium	10028-17-8	116	100	165	400	U	H

AARA 200-CW-3 Sampling and Analysis-Verification-QC

Lab id <u>EBRLNE</u>
Protocol <u>CHPRC</u>
Version <u>Ver 1.0</u>
Form <u>DVD-DS</u>
Version <u>3.06</u>
Report date <u>02/12/10</u>

EBERLINE ANALYTICAL/RICHMOND

REVISION 1

SAMPLE DELIVERY GROUP H4154

LAB METHOD SUMMARY

TRITIUM IN WATER

LIQUID SCINTILLATION COUNTING

Test H Matrix WATER
 SDG 7764
 Contact N. Joseph Verville

Client CHPRC
 Contract No. 33677
 Contract SDG H4154

RESULTS

LAB	RAW	SUF-			
SAMPLE ID	TEST FIX	PLANCHET	CLIENT SAMPLE ID	Tritium	
Preparation batch 7243-091					
S002032-03		7764-003	B23WMO	U	
S002032-04		7764-004	Lab Control Sample	ok	
S002032-05		7764-005	Method Blank	U	
S002032-07		7764-007	Duplicate (S002032-03)	-	U
S002032-08		7764-008	Spike (S002032-03)	ok	X

Nominal values and limits from method RDLs (pCi/L) 400

METHOD PERFORMANCE

LAB	RAW	SUF-	MDA	ALIQ	PREP	DILU-	YIELD	EFF	COUNT	FWHM	DRIFT	DAYS	ANAL-		
SAMPLE ID	TEST FIX	CLIENT SAMPLE ID	pCi/L	L	FAC	TION	%	%	min	keV	KeV	HELD	PREPARED	YIELD	DETECTOR
Preparation batch 7243-091 2σ prep error 10.0 % Reference Lab Notebook No. 7243, pg 091															
S002032-03		B23WMO	165	0.0100			100	100				12	02/10/10	02/10	LSC-007
S002032-04		Lab Control Sample	164	0.100			10	100					02/10/10	02/10	LSC-007
S002032-05		Method Blank	165	0.100			10	100					02/10/10	02/10	LSC-007
S002032-07		Duplicate (S002032-03)	165	0.0100			100	100				12	02/10/10	02/10	LSC-007
S002032-08		Spike (S002032-03)	165	0.0500			20	100				12	02/10/10	02/10	LSC-007

Nominal values and limits from method 400 0.0100 25 180

PROCEDURES REFERENCE 906.0_H3_LSC
 CP-210 Tritium in Water Samples by Distillation, rev 11

AVERAGES ± 2 SD MDA 165 ± 0.894
 FOR 5 SAMPLES YIELD 48 ± 95

METHOD SUMMARIES

Page 1

SUMMARY DATA SECTION

Page 16

Lab id: EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-LMS
 Version 3.06
 Report date 02/12/10

EBERLINE ANALYTICAL/RICHMOND

REVISION 1

SAMPLE DELIVERY GROUP H4154

LAB METHOD SUMMARY

NICKEL-63 IN LIQUID

LIQUID SCINTILLATION COUNTING

Test NI L Matrix WATER
 SDG 7764
 Contact N. Joseph Verville

Client CHPRC
 Contract No. 33677
 Contract SDG H4154

RESULTS

LAB	RAW	SUF-		
SAMPLE ID	TEST FIX	PLANCHET	CLIENT SAMPLE ID	Nickel 63
Preparation batch 7243-091				
S002032-01		7764-001	B23WL8	U
S002032-02		7764-002	B23WL9	U
S002032-04		7764-004	Lab Control Sample	ok
S002032-05		7764-005	Method Blank	U
S002032-06		7764-006	Duplicate (S002032-01)	- U

Nominal values and limits from method RDLs (pCi/L) 15.0

METHOD PERFORMANCE

LAB	RAW	SUF-	MDA	ALIQ	PREP	DILU-	YIELD	EFF	COUNT	FWHM	DRIFT	DAYS	ANAL-		
SAMPLE ID	TEST FIX	CLIENT SAMPLE ID	pCi/L	L	FAC	TION	%	%	min	keV	KeV	HELD	PREPARED	YZED	DETECTOR
Preparation batch 7243-091 2σ prep error 11.2 % Reference Lab Notebook No. 7243, pg 091															
S002032-01		B23WL8	2.90	0.500			100		50			12	02/10/10	02/10	LSC-006
S002032-02		B23WL9	2.83	0.500			100		50			12	02/10/10	02/10	LSC-006
S002032-04		Lab Control Sample	2.90	0.500			98		50				02/10/10	02/10	LSC-006
S002032-05		Method Blank	2.89	0.500			98		50				02/10/10	02/10	LSC-006
S002032-06		Duplicate (S002032-01)	2.84	0.500			99		50			12	02/10/10	02/10	LSC-006

Nominal values and limits from method 15.0 0.500 40-110 50 180

PROCEDURES REFERENCE NI63_LSC
 SPP-040 Environmental Water Dissolution, rev 2
 CP-280 Nickel-63 Purification, rev 5

AVERAGES ± 2 SD MDA 2.87 ± 0.068
 FOR 5 SAMPLES YIELD 99 ± 2

METHOD SUMMARIES

Page 2

SUMMARY DATA SECTION

Page 17

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-IMS
 Version 3.06
 Report date 02/12/10

SDG 7764
 Contact N. Joseph Verville

REPORT GUIDE

Client CHPRC
 Contract No. 33677
 Case no SDG H4154

SAMPLE SUMMARY

The Sample and QC Summary Reports show all samples, including QC samples, reported in one Sample Delivery Group (SDG).

The Sample Summary Report fully identifies client samples and gives the corresponding lab sample identification. The QC Summary Report shows at the sample level how the lab organized the samples into batches and generated QC samples. The Preparation Batch and Method Summary Reports show this at the analysis level.

The following notes apply to these reports:

- * LAB SAMPLE ID is the lab's primary identification for a sample.
- * DEPARTMENT SAMPLE ID is an alternate lab id, for example one assigned by a radiochemistry department in a lab.
- * CLIENT SAMPLE ID is the client's primary identification for a sample. It includes any sample preparation done by the client that is necessary to identify the sample.
- * QC BATCH is a lab assigned code that groups samples to be processed and QCed together. These samples should have similar matrices.

QC BATCH is not necessarily the same as SDG, which reflects samples received and reported together.

- * All Lab Control Samples, Method Blanks, Duplicates and Matrix Spikes are shown that QC any of the samples. Due to possible reanalyses, not all results for all these QC samples may be relevant to the SDG. The Lab Control Sample, Method Blank, Duplicate, Matrix Spike and Method Summary Reports detail these relationships.

REPORT GUIDES

Page 1

SUMMARY DATA SECTION

Page 18

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-RG
 Version 3.06
 Report date 02/12/10

SDG 7764
 Contact N. Joseph Verville

REPORT GUIDE

Client CHPRC
 Contract No. 33677
 Case no SDG_H4154

PREPARATION BATCH SUMMARY

The Preparation Batch Summary Report shows all preparation batches in one Sample Delivery Group (SDG) with information necessary to check the completeness and consistency of the SDG.

The following notes apply to this report:

- * The preparation batches are shown in the same order as the Method Summary Reports are printed.
- * Only analyses of planchets relevant to the SDG are included.
- * Each preparation batch should have at least one Method Blank and LCS in it to validate client sample results.
- * The QUALIFIERS shown are all qualifiers other than U, J, B, L and H that occur on any analysis in the preparation batch. The Method Summary Report has these qualifiers on a per sample basis.

These qualifiers should be reviewed as follows:

- X Some data has been manually entered or modified. Transcription errors are possible.
- P One or more results are 'preliminary'. The data is not ready for final reporting.
- 2 There were two or more results for one analyte on one planchet imported at one time. The results in DVD may not be the same as on the raw data sheets.

Other lab defined qualifiers may occur. In general, these should be addressed in the SDG narrative.

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-RG
 Version 3.06
 Report date 02/12/10

SDG 7764
 Contact N. Joseph Verville

REPORT GUIDE

Client CHPRC
 Contract No. 33677
 Case no SDG H4154

WORK SUMMARY

The Work Summary Report shows all samples, including QC samples, and all relevant analyses in one Sample Delivery Group (SDG). This report is often useful as supporting documentation for an invoice.

The following notes apply to this report:

- * TEST is a code for the method used to measure associated analytes. Results and related information for each analyte are on the Data Sheet Report. In special cases, a test code used in the summary data section is not the same as in associated raw data. In this case, both codes are shown on the Work Summary.
- * SUFFIX is the lab's code to distinguish multiple analyses (recounts, reworks, reanalyses) of a fraction of the sample. The suffix indicates which result is being reported. An empty suffix normally identifies the first attempt to analyze the sample.
- * The LAB SAMPLE ID, TEST and SUFFIX uniquely identify all supporting data for a result. The Method Summary Report for each TEST has method performance data, such as yield, for each lab sample id and suffix and procedures used in the method.
- * PLANCHET is an alternate lab identifier for work done for one test. It, combined with the TEST and SUFFIX, may be the best link to raw data.
- * For QC samples, only analyses that directly QC some regular sample are shown. The Lab Control Sample, Method Blank, Duplicate, Matrix Spike and Method Summary Reports detail these relationships.
- * The SAS (Special Analytical Services) Number is a client or lab assigned code that reflects special processing for samples, such as rapid turn around. Counts of tests done are lists by SAS number since it is likely to affect prices.

REPORT GUIDES

Page 3

SUMMARY DATA SECTION

Page 20

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-RG
 Version 3.06
 Report date 02/12/10

SDG 7764
 Contact N. Joseph Verville

REPORT GUIDE

Client CHPRC
 Contract No. 33677
 Case no SDG H4154

DATA SHEET

The Data Sheet Report shows all results and primary supporting information for one client sample or Method Blank. This report corresponds to both the CLP Inorganics and Organics Data Sheet.

The following notes apply to this report:

- * TEST is a code for the method used to measure an analyte. If the TEST is empty, no data is available; the analyte was not analyzed for.
- * The LAB SAMPLE ID and TEST uniquely identify work within the Summary Data Section of a Data Package. The Work Summary and Method Summary Reports further identify raw data that underlies this work.

The Method Summary Report for each TEST has method performance data, such as yield, for each Lab Sample ID and a list of procedures used in the method.

- * ERRORS can be labeled TOTAL or COUNT. TOTAL implies a preparation (non-counting method) error has been added, as square root of sum of squares, to the counting error denoted by COUNT. The preparation errors, which may vary by preparation batch, are shown on the Method Summary Report.
- * A RESULT can be 'N.R.' (Not Reported). This means the lab did this work but chooses not to report it now, possibly because it was reported at another time.
- * When reporting a Method Blank, a RESULT can be 'N.A.' (Not Applicable). This means there is no reported client sample work in the same preparation batch as the Blank's result. This is likely to occur when the Method Blank is associated with reanalyses of selected work for a few samples in the SDG.

The following qualifiers are defined by the DVD system:

- U The RESULT is less than the MDA (Minimum Detectable Activity).

Lab id EBRINE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-RG
 Version 3.06
 Report date 02/12/10

SDG 7764
Contact N. Joseph Verville

GUIDE, cont.

Client CHPRC
Contract No. 33677
Case no SDG H4154

DATA SHEET

If the MDA is blank, the ERROR is used as the limit.

- J The RESULT is less than the RDL (Required Detection Limit) and no U qualifier is assigned.
 - B A Method Blank associated with this sample had a result without a U flag and, after correcting for possibly different aliquots, that result is greater than or equal to the MDA for this sample.
- Normally, B is not assigned if U is. When method blank subtraction is shown on this report, B flags are assigned based on the unsubtracted values while U's are assigned based on the subtracted ones. Both flags can be assigned in this case.
- For each sample result, all Method Blank results in the same preparation batch are compared. The Method Summary Report documents this and other QC relationships.
- L Some Lab Control Sample that QC's this sample had a low recovery. The lab can disable assignment of this qualifier.
 - H Similar to 'L' except the recovery was high.
 - P The RESULT is 'preliminary'.
 - X Some data necessary to compute the RESULT, ERROR or MDA was manually entered or modified.
 - 2 There were two or more results available for this analyte. The reported result may not be the same as in the raw data.

Other qualifiers are lab defined. Definitions should be in the SDG narrative.

The following values are underlined to indicate possible problems:

- * An MDA is underlined if it is bigger than its RDL.

Lab id EBRINE
Protocol CHPRC
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 02/12/10

SDG 7764
 Contact N. Joseph Verville

GUIDE, cont.

Client CHPRC
 Contract No. 33677
 Case no SDG H4154

DATA SHEET

- * An ERROR is underlined if the 1.645 sigma counting error is bigger than both the MDA and the RESULT, implying that the MDA may not be a good estimate of the 'real' minimum detectable activity.
- * A negative RESULT is underlined if it is less than the negative of its 2 sigma counting ERROR.
- * When reporting a Method Blank, a RESULT is underlined if greater than its MDA. If the MDA is blank, the 2 sigma counting error is used in the comparison.

REPORT GUIDES

Page 6

SUMMARY DATA SECTION

Page 23

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-RG
 Version 3.06
 Report date 02/12/10

SDG 7764
Contact N. Joseph Verville

REPORT GUIDE

Client CHPRC
Contract No. 33677
Case no SDG_H4154

LAB CONTROL SAMPLE

The Lab Control Sample Report shows all results, recoveries and primary supporting information for one Lab Control Sample.

The following notes apply to this report:

- * All fields in common with the Data Sheet Report have similar usage. Refer to its Report Guide for details.
- * An amount ADDED is the lab's value for the actual amount spiked into this sample with its ERROR an estimate of the error of this amount.

An amount added is underlined if its ratio to the corresponding RDL is outside protocol specified limits.

- * REC (Recovery) is RESULT divided by ADDED expressed as a percent.
- * The first, computed limits for the recovery reflect:
 1. The error of RESULT, including that introduced by rounding the result prior to printing.
If the limits are labeled (TOTAL), they include preparation error in the result. If labeled (COUNT), they do not.
 2. The error of ADDED.
 3. A lab specified, per analyte bias. The bias changes the center of the computed limits.
- * The second limits are protocol defined upper and lower QC limits for the recovery.
- * The recovery is underlined if it is outside either of these ranges.

REPORT GUIDES

Page 7

SUMMARY DATA SECTION

Page 24

Lab id EBRINE
Protocol CHPRC
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 02/12/10

SDG 7764
 Contact N. Joseph Verville

REPORT GUIDE

Client CHPRC
 Contract No. 33677
 Case no SDG H4154

DUPLICATE

The Duplicate Report shows all results, differences and primary supporting information for one Duplicate and associated Original sample.

The following notes apply to this report:

- * All fields in common with the Data Sheet Report have similar usage. This applies both to the Duplicate and Original sample data. Refer to the Data Sheet Report Guide for details.

If the Duplicate has data for a TEST and the lab did not do this test to the Original, the Original's RESULTS are underlined.

- * The RPD (Relative Percent Difference) is the absolute value of the difference of the RESULTS divided by their average expressed as a percent.

If both RESULTS are less than their MDAs, no RPD is computed and a '-' is printed.

For an analyte, if the lab did work for both samples but has data for only one, the MDA from the sample with data is used as the other's result in the RPD.

- * The first, computed limit is the sum, as square root of sum of squares, of the errors of the results divided by the average result as a percent, hence the relative error of the difference rather than the error of the relative difference. The errors include those introduced by rounding the RESULTS prior to printing.

If this limit is labeled TOT, it includes the preparation error in the RESULTS. If labeled CNT, it does not.

This value reported for this limit is at most 999.

- * The second limit for the RPD is the larger of:
 1. A fixed percentage specified in the protocol.

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-RG
 Version 3.06
 Report date 02/12/10

SDG 7764
Contact N. Joseph Verville

GUIDE, cont.

Client CHPRC
Contract No. 33677
Case no SDG H4154

DUPLICATE

2. A protocol factor (typically 2) times the average MDA as a percent of the average result. This limit applies when the results are close to the MDAs.

- * The RPD is underlined if it is greater than either limit.
- * If specified by the lab, the second limit column is replaced by the Difference Error Ratio (DER), which is the absolute value of the difference of the results divided by the quadratic sum of their one sigma errors, the same errors as used in the first limit.

Except for differences due to rounding, the DER is the same as the RPD divided by the first RPD limit with the limit scaled to 1 sigma.

- * The DER is underlined if it is greater than the sigma factor, typically 2 or 3, shown in the header for the first RPD limit.

REPORT GUIDES

Page 9

SUMMARY DATA SECTION

Page 26

Lab id EBRLNE
Protocol CHPRC
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 02/12/10

SAMPLE DELIVERY GROUP H4154

SDG 7764
Contact N. Joseph Verville

REPORT GUIDE

Client CHPRC
Contract No. 33677
Case no SDG H4154

MATRIX SPIKE

The Matrix Spike Report shows all results, recoveries and primary supporting information for one Matrix Spike and associated Original sample.

The following notes apply to this report:

- * All fields in common with the Data Sheet Report have similar usage. This applies both to the Spiked and Original sample data. Refer to the Data Sheet Report Guide for details.

If the Spike has data for a TEST and the lab did not do this test to the Original, the Original's RESULTS are underlined.

- * An amount ADDED is the lab's value for the actual amount spiked into the Spike sample with its ERROR an estimate of the error of this amount.

An amount is underlined if its ratio to the corresponding RDL is outside protocol specified limits.

- * REC (Recovery) is the Spike RESULT minus the Original RESULT divided by ADDED expressed as a percent.

- * The first, computed limits for the recovery reflect:

1. The errors of the two RESULTS, including those introduced by rounding them prior to printing.

If the limits are labeled (TOTAL), they include preparation error in the result. If labeled (COUNT), they do not.

2. The error of ADDED.

3. A lab specified, per analyte bias. The bias changes the center of the computed limits.

- * The second limits are protocol defined upper and lower QC limits

REPORT GUIDES

Page 10

SUMMARY DATA SECTION

Page 27

Lab id EBRLNE
Protocol CHPRC
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 02/12/10

SDG 7764
Contact N. Joseph Verville

GUIDE, cont.

Client CHPRC
Contract No. 33677
Case no SDG H4154

MATRIX SPIKE

for the recovery.

These limits are left blank if the Original RESULT is more than a protocol defined factor (typically 4) times ADDED. This is a way of accounting for that when the spike is small compared to the amount in the original sample, the recovery is unreliable.

- * The recovery is underlined (out of spec) if it is outside either of these ranges.

REPORT GUIDES

Page 11

SUMMARY DATA SECTION

Page 28

Lab id EBRLNE
Protocol CHPRC
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 02/12/10

SDG 7764
 Contact N. Joseph Verville

REPORT GUIDE

Client CHPRC
 Contract No. 33677
 Case no SDG H4154

METHOD SUMMARY

The Method Summary Report has two tables. One shows up to five results measured using one method. The other has performance data for the method. There is one report for each TEST, as used on the Data Sheet Report.

The following notes apply to this report:

- * Each table is subdivided into sections, one for each preparation batch. A preparation batch is a group of aliquots prepared at roughly the same time in one work area of the lab using the same method.

There should be Lab Control Sample and Method Blank results in each preparation batch since this close correspondence makes the QC meaningful. Depending on lab policy, Duplicates need not occur in each batch since they QC sample dependencies such as matrix effects.

- * The RAW TEST column shows the test code used in the raw data to identify a particular analysis if it is different than the test code in the header of the report. This occurs in special cases due to method specific details about how the lab labels work.

The Lab Sample or Planchet ID combined with the (Raw) Test Code and Suffix uniquely identify the raw data for each analysis.

- * If a result is less than both its MDA and RDL, it is replaced by just 'U' on this report. If it is greater than or equal to the RDL but less than the MDA, the result is shown with a 'U' flag.

The J and X flags are as on the data sheet.

- * Non-U results for Method Blanks are underlined to indicate possible contamination of other samples in the preparation batch. The Method Blank Report has supporting data.

- * Lab Control Sample and Matrix Spike results are shown as: ok, No data, LOW or HIGH, with the last two underlined. 'No data'

REPORT GUIDES

Page 12

SUMMARY DATA SECTION

Page 29

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-RG
 Version 3.06
 Report date 02/12/10

SDG 7764
 Contact N. Joseph Verville

GUIDE, cont.

Client CHPRC
 Contract No. 33677
 Case no SDG H4154

METHOD SUMMARY

means no amount ADDED was specified. 'LOW' and 'HIGH' correspond to when the recovery is underlined on the Lab Control Sample or Matrix Spike Report. See these reports for supporting data.

- * Duplicate sample results are shown as: ok, No data, or OUT, with the last two underlined. 'No data' means there was no original sample data found for this duplicate. 'OUT' corresponds to when the RPD is underlined on the Duplicate Report. See this report for supporting data.
 - * If the MDA column is labeled 'MAX MDA', there was more than one result measured by the reported method and the MDA shown is the largest MDA. If not all these results have the same RDL, the MAX MDA reflects only those results with RDL equal to the smallest one.
- MDAs are underlined if greater than the printed RDL.
- * Aliquots are underlined if less than the nominal value specified for the method.
 - * Preparation factors are underlined if greater than the nominal value specified for the method.
 - * Dilution factors are underlined if greater than the nominal value specified for the method.
 - * Residues are underlined if outside the range specified for the method. Residues are not printed if yields are.
 - * Yields, which may be gravimetric, radiometric or some type of recovery depending on the method, are underlined if outside the range specified for the method.
 - * Efficiencies are underlined if outside the range specified for the method. Efficiencies are detector and geometry dependent so this test is only approximate.

REPORT GUIDES

Page 13

SUMMARY DATA SECTION

Page 30

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-RG
 Version 3.06
 Report date 02/12/10

SDG 7764
 Contact N. Joseph Verville

GUIDE, cont.

Client CHPRC
 Contract No. 33677
 Case no SDG H4154

METHOD SUMMARY

- * Count times are underlined if less than the nominal value specified for the method.
- * Resolutions (as FWHM; Full Width at Half Max) are underlined if greater than the method specified limit.
- * Tracer drifts are underlined if their absolute values are greater than the method specified limit. Tracer drifts are not printed if percent moistures are.
- * Days Held are underlined if greater than the holding time specified in the protocol.
- * Analysis dates are underlined if before their planchet's preparation date or, if a limit is specified, too far after it.

For some methods, ratios as percentages and error estimates for them are computed for pairs of results. A ratio column header like '1+3' means the ratio of the first result column and the third result column.

Ratios are not computed for Lab Control Sample, Method Blank or Matrix Spike results since their matrices are not necessarily similar to client samples'.

The error estimate for a ratio of results from one planchet reflects only counting errors since other errors should be correlated. For a ratio involving different planchets, if QC limits are computed based on total errors, the error for the ratio allows for the preparation errors for the planchets.

The ratio is underlined (out of spec) if the absolute value of its difference from the nominal value is greater than its error estimate. If no nominal value is specified, this test is not done.

For Gross Alpha or Gross Beta results, there may be a column showing the sum of other Alpha or Beta emitters. This sum includes all relevant

REPORT GUIDES

Page 14

SUMMARY DATA SECTION

Page 31

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-RG
 Version 3.06
 Report date 02/12/10

SDG 7764
Contact N. Joseph Verville

GUIDE, cont.

Client CHPRC
Contract No. 33677
Case no SDG H4154

METHOD SUMMARY

results in the DVD database, whether reported or not. Results in the sum are weighted by a particles/decay value specified by the lab for each relevant analyte. Results less than their MDA are not included. No sums are computed for Lab Control, Method Blank or Matrix Spike samples since their various planchets may not be physically related.

If a ratio of total isotopic to Gross Alpha or Beta is shown, the error for the ratio reflects both the error in the Gross result and the sum, as square root of sum of squares, of the errors in the isotopic results.

For total elemental uranium or thorium results, there may be a column showing the total weight computed from associated isotopic results. Ignoring results less than their MDAs, this is a weighted sum of the isotopic results. The weights depend on the molecular weight and half-life of each isotope so as to convert activities (decays) to weight (atoms).

If a ratio of total computed to measured elemental uranium or thorium is shown, the error for the ratio reflects the errors in all the measurements.

REPORT GUIDES

Page 15

SUMMARY DATA SECTION

Page 32

Lab id EBRLNE
Protocol CHPRC
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 02/12/10

COLLECTOR **KB Hulse**
CHPRC COMPANY CONTACT **WIDRIG, DL** *H4154* TELEPHONE NO. **376-2858** *(7764)* PROJECT COORDINATOR **WIDRIG, DL** PRICE CODE **7B** DATA TURNAROUND **7 Days / 15 Days**

SAMPLING LOCATION **216-N-1 COVERED POND; FTB** PROJECT DESIGNATION **AARA 200-CW-3 Sampling and Analysis - Verification - QC** SAF NO. **F10-034** AIR QUALITY METHOD OF SHIPMENT **FEDERAL EXPRESS**

ICE CHEST NO. **6WS-060** *W24 2/4/10* FIELD LOGBOOK NO. **HNP-N 507-8** ACTUAL SAMPLE DEPTH **N/A** COA **302222E510**

SHIPPED TO **Eberline Services** OFFSITE PROPERTY NO. **SEE PTR 02-04-10mm 00725** BILL OF LADING/AIR BILL NO. **SEE PTR 2-4-10mm 7932 4346 3982**

MATRIX* A=Air DL=Drum L=Liquids DS=Drum Solids L=Liquid O=Oil S=Soil SE=Sediment T=Tissue V=Vegetation W=Water WI=Wipe X=Other	POSSIBLE SAMPLE HAZARDS/ REMARKS Contains Radioactive Material at concentrations that may or may not be regulated for transportation per 49 CFR / IATA Dangerous Goods Regulations but are not releasable per DOE Order 5400.5 (1990/1993)	PRESERVATION HNO3 to pH <2	[Grid area with handwritten '1-29-10' and a large diagonal line through it]
	TYPE OF CONTAINER G/P		
	NO. OF CONTAINER(S) 1		
	VOLUME 1000mL		
	SPECIAL HANDLING AND/OR STORAGE SAMPLE ANALYSIS Method-63;		

SAMPLE NO.	MATRIX*	SAMPLE DATE	SAMPLE TIME	
B23WL8	WATER	JAN 29 2010	0710	✓ N/A
				✗ 1-29-10

CHAIN OF POSSESSION		SIGN/ PRINT NAMES		SPECIAL INSTRUCTIONS	
RELINQUISHED BY/REMOVED FROM KB Hulse CHPRC	DATE/TIME JAN 29 2010 1100	RECEIVED BY/STORED IN SSU #1	DATE/TIME JAN 29 2010 1100	** The CACN for all analytical work at WSCF is 401352. ** The 200 Area S&GRP Characterization and Monitoring Sampling and Analysis GKI applies to this SAF.	
RELINQUISHED BY/REMOVED FROM SSU #1	DATE/TIME FEB 04 2010 1100	RECEIVED BY/STORED IN M. McCoy	DATE/TIME 1100 FEB 04 2010		
RELINQUISHED BY/REMOVED FROM M. McCoy	DATE/TIME 2400 FEB 04 2010	RECEIVED BY/STORED IN Fed Ex	DATE/TIME		
RELINQUISHED BY/REMOVED FROM FED EX	DATE/TIME	RECEIVED BY/STORED IN TF. NATAWATON	DATE/TIME 02/05/10 0915		
RELINQUISHED BY/REMOVED FROM 7	DATE/TIME	RECEIVED BY/STORED IN	DATE/TIME		
RELINQUISHED BY/REMOVED FROM 0	DATE/TIME	RECEIVED BY/STORED IN	DATE/TIME		

ORIGINAL

REVISION 1

LABORATORY SECTION	RECEIVED BY	TITLE	DATE/TIME
FINAL SAMPLE DISPOSITION	DISPOSAL METHOD	DISPOSED BY	DATE/TIME

CH2MHill Plateau Remediation Company		CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST			F10-034-010	PAGE 1 OF 1
COLLECTOR K.J. Young CHPRC	COMPANY CONTACT WIDRIG, DL H4154	TELEPHONE NO. 376-2858 (7764)	PROJECT COORDINATOR WIDRIG, DL		PRICE CODE 7B	DATA TURNAROUND 7 Days / 15 Days
SAMPLING LOCATION 216-N-1 COVERED POND; EB	PROJECT DESIGNATION AARA 200-CW-3 Sampling and Analysis - Verification - QC		SAF NO. F10-034	AIR QUALITY <input type="checkbox"/>		
ICE CHEST NO. W05-060	FIELD LOGBOOK NO. HNF-N-507-8	ACTUAL SAMPLE DEPTH NA	COA 30222ES10	METHOD OF SHIPMENT FEDERAL EXPRESS		
SHIPPED TO Eberline Services	OFFSITE PROPERTY NO. SEE PTR 02-04-10 mm	00725		BILL OF LADING/AIR BILL NO. SEE PTR 2-4-10 mm 7932 4346 3982		
MATRIX* A=Air DL=Drum Liquids DS=Drum Solids L=Liquid O=Oil S=Soil SE=Sediment T=Tissue V=Vegetation W=Water WI=Wipe X=Other	POSSIBLE SAMPLE HAZARDS/ REMARKS Contains Radioactive Material at concentrations that may or may not be regulated for transportation per 49 CFR / IATA Dangerous Goods Regulations but are not releasable per DOE Order 5400.5 (1990/1993)	PRESERVATION HNO3 to pH < 2	TYPE OF CONTAINER G/P	NO. OF CONTAINER(S) 1	VOLUME 1000ml	SAMPLE ANALYSIS nickel-63
SPECIAL HANDLING AND/OR STORAGE		Discard				

SAMPLE NO.	MATRIX*	SAMPLE DATE	SAMPLE TIME						
B23WL9	WATER	JAN 29 2010	0755	NA					

LABORATORY SECTION	RECEIVED BY	TITLE	DATE/TIME
FINAL SAMPLE DISPOSITION	DISPOSAL METHOD	DISPOSED BY	DATE/TIME

CHAIN OF POSSESSION	SIGN/ PRINT NAMES	DATE/TIME
RELINQUISHED BY/REMOVED FROM K.J. Young CHPRC	RECEIVED BY/STORED IN SSU #1	JAN 29 2010 0700
RELINQUISHED BY/REMOVED FROM CHPRC	RECEIVED BY/STORED IN M. McCoy	1100 FEB 04 2010
RELINQUISHED BY/REMOVED FROM M. McCoy	RECEIVED BY/STORED IN Fedex	1400 FEB 04 2010
RELINQUISHED BY/REMOVED FROM Fedex	RECEIVED BY/STORED IN F. W. ...	02/05/10 0915
RELINQUISHED BY/REMOVED FROM 08	RECEIVED BY/STORED IN	
RELINQUISHED BY/REMOVED FROM 01	RECEIVED BY/STORED IN	
RELINQUISHED BY/REMOVED FROM 00	RECEIVED BY/STORED IN	

SPECIAL INSTRUCTIONS

** The CACN for all analytical work at WSCF is 401352.
** The 200 Area S&GRP Characterization and Monitoring Sampling and Analysis GKI applies to this SAF.

ORIGINAL

REVISION 1

CH2M HILL Plateau Remediation Company		CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST			F10-034-011	PAGE 1 OF 1
COLLECTOR KB Hulse CHPRC		COMPANY CONTACT WIDRIG, DL	TELEPHONE NO. 376-2858	PROJECT COORDINATOR WIDRIG, DL	PRICE CODE 78	DATA TURNAROUND 7 Days / 15 Days
SAMPLING LOCATION 216-N-1 COVERED POND; FTB		PROJECT DESIGNATION AARA 200-CW-3 Sampling and Analysis - Verification - QC		SAF NO. F10-034	AIR QUALITY <input type="checkbox"/>	
ICE CHEST NO. GWS-060 <i>MA KEH 2/4/10</i>		FIELD LOGBOOK NO. HNF-N-507-28	ACTUAL SAMPLE DEPTH NA	COA 302222ES10	METHOD OF SHIPMENT FEDERAL EXPRESS	
SHIPPED TO <i>2-4-10</i> Waste Sampling & Characterization <i>Eberline</i>		OFFSITE PROPERTY NO. SEE PTR <i>2-4-10 WMM</i>	00725	BILL OF LADING/AIR BILL NO. SEE PTR <i>2-4-10 WMM</i> 7932 4346 3982		
MATRIX* A=Air DL=Drum Liquids DS=Drum Solids L=Liquid O=Oil S=Soil SE=Sediment T=Tissue V=Vegetation W=Water WI=Wipe X=Other	POSSIBLE SAMPLE HAZARDS/ REMARKS Contains Radioactive Material at concentrations that may or may not be regulated for transportation per 49 CFR / IATA Dangerous Goods Regulations but are not releasable per DOE Order 5400.5 (1990/1993) <i>100219</i>	PRESERVATION None	TYPE OF CONTAINER G	NO. OF CONTAINER(S) 1	VOLUME 250ml	SAMPLE ANALYSIS Trilium - Ion Ex;
SPECIAL HANDLING AND/OR STORAGE						
SAMPLE NO.	MATRIX*	SAMPLE DATE	SAMPLE TIME			
B23WM0	WATER	JAN 29 2010	0710	✓ NA		
CHAIN OF POSSESSION	SIGN/ PRINT NAMES			SPECIAL INSTRUCTIONS		
RELINQUISHED BY/REMOVED FROM KB Hulse	DATE/TIME JAN 29 2010 1115	RECEIVED BY/STORED IN <i>M. Mazira</i>	DATE/TIME JAN 29 2010 1115	** The CACN for all analytical work at WSCF is 401352. ** The 200 Area S&GRP Characterization and Monitoring Sampling and Analysis GKI applies to this SAF.		
RELINQUISHED BY/REMOVED FROM <i>W. D. D...</i>	DATE/TIME 2/3/10 1500	RECEIVED BY/STORED IN <i>KE Hulse</i>	DATE/TIME 2/3/10 1500			
RELINQUISHED BY/REMOVED FROM <i>W. D. D...</i>	DATE/TIME 2/3/10 1515	RECEIVED BY/STORED IN SSU #1	DATE/TIME 2/3/10 1515			
RELINQUISHED BY/REMOVED FROM SSU #1	DATE/TIME 1100 FEB 04 2010	RECEIVED BY/STORED IN <i>M. Mazira</i>	DATE/TIME 1100 FEB 04 2010			
RELINQUISHED BY/REMOVED FROM <i>M. Mazira</i>	DATE/TIME 1400 FEB 04 2010	RECEIVED BY/STORED IN Fed Ex	DATE/TIME			
RELINQUISHED BY/REMOVED FROM <i>Fed Ex</i>	DATE/TIME FEB 04 2010	RECEIVED BY/STORED IN <i>J. W. W...</i>	DATE/TIME 02/05/10 0915			
RELINQUISHED BY/REMOVED FROM <i>Fed Ex</i>	DATE/TIME	RECEIVED BY/STORED IN	DATE/TIME			
LABORATORY SECTION	RECEIVED BY	TITLE	DATE/TIME			
FINAL SAMPLE DISPOSITION	DISPOSAL METHOD	DISPOSED BY	DATE/TIME			

 ORIGINAL

REVISION 1



RICHMOND, CA LABORATORY
SAMPLE RECEIPT CHECKLIST

REVISION 1

Client: CHPRC City RICHMOND State WA
Date/Time received 02/05/10 09:15 POC No. F10-084-008,010,011
Container I.D. No. SWS-060 Requested TAT (Days) 7/15 P.O. Received Yes [] No []

INSPECTION

1. Custody seals on shipping container intact? Yes [✓] No [] N/A []
2. Custody seals on shipping container dated & signed? Yes [✓] No [] N/A []
3. Custody seals on sample containers intact? Yes [✓] No [] N/A []
4. Custody seals on sample containers dated & signed? Yes [✓] No [] N/A []
5. Packing material is: Wet [] Dry [✓]
6. Number of samples in shipping container: 3 Sample Matrix _____
7. Number of containers per sample: 1 (Or see CoC _____)
8. Samples are in correct container Yes [✓] No []
9. Paperwork agrees with samples? Yes [✓] No []
10. Samples have: Tape [] Hazard labels [] Rad labels [] Appropriate sample labels [✓]
11. Samples are: In good condition [✓] Leaking [] Broken Container [] Missing []
12. Samples are: Preserved [✓] Not preserved [] pH 2/6 Preservative _____
13. Describe any anomalies:
RAD LABEL ON B23WNU - only

14. Was P.M. notified of any anomalies? Yes [] No [] Date _____
15. Inspected by RFM/AR Date: 02/05/10 Time: 10:30

Customer Sample No.	Beta/Gamma cpm	Ion Chamber mR/hr	Wipe	Customer Sample No.	Beta/Gamma cpm	Ion Chamber mR/hr	wipe
<u>All Sample</u>	<u>< 60</u>						

Ion Chamber Ser. No. _____ Calibration date _____
Alpha Meter Ser. No. _____ Calibration date _____
Beta/Gamma Meter Ser. No. 100482 Calibration date 5 Aug 2009