

SAF-RC-103
Remaining Sites Confirmation Sampling -
Other Liquid
FINAL DATA PACKAGE

COMPLETE COPY OF DATA PACKAGE TO:

Kathy Wendt H4-21

KW 6/18/08
INITIAL/DATE

COMMENTS:

SDG J00176A

SAF-RC-103

Rad only

Chem only

Rad & Chem

Complete

Partial

Waste Site: 100-H-28:3

RECEIVED
JUN 23 2008
EDMC

Analytical Data Package Prepared For
Washington Closure Hanford



Radiochemical Analysis By
TestAmerica

2800 G.W. Way, Richland Wa, 99354, (509)-375-3131.

Assigned Laboratory Code: TARL

Data Package Contains 18 Pages

Report No.: 39314

Results in this report relate only to the sample(s) analyzed.

SDG No.	Order No.	Client Sample ID (List Order)	Lot-Sa No.	Work Order	Report DB ID	Batch No.
J00176A	RC-103	J16VH6-A	J8F030154-1	KN9AC1AA	9KN9AC10	8155333

Certificate of Analysis

Washington Hanford Closure
2620 Fermi Avenue
Richland, WA 99354

June 17, 2008

Attention: Joan Kessner

SAF Number	:	RC-103
Date SDG Closed	:	June 2, 2008
Number of Samples	:	One (1)
Sample Type	:	Other Liquid
SDG Number	:	J00176A
Data Deliverable	:	15 -Day / Summary

CASE NARRATIVE

I. Introduction

On June 2, 2008 one other liquid sample was received at TestAmerica for chemistry analysis. Upon receipt, the sample was assigned the following laboratory ID number to correspond with the Washington Closure Hanford (WCH) specific ID:

<u>WCH ID#</u>	<u>STLR ID#</u>	<u>MATRIX</u>	<u>DATE OF RECEIPT</u>
J16VH6-A	KN9AC	OTHER LIQUID	6/02/08

II. Sample Receipt

The sample was received in good condition. There was no sample date on the COC. The sample date was taken from the sample bottle label. The client was contacted and the monitor's logbook confirmed the sample label had the correct sample date. No other anomalies were noted during check-in.

III. Analytical Results/Methodology

The analytical results for this report are presented by laboratory sample ID. Each set of data includes sample identification information, analytical results and the appropriate associated statistical errors.

The requested analyses were:

Chemical Analysis
Hexavalent Chromium by EPA method 7196A

IV. Quality Control

The analytical results for each analysis performed includes a minimum of one laboratory control sample (LCS), one method (reagent) blank, and one duplicate sample analysis. Any exceptions have been noted in the "Comments" section.

QC and sample results are reported in the same units.

V. Comments

Chemical Analysis

Hexavalent Chromium by EPA method 7196A:

Sample J16VH6 was approximately 75% other liquid and 25% other solid. The sample was centrifuged. The client was notified on 6/3/08. Instructions were as follows:

Liquid Fraction

Analyze the other liquid fraction as J16VH6-A in SDG J00176A. Use a 50 ml aliquot for the sample and sample duplicate. Do not analyze a matrix spike or matrix spike duplicate. Also include a blank and LCS.


Soil Fraction

Analyze the other solid fraction as J16VH6 in SDG J00176. Use 2.5 gram aliquot for the sample, sample duplicate and matrix spike. Also include a blank and LCS. Analyze percent moisture after sample has been analyzed, if there is sufficient sample volume.

The LCS, batch blank, sample and sample duplicate (J16VH6-A results are within contractual requirements.

I certify that this Certificate of Analysis is in compliance with the SOW, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the Laboratory Manager, or a designee as verified by the following signature.

Reviewed and approved:


Sandra Seger

Project Manager

Drinking Water Method Cross References

DRINKING WATER ASTM METHOD CROSS REFERENCES		
Referenced Method	Isotope(s)	TestAmerica Richland's SOP No.
EPA 901.1	Cs-134, I-131	RICH-RC-5017
EPA 900.0	Alpha & Beta	RICH-RC-5014
EPA 00-02	Gross Alpha (Coprecipitation)	RICH-RC-5021
EPA 903.0	Total Alpha Radium (Ra-226)	RICH-RC-5027
EPA 903.1	Ra-226	RICH-RC-5005
EPA 904.0	Ra-228	RICH-RC-5005
EPA 905.0	Sr-89/90	RICH-RC-5006
ASTM D5174	Uranium	RICH-RC-5058
EPA 906.0	Tritium	RICH-RC-5007

Results in this report relate only to the sample(s) analyzed.

Uncertainty Estimation

TestAmerica Richland has adopted the internationally accepted approach to estimating uncertainties described in "NIST Technical Note 1297, 1994 Edition". The approach, "Law of Propagation of Errors", involves the identification of all variables in an analytical method which are used to derive a result. These variables are related to the analytical result (R) by some functional relationship, $R = \text{constants} * f(x,y,z,\dots)$. The components (x,y,z) are evaluated to determine their contribution to the overall method uncertainty. The individual component uncertainties (u_i) are then combined using a statistical model that provides the most probable overall uncertainty value. All component uncertainties are categorized as type A, evaluated by statistical methods, or type B, evaluated by other means. Uncertainties not included in the components, such as sample homogeneity, are combined with the component uncertainty as the square root of the sum-of-the-squares of the individual uncertainties. The uncertainty associated with the derived result is the combined uncertainty (u_c) multiplied by the coverage factor (1,2, or 3).

When three or more sample replicates are used to derive the analytical result, the type A uncertainty is the standard deviation of the mean value (S/\sqrt{n}), where S is the standard deviation of the derived results. The type B uncertainties are all other random or non-random components that are not included in the standard deviation.

The derivation of the general "Law of Propagation of Errors" equations and specific example are available on request.

Report Definitions

Action Lev	An agreed upon activity level used to trigger some action when the final result is greater than or equal to the Action Level. Often the Action Level is related to the Decision Limit.
Batch	The QC preparation batch number that relates laboratory samples to QC samples that were prepared and analyzed together.
Bias	Defined by the equation (Result/Expected)-1 as defined by ANSI N13.30.
COC No	Chain of Custody Number assigned by the Client or TestAmerica.
Count Error (#s)	Poisson counting statistics of the gross sample count and background. The uncertainty is absolute and in the same units as the result. For Liquid Scintillation Counting (LSC) the batch blank count is the background.
Total Uncert (#s) u_c - Combined Uncertainty.	All known uncertainties associated with the preparation and analysis of the sample are propagated to give a measure of the uncertainty associated with the result, u_c the combined uncertainty. The uncertainty is absolute and in the same units as the result.
(#s), Coverage Factor	The coverage factor defines the width of the confidence interval, 1, 2 or 3 standard deviations.
CRDL (RL)	Contractual Required Detection Limit as defined in the Client's Statement Of Work or TestAmerica "default" nominal detection limit. Often referred to the reporting level (RL)
Lc	Decision Level based on instrument background or blank, adjusted by the Efficiency, Chemical Yield, and Volume associated with the sample. The Type I error probability is approximately 5%. $Lc = (1.645 * \text{Sqrt}(2 * (\text{BkgndCnt}/\text{BkgndCntMin})/\text{SCntMin})) * (\text{ConvFct}/(\text{Eff} * \text{Yld} * \text{Abn} * \text{Vol}) * \text{IngrFct})$. For LSC methods the batch blank is used as a measure of the background variability. Lc cannot be calculated when the background count is zero.
Lot-Sample No	The number assigned by the LIMS software to track samples received on the same day for a given client. The sample number is a sequential number assigned to each sample in the Lot.
MDC MDA	Detection Level based on instrument background or blank, adjusted by the Efficiency, Chemical Yield, and Volume with a Type I and II error probability of approximately 5%. $MDC = (4.65 * \text{Sqrt}((\text{BkgndCnt}/\text{BkgndCntMin})/\text{SCntMin}) + 2.71/\text{SCntMin}) * (\text{ConvFct}/(\text{Eff} * \text{Yld} * \text{Abn} * \text{Vol}) * \text{IngrFct})$. For LSC methods the batch blank is used as a measure of the background variability.
Primary Detector	The instrument identifier associated with the analysis of the sample aliquot.
Ratio U-234/U-238	The U-234 result divided by the U-238 result. The U-234/U-238 ratio for natural uranium in NIST SRM 4321C is 1.038.
Rst/MDC	Ratio of the Result to the MDC. A value greater than 1 may indicate activity above background at a high level of confidence. Caution should be used when applying this factor and it should be used in concert with the qualifiers associated with the result.
Rst/TotUcert	Ratio of the Result to the Total Uncertainty. If the uncertainty has a coverage factor of 2 a value greater than 1 may indicate activity above background at approximately the 95% level of confidence assuming a two-sided confidence interval. Caution should be used when applying this factor and it should be used in concert with the qualifiers associated with the result.
Report DB No	Sample Identifier used by the report system. The number is based upon the first five digits of the Work Order Number.
RER	The equation Replicate Error Ratio = $(S-D)/[\text{sqrt}(\text{TPUs}^2 + \text{TPUd}^2)]$ as defined by ICPT BOA where S is the original sample result, D is the result of the duplicate, TPUs is the total uncertainty of the original sample and TPUd is the total uncertainty of the duplicate sample.
SDG	Sample Delivery Group Number assigned by the Client or assigned by TestAmerica upon sample receipt.
Sum Rpt Alpha Spec Rst(s)	The sum of the reported alpha spec results for tests derived from the same sample excluding duplicate result where the results are in the same units.
Work Order	The LIMS software assign test specific identifier.
Yield	The recovery of the tracer added to the sample such as Pu-242 used to trace a Pu-239/40 method.

Sample Results Summary

Date: 17-Jun-08

TestAmerica TARL

Ordered by Method, Batch No., Client Sample ID.

Report No. : 39314

SDG No: J00176A

Batch	Client Id Work Order	Parameter	Result +/- Uncertainty (2s)	Qual	Units	Tracer Yield	MDC or MDA	CRDL	RPD
8155333	7196_CR6								
	J16VH6-A								
	KN9AC1AA	HEXCHROME	2.00E-03 +/- 0.00E+00	U	mg/L	N/A	2.00E-03	2.00E-03	
	KN9AC1AC	HEXCHROME	2.00E-03 +/- 0.00E+00	U	mg/L	N/A	2.00E-03	2.00E-03	0.0
No. of Results:		2							

TestAmerica

RPD - Relative Percent Difference.

rptSTLRchSaSum
mary2 V5.1.6
A2002

U Qual - Analyzed for but not detected above limiting criteria. Limit criteria is less than the Mdc/Mda or Total Uncert or not identified by gamma scan software.

QC Results Summary
TestAmerica TARL
 Ordered by Method, Batch No, QC Type,.

Date: 17-Jun-08

Report No. : 39314

SDG No.: J00176A

Batch	Work Order	Parameter	Result +- Uncertainty (2s)	Qual	Units	Tracer Yield	LCS Recovery	Bias	MDC MDA
7196_CR6	8155333 LCS,								
	KN9MN1AC	HEXCHROME	5.08E-01 +- 0.00E+00		mg/L	N/A	102%	0.0	2.00E-03
	8155333 BLANK QC,								
	KN9MN1AA	HEXCHROME	2.00E-03 +- 0.00E+00	U	mg/L	N/A			2.00E-03
No. of Results: 2									

TestAmerica Bias - (Result/Expected)-1 as defined by ANSI N13.30.
 rptSTLRchQcSummary V5.1.6 A2002 U Qual - Analyzed for but not detected above limiting criteria. Limit criteria is less than the Mdc/Mda or Total Uncert or not identified by gamma scan software.

FORM I

Date: 17-Jun-08

SAMPLE RESULTS

Lab Name: TestAmerica
 Lot-Sample No.: J8F030154-1
 Client Sample ID: J16VH6-A

SDG: J00176A
 Report No.: 39314
 COC No.: RC-103-83

Collection Date: 6/2/2008 1:00:00 PM
 Received Date: 6/2/2008 2:55:00 PM
 Matrix: WATER

Ordered by Client Sample ID, Batch No.

Parameter	Result	Qual	Count Error (2 s)	Total Uncert(2 s)	MDC MDA, Action Lev	Rpt Unit, Lc	Yield CRDL(RL)	Rst/MDC, Rst/TotUcert	Analysis, Prep Date	Total Sa Size	Aliquot Size	Primary Detector
Batch: 8155333	7196_CR6				Work Order: KN9AC1AA		Report DB ID: 9KN9AC10					
HEXCHROME	2.00E-03	U		0.0E+00	2.00E-03	mg/L	N/A	1.	6/3/08		100.0	
							2.00E-03	N/A			ML	
No. of Results:	1	Comments:										

FORM II

Date: 17-Jun-08

DUPLICATE RESULTS

Lab Name: TestAmerica

SDG: J00176A

Collection Date: 6/2/2008 1:00:00 PM

Lot-Sample No.: J8F030154-1

Report No.: 39314

Received Date: 6/2/2008 2:55:00 PM

Client Sample ID: J16VH6-A

COC No.: RC-103-83

Matrix: WATER

Parameter	Result, Orig Rst	Qual	Count Error (2 s)	Total Uncert(2 s)	MDC MDA, Action Lev	Rpt Unit, CRDL	Yield	Rst/MDC, Rst/TotUcert	Analysis, Prep Date	Total Sa Size	Aliquot Size	Primary Detector
Batch: 8155333	7196_CR6			Work Order: KN9AC1AC		Report DB ID: KN9AC1ER			Orig Sa DB ID: 9KN9AC10			
HEXCHROME	2.00E-03	U		0.0E+00	2.00E-03	mg/L	N/A	1.	6/3/08		100.0	
	2.00E-03	U	RPD	0.0		2.00E-03		N/A			ML	

No. of Results: 1 Comments:

TestAmerica RPD - Relative Percent Difference.

rptSTLRchDupV5.1 MDC|MDA,Lc - Detection, Decision Level based on instrument background or blank, adjusted by the sample Efficiency, Yield, and Volume.

.6 A2002 U Qual - Analyzed for but not detected above limiting criteria. Limit criteria is less than the Mdc/Mda or Total Uncert or not identified by gamma scan software.

FORM II
BLANK RESULTS

Date: 17-Jun-08

Lab Name: TestAmerica

SDG: J00176A

Matrix: WATER

Report No. : 39314

Parameter	Result	Qual	Count Error (2 s)	Total Uncert(2 s)	MDC MDA ,	Rpt Unit, CRDL	Yield	Rst/MDC, Rst/TotUcert	Analysis, Prep Date	Total Sa Size	Aliquot Size	Primary Detector
Batch: 8155333	7196_CR6				Work Order: KN9MN1AA			Report DB ID: KN9MN1AB				
HEXCHROME	2.00E-03	U		0.0E+00	2.00E-03	mg/L	N/A	1.	6/3/08		100.0	
						2.00E-03		N/A			ML	

No. of Results: 1

Comments:

TestAmerica
rptSTLRchBlank
V5.1.6 A2002

MDC|MDA,Lc - Detection, Decision Level based on Instrument background or blank, adjusted by the sample Efficiency, Yield, and Volume.
U Qual - Analyzed for but not detected above limiting criteria. Limit criteria is less than the Mdc/Mda or Total Uncert or not identified by gamma scan software.

FORM II
LCS RESULTS

Date: 17-Jun-08

Lab Name: TestAmerica

SDG: J00176A

Matrix: WATER

Report No.: 39314

Parameter	Result	Count Qual Error (2 s)	Total Uncert(2 s)	MDC MDA	Report Unit	Yield	Expected	Expected Uncert	Recovery, Bias	Analysis, Prep Date	Aliquot Size	Primary Detector
Batch: 8155333	7196_CR6				Work Order: KN9MN1AC		Report DB ID: KN9MN1AS					
HEXCHROME	5.08E-01		0.0E+00	2.00E-03	mg/L	N/A	5.00E-01		102%	6/3/08	100.0	
						Rec Limits:	85	115	0.0		ML	
No. of Results:	1	Comments:										

TestAmerica Bias - (Result/Expected)-1 as defined by ANSI N13.30.

rptSTLRchLcs
V5.1.6 A2002

Batch Number(s): 8155333 J8F 030154 J00176A Due 6/17				
Lab Sample Numbers or				
Method/Test/Parameter: Cr+6 in Water / RICH-WC-5003				
Review Item	Yes (✓)	No (✓)	N/A (✓)	2 nd Level Review (✓)
A. Initial Calibration	✓			✓
1. Performed at required frequency with required number of levels?	✓			✓
2. Correlation coefficient within QC limits?	✓			✓
3. Initial calibration verification (ICV) analyzed immediately after calibration and results within QC limits?	✓			✓
4. Initial calibration blank (ICB) analyzed immediately after ICV and concentrations of all parameters ≤ reporting limit?	✓			✓
B. Continuing Calibration	✓			✓
1. CCV analyzed at required frequency and all parameters within QC limits?	✓			✓
2. CCB analyzed at required frequency and all results ≤ reporting limit?	✓			✓
C. Sample Analysis	✓			✓
1. Were any samples with concentrations above the linear range for any parameter diluted and reanalyzed?	✓			✓
2. Were all sample holding times met?	✓			✓
D. QC Samples	✓			✓
1. All results for the preparation blank below limits?	✓			✓
2. MS or MS/MSD recoveries within QC limits and %RPD (for MSD) acceptable?	✓			✓
3. LCS percent recovery within QC limits and %RPD (for LCSD) acceptable?	✓			✓
4. Analytical spikes within QC limits where applicable?			✓	N/A ✓
5. ICP only: One serial dilution performed per SDG?			✓	✓
6. ICP only: CRDL standard (CRI or CRA) analyzed at required frequency?			✓	✓
7. ICP only: Interference check samples (ICSA, ICSAB) and HICAL analyzed at the required frequencies and within QC limits?			✓	✓

Review Item	Yes (✓)	No (✓)	N/A (✓)	2 nd Level Review (✓)
E. Other	✓			NO NCM ✓
1. Are all nonconformances included and noted?				✓
2. Is the correct date and time of analysis shown?	✓			✓
3. Did the analyst sign and date the front page of the analytical run?	✓			✓
4. Correct methodology used?	✓			✓
5. Transcriptions checked?	✓			✓
6. Calculations checked at minimum frequency?	✓			✓
7. Units checked?	✓			✓

Comments on any "No" response:

Analyst: *Leah Ait*
 Second-Level Review: *David Gje*

Date: 6/4/08
 Date: 6/16/08



Sample Check-in List

Date/Time Received: 6208 1455 GM Screen Result 0.2K

Client: WCH SDG #: J00176A NA [] SAF #: RC-103 NA []

Work Order Number: J8F020187 Chain of Custody # RC-103-083
SPS 6/3/08

Shipping Container ID: N/A Air Bill # N/A

1. Custody Seals on shipping container intact? NA [] Yes No []
2. Custody Seals dated and signed? NA [] Yes No []
3. Chain of Custody record present? NA [] Yes No []
4. Cooler Temperature: _____ NA 5. Vermiculite/packing materials is NA Wet [] Dry []

6. Number of samples in shipping container: 1

7. Sample holding times exceeded? NA Yes [] No []

8. Samples have:
 Tape Hazard Lables
 Custody Seals Appropriate Sample Lables


9. Samples are:
 In Good Condition Leaking
 Broken Have Air Bubbles
(Only for samples requiring no head space.)

10. Sample pH taken? NA [] pH < 2 [] pH > 2 pH > 9 [] Amount HNO₃ Added _____

11. Sample Location, Sample Collector Listed? *
 *For documentation only. No corrective action needed.

12. Were any anomalies identified in sample receipt? Yes [] No

13. Description of anomalies (include sample numbers): _____

Sample Custodian:  Date: 6208

Client Sample ID	Analysis Requested	Condition	Comments/Action

Client Informed on _____ by _____ Person Contacted _____

[] No action necessary; process as is.

Project Manager _____ Date _____

TESTAMERICA

6/3/2008 11:54:40 AM

Sample Preparation/Analysis

Balance Id:

127642, Washington Closure Hanford
Bechtel Hanford, Inc.

88 NO SAMPLE PREPARATION PERFORMED / DIRECT INJECTION
EA Chromium, Hexavalent (7196A)
5I CLIENT: HANFORD

Pipet #:

AnalyDueDate: 06/17/2008

Sep1 DT/Tm Tech:

Batch: 8155333

mg/L

PM, Quote: SS , 27023

Sep2 DT/Tm Tech:

SEO Batch, Test: None All Tests: 8155333 88EA,

Prep Tech:



Work Order, Lot, Sample Date/Time	Total Amt/Unit	Initial Aliquot Amt/Unit	QC Tracer Prep Date	Count Time Min	Detector Id	Count On Off (24hr) Circle	CR Analyst, Init/Date	Comments:
-----------------------------------	----------------	--------------------------	---------------------	----------------	-------------	------------------------------	-----------------------	-----------

1 KN9AC-1-AA

J8F030154-1-SAMP

06/02/2008 13:00		AmtRec: 200 ML	#Containers: 1			Scr:	Alpha:	Beta:
------------------	--	----------------	----------------	--	--	------	--------	-------

2 KN9AC-1-AC-X

J8F030154-1-DUP

06/02/2008 13:00		AmtRec: 200 ML	#Containers: 1			Scr:	Alpha:	Beta:
------------------	--	----------------	----------------	--	--	------	--------	-------

3 KN9MN-1-AA-B

J8F030000-333-BLK

06/02/2008 13:00		AmtRec:	#Containers: 1			Scr:	Alpha:	Beta:
------------------	--	---------	----------------	--	--	------	--------	-------

4 KN9MN-1-AC-C

J8F030000-333-LCS

06/02/2008 13:00		AmtRec:	#Containers: 1			Scr:	Alpha:	Beta:
------------------	--	---------	----------------	--	--	------	--------	-------

Comments:

All Clients for Batch:

127642, Washington Closure Hanford

Bechtel Hanford, Inc.

, SS , 27023

KN9AC1AA-SAMP Constituent List:

KN9MN1AA-BLK:

KN9MN1AC-LCS:

KN9AC1AA-SAMP Calc Info:

TESTAMERICA

6/3/2008 11:54:41 AM

Sample Preparation/Analysis

Balance Id: _____

88 NO SAMPLE PREPARATION PERFORMED / DIRECT INJECTION
EA Chromium, Hexavalent (7196A)
5I CLIENT: HANFORD

Pipet #: _____

AnalyDueDate: 06/17/2008

Sep1 DT/Tm Tech: _____

Batch: 8155333
 SEQ Batch, Test: None

mg/L

Sep2 DT/Tm Tech: _____

Prep Tech: _____



Work Order, Lot, Sample DateTime	Total Amt/Unit	Initial Aliquot Amt/Unit	QC Tracer Prep Date	Count Time Min	Detector Id	Count On Off (24hr) Circle	CR Analyst, Init/Date	Comments:
Uncert Level (#s) : 2	Decay to SaDt: Y	Blk Subt.: N	Sci.Not.: Y	ODRs: B				
KN9MN1AA-BLK:								
Uncert Level (#s) : 2	Decay to SaDt: Y	Blk Subt.: N	Sci.Not.: Y	ODRs: B				
KN9MN1AC-LCS:								
Uncert Level (#s) : 2	Decay to SaDt: Y	Blk Subt.: N	Sci.Not.: Y	ODRs: B				

Approved By _____ Date: _____

17

Analyst:		Calibration Curve Information				SOP Information		BATCH #		
L. Dinh						RICH-WC-5003		8155333		
Start Date: 6/3/2008						Revision 7		SDG # J00176A		
Start Time:		Blank	0.000	0.000	0.000			Matrix Water		
End Date: 6/3/2008		Std. 1	0.100	0.050	0.100					
End Time:		Std. 2	0.500	0.250	0.495					
Analyst Signature: <i>L. Dinh</i>		Std. 3	0.750	0.375	0.731	MDL (mg/L) 0.002		Instrument Information		
Date: 6/4/08		Std. 4	1.500	0.750	1.439			Instrument: Hach DR2010		
		Std 5	2.000	1.000	1.888			Wavelength: 540		
		Standard Volume (mL):			100.000			R Squared 0.99976		
		Date of Curve:			6/3/2008			Slope: 1.89023		
								Intercept: 0.01153		
		Calibration Information:		ICV Information:		LCS Information:		Matrix Spike Information:		
Dilution ID #		Cr-08-00115		Cr-08-00116		Cr-08-00115		Cr-08-00115		
Prep Date:		06/03/08		06/03/08		06/03/08		06/03/08		
Concentration (mg/L)		50		50		50		50		
Expiration Date:		06/04/08		06/04/08		06/04/08		06/04/08		
Pipettor(s)		70,190		190		190		190		
Volume Used (Expected Value				1.000	0.50000	1.00	0.50000	0.50	0.26316	
Expected values are only amounts added in mg and not final concentrations										
Sample ID	Client ID	Type	Sample Volume (mL)	Sample ABS.	Blank ABS.	Corrected ABS.	Dilution Factor	Curve Conc. (mg/L)	Final Conc. (mg/L)	% Rec.
n/a	n/a	ICV	100.000	0.982	0.000	0.982	1	0.5134	0.513	102.68%
n/a	n/a	ICB	100.000	0.000	0.000	0.000	1	<MDL	<MDL	
KN9MN1AA	n/a	Prep Blank	100.000	-0.001		-0.001	1	<MDL	<MDL	
KN9MN1AC	n/a	LCS	100.000	0.971		0.971	1	0.5076	0.508	101.52%
KN9AC1AA	J16VH6-A	Sample	100.000	0.054	0.072	-0.018	1	<MDL	<MDL	
KN9AC1AC-S	J16VH6-MS	MS	100.000				1			
KN9AC1AD-D	J16VH6-MSD	MSD	100.000				1			
KN9AC1AE-X	J16VH6-DUP	Duplicate	100.000	0.053	0.071	-0.018	1	<MDL	<MDL	
			100.000				1			
			100.000				1			
			100.000				1			
			100.000				1			
n/a	n/a	CCV	100.000	0.982		0.982	1	0.5134	0.513	102.68%
n/a	n/a	CCB	100.000	0.001		0.001	1	-0.0056	<MDL	
			100.000				1			
			100.000				1			
			100.000				1			
			100.000				1			