

START

9613478.1103

BOOHY7-K25-086
MARTIN MARIETTA

0045335

6/12

MARTIN MARIETTA ENERGY SYSTEMS, INC.

POST OFFICE BOX 2003
OAK RIDGE, TENNESSEE 37831-7440

June 12, 1991

Ms. Joan Kessner
Westinghouse Hanford Company
2355 Stevens Drive
Richland, Washington 99352

Dear Ms. Kessner:

AnaLis Data Reports for Project 91-013: 300-FF-1 Nonhazardous Source Characterization

Attached are the AnaLis data reports for two (2) 300-FF-1 Nonhazardous Source Characterization samples, Project 91-013, received into the K-25 Site Analytical Chemistry Department (ACD) laboratories on March 19, and April 1, 1991. Full CLP format was requested for the data deliverables package and is being generated at this time. However, since the CLP packaging of data requires a substantial delay in returning data, I am releasing the data available at this time in AnaLis report format. The AA metals and the mercury data is not entered into the AnaLis database at all for analyses performed to CLP protocol. That data will not be available until the full CLP package is completed.

Sincerely,

Deborah L. Amburgey

Deborah L. Amburgey
Program Manager
Hanford Support Program



Attachments

cc/attach: D.L.Amburgey
S.R.Smith - RC



9613478.1104

Oak Ridge K-25 Site
Analytical Chemistry Department
Results of Analyses

Date Printed:
7-JUN-1991 10:58

AnalIS ID: 910320-106 Project: G132 013 Customer Sample ID: BOOHY7
Customer: KESSNER/KRAMER Requisition Number:
Date Sampled: 14-MAR-1991 Date Sample Received: 19-MAR-1991
Sampled By: Date Sample Completed: 31-MAY-1991
Material Description: SOIL FOR CLP [] : Result has been Corrected for Spike

Activ. Number	Procedure No.	Analysis	Result	Units	Analyst	QA File Number	Date Completed
	EPA CLP SOW	Arsenic	-----	mg/Kg	KM COOLEY	10403Y	11-APR-1991
	EPA-204.2(CLP)						
	EPA CLP SOW	Lead	-----	mg/Kg	KM COOLEY	10403Z1	20-APR-1991
	EPA-239.2(CLP)						
	EPA-245.5(CLP)	Mercury	-----	mg/Kg	SA BURGESS	10403AC	23-APR-1991
	EPA CLP SOW	Selenium	-----	mg/Kg	KM COOLEY	10403AA1	20-APR-1991
	EPA-270.2(CLP)						
	EPA CLP SOW	Thallium	-----	mg/Kg	KM COOLEY	10403AB	18-APR-1991
	EPA-279.2(CLP)						
090703	EPA-3050	Boron	111	mg/Kg	VF BELT	10531Y	31-MAY-1991
	EPA-200.7						
	EPA-3050	Uranium	<6.0	mg/Kg	VF BELT	10531Y	31-MAY-1991
	EPA-200.7						

135003 CLP10/86 SVD10 Prep (BNA- CLP) R CC JM MAHER 2646/2745 9-APR-1991

Prep (BNA- CLP)

Analyst = JM MAHER
Date Extracted = 9-APR-1991
Sample Weight Extracted (g) = 10
Percent Solids = 56.43
Calculated Dried Weight (g) = 5.64
Extraction Method = Soxhlet
Extraction Solvent = Methylene Chloride/Acetone
Extraction Cleanup = Sodium Sulfate
Final Volume of Extract (mL) = 1
Associated Blank = 910401-175

Spike Recovery Data

Analysis	Unspike Result	Amount Spike	Spike Result	Units	Amount Recovered	Percent Recovered
ANTIMONY	0.0123	0.120	0.108	mg/L	0.096	79.8
SILVER	0.00200	0.0500	0.0490	mg/L	0.0470	94.0
BARIUM	1.94	2.00	3.75	mg/L	1.81	90.5
BERYLLIUM	0.00280	0.0500	0.0483	mg/L	0.0455	91.0
CADMIUM	0.0	0.0500	0.0454	mg/L	0.0454	90.8
COBALT	0.0130	0.500	0.477	mg/L	0.464	92.8
CHROMIUM	0.0210	0.200	0.211	mg/L	0.190	95.0
COPPER	0.121	0.250	0.345	mg/L	0.224	89.6
MANGANESE	0.0816	0.500	0.541	mg/L	0.459	91.9

9613478.1105

NICKEL	0.114	0.500	0.566 mg/L	0.452	90.4
VANADIUM	0.378	0.500	0.808 mg/L	0.430	86.0
ZINC	0.113	0.500	0.555 mg/L	0.442	88.4

- Program Manager: D. L. Amburgey
Date Approved:

AnalIS ID: 910320-106
 Laboratory: Organic Mass Spectroscopy Laboratory
 File ID: >11285
 Instrument ID: 5970-3
 Authorized By: D. C. Canada

Customer Sample ID: BOOHY7
 Customer: KESSNER/KRAMER
 Sample Matrix: SOIL
 Requisition Number:
 Date Sample Received: 19-MAR-1991

BNA - Base/Neutral/Acid Compounds (CLP)

Date Extracted/Prepared: 8-APR-1991
 Preparation Procedure Number: CLP10/86 SVD10
 Percent Moisture: 43.6
 Percent Moisture (decanted):
 Associated Blank: 910401-175
 [] : Result has been Corrected for Spike

Date Analyzed: 11-APR-1991
 Analysis Procedure Number: ACD 2470
 Dilution Factor: 2.0
 Analyst: C MEEHAN
 QA File Number: 910411CM-0088

CAS		ug/Kg	CAS		ug/Kg
108-95-2	Phenol	3500U	106-47-8	4-Chloroaniline	3500U
111-44-4	bis(2-Chloroethyl)ether	3500U	87-68-3	Hexachlorobutadiene	3500U
95-57-8	2-Chlorophenol	3500U	59-50-7	4-Chloro-3-methylphenol	3500U
541-73-1	1,3-Dichlorobenzene	3500U	91-57-6	2-Methylnaphthalene	1800 J
106-46-7	1,4-Dichlorobenzene	3500U	77-47-4	Hexachlorocyclopentadiene	3500U
100-51-6	Benzyl Alcohol	3500U	88-06-2	2,4,6-Trichlorophenol	3500U
95-50-1	1,2-Dichlorobenzene	3500U	95-95-4	2,4,5-Trichlorophenol	17000U
95-48-7	2-Methylphenol	3500U	91-58-7	2-Chloronaphthalene	3500U
108-60-1	bis(2-Chloroisopropyl)ether	3500U	88-74-4	2-Nitroaniline	17000U
106-44-5	4-Methylphenol	3500U	131-11-3	Dimethylphthalate	3500U
621-64-7	N-Nitroso-di-n-propylamine	3500U	208-96-8	Acenaphthylene	3500U
67-72-1	Hexachloroethane	3500U	99-09-2	3-Nitroaniline	17000U
98-95-3	Nitrobenzene	3500U	83-32-9	Acenaphthene	3500U
78-59-1	Isophorone	3500U	51-28-5	2,4-Dinitrophenol	17000U
88-75-5	2-Nitrophenol	3500U	100-02-7	4-Nitrophenol	17000U
105-67-9	2,4-Dimethylphenol	3500U	132-64-9	Dibenzofuran	3500U
65-85-0	Benzoic Acid	17000U	121-14-2	2,4-Dinitrotoluene	3500U
111-91-1	bis(2-Chloroethoxy)methane	3500U	606-20-2	2,6-Dinitrotoluene	3500U
120-83-2	2,4-Dichlorophenol	3500U	84-66-2	Diethylphthalate	3500U
120-82-1	1,2,4-Trichlorobenzene	3500U	7005-72-3	4-Chlorophenyl-phenylether	3500U
91-20-3	Naphthalene	920 J	86-73-7	Fluorene	3500U

Data Reporting Qualifiers:

- U - Compound was analyzed for but not detected. The number is the attainable detection limit for the sample.
- B - Analyte was found in the reagent blank as well as the sample.
- J - Indicates an estimated value.
- ND - Not detected.
- A - Aldol condensation product.
- D - Secondary dilution.
- E - Exceeds initial calibration range.

ANALIS ID: 910320-106
 Laboratory: Organic Mass Spectroscopy Laboratory
 File ID: >11285
 Instrument ID: 5970-3
 Authorized By: D. C. Canada

Customer Sample ID: BOOHY7
 Customer: KESSNER/KRAMER
 Sample Matrix: SOIL
 Requisition Number:
 Date Sample Received: 19-MAR-1991

BNA - Base/Neutral/Acid Compounds (CLP)

Date Extracted/Prepared: 8-APR-1991
 Preparation Procedure Number: CLP10/86 SVD10
 Percent Moisture: 43.6
 Percent Moisture (decanted):
 Associated Blank: 910401-175
 [] : Result has been Corrected for Spike

Date Analyzed: 11-APR-1991
 Analysis Procedure Number: ACD 2470
 Dilution Factor: 2.0
 Analyst: C MEEHAN
 QA File Number: 910411CM-0088

CAS		ug/Kg	CAS		ug/Kg
100-01-6	4-Nitroaniline	17000U	53-70-3	Dibenz(a,h)anthracene	3500U
534-52-1	4,6-Dinitro-2-methylphenol	17000U	191-24-2	Benzo(g,h,i)perylene	3500U
86-30-6	N-Nitrosodiphenylamine	3500U			
101-55-3	4-Bromophenyl-phenylether	3500U			
118-74-1	Hexachlorobenzene	3500U			
87-86-5	Pentachlorophenol	17000U			
85-01-8	Phenanthrene	3500U			
120-12-7	Anthracene	3500U			
84-74-2	Di-n-butylphthalate	3500U			
206-44-0	Fluoranthene	3500U			
129-00-0	Pyrene	3500U			
85-68-7	Butylbenzylphthalate	3500U			
91-94-1	3,3'-Dichlorobenzidine	7000U			
56-55-3	Benzo(a)anthracene	3500U			
117-81-7	bis(2-Ethylhexyl)phthalate	530 J8			
218-01-9	Chrysene	3500U			
117-84-0	Di-n-octylphthalate	3500U			
205-99-2	Benzo(b)fluoranthene	3500U			
207-08-9	Benzo(k)fluoranthene	3500U			
50-32-8	Benzo(a)pyrene	3500U			
193-39-5	Indeno(1,2,3-cd)pyrene	3500U			

Data Reporting Qualifiers:

- U - Compound was analyzed for but not detected. The number is the attainable detection limit for the sample.
- B - Analyte was found in the reagent blank as well as the sample.
- J - Indicates an estimated value.
- ND - Not detected.
- A - Aldol condensation product.
- D - Secondary dilution.
- E - Exceeds initial calibration range.

Oak Ridge K-25 Site
Analytical Chemistry Department

Date Printed:
7-JUN-1991 10:58

BNA (CLP) SOIL ORGANICS ANALYSIS DATA SHEETS
TENTATIVELY IDENTIFIED COMPOUNDS

AnalIS ID: 910320-106
Laboratory: Organic Mass Spectroscopy Laboratory
Sample Matrix: SOIL
Level: (low/med): LOW
Dilution Factor: 2.0
% Moisture: not dec. 43.6 dec. _____
Extraction: (SepF/Cont/Sonc) SoxH
GPC Cleanup: (Y/N) N

Customer Sample ID: BOOHY7
Customer: KESSNER/KRAMER
File ID: >11285
Date Received: 19-MAR-1991
Date Analyzed: 11-APR-1991
Date Extracted: 8-APR-1991
pH: _____

Number TICs found: 19

Concentration Units
(ug/L or ug/Kg): ug/Kg

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1. 123-42-2	2-Pentanone, 4-hydroxy-4-methy	6.33	150000	JAB
2.	NAPHTHALENE, DIMETHYL	16.37	4500	JY
3.	SAT'D HYDROCARBON	16.72	5200	J
4.	NAPHTHALENE, TRIMETHYL	18.25	8000	J
5.	NAPHTHALENE, TRIMETHYL	18.59	3000	J
6.	UNKNOWN (ALKYL AZULENE?)	19.65	6800	J
7.	ALKYL HYDROCARBON	19.77	15000	J
8.	UNKNOWN (ALKYL AZULENE)	20.38	3800	J
9. 57-10-3	Hexadecanoic acid	22.56	3400	JB
10.	SAT'D HYDROCARBON	25.78	4700	J
11.	ALKYL HYDROCARBON	26.65	5700	J
12.	ALKYL HYDROCARBON	27.49	7700	J
13.	ALKYL HYDROCARBON	28.29	6000	J
14.	SAT'D HYDROCARBON	29.06	12000	J
15.	SAT'D HYDROCARBON	29.86	4700	J
16.	SAT'D HYDROCARBON	30.71	7000	J
17.	SAT'D HYDROCARBON	32.77	5700	J
18.	UNKNOWN (BENZENAMINE?)	33.73	4700	J
19.	UNKNOWN	35.00	4000	J
20.				
21.				
22.				
23.				
24.				
25.				
26.				
27.				
28.				
29.				
30.				

Data Reporting Qualifiers:

- U - Compound was analyzed for but not detected. The number is the attainable detection limit for the sample.
- B - Analyte was found in the reagent blank as well as the sample.
- J - Indicates an estimated value.
- ND - Not detected.
- A - Aldol condensation product.
- D - Secondary dilution.
- E - Exceeds initial calibration range.

9613478.1109

P - Probable Identification.

ANALYSIS DATA REPORT

Date Printed:
7-JUN-1991 10:58

AnalIS ID: 910320-106
Laboratory: Inductively Coupled Plasma Laboratory
File ID:
Instrument ID:
Authorized By: M. J. Scheuer

Customer Sample ID: BOOHY7
Customer: KESSNER/KRAMER
Sample Matrix: SOIL
Requisition Number:
Date Sample Received:

18 Element Scan by ICP Spectrometry - CLP

Date Extracted/Prepared:
Preparation Procedure Number: EPA-3050
Percent Moisture:
Percent Moisture (decanted):
Associated Blank:
[] : Result has been Corrected for Spike

Date Analyzed: 31-MAY-1991
Analysis Procedure Number: EPA-200.7
Dilution Factor:
Analyst: VF BELT
QA File Number: 10531Y

CAS		mg/Kg	CAS		mg/Kg
7429-90-5	Aluminum	6810			
7440-36-0	Antimony	10.0U			
7440-39-3	Barium	388			
7440-41-7	Beryllium	0.56			
7440-43-9	Cadmium	0.60U			
7440-70-2	Calcium	4680			
7440-47-3	Chromium	4.2			
7440-48-4	Cobalt	2.6			
7440-50-8	Copper	24.1			
7439-89-6	Iron	5750			
7439-95-4	Magnesium	627			
7439-96-5	Manganese	16.3			
7440-02-0	Nickel	22.7			
7440-09-7	Potassium	371			
7440-22-4	Silver	1.2U			
7440-23-5	Sodium	703			
7440-62-2	Vanadium	75.5			
7440-66-6	Zinc	22.6			

Oak Ridge K-25 Site
Analytical Chemistry Department
Results of Analyses

Date Printed:
7-JUN-1991 11:08

AnalIS ID: 910408-131 Project: G132 013 Customer Sample ID: BOOH23
Customer: KESSNER/KRAMER Requisition Number:
Date Sampled: 1-APR-1991 Date Sample Received: 6-APR-1991
Sampled By: Date Sample Completed:
Material Description: SOIL FOR CLP ☐ : Result has been Corrected for Spike

Activ. Number	Procedure No.	Analysis	Result	Units	Analyst	QA File Number	Date Completed
	EPA CLP SOW	Arsenic	-----	mg/Kg			
	EPA CLP SOW	Lead	-----	mg/Kg			
	EPA-245.5(CLP)	Mercury	-----	mg/Kg	SA BURGESS	10411U	23-APR-1991
	EPA CLP SOW	Selenium	-----	mg/Kg			
	EPA CLP SOW	Thallium	-----	mg/Kg			
090703	EPA-3050	Boron	1.0	mg/Kg	VF BELT	10531Y	31-MAY-1991
	EPA-200.7						
	EPA-3050	Uranium	6.2	mg/Kg	VF BELT	10531Y	31-MAY-1991
	EPA-200.7						
135003	CLP10/86 SVD10 Prep (BNA- CLP)		C		JM MAHER	2745	9-APR-1991

Prep (BNA- CLP)

Analyst = JM MAHER
Date Extracted = 9-APR-1991
Sample Weight Extracted (g) = 10
Percent Solids = 77.36
Calculated Dried Weight (g) = 7.74
Extraction Method = Soxhlet
Extraction Solvent = Methylene Chloride/Acetone
Extraction Cleanup = Sodium Sulfate
Final Volume of Extract (mL) = 1
Associated Blank = 910401-175

Program Manager: D. L. Amburgey
Date Approved:

ANALYSIS DATA REPORT

Date Printed:
7-JUN-1991 11:08

AnalIS ID: 910408-131
 Laboratory: Organic Mass Spectroscopy Laboratory
 File ID: >11288
 Instrument ID: 5970-3
 Authorized By: D. C. Canada

Customer Sample ID: BOOH23
 Customer: KESSNER/KRAMER
 Sample Matrix: SOIL
 Requisition Number:
 Date Sample Received: 6-APR-1991

BNA - Base/Neutral/Acid Compounds (CLP)

Date Extracted/Prepared: 8-APR-1991
 Preparation Procedure Number: CLP10/86 SVD10
 Percent Moisture: 32.6
 Percent Moisture (decanted):

Date Analyzed: 11-APR-1991
 Analysis Procedure Number: ACD 2470
 Dilution Factor: 1.0
 Analyst: C MEEHAN
 QA File Number: 910411CM-0088

Associated Blank: 910401-175

[] : Result has been Corrected for Spike

CAS		ug/Kg	CAS		ug/Kg
108-95-2	Phenol	1500U	106-47-8	4-Chloroaniline	1500U
111-44-4	bis(2-Chloroethyl)ether	1500U	87-68-3	Hexachlorobutadiene	1500U
95-57-8	2-Chlorophenol	1500U	59-50-7	4-Chloro-3-methylphenol	1500U
541-73-1	1,3-Dichlorobenzene	1500U	91-57-6	2-Methylnaphthalene	1500U
106-46-7	1,4-Dichlorobenzene	1500U	77-47-4	Hexachlorocyclopentadiene	1500U
100-51-6	Benzyl Alcohol	1500U	88-06-2	2,4,6-Trichlorophenol	1500U
95-50-1	1,2-Dichlorobenzene	1500U	95-95-4	2,4,5-Trichlorophenol	7100U
95-48-7	2-Methylphenol	1500U	91-58-7	2-Chloronaphthalene	1500U
108-60-1	bis(2-Chloroisopropyl)ether	1500U	88-74-4	2-Nitroaniline	7100U
106-44-5	4-Methylphenol	1500U	131-11-3	Dimethylphthalate	1500U
621-64-7	N-Nitroso-di-n-propylamine	1500U	208-96-8	Acenaphthylene	1500U
67-72-1	Hexachloroethane	1500U	99-09-2	3-Nitroaniline	7100U
98-95-3	Nitrobenzene	1500U	83-32-9	Acenaphthene	1500U
78-59-1	Isophorone	1500U	51-28-5	2,4-Dinitrophenol	7100U
88-75-5	2-Nitrophenol	1500U	100-02-7	4-Nitrophenol	7100U
105-67-9	2,4-Dimethylphenol	1500U	132-64-9	Dibenzofuran	1500U
65-85-0	Benzoic Acid	7100U	121-14-2	2,4-Dinitrotoluene	1500U
111-91-1	bis(2-Chloroethoxy)methane	1500U	606-20-2	2,6-Dinitrotoluene	1500U
120-83-2	2,4-Dichlorophenol	1500U	84-66-2	Diethylphthalate	1500U
120-82-1	1,2,4-Trichlorobenzene	1500U	7005-72-3	4-Chlorophenyl-phenylether	1500U
91-20-3	Naphthalene	1500U	86-73-7	Fluorene	1500U

Data Reporting Qualifiers:

- U - Compound was analyzed for but not detected. The number is the attainable detection limit for the sample.
- B - Analyte was found in the reagent blank as well as the sample.
- J - Indicates an estimated value.
- ND - Not detected.
- A - Aldol condensation product.
- D - Secondary dilution.
- E - Exceeds initial calibration range.

ANALYSIS DATA REPORT

Date Printed:
7-JUN-1991 11:08

AnalIS ID: 910408-131
 Laboratory: Organic Mass Spectroscopy Laboratory
 File ID: >11288
 Instrument ID: 5970-3
 Authorized By: D. C. Canada

Customer Sample ID: BOOH23
 Customer: KESSNER/KRAMER
 Sample Matrix: SOIL
 Requisition Number:
 Date Sample Received: 6-APR-1991

BNA - Base/Neutral/Acid Compounds (CLP)

Date Extracted/Prepared: 8-APR-1991
 Preparation Procedure Number: CLP10/86 SVD10
 Percent Moisture: 32.6
 Percent Moisture (decanted):
 Associated Blank: 910401-175
 [] : Result has been Corrected for Spike

Date Analyzed: 11-APR-1991
 Analysis Procedure Number: ACD 2470
 Dilution Factor: 1.0
 Analyst: C MEEHAN
 QA File Number: 910411CM-0088

CAS		ug/Kg	CAS		ug/Kg
100-01-6	4-Nitroaniline	7100U	53-70-3	Dibenz(a,h)anthracene	1500U
534-52-1	4,6-Dinitro-2-methylphenol	7100U	191-24-2	Benzo(g,h,i)perylene	1500U
86-30-6	N-Nitrosodiphenylamine	1500U			
101-55-3	4-Bromophenyl-phenylether	1500U			
118-74-1	Hexachlorobenzene	1500U			
87-86-5	Pentachlorophenol	7100U			
85-01-8	Phenanthrene	1500U			
120-12-7	Anthracene	1500U			
84-74-2	Di-n-butylphthalate	1500U			
206-44-0	Fluoranthene	1500U			
129-00-0	Pyrene	1500U			
85-68-7	Butylbenzylphthalate	1500U			
91-94-1	3,3'-Dichlorobenzidine	2900U			
56-55-3	Benzo(a)anthracene	1500U			
117-81-7	bis(2-Ethylhexyl)phthalate	660 J			
218-01-9	Chrysene	1500U			
117-84-0	Di-n-octylphthalate	1500U			
205-99-2	Benzo(b)fluoranthene	1500U			
207-08-9	Benzo(k)fluoranthene	1500U			
50-32-8	Benzo(a)pyrene	1500U			
193-39-5	Indeno(1,2,3-cd)pyrene	1500U			

Data Reporting Qualifiers:

- U - Compound was analyzed for but not detected. The number is the attainable detection limit for the sample.
- B - Analyte was found in the reagent blank as well as the sample.
- J - Indicates an estimated value.
- ND - Not detected.
- A - Aldol condensation product.
- D - Secondary dilution.
- E - Exceeds initial calibration range.

9517470.1114

Oak Ridge K-25 Site
Analytical Chemistry DepartmentDate Printed:
7-JUN-1991 11:08BNA (CLP) SOIL ORGANICS ANALYSIS DATA SHEETS
TENTATIVELY IDENTIFIED COMPOUNDS

AnalIS ID: 910408-131
Laboratory: Organic Mass Spectroscopy Laboratory
Sample Matrix: SOIL
Level: (low/med): LOW
Dilution Factor: 1.0
% Moisture: not dec. 32.6 dec.
Extraction: (SepF/Cont/Sonc) SoxH
GPC Cleanup: (Y/N) N

Customer Sample ID: BOOHZ3
Customer: KESSNER/KRAMER
File ID: >11288
Date Received: 6-APR-1991
Date Analyzed: 11-APR-1991
Date Extracted: 8-APR-1991
pH:

Number TICs found: 8Concentration Units
(ug/L or ug/Kg): ug/Kg

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1. 141-79-7	3-Penten-2-one, 4-methyl-	5.35	690	JAB
2. 123-42-2	2-Pentanone, 4-hydroxy-4-methy	6.49	120000	JAB
3.	UNKNOWN (HYDROCARBON)	22.07	3500	JB
4. 57-10-3	Hexadecanoic acid	22.57	1300	JB
5.	UNKNOWN	24.34	1700	JB
6.	UNKNOWN	26.95	8400	J
7.	UNKNOWN	30.18	750	JB
8.	UNKNOWN	31.26	650	J
9.				
10.				
11.				
12.				
13.				
14.				
15.				
16.				
17.				
18.				
19.				
20.				
21.				
22.				
23.				
24.				
25.				
26.				
27.				
28.				
29.				
30.				

Data Reporting Qualifiers:

- U - Compound was analyzed for but not detected. The number is the attainable detection limit for the sample.
- B - Analyte was found in the reagent blank as well as the sample.
- J - Indicates an estimated value.
- ND - Not detected.
- A - Aldol condensation product.
- D - Secondary dilution.
- E - Exceeds initial calibration range.

9613478.1115

P - Probable Identification.

AnalIS ID: 910408-131
Laboratory: Inductively Coupled Plasma Laboratory
File ID:
Instrument ID:
Authorized By: M. J. Scheuer

Customer Sample ID: BOOHZ3
Customer: KESSNER/KRAMER
Sample Matrix: SOIL
Requisition Number:
Date Sample Received:

18 Element Scan by ICP Spectrometry - CLP

Date Extracted/Prepared:
Preparation Procedure Number: EPA-3050
Percent Moisture:
Percent Moisture (decanted):
Associated Blank:
[] : Result has been Corrected for Spike

Date Analyzed: 31-MAY-1991
Analysis Procedure Number: EPA-200.7
Dilution Factor:
Analyst: VF BELT
QA File Number: 10531Y

CAS		mg/Kg	CAS		mg/Kg
7429-90-5	Aluminum	9890			
7440-36-0	Antimony	8.7U			
7440-39-3	Barium	69.5			
7440-41-7	Beryllium	0.37			
7440-43-9	Cadmium	0.52U			
7440-70-2	Calcium	3960			
7440-47-3	Chromium	11.9			
7440-48-4	Cobalt	8.9			
7440-50-8	Copper	40.2			
7439-89-6	Iron	15800			
7439-95-4	Magnesium	4500			
7439-96-5	Manganese	350			
7440-02-0	Nickel	9.2			
7440-09-7	Potassium	982			
7440-22-4	Silver	1.0U			
7440-23-5	Sodium	206			
7440-62-2	Vanadium	36.9			
7440-66-6	Zinc	68.8			

MARTIN MARIETTA ENERGY SYSTEMS, INC.

POST OFFICE BOX 2003
OAK RIDGE, TENNESSEE 37831

December 6, 1991

Ms. Joan Kessner
Westinghouse Hanford Company
Office of Sample Management
2355 Stevens Drive
Richland, Washington, 99352

Dear Ms. Kessner:

Inorganic Analysis CLP Package on Project 91-013: 300-FF-1 Nonhazardous Source Characterization

Attached are the results of the inorganic analyses on the 300-FF-1 Nonhazardous Source Characterization samples, SDG# BOOHY7/BOOHZ3, Project 91-013, received into the K-25 Site Analytical Chemistry Department (ACD) laboratories on March 19, 1991, and April 6, 1991. Also attached are copies of the Chain of Custody records for the samples, and a sample identification table. The results are reported in CLP format for the inorganic analyses. All data quality objectives were satisfied on this project.

I certify that this data package is in compliance with the terms and conditions of the OSM's revised Statement of Work and letter dated December 20, 1990, both technically and for completeness, for other than any conditions detailed in the forms. Release of the data contained in this hardcopy data package has been authorized by the Laboratory Manager or his designee, as verified by the following signature.

Sincerely,

Deborah L. Amburgey
Deborah L. Amburgey
Program Manager
Hanford Support Program

Clarence R. Kirkpatrick
Clarence R. Kirkpatrick
Program Manager
Waste Management Analysis

Roy W. Morrow
Roy W. Morrow

Department Manager
K-25 Site Analytical Chemistry Department

Attachments

cc/attach: D.L.Amburgey
S.R.Smith - RC
cc: N.P.Buddin
H.H.Sullivan

*Copies w/ missing
signatures received by
OSM on Dec. 7, 1991.*

*These pages are attached
to the letter dated
Jan. 23, 1992.*

Dois Ayres
3-15-94

December 05, 1991

Joan Kessner
Project Manager
Westinghouse Hanford Company
Richland, WA 99336

Dear Ms. Kessner:

Between March 19 and April 6, 1991 the K-25 (ORK25P) Analytical Chemistry Department received the following soil samples for analysis:


<u>Laboratory ID</u>	<u>Customer Sample ID</u>	<u>EPA ID</u>
910320-106	BOOHY7	BOOHY7
910408-131	BOOHZ3	BOOHZ3

The above samples were designated as ICP SDG number BOOHY7.

ICP calibration verifications	SPEX Multielement Standards (SPEX QC-19, SPEX QC-7)
ICP interference check standards	UNLV-QAL ICS-A(1084) and ICS-B(1084)
ICP soil laboratory control std	UNLV-QAL LCS(0287)
ICP CRDL standard	Perkin Elmer Multielement Standard (PE Pure No. N930-0225)
ICP Calibration Standards	
A. STD1 = Matrix Matched Standard Blank	
B. STD2 = SPEX Multielement Standard (XORN130)	
C. STD3 = SPEX Multielement Standard (XORN131)	
D. STD4 = SPEX Multielement Standard (XORN119, XORN132)	
E. STD5 = SPEX Multielement Standard (XORN123)	
F. STD6 = SPEX Multielement Standard (XORN125)	
G. STD7 = SPEX Multielement Standard (XORN126)	
H. STD8 = SPEX Multielement Standard (XORN121)	

All associated QA/QC was within the criteria specified by CLP. However these two samples were incorrectly digested and analyzed as one Sample Delivery Group(SDG). CLP protocol requires that samples received more than fourteen days apart must be grouped in separate SDG's. These samples were received within a period of eighteen days.


Sincerely




M.J. Scheuer

cc: M.S. Dill
R.W. Morrow
File - MJS - NoRC

This data has been reviewed and is approved for release.



R.W. Morrow, Dept. Head



M.S. Dill, Methodology and QA

December 5, 1991

Joan Kessner
Project Manager
Westinghouse Hanford Company
Richland, WA 99336

Dear Ms. Kessner,

On April 6, 1991 the Oak Ridge K-25 Site (K-25) Analytical Chemistry Department (ACD) received 1 soil sample from Westinghouse Hanford Company. This sample was grouped into a Sample Delivery Group following EPA CLP protocols. The K-25 sample identification number was assigned as follows:

Laboratory ID	Customer ID	EPA Sample Number
910408-131	BOOHZ3	BOOHZ3

A narrative of the Atomic Spectrometry and Mercury Analysis laboratories experiences and problems in the preparation and analysis of the sample is given below:

The one (1) sample within this Sample Delivery Group was prepared and analyzed following EPA-CLP protocols. The Inorganic Analysis Data Sheet, Form I, the arsenic and lead analyte results are flagged with "S" and "N" Q flags, respectively. The arsenic result was derived from method of standard additions and is therefore flagged "S", the lead pre-digest spike recovery is not within the 25% QC range and is flagged "N" according to CLP protocol. There is no analytical explanation for the low lead recovery, as other QC samples recovered within required ranges. Due to another problem with the form generation software, the Analysis Run Log, Form XIV, graphite furnace analytical spike recoveries were hand entered. The analytical spike recoveries have been transcribed from the software directly on the forms for all the appropriate Form XIV's. No other unusual problems were encountered with this sample set.


The following quality control solutions were used for the analysis of these materials:

GFAAS initial calibration verification	EPA ICV-2 (0590) and ICV-4 (0389)
GFAAS continuing calibration verification	Perkin-Elmer CLP standard N930-0221
GFAAS solid laboratory control sample	EPA LCS (0287)
Hg calibration verification	EPA ICV-5 (0788)
Hg continuing calibration verification	J.T. Baker Hg standard 6934-01
Hg solid laboratory control sample	EPA LCS (0287)

All values on all forms have been rounded to the appropriate number of significant figures in

accordance with the 10/89 revision of the EPA CLP statement of work, SOW 7/88. All data qualifier flags, C and Q field, are consistent with requirements of SOW 7/88. All calculated results shown on forms are derived from the rounded values given on the forms, not from the original raw data.


Sincerely,




Thomas J. Oatts

cc: M. S. Dill
R. W. Morrow
File - TJO - NoRC

This data has been reviewed and is approved for release.



R. W. Morrow, Dept. Head



M. S. Dill, Methodology and QA

December 5, 1991

Joan Kessner
Project Manager
Westinghouse Hanford Company
Richland, WA 99336

Dear Ms. Kessner,

On March 19, 1991 the Oak Ridge K-25 Site (K-25) Analytical Chemistry Department (ACD) received 1 soil sample from Westinghouse Hanford Company. This sample was grouped into a Sample Delivery Group following EPA CLP protocols. The K-25 sample identification number was assigned as follows:

Laboratory ID	Customer ID	EPA Sample Number
910320-106	BOOHY7	BOOHY7

A narrative of the Atomic Spectrometry and Mercury Analysis laboratories experiences and problems in the preparation and analysis of the sample is given below:

The one (1) sample within this Sample Delivery Group was prepared and analyzed following EPA-CLP protocols. The Inorganic Analysis Data Sheet, Form I, the lead and selenium analyte results are flagged with Q flags. The lead result was derived from method of standard additions and is therefore flagged "S" and the duplicate precision was outside the CLP allowable range which generates a "***" flag. The selenium pre-digest spike recovery was outside the 25% QC range and is flagged "N" according to CLP protocol and the sample result was derived from a MSA run which creates the "S" flag. Because CLP run protocols are specific, the analytical sample and duplicate are subject to possible MSA determinations, whereas the pre-digest spike sample is not. Therefore the Form I values which are derived from MSA determinations are compared to non-MSA results for the spike. This may yield low pre-digest spike recoveries like those reported with this sample. The lead pre-digest spike did not recover, however due to the high analyte concentration no Q flag is required.

Form II (Part 1), Initial and Continuing Calibration Verification, shows that two different initial calibration verification (ICV) standards were used for lead determinations. The percent recoveries on Form II are correct for the two Pb ICV's. Form VI, Duplicates, shows Q flags for lead and mercury duplicates. The lead value is flagged correctly, however the mercury is flagged with a "***" flag incorrectly. The Q flag on Form I has been removed to reflect the correct analytical result however the flag on Form VI could not be altered in the software. Please disregard this flag. Due to another problem with the form generation software, the Analysis Run Log, Form XIV, graphite furnace analytical spike

9615479.1113

recoveries were hand entered. The analytical spike recoveries have been transcribed from the software directly on the forms for all the appropriate Form XIV's. No other unusual problems were encountered with this sample set.

The following quality control solutions were used for the analysis of these materials:

GFAAS initial calibration verification	EPA ICV-2 (0590), ICV-4 (0389), and Perkin-Elmer CLP standard N930-0224
GFAAS continuing calibration verification	Perkin-Elmer CLP standard N930-0221
GFAAS solid laboratory control sample	EPA LCS (0287)
Hg calibration verification	EPA ICV-5 (0788)
Hg continuing calibration verification	J.T. Baker Hg standard 6934-01
Hg solid laboratory control sample	EPA LCS (0287)

All values on all forms have been rounded to the appropriate number of significant figures in accordance with the 10/89 revision of the EPA CLP statement of work, SOW 7/88. All data qualifier flags, C and Q field, are consistent with requirements of SOW 7/88. All calculated results shown on forms are derived from the rounded values given on the forms, not from the original raw data.


Sincerely,




Thomas J. Oatts

cc: M. S. Dill
R. W. Morrow
File - TJO - NoRC

This data has been reviewed and is approved for release.



R. W. Morrow, Dept. Head



M. S. Dill, Methodology and QA

9613978.124

SAMPLE IDENTIFICATION TABLE FOR SDG# BOOHY7/BOOHZ3
300-FF-1 NONHAZARDOUS SOURCE CHARACTERIZATION SAMPLES

Date Sample Received	OSM Sample ID	Lab Sample ID	Matrix	Comments
3/19/91	BOOHY7	910320-106	fly ash	Matrix determination obtained from Chain of Custody information
4/06/91	BOOHZ3	910408-131	soil	Matrix determination obtained from Chain of Custody information

COVER PAGE - INORGANIC ANALYSES DATA PACKAGE

Lab Name: MARTIN MARIETTA

Contract:

Lab Code: K25

Case No. :

SAS No.:

SDG No.: BOOHY7

SOW No.: 7/88

Sample No.

BOOHY7

BOOHY7D

BOOHY7S

BOOHZ3

Lab Sample ID.

910320-106

910320-106

910320-106

910408-131

Were ICP interelement corrections applied?

Yes/No YES

Were ICP background corrections applied?

Yes/No YES

If yes, were raw data generated before application of background corrections?

Yes/No NO

Comments:

I certify that this data package is in compliance with the terms, and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hardcopy data package and in the computer-readable data submitted on floppy diskette has been authorized by the Laboratory Manager or the Manager's designee, as verified by the following signature.

Signature: Mark J. Scherer

Name: Mark J. Scheuer

Date: 12/5/91

Title: ICP Lab Supervisor

COVER PAGE - INORGANIC ANALYSES DATA PACKAGE

Lab Name: Oak Ridge K-25 Site

Contract:

Lab Code: SPEC

Case No.:

SAS No.:

SDG No.: BOOHY7

SOW No.: 7/88

Sample No.

BOOHY7

BOOHY7D

BOOHY7S

Lab Sample ID.

910320-106

Duplicate

13

Were ICP interelement corrections applied?

Yes/No NO

Were ICP background corrections applied?

Yes/No NO

If yes, were raw data generated before
application of background corrections?

Yes/No NO

Comments:

I certify that this data package is in compliance with the terms, and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hardcopy data package and in the computer-readable data submitted on floppy diskette has been authorized by the Laboratory Manager or the Manager's designee, as verified by the following signature.

Signature:

Thomas J. Ertts

Name:

Thomas J. Ertts

Date:

12-5-91

Title:

Supervisor

COVER PAGE - INORGANIC ANALYSES DATA PACKAGE

Lab Name: Oak Ridge K-25 Site

Contract:

Lab Code: SPEC

Case No. :

SAS No.:

SDG No.: BOOHZ3

SOW No.: 7/88

Sample No.

BOOHZ3

BOOHZ3D

BOOHZ3S

Lab Sample ID.

910408-131

duplicate

spike

Were ICP interelement corrections applied?

Yes/No NO

Were ICP background corrections applied?

Yes/No NO

If yes, were raw data generated before application of background corrections?

Yes/No NO

Comments:

I certify that this data package is in compliance with the terms, and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hardcopy data package and in the computer-readable data submitted on floppy diskette has been authorized by the Laboratory Manager or the Manager's designee, as verified by the following signature.

Signature:

Thomas J. Outh

Name:

Thomas J. Oatts

Date:

12-5-91

Title:

Supervisor

1
INORGANIC ANALYSIS DATA SHEET

BOOHY7

Lab Name: MARTIN MARIETTA

Contract:

Lab Code: K25

Case No.:

SAS No.:

SDG No.: BOOHY7

Matrix (soil/water): SOIL

Lab Sample ID: 910320-106

Level (low/med): LOW

Date Received: 03/19/91

% Solids: 70.6

Concentration Units (ug/L or mg/kg dry weight): MG/KG

CAS No.	Analyte	Concentration	C	Q	M
7429-90-5	Aluminum	9650	—		P
7440-36-0	Antimony	14.1	U	N	P
7440-39-3	Barium	550	—		P
7440-41-7	Beryllium	0.79	B		P
7440-43-9	Cadmium	0.85	U		P
7440-70-2	Calcium	6640	—		P
7440-47-3	Chromium	5.9	—		P
7440-48-4	Cobalt	3.7	B		P
7440-50-8	Copper	34.1	—		P
7439-89-6	Iron	8140	—		P
7439-95-4	Magnesium	892	B		P
7439-96-5	Manganese	23.1	—		P
7440-02-0	Nickel	32.1	—		P
7440-09-7	Potassium	526	B		P
7440-22-4	Silver	1.7	U		P
7440-23-5	Sodium	1040	B		P
7440-62-2	Vanadium	107	—		P
7440-66-6	Zinc	32.9	—	E	P
	Uranium	8.5	U		P
	Boron	158	—		P

Color Before: BLACK

Clarity Before:

Texture: COURSE

Color After: GREY

Clarity After: CLEAR

Artifacts:

Comments:

K-25 Analytical Chemistry Department ANALIS ID; 910320-106

ENVIROFORMS/CLP 788

SAMPLE NO.

1
INORGANIC ANALYSIS DATA SHEET

BOOHY7

Lab Name: Oak Ridge K-25 Site

Contract:

Lab Code: SPEC

Case No.:

SAS No.:

SDG No.: BOOHY7

Matrix (soil/water): SOIL

Lab Sample ID: 910320-106

Level (low/med): LOW

Date Received: 03/19/91

% Solids: 70.6

Concentration Units (ug/L or mg/kg dry weight): MG/KG

CAS No.	Analyte	Concentration	C	Q	M
7440-38-2	Arsenic	3.8	-		F
7439-92-1	Lead	47.4	-	S*	F
7782-49-2	Selenium	2.7	-	NS	F
7440-28-0	Thallium	1.5	U		F
7439-97-6	Mercury	0.14	U		CV

Color Before: BLACK

Clarity Before: N/A

Texture: COARSE

Color After: BLACK

Clarity After: N/A

Artifacts: YES

Comments:

Sample contained rocks

K-25 ACD AnaLIS ID number 910320-106

Run under EPA CLP QC protocol. Package generated following CLP format

ENVIROFORMS/CLP 788

SAMPLE NO.

1
INORGANIC ANALYSIS DATA SHEET

BOOHZ3

Lab Name: MARTIN MARIETTA

Contract:

Lab Code: K25

Case No.:

SAS No.:

SDG No.: BOOHY7

Matrix (soil/water): SOIL

Lab Sample ID: 910408-131

Level (low/med): LOW

Date Received: 04/06/91

% Solids: 80.9

Concentration Units (ug/L or mg/kg dry weight): MG/KG

CAS No.	Analyte	Concentration	C	Q	M
7429-90-5	Aluminum	12100	—		P
7440-36-0	Antimony	10.8	U	N	P
7440-39-3	Barium	85.9	—		P
7440-41-7	Beryllium	0.47	B		P
7440-43-9	Cadmium	0.65	U		P
7440-70-2	Calcium	4900	—		P
7440-47-3	Chromium	14.8	—		P
7440-48-4	Cobalt	11.0	—		P
7440-50-8	Copper	49.7	—		P
7439-89-6	Iron	19500	—		P
7439-95-4	Magnesium	5570	—		P
7439-96-5	Manganese	432	—		P
7440-02-0	Nickel	11.4	—		P
7440-09-7	Potassium	1210	—		P
7440-22-4	Silver	1.3	U		P
7440-23-5	Sodium	288	B		P
7440-62-2	Vanadium	45.6	—		P
7440-66-6	Zinc	85.8	—	E	P
	Uranium	7.6	—		P
	Boron	1.3	—		P

Color Before: BROWN

Clarity Before:

Texture: COURSE

Color After: GREY

Clarity After: CLOUDY

Artifacts:

Comments:

K-25 Analytical Chemistry Department ANALIS ID; 910408-131

ENVIROFORMS/CLP 788

SAMPLE NO.

1
INORGANIC ANALYSIS DATA SHEET

BOOHZ3

Lab Name: Oak Ridge K-25 Site

Contract:

Lab Code: SPEC

Case No.:

SAS No.:

SDG No.: BOOHZ3

Matrix (soil/water): SOIL

Lab Sample ID: 910408-131

Level (low/med): LOW

Date Received: 04/06/91

% Solids: 80.9

Concentration Units (ug/L or mg/kg dry weight): MG/KG

CAS No.	Analyte	Concentration	C	Q	M
7440-38-2	Arsenic	4.1	-	S	F
7439-92-1	Lead	6.4	-	N	F
7782-49-2	Selenium	0.49	U		F
7440-28-0	Thallium	1.4	U		F
7439-97-6	Mercury	0.12	U		CV

Color Before: BROWN

Clarity Before: N/A

Texture: COARSE

Color After: BROWN

Clarity After: N/A

Artifacts: YES

Comments:

Sample contained rocks

K-25 ACD AnaLIS ID number 910408-131

Run under EPA-CLP Quality Control protocol, reported as CLP package

CASE NARRATIVE
for GC/MS data

Laboratory Name: MARTIN MARIETTA (K-25 SITE)
Case # : G132-013
BNA SDG # : BOOHY7
Contract # : 0288

Sample Identification:

EPA SAMPLE ID	LAB SAMPLE ID	LAB FILE ID	MATRIX	COMMENT
=====	=====	=====	=====	=====

Semivolatile soil samples (SDG # BOOHY7)

SBLKJM	910401-175	>11231	soil	method blank
BOOHY7	910320-106	>11232	soil	sample
BOOHYMS	910401-176	>11233	soil	sample matrix spike
BOOHYMSD	910401-177	>11234	soil	sample matrix spike dup
BOOHZ3	910408-131	>11288	soil	sample
SBLKJM2	910401-175	>11284	soil	rerun method blank
BOOHY7RE	910320-106	>11285	soil	rerun sample
BOOHY7REMS	910401-176	>11286	soil	rerun matrix spike
BOOHY7REMSD	910401-177	>11287	soil	rerun matrix spike dup

SEMIVOLATILE WATER SDG # BOOHY7 COMMENTS:

The values of the surrogates for samples BOOHY7, BOOHY7MS and BOOHY7MSD with the original prep were outside of contract required QC limits (see Form 2D). Therefore the samples were reprepared and reanalyzed. All surrogate standards for the reanalyzed samples were within percent recovery acceptance criteria. No pH analyses were performed or reported for the samples. All other QC, DFTPP tune, "CCC" and "SPCC" components and internal standard areas were within acceptance criteria.

This GC/MS hardcopy data package is in compliance with the terms and conditions of the 2/88 organic CLP SOW, both technically and for completeness, to the best of my knowledge, for other than the conditions detailed above. Release of the data contained in this hardcopy data package has been authorized by the Laboratory Manager or his designee, as verified by the following signature.

SIGNATURE

Cyrus Meehan
Cyrus Meehan, Ph.D
Chemist, GC/MS Laboratory

DATE

11/3/91
November 3, 1991

15

EPA SAMPLE NO.

1B
SEMIVOLATILE ORGANICS ANALYSIS DATA SHEET

BOOHY7

Lab Name: MARTIN MARIETTA

Contract: 0288

Lab Code: K25

Case No.: G132-013SAS No.: NA

SDG No.: ~~NA~~ ^{BOOHY7} 5/15/91

Matrix: (soil/water) SOIL

Lab Sample ID: 910320-106

Sample wt/vol: 10 (g/mL) G

Lab File ID: >11232

Level: (low/med) LOW

Date Received: 03/19/91

% Moisture: not dec. 43.6 dec. NA

Date Extracted: 03/27/91

Extraction: (Sepf/Cont/Sonc) CONT

Date Analyzed: 4/02/91

GPC Cleanup: (Y/N) N pH: NA

Dilution Factor: 1.00000

CAS NO.	COMPOUND	CONCENTRATION UNITS: (ug/L or ug/Kg) ug/KG	Q
108-95-2	Phenol	1800.	IU
111-44-4	bis(2-Chloroethyl)Ether	1800.	IU
95-57-8	2-Chlorophenol	1800.	IU
541-73-1	1,3-Dichlorobenzene	1800.	IU
106-46-7	1,4-Dichlorobenzene	1800.	IU
100-51-6	Benzyl alcohol	1800.	IU
95-50-1	1,2-Dichlorobenzene	1800.	IU
95-48-7	2-Methylphenol	1800.	IU
39638-32-9	bis(2-chloroisopropyl)ether	1800.	IU
106-44-5	4-Methylphenol	1800.	IU
621-64-7	N-Nitroso-Di-n-propylamine	1800.	IU
67-72-1	Hexachloroethane	1800.	IU
98-95-3	Nitrobenzene	1800.	IU
78-59-1	Isophorone	1800.	IU
88-75-5	2-Nitrophenol	1800.	IU
105-67-9	2,4-Dimethylphenol	1800.	IU
65-85-0	Benzoic acid	8900.	IU
111-91-1	bis(2-Chloroethoxy)methane	1800.	IU
120-83-2	2,4-Dichlorophenol	1800.	IU
120-82-1	1,2,4-Trichlorobenzene	1800.	IU
91-20-3	Naphthalene	360.	IJ
106-47-8	4-Chloroaniline	1800.	IU
87-68-3	Hexachlorobutadiene	1800.	IU
59-50-7	4-Chloro-3-methylphenol	1800.	IU
91-57-6	2-Methylnaphthalene	670.	IJ
77-47-4	Hexachlorocyclopentadiene	1800.	IU
88-06-2	2,4,6-Trichlorophenol	1800.	IU
95-95-4	2,4,5-Trichlorophenol	8900.	IU
91-58-7	2-Chloronaphthalene	1800.	IU
88-74-4	2-Nitroaniline	8900.	IU
131-11-3	Dimethylphthalate	1800.	IU
208-96-8	Acenaphthylene	1800.	IU
606-20-2	2,6-Dinitrotoluene	1800.	IU

1C
 SEMIVOLATILE ORGANICS ANALYSIS DATA SHEET

BOOHY7

Lab Name: MARTIN MARIETTA

Contract: 0288

Lab Code: K25

Case No.: G132-013SAS No.: NA

SDG No.: ~~NA~~ ^{BOOHY7} _{5/25/91}

Matrix: (soil/water) SOIL

Lab Sample ID: 910320-106

Sample wt/vol: 10 (g/mL) G

Lab File ID: >11232

Level: (low/med) LOW

Date Received: 03/19/91

% Moisture: not dec. 43.6 dec. NA

Date Extracted: 03/27/91

Extraction: (Sepf/Cont/Sonc) CONT

Date Analyzed: 4/02/91

GPC Cleanup: (Y/N) N pH: NA

Dilution Factor: 1.00000

CAS NO.	COMPOUND	CONCENTRATION UNITS: (ug/L or ug/Kg) ug/KG	Q
99-09-2-----	3-Nitroaniline	8900.	IU
83-32-9-----	Acenaphthene	1800.	IU
51-28-5-----	2,4-Dinitrophenol	8900.	IU
100-02-7-----	4-Nitrophenol	8900.	IU
132-64-9-----	Dibenzofuran	1800.	IU
121-14-2-----	2,4-Dinitrotoluene	1800.	IU
84-66-2-----	Diethylphthalate	1800.	IU
7005-72-3-----	4-Chlorophenyl-phenylether	1800.	IU
86-73-7-----	Fluorene	1800.	IU
100-01-6-----	4-Nitroaniline	8900.	IU
534-52-1-----	4,6-Dinitro-2-methylphenol	8900.	IU
86-30-6-----	N-Nitrosodiphenylamine (1)	1800.	IU
101-55-3-----	4-Bromophenyl-phenylether	1800.	IU
118-74-1-----	Hexachlorobenzene	1800.	IU
87-86-5-----	Pentachlorophenol	8900.	IU
85-01-8-----	Phenanthrene	1800.	IU
120-12-7-----	Anthracene	1800.	IU
84-74-2-----	Di-n-butylphthalate	1800.	IU
206-44-0-----	Fluoranthene	1800.	IU
129-00-0-----	Pyrene	1800.	IU
85-68-7-----	Butylbenzylphthalate	1800.	IU
91-94-1-----	3,3'-Dichlorobenzidine	3500.	IU
56-55-3-----	Benzo(a)anthracene	1800.	IU
218-01-9-----	Chrysene	1800.	IU
117-81-7-----	bis(2-Ethylhexyl)phthalate	1800.	IU
117-84-0-----	Di-n-octylphthalate	1800.	IU
205-99-2-----	Benzo(b)fluoranthene	1800.	IU
207-08-9-----	Benzo(k)fluoranthene	1800.	IU
50-32-8-----	Benzo(a)pyrene	1800.	IU
193-39-5-----	Indeno(1,2,3-cd)pyrene	1800.	IU
53-70-3-----	Dibenz(a,h)anthracene	1800.	IU
191-24-2-----	Benzo(g,h,i)perylene	1800.	IU

(1) - Cannot be separated from Diphenylamine

1F
SEMIVOLATILE ORGANICS ANALYSIS DATA SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

EPA SAMPLE NO.

BOOHY7

Lab Name: MARTIN MARIETTA

Contract: 0288

Lab Code: K25

Case No.: G132-013SAS No.: NA

SDG No.: ~~NA~~ ^{BOOHY7} 5/15/91

Matrix: (soil/water) SOIL

Lab Sample ID: 910320-106

Sample wt/vol: 10 (g/mL) G

Lab File ID: >11232

Level: (low/med) LOW

Date Received: 03/19/91

% Moisture: not dec. 43.6 dec. NA

Date Extracted: 03/27/91

Extraction: (Sepf/Cont/Sonc) CONT

Date Analyzed: 4/02/91

GPC Cleanup: (Y/N) N pH: NA

Dilution Factor: 1.00000

CONCENTRATION UNITS:
(ug/L or ug/Kg) ug/KG

Number TICs found: 20

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1. 123-42-2	2-Pentanone, 4-Hydroxy-4-methyl-	6.49	64000	JAB
2.	Naphthalene, dimethyl	16.41	1100	JY
3.	UNKNOWN (Alkyl Hydrocarbon)	16.75	1400	J
4.	Naphthalene, Trimethyl	18.30	2000	JY
5.	Naphthalene, Alkyl	19.30	800	J
6. 489849	Azulene, 1,4-dimethyl-7-(1-methyl)	19.68	1700	J
7.	UNKNOWN (Alkyl Hydrocarbon)	19.80	3600	J
8.	Azulene, 7-ethyl-1,4-dimethyl-	20.41	830	J
9.	UNKNOWN (Sat'd Hydrocarbon)	25.82	2000	JB
10.	UNKNOWN	26.45	1500	JB
11.	UNKNOWN (Sat'd Hydrocarbon)	26.70	2200	JB
12.	UNKNOWN (Sat'd Hydrocarbon)	27.53	3600	JB
13.	UNKNOWN (Sat'd Hydrocarbon)	28.33	2700	JB
14.	UNKNOWN (Sat'd Hydrocarbon)	29.10	4600	JB
15.	UNKNOWN (Sat'd Hydrocarbon)	29.89	2800	J
16.	UNKNOWN (Sat'd Hydrocarbon)	30.77	4200	J
17.	UNKNOWN (Sat'd Hydrocarbon)	32.83	3300	J
18.	UNKNOWN (Mixture of two?)	33.38	1800	J
19.	UNKNOWN	33.78	1800	J
20.	UNKNOWN	35.07	2100	J
21.				
22.			CM	
23.			11/2/91	
24.				
25.				
26.				
27.				
28.				
29.				
30.				

48

EPA SAMPLE NO.

1B
SEMIVOLATILE ORGANICS ANALYSIS DATA SHEET

Lab Name: MARTIN MARIETTA

Contract: 0288

BOOHY7RE

Lab Code: K25

Case No.: G132-013SAS No.: NA

SDG No.: ~~NA~~ ^{BOOHY7} 5/25/91

Matrix: (soil/water) SOIL

Lab Sample ID: 910320-106

Sample wt/vol: 10 (g/mL) G

Lab File ID: >11285

Level: (low/med) LOW

Date Received: 03/19/91

% Moisture: not dec. 43.6 dec. NA

Date Extracted: 04/08/91

Extraction: (Sepf/Cont/Sonc) SONC

Date Analyzed: 4/11/91

GPC Cleanup: (Y/N) N pH: NA

Dilution Factor: 2.0

BLK : 910401-175

CONCENTRATION UNITS:

CAS NO.

COMPOUND

(ug/L or ug/Kg) ug/KG

Q

108-95-2-----	Phenol	3600.	IU	
111-44-4-----	bis(2-Chloroethyl)Ether	3600.	IU	
95-57-8-----	2-Chlorophenol	3600.	IU	
541-73-1-----	1,3-Dichlorobenzene	3600.	IU	
106-46-7-----	1,4-Dichlorobenzene	3600.	IU	
100-51-6-----	Benzyl alcohol	3600.	IU	
95-50-1-----	1,2-Dichlorobenzene	3600.	IU	
95-48-7-----	2-Methylphenol	3600.	IU	
39638-32-9-----	bis(2-chloroisopropyl)ether	3600.	IU	
106-44-5-----	4-Methylphenol	3600.	IU	
621-64-7-----	N-Nitroso-Di-n-propylamine	3600.	IU	
67-72-1-----	Hexachloroethane	3600.	IU	
98-95-3-----	Nitrobenzene	3600.	IU	
78-59-1-----	Isophorone	3600.	IU	
88-75-5-----	2-Nitrophenol	3600.	IU	
105-67-9-----	2,4-Dimethylphenol	3600.	IU	
65-85-0-----	Benzoic acid	18000.	IU	
111-91-1-----	bis(2-Chloroethoxy)methane	3600.	IU	
120-83-2-----	2,4-Dichlorophenol	3600.	IU	
120-82-1-----	1,2,4-Trichlorobenzene	3600.	IU	
91-20-3-----	Naphthalene	3300 920.	IJ	05/12/91
106-47-8-----	4-Chloroaniline	3600.	IU	
87-68-3-----	Hexachlorobutadiene	3600.	IU	
59-50-7-----	4-Chloro-3-methylphenol	3600.	IU	
91-57-6-----	2-Methylnaphthalene	6300 1800.	IJ	05/12/91
77-47-4-----	Hexachlorocyclopentadiene	3600.	IU	
88-06-2-----	2,4,6-Trichlorophenol	3600.	IU	
95-95-4-----	2,4,5-Trichlorophenol	18000.	IU	
91-58-7-----	2-Chloronaphthalene	3600.	IU	
88-74-4-----	2-Nitroaniline	18000.	IU	
131-11-3-----	Dimethylphthalate	3600.	IU	
208-96-8-----	Acenaphthylene	3600.	IU	
606-20-2-----	2,6-Dinitrotoluene	3600.	IU	

49

EPA SAMPLE NO.

1C
SEMIVOLATILE ORGANICS ANALYSIS DATA SHEET

Lab Name: MARTIN MARIETTA

Contract: 0288

BOOHY7RE

Lab Code: K25

Case No.: G132-013SAS No.: NA

SDG No.: NA

Boohy7

SAS
12/5/91

Matrix: (soil/water) SOIL

Lab Sample ID: 910320-106

Sample wt/vol: 10 (g/mL) G

Lab File ID: >11285

Level: (low/med) LOW

Date Received: 03/19/91

% Moisture: not dec. 43.6 dec. NA

Date Extracted: 04/08/91

Extraction: (Sepf/Cont/Sonc) SONC

Date Analyzed: 4/11/91

GPC Cleanup: (Y/N) N pH: NA

Dilution Factor: 2.0

CAS NO.	COMPOUND	CONCENTRATION UNITS: (ug/L or ug/Kg) ug/KG	Q
99-09-2-----	3-Nitroaniline	18000.	IU
83-32-9-----	Acenaphthene	3600.	IU
51-28-5-----	2,4-Dinitrophenol	18000.	IU
100-02-7-----	4-Nitrophenol	18000.	IU
132-64-9-----	Dibenzofuran	3600.	IU
121-14-2-----	2,4-Dinitrotoluene	3600.	IU
84-66-2-----	Diethylphthalate	3600.	IU
7005-72-3-----	4-Chlorophenyl-phenylether	3600.	IU
86-73-7-----	Fluorene	3600.	IU
100-01-6-----	4-Nitroaniline	18000.	IU
534-52-1-----	4,6-Dinitro-2-methylphenol	18000.	IU
86-30-6-----	N-Nitrosodiphenylamine (1)	3600.	IU
101-55-3-----	4-Bromophenyl-phenylether	3600.	IU
118-74-1-----	Hexachlorobenzene	3600.	IU
87-86-5-----	Pentachlorophenol	18000.	IU
85-01-8-----	Phenanthrene	3600.	IU
120-12-7-----	Anthracene	3600.	IU
84-74-2-----	Di-n-butylphthalate	3600.	IU
206-44-0-----	Fluoranthene	3600.	IU
129-00-0-----	Pyrene	3600.	IU
85-68-7-----	Butylbenzylphthalate	3600.	IU
91-94-1-----	3,3'-Dichlorobenzidine	7200.	IU
56-55-3-----	Benzo(a)anthracene	3600.	IU
218-01-9-----	Chrysene	3600.	IU
117-81-7-----	bis(2-Ethylhexyl)phthalate	1900. 530.	IJB
117-84-0-----	Di-n-octylphthalate	3600.	IU
205-99-2-----	Benzo(b)fluoranthene	3600.	IU
207-08-9-----	Benzo(k)fluoranthene	3600.	IU
50-32-8-----	Benzo(a)pyrene	3600.	IU
193-39-5-----	Indeno(1,2,3-cd)pyrene	3600.	IU
53-70-3-----	Dibenz(a,h)anthracene	3600.	IU
191-24-2-----	Benzo(g,h,i)perylene	3600.	IU

(1) - Cannot be separated from Diphenylamine

1F
SEMIVOLATILE ORGANICS ANALYSIS DATA SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

Lab Name: MARTIN MARIETTA

Contract: 0288

BOOHY7RE

Lab Code: K25

Case No.: G132-013SAS No.: NA

SDG No.: NA

Boohy7

2/5/91

Matrix: (soil/water) SOIL

Lab Sample ID: 910320-106

Sample wt/vol: 10 (g/mL) G

Lab File ID: >11285

Level: (low/med) LOW

Date Received: 03/19/91

% Moisture: not dec. 43.6 dec. NA

Date Extracted: 04/08/91

Extraction: (Sepf/Cont/Sonc) SONC

Date Analyzed: 4/11/91

GPC Cleanup: (Y/N) N pH: NA

Dilution Factor: 2.00000

CONCENTRATION UNITS:

Number TICs found: 19

(ug/L or ug/Kg) ug/KG

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1. 123-42-2	Diacetone Alcohol	6.33	150,000	JAB
2.	Naphthalene, Dimethyl	16.37	4500	JY
3.	SAT'D Hydrocarbon	16.72	5200	J
4.	Naphthalene, Trimethyl	18.25	8000	J
5.	" "	18.59	3000	J
6.	UNKNOWN (Alkyl Azulene?)	19.65	6800	J
7.	Alkyl Hydrocarbon	19.77	15000	J
8.	UNKNOWN (Alkyl Azulene)	20.38	3800	J
9. 57-10-3	Hexadecanoic Acid	22.56	3400	JB
10.	SAT'D Hydrocarbon	25.78	4700	J
11.	Alkyl Hydrocarbon	26.65	5700	J
12.	Alkyl Hydrocarbon	27.49	7700	J
13.	Alkyl Hydrocarbon	28.29	6000	J
14.	SAT'D Hydrocarbon	29.06	12000	J
15.	SAT'D Hydrocarbon	29.86	4700	J
16.	SAT'D Hydrocarbon	30.71	7000	J
17.	SAT'D Hydrocarbon	32.77	5700	J
18.	UNKNOWN (Benzenamine?)	33.73	4700	J
19.	UNKNOWN	35.00	4000	J
20.				
21.			CM	
22.			4/16/91	
23.				
24.				
25.				
26.				
27.				
28.				
29.				
30.				

82

EPA SAMPLE NO.

1B
SEMIVOLATILE ORGANICS ANALYSIS DATA SHEET

BOOHZ3

BooHY7

Lab Name: MARTIN MARIETTA

Contract: 0288

Lab Code: K25

Case No.: G132-013SAS No.: NA

SDG No.: ~~NA~~ 225/9

Matrix: (soil/water) SOIL

Lab Sample ID: 910408-131

Sample wt/vol: 10 (g/mL) G

Lab File ID: >11288

Level: (low/med) LOW

Date Received: 04/06/91

% Moisture: not dec. 32.6 dec. NA

Date Extracted: 04/08/91

Extraction: (Sepf/Cont/Sonc) SONC

Date Analyzed: 4/11/91

GPC Cleanup: (Y/N) N pH: NA

Dilution Factor: 1.00000

BLK : 910401-175

CONCENTRATION UNITS:
(ug/L or ug/Kg) ug/KG

CAS NO.	COMPOUND		Q
108-95-2	Phenol	1500.	IU
111-44-4	bis(2-Chloroethyl)Ether	1500.	IU
95-57-8	2-Chlorophenol	1500.	IU
541-73-1	1,3-Dichlorobenzene	1500.	IU
106-46-7	1,4-Dichlorobenzene	1500.	IU
100-51-6	Benzyl alcohol	1500.	IU
95-50-1	1,2-Dichlorobenzene	1500.	IU
95-48-7	2-Methylphenol	1500.	IU
39638-32-9	bis(2-chloroisopropyl)ether	1500.	IU
106-44-5	4-Methylphenol	1500.	IU
621-64-7	N-Nitroso-Di-n-propylamine	1500.	IU
67-72-1	Hexachloroethane	1500.	IU
98-95-3	Nitrobenzene	1500.	IU
78-59-1	Isophorone	1500.	IU
88-75-5	2-Nitrophenol	1500.	IU
105-67-9	2,4-Dimethylphenol	1500.	IU
65-85-0	Benzoic acid	7400.	IU
111-91-1	bis(2-Chloroethoxy)methane	1500.	IU
120-83-2	2,4-Dichlorophenol	1500.	IU
120-82-1	1,2,4-Trichlorobenzene	1500.	IU
91-20-3	Naphthalene	1500.	IU
106-47-8	4-Chloroaniline	1500.	IU
87-68-3	Hexachlorobutadiene	1500.	IU
59-50-7	4-Chloro-3-methylphenol	1500.	IU
91-57-6	2-Methylnaphthalene	1500.	IU
77-47-4	Hexachlorocyclopentadiene	1500.	IU
88-06-2	2,4,6-Trichlorophenol	1500.	IU
95-95-4	2,4,5-Trichlorophenol	7400.	IU
91-58-7	2-Chloronaphthalene	1500.	IU
88-74-4	2-Nitroaniline	7400.	IU
131-11-3	Dimethylphthalate	1500.	IU
208-96-8	Acenaphthylene	1500.	IU
606-20-2	2,6-Dinitrotoluene	1500.	IU

1C
SEMIVOLATILE ORGANICS ANALYSIS DATA SHEET

83
EPA SAMPLE NO.

BOOHZ3

Lab Name: MARTIN MARIETTA

Contract: 0288

Boohy 7

Lab Code: K25

Case No.: G132-013SAS No.: NA

SDG No.: ~~NA~~

*SAS
12/5/9*

Matrix: (soil/water) SOIL

Lab Sample ID: 910408-131

Sample wt/vol: 10 (g/mL) G

Lab File ID: >11288

Level: (low/med) LOW

Date Received: 04/06/91

% Moisture: not dec. 32.6 dec. NA

Date Extracted: 04/08/91

Extraction: (Sepf/Cont/Sonc) SONC

Date Analyzed: 4/11/91

GPC Cleanup: (Y/N) N pH: NA

Dilution Factor: 1.00000

CAS NO.	COMPOUND	CONCENTRATION UNITS: (ug/L or ug/Kg) ug/KG	Q
99-09-2-----	3-Nitroaniline_____	7400.	IU
83-32-9-----	Acenaphthene_____	1500.	IU
51-28-5-----	2,4-Dinitrophenol_____	7400.	IU
100-02-7-----	4-Nitrophenol_____	7400.	IU
132-64-9-----	Dibenzofuran_____	1500.	IU
121-14-2-----	2,4-Dinitrotoluene_____	1500.	IU
84-66-2-----	Diethylphthalate_____	1500.	IU
7005-72-3-----	4-Chlorophenyl-phenylether_____	1500.	IU
86-73-7-----	Fluorene_____	1500.	IU
100-01-6-----	4-Nitroaniline_____	7400.	IU
534-52-1-----	4,6-Dinitro-2-methylphenol_____	7400.	IU
86-30-6-----	N-Nitrosodiphenylamine (1)_____	1500.	IU
101-55-3-----	4-Bromophenyl-phenylether_____	1500.	IU
118-74-1-----	Hexachlorobenzene_____	1500.	IU
87-86-5-----	Pentachlorophenol_____	7400.	IU
85-01-8-----	Phenanthrene_____	1500.	IU
120-12-7-----	Anthracene_____	1500.	IU
84-74-2-----	Di-n-butylphthalate_____	1500.	IU
206-44-0-----	Fluoranthene_____	1500.	IU
129-00-0-----	Pyrene_____	1500.	IU
85-68-7-----	Butylbenzylphthalate_____	1500.	IU
91-94-1-----	3,3'-Dichlorobenzidine_____	3000.	IU
56-55-3-----	Benzo(a)anthracene_____	1500.	IU
218-01-9-----	Chrysene_____	1500.	IU
117-81-7-----	bis(2-Ethylhexyl)phthalate_____	6601500.	IU JB
117-84-0-----	Di-n-octylphthalate_____	1500.	IU
205-99-2-----	Benzo(b)fluoranthene_____	1500.	IU
207-08-9-----	Benzo(k)fluoranthene_____	1500.	IU
50-32-8-----	Benzo(a)pyrene_____	1500.	IU
193-39-5-----	Indeno(1,2,3-cd)pyrene_____	1500.	IU
53-70-3-----	Dibenz(a,h)anthracene_____	1500.	IU
191-24-2-----	Benzo(g,h,i)perylene_____	1500.	IU

(1) - Cannot be separated from Diphenylamine

1F
SEMIVOLATILE ORGANICS ANALYSIS DATA SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

Lab Name: MARTIN MARIETTA

Contract: 0288

BOOHZ3

BOOHY 7

Lab Code: K25

Case No.: G132-013SAS No.: NA

SDG No.: NA 5/12/91

Matrix: (soil/water) SOIL

Lab Sample ID: 910408-131

Sample wt/vol: 10 (g/mL) G

Lab File ID: >11288

Level: (low/med) LOW

Date Received: 04/06/91

% Moisture: not dec. 32.6 dec. NA

Date Extracted: 04/08/91

Extraction: (Sepf/Cont/Sonc) SONC

Date Analyzed: 4/11/91

GPC Cleanup: (Y/N) N pH: NA

Dilution Factor: 1.00000

Number TICs found: 8

CONCENTRATION UNITS:
(ug/L or ug/Kg) ug/Kg

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1. 141-79-7	3-Penten-2-one, 4-methyl-	5.35	690	JAB
2. 123-42-2	Diacetone Alcohol	6.49	120,000	JAB
3.	UNKNOWN (Hydrocarbon)	22.07	3500	JB
4.	Hexadecanoic Acid	22.57	1300	JB
5.	UNKNOWN	24.34	1700	JB
6.	"	26.95	840	J
7.	"	30.18	750	JB
8.	"	31.26	650	J
9.				
10.			91	
11.			411691	
12.				
13.				
14.				
15.				
16.				
17.				
18.				
19.				
20.				
21.				
22.				
23.				
24.				
25.				
26.				
27.				
28.				
29.				
30.				

CHAIN OF CUSTODY RECORD

SAMPLING AND ENVIRONMENTAL SUPPORT DEPARTMENT

SES/COC NO. 000187

[illegible]

SSS

9613478.0443

~~"Next Day Service"~~ Overnight

218

Contractor Westinghouse	OFF-SITE PROPERTY CONTROL	CONTROL NUMBER (To be obtained from PROPERTY MANAGEMENT) W91-0-0240
----------------------------	------------------------------	---

PART I - TO BE COMPLETED BY ORIGINATOR

Department Environmental Eng.	Section Technical Baseline	Unit
The following items are to be shipped from		☒ Contractor ☐ Vendor
Routing		☐ Contractor ☐ Vendor
Shipped to U.S. Department of Energy c/o Martin Marietta Energy Systems Oak Ridge Gaseous Diffusion Plant Blair Road Highway 58 Oak Ridge, TN 37831-7440		Off-site Custodian C. R. Kirkpatrick, K-1004A Drop Point A20 Full Title
Quantity	Description (Include Serial and any Government Tag Numbers)	Original Cost
1 23 lbs.	• Metal Ice Chest - Two glass jars of fly ash packed in vermiculite and blue ice - Sample BOOHY7	
☐ Classified ☒ Unclassified ☐ Shipped Under DOE Contract ☐ Shipped Under Contractor's Use Permit Contract		

Necessity for the Off-Site Use of this Property

Program not available on site.

Bill of Lading # PSC 2426159120

CERTIFICATION OF THE RADIATION MONITORING RELEASE MUST BE SECURED THE SAME DAY THAT MATERIAL IS DELIVERED TO SHIPPING.

RM Clearance for Public Release Loria Miller	RM Survey No 910084	Date 3/18/91
Location of Property (Area & Bldg) 300-FF-1 Fly Ash Pits	Contact Chris Kramer	Phone 6-4186
Date Ready for Shipment	Cost Code to be Charged EC3B2	Approximate Date This Property will be Returned n/a
Originated By Chris Kramer	Date 3/18/91	Authorized By [Signature]
Signature and Name of Property Control	Custodian Date 3/18/91	Property Management Approval [Signature]

PART II - TO BE COMPLETED BY SHIPPING

Signature of Recipient J. Lemawst	Return Order No.	Date Issued	Purchase Order No.	Date Issued
Date 3-18-91				

DISTRIBUTION

By Originator White, Green, Yellow, Pink - Property Management Goldenrod - Retain	Shipping Operation - Sign all Copies and Forward to: White - Property Management Green - Property Control Custodian (Issuing Office) Yellow - Retain Pink - Originator
---	--

EMERY
WORLDWIDE



SIGNATURE AND TALLY RECORD

80029-46 (8/89) Litho USA

DATE: 3-18-91 SHIPMENT NO.: PSC242615912 0
SHIPPER: WESTINGHOUSE SHIPPING DEPT (509) 376-6665
REFERENCE NO.: W91-0-0240

SHIPPER NAME AND ADDRESS

WESTINGHOUSE HANFORD SHIPPING (509) 376-6665
G2-06 US DEPARTMENT OF ENERGY C/O
WESTINGHOUSE HANFORD COMPANY
2355 STEVENS DR
PO BOX 1970
RICHLAND WA 99352

CONSIGNEE NAME AND ADDRESS

C.R. KIRKPATRICK K-1004A DROP POINT A20
U.S. DEPARTMENT OF ENERGY
C/O MARTIN MARIETTA ENERGY SYSTEMS
OAK RIDGE GASEOUS DIFFUSION PLANT
BLAIR ROAD HIGHWAY 58
OAK RIDGE TN 37831-7440

Pieces	Weight	Description/Marks	Emery Authorization No.
1	23 LBS	W91-0-0240 METAL ICE CHEST SAMPLE BOUHY7	

EACH PERSON HANDLING OR TAKING CUSTODY OF THIS SHIPMENT MUST SIGN AND COMPLETE THE INFORMATION BELOW

Name of Person/Company	Transship Point/Destination	Signature of Person Accepting Custody	Time/Date
WESTINGHOUSE SHIPPING			
1. JOYCE DEMAREST 1167/1100	RICHLAND WA		103-18-91
2. T. L. VANCE	LA	T. L. VANCE	3/14/91
3. S. R. RAY	LA	S. R. RAY	3/14/91
4. T. L. DUNCAN	LA	T. L. DUNCAN	3/14/91
5. T. L. DUNCAN	LA	T. L. DUNCAN	3/14/91
6. T. L. DUNCAN	LA	T. L. DUNCAN	3/14/91
7. T. L. DUNCAN	LA	T. L. DUNCAN	3/14/91
8. T. L. DUNCAN	LA	T. L. DUNCAN	3/14/91

SPECIAL HANDLING INSTRUCTIONS

CONSIGNEE COPY

Contractor Westinghouse Hanford Co.	OFF-SITE PROPERTY CONTROL	CONTROL NUMBER (To be obtained from PROPERTY MANAGEMENT) W91-0271 W91-0271
---	------------------------------	---

PART I - TO BE COMPLETED BY ORIGINATOR

Department Environmental Engineering	Section Technical Baseline	Unit
The following items are to be shipped from		☒ Contractor ☐ Vendor
Routing Emery		☐ Contractor ☐ Vendor
Shipped to K-25 U.S. Department of Energy c/o Martin Marietta Energy Systems Oak Ridge Gaseous Diffusion Plant Blair Road, Highway 58 Oak Ridge, TN 37831-740		Off-site Custodian ATTN: C.R. Kirkpatrick, K-1004A droppoint A 2c
Full Title		

Quantity	Description (Include Serial and any Government Tag Numbers)	Original Cost
1	Poly Ice chest - Glass jars of soil, packed in vermiculite and blue ice. - Sample # BOOJ 75, BOOJ 76 - Sample # BOOJ HZ 3 (Separate project, 91-013) Ice Chest EA012 (37 lbs)	

☐ Classified ☐ Unclassified ☐ Shipped Under DOE Contract ☐ Shipped Under Contractor's Use Permit Contract

Necessity for the Off-Site Use of this Property

Program not available on-site.

Bill of Lading # 247425669 7

CERTIFICATION OF THE RADIATION MONITORING RELEASE MUST BE SECURED THE SAME DAY THAT MATERIAL IS DELIVERED TO SHIPPING.

RM Clearance for Public Release	RM Survey No	Date
Location of Property (Area & Bldg) 300 BP-1 / 300-FF-1	Contact Ron Mitchell	Phone 376-5122
Date Ready for Shipment	Cost Code to be Charged E32NA 81220	Approximate Date This Property will be Returned n/a
Originated By Ron Mitchell	Date	Authorized By RPNH
Signature and Name of Property Control	Custodian Date	Property Management Approval Date 4/4/91

PART II - TO BE COMPLETED BY SHIPPING

Signature of Recipient J. Demarest	Return Order No	Date Issued	Purchase Order No.	Date Issued
Date 4-4-91				

DISTRIBUTION

By Originator White, Green, Yellow, Pink - Property Management Goldenrod - Retain	Shipping Operation - Sign all Copies and Forward to: White - Property Management Green - Property Control Custodian (Issuing Office) Yellow - Retain Pink - Originator
---	--

CHAIN OF CUSTODY RECORD

SAMPLING AND ENVIRONMENTAL SUPPORT DEPARTMENT

000188

UCN-16691A (1 8-90)

DISTRIBUTION: ORIGINAL ACCOMPANIES SHIPMENT: COPY TO COORDINATOR OF FIELD FILES