



February 22, 2017

Mr. Scot Fitzgerald
CH2MHill Plateau Remediation Company
MSIN R3-50 CHPRC
PO Box 1600
Richland, Washington 99352

Re: CHPRC SAF F16-028
Work Order: 415150
SDG: GEL415150

Dear Mr. Fitzgerald:

GEL Laboratories, LLC (GEL) appreciates the opportunity to provide the enclosed analytical results for the sample(s) we received on January 27, 2017. This original data report has been prepared and reviewed in accordance with GEL's standard operating procedures.

Our policy is to provide high quality, personalized analytical services to enable you to meet your analytical needs on time every time. We trust that you will find everything in order and to your satisfaction. If you have any questions, please do not hesitate to call me at (843) 556-8171, ext. 4505.

Sincerely,

B Luthman
Brielle Luthman for
Heather Shaffer
Project Manager

Purchase Order: 300192 - 8H
Chain of Custody: F16-028-065
Enclosures



Table of Contents

Case Narrative.....	1
Chain of Custody and Supporting Documentation.....	8
Data Review Qualifier Definitions.....	13
Laboratory Certifications.....	15
Metals Analysis.....	17
Case Narrative.....	18
Sample Data Summary.....	22
Quality Control Summary.....	24
General Chem Analysis.....	27
Case Narrative.....	28
Sample Data Summary.....	32
Quality Control Summary.....	34
Radiological Analysis.....	38
Case Narrative.....	39
Sample Data Summary.....	49
Quality Control Summary.....	63

Case Narrative

**General Narrative
for
CH2MHill Plateau Remediation Company
CHPRC SAF F16-028
SDG: GEL415150**

February 22, 2017

Laboratory Identification:

GEL Laboratories LLC
2040 Savage Road
Charleston, South Carolina 29407
(843) 556-8171

Summary

Sample receipt

The sample(s) arrived at GEL Laboratories, LLC, Charleston, South Carolina on January 27, 2017, for analysis. The sample was delivered with proper chain of custody documentation and signatures. All sample containers arrived without any visible signs of tampering or breakage. There are no additional comments concerning sample receipt.

Items of Note All efforts were made by the lab to meet any short hold times. Samples that were analyzed outside of the initial hold time but still within 2X hold time will be noted in the lab case narrative.

Sample Identification

The laboratory received the following sample:

Laboratory Identification	Sample Description
415150001	B36M10

Case Narrative

Sample analyses were conducted using methodology as outlined in GEL Laboratories, LLC (GEL) Standard Operating Procedures. Any technical or administrative problems during analysis, data review, and reduction are contained in the analytical case narratives in the enclosed data package.

Data Package

The enclosed data package contains the following sections: General Narrative, Chain of Custody and Supporting Documentation, and data from the following fractions: General Chemistry, Metals and Radiochemistry.

We certify that this package is in compliance with the SOW, both technically and for completeness, including a full description of, explanation of, and corrective actions for, any and all deviations, from either the analyses requested or the case narrative requested. Release of the data contained in this hard copy data package has been authorized by the Laboratory Analytical Manager (or designee) and the laboratory's client services representative as verified by their signatures on this report.

2/23/2017

REV.0

B. Luthman
Brielle Luthman for
Heather Shaffer
Project Manager

Technical Case Narrative
CH2M Hill Plateau Remediation Company (CPRC)
SDG #: GEL415150
Work Order #: 415150

Metals

Determination of Metals by ICP-MS

All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable, with the following exceptions.

Technical Information

Sample Dilutions

The ICPMS solid samples in this SDG were diluted the standard two times.

Analyte	415150
	001
Uranium	2X

General Chemistry

Ion Chromatography

All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable, with the following exceptions.

Miscellaneous Information

Manual Integrations

Samples 1203719658 (B36M10DUP) and 415150001 (B36M10) were manually integrated to correctly position the baseline as set in the calibration standards.

pH

All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable, with the following exceptions.

Technical Information

Holding Times

Samples (See Below) were received by the laboratory outside of the method specified holding time. The data is qualified.

Sample	Analyte	Value
1203716728 (B36M10DUP)	pH	Received 27-JAN-17, out of holding 25-JAN-17
415150001 (B36M10)	pH	Received 27-JAN-17, out of holding 25-JAN-17

Radiochemistry**AMCMISO_EIE_PRECIP_AEA: COMMON**

All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable, with the following exceptions.

Technical Information**Recounts**

Sample 1203718268 (LCS) was recounted due to high carrier/tracer yield. The recount is reported.

NP237_IE_PRECIP_AEA: COMMON

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

PUISO_PRECIP_AEA:COMMON

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

THISO_IE_PLATE_AEA: COMMON

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

UIISO_IE_PRECIP_AEA:COMMON

All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable, with the following exceptions.

Technical Information**Recounts**

Sample 415150001 (B36M10) was recounted due to a suspected false positive. The recount is reported.

Miscellaneous Information**Dry Weight**

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

GAMMA_GS:COMMON + (Add-on)

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

I129_SEP_LEPS_GS

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

SRTOT_SEP_PRECIP_GPC: COMMON

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

TC99_SEP_GPC

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

TRITIUM_DIST_LSC: COMMON

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

C14_LSC: COMMON

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

NI63_LSC

All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable, with the following exceptions.

Technical Information

Sample Re-prep/Re-analysis

Sample 415150001 (B36M10) was repped due to high recovery. The re-analysis is being reported.

Certification Statement

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless otherwise noted in the analytical case narrative.

Chain of Custody and Supporting Documentation

CH2M Hill Plateau Remediation Company		CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST		F16-028-065	PAGE 2 OF 2
COLLECTOR	Jeff Tuckesen CHPRC	COMPANY CONTACT	TELEPHONE NO.	PROJECT COORDINATOR	PRICE CODE
SAMPLING LOCATION	C9567, I-005	TODAK, D	376-6427	TODAK, D	8H
ICE CHEST NO.		PROJECT DESIGNATION		SAF NO.	AIR QUALITY
		200-WA-1 Opportunistic sampling - soil		F16-028	<input type="checkbox"/>
		FIELD LOGBOOK NO.	ACTUAL SAMPLE DEPTH	COA	METHOD OF SHIPMENT
		HNF-N-645 <u>1-62</u>	<u>245.82 - 247.35</u>	300192	FEDERAL EXPRESS
SHIPPED TO		OFFSITE PROPERTY NO.		BILL OF LADING/AIR BILL NO.	
GEL Laboratories, LLC		<u>7486</u>		<u>7782 8683 1215</u>	

SPECIAL INSTRUCTIONS

TRVL-16-060


(1) 6020_METALS_ICPMS: COMMON (Add-on) {Uranium}; 9056_ANIONS_IC: COMMON {Chloride, Fluoride, Nitrogen in Nitrate, Nitrogen in Nitrite, Sulfate}; 9056_ANIONS_IC: COMMON (Add-on) {Phosphorus in phosphate};

(2) GAMMA_GS: COMMON; GAMMA_GS: COMMON (Add-on) {Radium-226, Radium-228};

(3) AMCMISO_IE_PRECIP_AEA: COMMON {Americium-241}; C14_LSC: COMMON; I129_SEP_LEPS_GS: COMMON; NI63_LSC: COMMON; NP237_IE_PRECIP_AEA: COMMON; SRTOT_SEP_PRECIP_GPC: COMMON; TC99_EIE_LSC: COMMON; THISO_IE_PLATE_AEA: COMMON {Thorium-232}; UJISO_IE_PRECIP_AEA: COMMON; NP237_IE_PRECIP_AEA: COMMON; TRITIUM_DIST_LSC: COMMON;

(4) Moisture Content - D2216 {Percent moisture (wet sample)};

(5) 9045_pH (Non-Aqueous): COMMON {pH Measurement};



SAMPLE RECEIPT & REVIEW FORM

Client: <u>CPRC</u>		SDG/AR/COC/Work Order: <u>415150</u>	
Received By: <u>AG</u>		Date Received: <u>1/27/17</u>	
Suspected Hazard Information	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	*If Net Counts > 100cpm on samples not marked "radioactive", contact the Radiation Safety Group for further investigation.
COC/Samples marked as radioactive?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	Maximum Net Counts Observed* (Observed Counts - Area Background Counts): <u>0CPM</u>
Classified Radioactive II or III by RSO?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	If yes, Were swipes taken of sample containers < action levels?
COC/Samples marked containing PCBs?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Package, COC, and/or Samples marked as beryllium or asbestos containing?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	If yes, samples are to be segregated as Safety Controlled Samples, and opened by the GEL Safety Group.
Shipped as a DOT Hazardous?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	Hazard Class Shipped: UN#:
Samples identified as Foreign Soil?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	

Sample Receipt Criteria		Yes	NA	No	Comments/Qualifiers (Required for Non-Conforming Items)
1	Shipping containers received intact and sealed?	<input checked="" type="checkbox"/>			Circle Applicable: Seals broken Damaged container Leaking container Other (describe)
2	Samples requiring cold preservation within (0 ≤ deg. C)?*	<input checked="" type="checkbox"/>			Preservation Method: ice bags Blue ice Dry ice None Other (describe) <u>1°C</u> *all temperatures are recorded in Celsius
2a	Daily check performed and passed on IR temperature gun?	<input checked="" type="checkbox"/>			Temperature Device Serial #: Secondary Temperature Device Serial # (If Applicable): <u>E5102009185</u>
3	Chain of custody documents included with shipment?	<input checked="" type="checkbox"/>			
4	Sample containers intact and sealed?	<input checked="" type="checkbox"/>			Circle Applicable: Seals broken Damaged container Leaking container Other (describe)
5	Samples requiring chemical preservation at proper pH?	<input type="checkbox"/>		<input checked="" type="checkbox"/>	Sample ID's, containers affected and observed pH: <u>See Continuation Form</u>
6	Do Low Level Perchlorate samples have headspace as required?	<input type="checkbox"/>		<input checked="" type="checkbox"/>	If Preservation added, Lot#: Sample ID's and containers affected:
7	VOA vials contain acid preservation?	<input checked="" type="checkbox"/>			(If unknown, select No)
8	VOA vials free of headspace (defined as < 6mm bubble)?	<input checked="" type="checkbox"/>			Sample ID's and containers affected:
9	Are Encore containers present?	<input type="checkbox"/>		<input checked="" type="checkbox"/>	(If yes, immediately deliver to Volatiles laboratory)
10	Samples received within holding time?	<input checked="" type="checkbox"/>			ID's and tests affected:
11	Sample ID's on COC match ID's on bottles?	<input checked="" type="checkbox"/>			Sample ID's and containers affected:
12	Date & time on COC match date & time on bottles?	<input checked="" type="checkbox"/>			Sample ID's affected:
13	Number of containers received match number indicated on COC?	<input checked="" type="checkbox"/>			Sample ID's affected:
14	Are sample containers identifiable as GEL provided?	<input type="checkbox"/>		<input checked="" type="checkbox"/>	
15	COC form is properly signed in relinquished/received sections?	<input checked="" type="checkbox"/>			
16	Carrier and tracking number.				Circle Applicable: FedEx Air <u> </u> FedEx Ground <u> </u> UPS <u> </u> Field Services <u> </u> Courier <u> </u> Other <u> </u> <u>7782 8683 1324</u> <u>7782 8683 0679</u> <u>7782 8683 1362</u> <u>7782 8683 1248</u> <u>7782 8683 1215</u>

Comments (Use Continuation Form if needed):

PM (or PMA) review: Initials DS Date 1/31/17 Page 1 of 2

GEL Laboratories LLC SAMPLE RECEIPT & REVIEW CONTINUATION FORM

Client: CPRC Received By: AG Date Received: 1/27/17 SDG/AR/COC/Work Order: 415150

Containers should contain HNO₃ as preservative,
 not they do WOT!

IDs affected: B380T8
 B380T5
 B38046
 B38042
 B386M4
 B386m2

PM (or PMA) review: Initials NS Date 1/31/17 Page 2 of 2

Data Review Qualifier Definitions

Project Specific Qualifier Definitions for GEL Client Code: CPRC

Qualifier	Qualifier Definition	Department	Fraction
U	Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.		
J	The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate). Value is estimated	Organics	
P	Aroclor target analyte with greater than 25% difference between column analyses.	Organics	
C	Analyte has been confirmed by GC/MS analysis	Organics	Pesticide
B	The analyte was detected in both the associated QC blank and in the sample.	Organics	
E	Concentration exceeds the calibration range of the instrument	Organics	
A	The TIC is a suspected aldol-condensation product	Organics	Semi-Volatile
X	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier		
N	Spike Sample recovery is outside control limits.		
*	Duplicate analysis not within control limits	Inorganics	
>	Result greater than quantifiable range or greater than upper limit of the analysis range	General Chemistry	
Z	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier		
B	The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate).	Inorganics	Metals
D	Results are reported from a diluted aliquot of sample.		
E	Reported value is estimated due to interferences. See comment in narrative.	Inorganics	Metals
M	Duplicate precision not met.	Inorganics	Metals
o	Analyte failed to recover within LCS limits (Organics only)	Organics	
S	Reported value determined by the Method of Standard Additions (MSA)	Inorganics	
T	Spike and/or spike duplicate sample recovery is outside control limits.	Organics	
W	Post-digestion spike recovery for GFAA out of control limit. Sample absorbency < 50% of spike absorbency.	Inorganics	
B	The associated QC sample blank has a result $\geq 2X$ the MDA and, after corrections, result is \geq MDA for this sample	Radiological	
Y	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier		
+	Correlation coefficient for Method of Standard Additions (MSA) is < 0.995	Inorganics	
B	The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate).	General Chemistry	
C	Target analyte was detected in the sample and the associated blank. The associated blank concentration is \geq EQL or is > 5% of the measured concentration and/or decision level for associated samples.	Inorganics	Metals
C	Target analyte was detected in the sample and the associated blank. The associated blank concentration is \geq EQL or is > 5% of the measured concentration and/or decision level for associated samples.	General Chemistry	
<	Sample is below the EPA guidance level for Reactive Releasable Cyanide and/or Reactive Releasable Sulfide	General Chemistry	
UX	Gamma Spectroscopy--Uncertain identification	Radiological	

Laboratory Certifications

List of current GEL Certifications as of 22 February 2017

State	Certification
Alaska	UST-0110
Arkansas	88-0651
CLIA	42D0904046
California	2940
Colorado	SC00012
Connecticut	PH-0169
Delaware	SC00012
DoD ELAP/ ISO17025 A2LA	2567.01
Florida NELAP	E87156
Foreign Soils Permit	P330-15-00283, P330-15-00253
Georgia	SC00012
Georgia SDWA	967
Hawaii	SC00012
Idaho Chemistry	SC00012
Idaho Radiochemistry	SC00012
Illinois NELAP	200029
Indiana	C-SC-01
Kansas NELAP	E-10332
Kentucky SDWA	90129
Kentucky Wastewater	90129
Louisiana NELAP	03046 (AI33904)
Louisiana SDWA	LA170010
Maryland	270
Massachusetts	M-SC012
Michigan	9976
Mississippi	SC00012
Nebraska	NE-OS-26-13
Nevada	SC000122016-1
New Hampshire NELAP	205415
New Jersey NELAP	SC002
New Mexico	SC00012
New York NELAP	11501
North Carolina	233
North Carolina SDWA	45709
North Dakota	R-158
Oklahoma	9904
Pennsylvania NELAP	68-00485
S.Carolina Radchem	10120002
South Carolina Chemistry	10120001
Tennessee	TN 02934
Texas NELAP	T104704235-17-12
Utah NELAP	SC000122016-21
Vermont	VT87156
Virginia NELAP	460202
Washington	C780
West Virginia	997404

Metals Analysis

Case Narrative

Metals
Technical Case Narrative
CH2MHill Plateau Remediation Company (CPRC)
SDG #: GEL415150
Work Order #: 415150

Product: Determination of Metals by ICP-MS**Analytical Method:** SW846 3050B/6020B**Analytical Procedure:** GL-MA-E-014 REV# 28**Analytical Batch:** 1634971**Preparation Method:** SW846 3050B**Preparation Procedure:** GL-MA-E-009 REV# 26**Preparation Batch:** 1634970

The following samples were analyzed using the above methods and analytical procedure(s).

<u>GEL Sample ID#</u>	<u>Client Sample Identification</u>
415150001	B36M10
1203717290	Method Blank (MB)ICP-MS
1203717291	Laboratory Control Sample (LCS)
1203717294	415150001(B36M10L) Serial Dilution (SD)
1203717292	415150001(B36M10D) Sample Duplicate (DUP)
1203717293	415150001(B36M10S) Matrix Spike (MS)

The samples in this SDG were analyzed on a "dry weight" basis.

Data Summary:

All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable, with the following exceptions.

Technical Information**Sample Dilutions**

Dilutions may be required for many reasons, including to minimize matrix interferences or to bring over range target analyte concentrations into the linear calibration range. The ICPMS solid samples in this SDG were diluted the standard two times.

Analyte	415150
	001
Uranium	2X

Certification Statement

Where the analytical method has been performed under NELAP certification, the analysis has met all of the

requirements of the NELAC standard unless otherwise noted in the analytical case narrative.

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

**Qualifier Definition Report
for**

CPRC001 CH2MHill Plateau Remediation Company

Client SDG: GEL415150 GEL Work Order: 415150

The Qualifiers in this report are defined as follows:

* Duplicate analysis not within control limits

B The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate).

D Results are reported from a diluted aliquot of sample.

U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.

Review/Validation

GEL requires all analytical data to be verified by a qualified data reviewer. In addition, all CLP-like deliverables receive a third level review of the fractional data package.

The following data validator verified the information presented in this data report:

Signature:**Name: Nik-Cole Elmore****Date: 01 FEB 2017****Title: Data Validator**

Sample Data Summary

METALS
-1-
INORGANICS ANALYSIS DATA PACKAGE

SDG No: GEL415150

CONTRACT: CPRC0F16028

METHOD TYPE: SW846

SAMPLE ID: 415150001

BASIS: Dry Weight

DATE COLLECTED 25-JAN-17

CLIENT ID: B36M10

LEVEL: Low

DATE RECEIVED 27-JAN-17

MATRIX: SOIL

%SOLIDS: 87

CAS No.	Analyte	Result	Units	Qual	MDL	PQL	CRDL	DF	M*	Analyst	Run Date	Analytical Run	Analytical Batch
7440-61-1	Uranium	662	ug/kg	D	14.4	43.6	43.6	2	MS	SKJ	01/31/17 08:54	170130-1	1634971

Prep Information:

Analytical Batch	Prep Batch	Prep Method	Initial wt./vol.	Units	Final wt./vol.	Units	Date	Analyst
1634971	1634970	SW846 3050B	0.525	g	50	mL	01/30/17	SXW1

***Analytical Methods:**

MS SW846 3050B/6020B

Quality Control Summary

GEL LABORATORIES LLC

2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Report Date: February 1, 2017

Page 1 of 2

CH2M Hill Plateau Remediation Company

MSIN R3-50 CHPRC

PO Box 1600

Richland, Washington

Contact: Mr. Scot Fitzgerald

Workorder: 415150

Parmname	NOM	Sample	Qual	QC	Units	RPD/D%	REC%	Range	Anlst	Date	Time
Metals Analysis - ICPMS											
Batch	1634971										
QC1203717292	415150001	DUP									
Uranium		D	662	D	585	ug/kg	12.5	(0%-20%)	SKJ	01/31/17	08:58
QC1203717291	LCS										
Uranium	4840			D	5090	ug/kg		(80%-120%)		01/31/17	08:50
QC1203717290	MB										
Uranium				DU	13.1	ug/kg				01/31/17	08:46
QC1203717293	415150001	MS									
Uranium	5480	D	662	D	6220	ug/kg		(75%-125%)		01/31/17	09:02
QC1203717294	415150001	SDILT									
Uranium		D	3.04	D	0.612	ug/L	.592	(0%-20%)		01/31/17	09:06

Notes:

The Qualifiers in this report are defined as follows:

- * Duplicate analysis not within control limits
- + Correlation coefficient for Method of Standard Additions (MSA) is < 0.995
- B The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate).
- C Target analyte was detected in the sample and the associated blank. The associated blank concentration is >= EQL or is > 5% of the measured concentration and/or decision level for associated samples.
- D Results are reported from a diluted aliquot of sample.
- E Reported value is estimated due to interferences. See comment in narrative.
- M Duplicate precision not met.
- N Spike Sample recovery is outside control limits.
- S Reported value determined by the Method of Standard Additions (MSA)
- U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
- W Post-digestion spike recovery for GFAA out of control limit. Sample absorbency < 50% of spike absorbency.
- X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
- Y Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier

GEL LABORATORIES LLC

2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 415150

Page 2 of 2

Parmname	NOM	Sample	Qual	QC	Units	RPD/D%	REC%	Range	Anlst	Date	Time
-----------------	------------	---------------	-------------	-----------	--------------	---------------	-------------	--------------	--------------	-------------	-------------

Z Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier

N/A indicates that spike recovery limits do not apply when sample concentration exceeds spike conc. by a factor of 4 or more or %RPD not applicable.

^ The Relative Percent Difference (RPD) obtained from the sample duplicate (DUP) is evaluated against the acceptance criteria when the sample is greater than five times (5X) the contract required detection limit (RL). In cases where either the sample or duplicate value is less than 5X the RL, a control limit of +/- the RL is used to evaluate the DUP result.

* Indicates that a Quality Control parameter was not within specifications.

For PS, PSD, and SDILT results, the values listed are the measured amounts, not final concentrations.

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless qualified on the QC Summary.

General Chem Analysis

Case Narrative

**General Chemistry
 Technical Case Narrative
 CH2MHill Plateau Remediation Company (CPRC)
 SDG #: GEL415150
 Work Order #: 415150**

Product: Ion Chromatography**Analytical Method:** 9056_ANIONS_IC**Analytical Procedure:** GL-GC-E-086 REV# 25**Analytical Batches:** 1635879 and 1635878

The following samples were analyzed using the above methods and analytical procedure(s).

<u>GEL Sample ID#</u>	<u>Client Sample Identification</u>
415150001	B36M10
1203719656	Method Blank (MB)
1203719657	Laboratory Control Sample (LCS)
1203719658	415150001(B36M10) Sample Duplicate (DUP)
1203719659	415150001(B36M10) Matrix Spike (MS)

The samples in this SDG were analyzed on a "dry weight" basis.

Data Summary:

All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable, with the following exceptions.

Miscellaneous Information**Manual Integrations**

Samples 1203719658 (B36M10DUP) and 415150001 (B36M10) were manually integrated to correctly position the baseline as set in the calibration standards.

Product: pH**Analytical Method:** SW846 9045D**Analytical Procedure:** GL-GC-E-008 REV# 22**Analytical Batch:** 1634765

The following samples were analyzed using the above methods and analytical procedure(s).

GEL Sample ID#	Client Sample Identification
415150001	B36M10
1203716727	Laboratory Control Sample (LCS)
1203716728	415150001(B36M10) Sample Duplicate (DUP)

The samples in this SDG were analyzed on an "as received" basis.

Data Summary:

All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable, with the following exceptions.

Technical Information**Holding Times**

Samples (See Below) were received by the laboratory outside of the method specified holding time. The data is qualified.

Sample	Analyte	Value
1203716728 (B36M10DUP)	pH	Received 27-JAN-17, out of holding 25-JAN-17
415150001 (B36M10)	pH	Received 27-JAN-17, out of holding 25-JAN-17

Certification Statement

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless otherwise noted in the analytical case narrative.

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

**Qualifier Definition Report
for**

CPRC001 CH2MHill Plateau Remediation Company

Client SDG: GEL415150 GEL Work Order: 415150

The Qualifiers in this report are defined as follows:

B The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate).

U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.

X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier

Review/Validation

GEL requires all analytical data to be verified by a qualified data reviewer. In addition, all CLP-like deliverables receive a third level review of the fractional data package.

The following data validator verified the information presented in this data report:

Signature: **Name:** Aubrey Kingsbury**Date:** 13 FEB 2017**Title:** Analyst I

Sample Data Summary

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis

Report Date: February 13, 2017

Company : CH2MHill Plateau Remediation Company
 Address : MSIN R3-50 CHPRC
 PO Box 1600
 Richland, Washington 99352
 Contact: Mr. Scot Fitzgerald
 Project: CHPRC SAF F16-028

Client Sample ID: B36M10 Project: CPRC0F16028
 Sample ID: 415150001 Client ID: CPRC001
 Matrix: SOIL
 Collect Date: 25-JAN-17 13:30
 Receive Date: 27-JAN-17
 Collector: Client
 Moisture: 12.5%

Parameter	Qualifier	Result	DL	RL	Units	PF	DF	Analyst	Date	Time	Batch	Method
Ion Chromatography												
9056_ANIONS_IC:COMMON + (Add-on) "Dry Weight Corrected"												
Chloride		2530	823	2290	ug/Kg	10.0	1	MAR1	02/09/17	1531	1635879	1
Fluoride	B	830	389	1140	ug/Kg	10.0	1					
Nitrate-N	B	783	377	1140	ug/Kg	10.0	1					
Nitrite-N	U	377	377	1140	ug/Kg	10.0	1					
Phosphorus in phosphate	U	766	766	2290	ug/Kg	10.0	1					
Sulfate		7490	1520	4570	ug/Kg	10.0	1					

Titration and Ion Analysis

9045_pH (Non-Aqueous):COMMON "As Received"

pH at Temp 20.8C	X	8.65	0.010	0.100	SU		1	RXB5	01/27/17	1742	1634765	2
------------------	---	------	-------	-------	----	--	---	------	----------	------	---------	---

The following Prep Methods were performed:

Method	Description	Analyst	Date	Time	Prep Batch
SW846 9056A	SW846 9056A Total Anions in Soil	MXL2	02/09/17	1023	1635878

The following Analytical Methods were performed:

Method	Description	Analyst Comments
1	9056_ANIONS_IC	
2	SW846 9045D	

Notes:

Column headers are defined as follows:

DF: Dilution Factor Lc/LC: Critical Level
 DL: Detection Limit PF: Prep Factor
 MDA: Minimum Detectable Activity RL: Reporting Limit
 MDC: Minimum Detectable Concentration SQL: Sample Quantitation Limit

Quality Control Summary

GEL LABORATORIES LLC

2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Report Date: February 13, 2017

CH2MHill Plateau Remediation Company

MSIN R3-50 CHPRC

PO Box 1600

Richland, Washington

Contact: Mr. Scot Fitzgerald

Workorder: 415150

Parmname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date	Time
Ion Chromatography											
Batch	1635879										
QC1203719658	415150001	DUP									
Chloride		2530		2460	ug/Kg	3.07 ^		(+/-2290)	MAR1	02/09/17	16:03
Fluoride	B	830	B	814	ug/Kg	1.95 ^		(+/-1140)			
Nitrate-N	B	783	B	736	ug/Kg	6.17 ^		(+/-1140)			
Nitrite-N	U	377	U	377	ug/Kg	N/A					
Phosphorus in phosphate	U	766	U	766	ug/Kg	N/A					
Sulfate		7490		6930	ug/Kg	7.71 ^		(+/-4570)			
QC1203719657	LCS										
Chloride	50000			48500	ug/Kg		97.1	(80%-120%)		02/09/17	14:59
Fluoride	25000			24800	ug/Kg		99.2	(80%-120%)			
Nitrate-N	25000			24500	ug/Kg		97.9	(80%-120%)			
Nitrite-N	25000			24300	ug/Kg		97.3	(80%-120%)			
Phosphorus in phosphate	12500			13700	ug/Kg		109	(80%-120%)			
Sulfate	100000			98100	ug/Kg		98.1	(80%-120%)			
QC1203719656	MB										
Chloride			U	720	ug/Kg					02/09/17	14:27

GEL LABORATORIES LLC

2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 415150

Page 2 of 3

Parmname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date	Time
Ion Chromatography											
Batch	1635879										
Fluoride			U	340	ug/Kg				MAR1	02/09/17	14:27
Nitrate-N			U	330	ug/Kg						
Nitrite-N			U	330	ug/Kg						
Phosphorus in phosphate			U	670	ug/Kg						
Sulfate			U	1330	ug/Kg						
QC1203719659 415150001 MS											
Chloride	57000	2530		55600	ug/Kg		93.1	(48%-145%)		02/09/17	16:35
Fluoride	28500	B	830	27000	ug/Kg		91.7	(30%-135%)			
Nitrate-N	28500	B	783	27200	ug/Kg		92.8	(70%-125%)			
Nitrite-N	28500	U	377	28100	ug/Kg		98.6	(70%-120%)			
Phosphorus in phosphate	14300	U	766	13300	ug/Kg		89.8	(35%-134%)			
Sulfate	114000		7490	116000	ug/Kg		95.3	(45%-162%)			
Titration and Ion Analysis											
Batch	1634765										
QC1203716728 415150001 DUP											
pH		X	8.65	X	8.69	SU	0.461	(0%-30%)	RXB5	01/27/17	17:43
QC1203716727 LCS											
pH	7.00				6.99	SU	99.9	(70%-130%)		01/27/17	17:40

Notes:

GEL LABORATORIES LLC

2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 415150

Page 3 of 3

Parmname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date	Time
----------	-----	--------	------	----	-------	------	------	-------	-------	------	------

The Qualifiers in this report are defined as follows:

- < Sample is below the EPA guidance level for Reactive Releasable Cyanide and/or Reactive Releasable Sulfide
- > Result greater than quantifiable range or greater than upper limit of the analysis range
- B The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate).
- C Target analyte was detected in the sample and the associated blank. The associated blank concentration is \geq EQL or is $>$ 5% of the measured concentration and/or decision level for associated samples.
- D Results are reported from a diluted aliquot of sample.
- N Spike Sample recovery is outside control limits.
- U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
- X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
- Y Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
- Z Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier

N/A indicates that spike recovery limits do not apply when sample concentration exceeds spike conc. by a factor of 4 or more or %RPD not applicable.

^ The Relative Percent Difference (RPD) obtained from the sample duplicate (DUP) is evaluated against the acceptance criteria when the sample is greater than five times (5X) the contract required detection limit (RL). In cases where either the sample or duplicate value is less than 5X the RL, a control limit of +/- the RL is used to evaluate the DUP result.

* Indicates that a Quality Control parameter was not within specifications.

For PS, PSD, and SDILT results, the values listed are the measured amounts, not final concentrations.

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless qualified on the QC Summary.

Radiological Analysis

Case Narrative

Radiochemistry
Technical Case Narrative
CH2MHill Plateau Remediation Company (CPRC)
SDG #: GEL415150
Work Order #: 415150

Product: AMCMISO_EIE_PRECIP_AEA: COMMON
Analytical Method: AMCMISO_EIE_PREC_AEA
Analytical Procedure: GL-RAD-A-011 REV# 26
Analytical Batch: 1635302

Preparation Method: ASTM D 2216 (Modified)
Preparation Procedure: GL-RAD-A-021 REV# 20
Preparation Batch: 1634908

The following samples were analyzed using the above methods and analytical procedure(s).

<u>GEL Sample ID#</u>	<u>Client Sample Identification</u>
415150001	B36M10
1203718266	Method Blank (MB)
1203718267	415150001(B36M10) Sample Duplicate (DUP)
1203718268	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on a "dry weight" basis.

Data Summary:

All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable, with the following exceptions.

Technical Information

Recounts

Sample 1203718268 (LCS) was recounted due to high carrier/tracer yield. The recount is reported.

Product: NP237_IE_PRECIP_AEA: COMMON
Analytical Method: ASTM C 1475-00 Modified
Analytical Procedure: GL-RAD-A-032 REV# 21
Analytical Batch: 1635305

Preparation Method: ASTM D 2216 (Modified)
Preparation Procedure: GL-RAD-A-021 REV# 20
Preparation Batch: 1634908

The following samples were analyzed using the above methods and analytical procedure(s).

<u>GEL Sample ID#</u>	<u>Client Sample Identification</u>
415150001	B36M10
1203718269	Method Blank (MB)
1203718270	415150001(B36M10) Sample Duplicate (DUP)
1203718271	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on a "dry weight" basis.

Data Summary:

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

Product: PUISO_PRECIP_AEA:COMMON
Analytical Method: PUISO_PRECIP_AEA
Analytical Procedure: GL-RAD-A-011 REV# 26
Analytical Batch: 1635307

Preparation Method: ASTM D 2216 (Modified)
Preparation Procedure: GL-RAD-A-021 REV# 20
Preparation Batch: 1634908

The following samples were analyzed using the above methods and analytical procedure(s).

<u>GEL Sample ID#</u>	<u>Client Sample Identification</u>
415150001	B36M10
1203718272	Method Blank (MB)
1203718273	415150001(B36M10) Sample Duplicate (DUP)
1203718274	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on a "dry weight" basis.

Data Summary:

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

Product: THISO_IE_PLATE_AEA: COMMON
Analytical Method: THISO_IE_PRECIP_AEA
Analytical Procedure: GL-RAD-A-038 REV# 17

Analytical Batch: 1635308

Preparation Method: ASTM D 2216 (Modified)

Preparation Procedure: GL-RAD-A-021 REV# 20

Preparation Batch: 1634908

The following samples were analyzed using the above methods and analytical procedure(s).

<u>GEL Sample ID#</u>	<u>Client Sample Identification</u>
415150001	B36M10
1203718275	Method Blank (MB)
1203718276	415150001(B36M10) Sample Duplicate (DUP)
1203718277	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on a "dry weight" basis.

Data Summary:

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

Product: UISO_IE_PRECIP_AEA:COMMON

Analytical Method: UISO_IE_PRECIP_AEA

Analytical Procedure: GL-RAD-A-011 REV# 26

Analytical Batch: 1635309

Preparation Method: ASTM D 2216 (Modified)

Preparation Procedure: GL-RAD-A-021 REV# 20

Preparation Batch: 1634908

The following samples were analyzed using the above methods and analytical procedure(s).

<u>GEL Sample ID#</u>	<u>Client Sample Identification</u>
415150001	B36M10
1203718278	Method Blank (MB)
1203718279	415150001(B36M10) Sample Duplicate (DUP)
1203718280	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on a "dry weight" basis.

Data Summary:

All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable, with the following exceptions.

Technical Information

Recounts

Sample 415150001 (B36M10) was recounted due to a suspected false positive. The recount is reported.

Miscellaneous Information

1. The U-232 tracer peak centroid for sample 415150001 is greater than 50 keV from the expected energy value of 5302 keV. 1. The sample does meet the tracer yield requirement, the detection limits, and its tracer peak is within the U-232 region of interest. Reporting results.

Product: Dry Weight

Preparation Method: ASTM D 2216 (Modified)

Preparation Procedure: GL-RAD-A-021 REV# 20

Preparation Batch: 1634908

The following samples were analyzed using the above methods and analytical procedure(s).

<u>GEL Sample ID#</u>	<u>Client Sample Identification</u>
415150001	B36M10
1203717112	415150001(B36M10) Sample Duplicate (DUP)

The samples in this SDG were analyzed on an "as received" basis.

Data Summary:

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

Product: GAMMA_GS:COMMON + (Add-on)

Analytical Method: GAMMA_GS

Analytical Procedure: GL-RAD-A-013 REV# 25

Analytical Batch: 1635217

Preparation Method: ASTM D 2216 (Modified)

Preparation Procedure: GL-RAD-A-021 REV# 20

Preparation Batch: 1634908

The following samples were analyzed using the above methods and analytical procedure(s).

<u>GEL Sample ID#</u>	<u>Client Sample Identification</u>
415150001	B36M10
1203718016	Method Blank (MB)
1203718017	415150001(B36M10) Sample Duplicate (DUP)
1203718018	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on a "dry weight" basis.

Data Summary:

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

Qualifier Information

Qualifier	Reason	Analyte	Sample	Client Sample
X	Data rejected due to high counting uncertainty.	Europium-155	415150001	B36M10

Product: I129_SEP_LEPS_GS**Analytical Method:** I129_SEP_LEPS_GS**Analytical Procedure:** GL-RAD-A-006 REV# 21**Analytical Batch:** 1635249

The following samples were analyzed using the above methods and analytical procedure(s).

<u>GEL Sample ID#</u>	<u>Client Sample Identification</u>
415150001	B36M10
1203718103	Method Blank (MB)
1203718104	415150001(B36M10) Sample Duplicate (DUP)
1203718105	415150001(B36M10) Matrix Spike (MS)
1203718106	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

Data Summary:

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

Product: SRTOT_SEP_PRECIP_GPC: COMMON**Analytical Method:** SRTOT_SEP_PRECIP_GPC**Analytical Procedure:** GL-RAD-A-004 REV# 17**Analytical Batch:** 1635085**Preparation Method:** ASTM D 2216 (Modified)**Preparation Procedure:** GL-RAD-A-021 REV# 20**Preparation Batch:** 1634908

The following samples were analyzed using the above methods and analytical procedure(s).

<u>GEL Sample ID#</u>	<u>Client Sample Identification</u>
415150001	B36M10
1203717596	Method Blank (MB)
1203717597	415150001(B36M10) Sample Duplicate (DUP)
1203717598	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on a "dry weight" basis.

Data Summary:

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

Product: TC99_SEP_GPC

Analytical Method: TC99_EIE_LSC

Analytical Procedure: GL-RAD-A-059 REV# 4

Analytical Batch: 1635477

The following samples were analyzed using the above methods and analytical procedure(s).

<u>GEL Sample ID#</u>	<u>Client Sample Identification</u>
415150001	B36M10
1203718729	Method Blank (MB)
1203718730	415150001(B36M10) Sample Duplicate (DUP)
1203718731	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

Data Summary:

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

Product: TRITIUM_DIST_LSC: COMMON

Analytical Method: TRITIUM_DIST_LSC

Analytical Procedure: GL-RAD-A-002 REV# 21

Analytical Batch: 1636812

The following samples were analyzed using the above methods and analytical procedure(s).

<u>GEL Sample ID#</u>	<u>Client Sample Identification</u>
415150001	B36M10

1203722046	Method Blank (MB)
1203722047	415150001(B36M10) Sample Duplicate (DUP)
1203722048	415150001(B36M10) Matrix Spike (MS)
1203722049	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

Data Summary:

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

Product: C14_LSC: COMMON

Analytical Method: C14_LSC

Analytical Procedure: GL-RAD-A-003 REV# 15

Analytical Batch: 1636853

The following samples were analyzed using the above methods and analytical procedure(s).

<u>GEL Sample ID#</u>	<u>Client Sample Identification</u>
415150001	B36M10
1203722176	Method Blank (MB)
1203722177	415150001(B36M10) Sample Duplicate (DUP)
1203722178	415150001(B36M10) Matrix Spike (MS)
1203722179	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

Data Summary:

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

Product: NI63_LSC

Analytical Method: NI63_LSC

Analytical Procedure: GL-RAD-A-022 REV# 18

Analytical Batch: 1639089

Preparation Method: ASTM D 2216 (Modified)

Preparation Procedure: GL-RAD-A-021 REV# 20

Preparation Batch: 1634908

The following samples were analyzed using the above methods and analytical procedure(s).

<u>GEL Sample ID#</u>	<u>Client Sample Identification</u>
415150001	B36M10
1203727697	Method Blank (MB)
1203727698	415150001(B36M10) Sample Duplicate (DUP)
1203727699	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on a "dry weight" basis.

Data Summary:

All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable, with the following exceptions.

Quality Control (QC) Information

QC Information

All of the QC samples meet the required acceptance limits with the following exceptions: The sample and the duplicate, 1203727698 (B36M10DUP) and 415150001 (B36M10), relative error ratio is greater than 2; however, both results are less than their respective MDCs.

Technical Information

Sample Re-prep/Re-analysis

Sample 415150001 (B36M10) was re-prepped due to high recovery. The re-analysis is being reported.

Certification Statement

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless otherwise noted in the analytical case narrative.

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

**Qualifier Definition Report
for**

CPRC001 CH2MHill Plateau Remediation Company

Client SDG: GEL415150 GEL Work Order: 415150

The Qualifiers in this report are defined as follows:

U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.

X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier

Review/Validation

GEL requires all analytical data to be verified by a qualified data reviewer. In addition, all CLP-like deliverables receive a third level review of the fractional data package.

The following data validator verified the information presented in this data report:

Signature:**Name: Theresa Austin****Date: 22 FEB 2017****Title: Group Leader**

Sample Data Summary

Rad
Certificate of Analysis
Sample Summary

SDG Number: GEL415150	Client: CPRC001	Project: CPRC0F16028
Lab Sample ID: 415150001	Date Collected: 01/25/2017 13:30	Matrix: SOIL
	Date Received: 01/27/2017 08:55	%Moisture: 12.5
Client ID: B36M10		Prep Basis: "Dry Weight Corrected"
Batch ID: 1635302	Method: AMCMISO_EIE_PREC_AEA	SOP Ref: GL-RAD-A-011
Run Date: 02/04/2017 11:47	Analyst: KXB2	Instrument: 1079
Data File: S0415150001_AM.1A.gcnf	Aliquot: 0.11 g	Count Time: 240 min
Prep Batch: 1635302	Prep Method: DOE EML HASL-300, Am-05	Prep SOP Ref: GL-RAD-A-021
Prep Date: 01/31/2017 00:00		

CAS No.	Parmname	Qual	Result	Units	Uncert	TPU	MDC	RDL
14596-10-2	Americium-241	U	0.00258	pCi/g	+/-0.192	0.192	0.426	1.00

Surrogate/Tracer recovery	Result	Nominal	Units	Recovery%	Acceptable Limits
Americium-243 Tracer	18.1	19.4	pCi/g	93.4	(30%-105%)

Comments:

- U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
 - X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
- TPU and Counting Uncertainty are calculated at the 95% confidence level (1.96-sigma).
 The MDC is a sample specific MDC.

**Rad
Certificate of Analysis
Sample Summary**

SDG Number: GEL415150	Client: CPRC001	Project: CPRC0F16028
Lab Sample ID: 415150001	Date Collected: 01/25/2017 13:30	Matrix: SOIL
	Date Received: 01/27/2017 08:55	%Moisture: 12.5
Client ID: B36M10		Prep Basis: "Dry Weight Corrected"
Batch ID: 1635305	Method: ASTM C 1475-00 Modified	SOP Ref: GL-RAD-A-032
Run Date: 02/02/2017 09:39	Analyst: KXB2	Instrument: 1167
Data File: S0415150001_NP.1A.gcnf	Aliquot: 0.106 g	Count Time: 240 min
Prep Batch: 1635305	Prep Method: ASTM C 1475-00 Modified	Prep SOP Ref: GL-RAD-A-021
Prep Date: 01/31/2017 00:00		

CAS No.	Parmname	Qual	Result	Units	Uncert	TPU	MDC	RDL
13994-20-2	Neptunium-237	U	0.107	pCi/g	+/-0.301	0.302	0.322	1.00

Surrogate/Tracer recovery	Result	Nominal	Units	Recovery%	Acceptable Limits
Americium-243 Tracer	1380	2020	pCi/g	68.2	(30%-105%)

Comments:

U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
 TPU and Counting Uncertainty are calculated at the 95% confidence level (1.96-sigma).
 The MDC is a sample specific MDC.

Rad
Certificate of Analysis
Sample Summary

SDG Number: GEL415150	Client: CPRC001	Project: CPRC0F16028
Lab Sample ID: 415150001	Date Collected: 01/25/2017 13:30	Matrix: SOIL
	Date Received: 01/27/2017 08:55	%Moisture: 12.5
Client ID: B36M10		Prep Basis: "Dry Weight Corrected"
Batch ID: 1635307	Method: PUIISO_PRECIP_AEA	SOP Ref: GL-RAD-A-011
Run Date: 02/04/2017 12:10	Analyst: KXB2	Instrument: 1072
Data File: S0415150001_PU.1A.gcnf	Aliquot: 0.11 g	Count Time: 240 min
Prep Batch: 1635307	Prep Method: DOE EML HASL-300, Pu-11-	Prep SOP Ref: GL-RAD-A-021
Prep Date: 01/31/2017 00:00		

CAS No.	Parmname	Qual	Result	Units	Uncert	TPU	MDC	RDL
I3981-16-3	Plutonium-238	U	0.0338	pCi/g	+/-0.188	0.188	0.360	1.00
OER-100-70	Plutonium-239/240	U	0.0676	pCi/g	+/-0.231	0.231	0.428	1.00

Surrogate/Tracer recovery	Result	Nominal	Units	Recovery%	Acceptable Limits
Plutonium-242 Tracer	14.8	17.9	pCi/g	82.6	(30%-105%)

Comments:

- U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
 - X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
- TPU and Counting Uncertainty are calculated at the 95% confidence level (1.96-sigma).
The MDC is a sample specific MDC.

**Rad
Certificate of Analysis
Sample Summary**

SDG Number: GEL415150	Client: CPRC001	Project: CPRC0F16028
Lab Sample ID: 415150001	Date Collected: 01/25/2017 13:30	Matrix: SOIL
	Date Received: 01/27/2017 08:55	%Moisture: 12.5
Client ID: B36M10		Prep Basis: "Dry Weight Corrected"
Batch ID: 1635308	Method: THISO_IE_PRECIP_AEA	SOP Ref: GL-RAD-A-038
Run Date: 02/04/2017 11:41	Analyst: KXB2	Instrument: 1033
Data File: S0415150001_TH.1A.gcnf	Aliquot: 0.107 g	Count Time: 239.9998 min
Prep Batch: 1635308	Prep Method: DOE EML HASL-300, Th-01-	Prep SOP Ref: GL-RAD-A-021
Prep Date: 01/31/2017 00:00		

CAS No.	Parmname	Qual	Result	Units	Uncert	TPU	MDC	RDL
TH-232 <small>7440-29-1</small>	Thorium-232		0.425	pCi/g	+/-0.344	0.349	0.256	1.00

Surrogate/Tracer recovery	Result	Nominal	Units	Recovery%	Acceptable Limits
Thorium-229 Tracer	16.9	19.3	pCi/g	87.3	(30%-105%)

Comments:

- U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
 - X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
- TPU and Counting Uncertainty are calculated at the 95% confidence level (1.96-sigma).
The MDC is a sample specific MDC.

**Rad
Certificate of Analysis
Sample Summary**

SDG Number: GEL415150
Lab Sample ID: 415150001

Client: CPRC001
Date Collected: 01/25/2017 13:30
Date Received: 01/27/2017 08:55

Project: CPRC0F16028
Matrix: SOIL
%Moisture: 12.5

Client ID: B36M10
Batch ID: 1635309
Run Date: 02/08/2017 16:13
Data File: S0415150001_UU.1B.gcnf
Prep Batch: 1635309
Prep Date: 01/31/2017 00:00

Method: UIISO_IE_PRECIP_AEA
Analyst: KXB2
Aliquot: 0.11 g
Prep Method: DOE EML HASL-300, U-02-R

Prep Basis: "Dry Weight Corrected"
SOP Ref: GL-RAD-A-011
Instrument: 1115
Count Time: 240 min
Prep SOP Ref: GL-RAD-A-021

CAS No.	Parmname	Qual	Result	Units	Uncert	TPU	MDC	RDL
U-233/234 <small>13968-55-3/13966-29-5</small>	Uranium-233/234	U	0.229	pCi/g	+/-0.319	0.320	0.475	1.00
15117-96-1/13982-7	Uranium-235/236	U	-0.0572	pCi/g	+/-0.173	0.173	0.485	1.00
7440-61-1	Uranium-238	U	-0.0514	pCi/g	+/-0.314	0.314	0.709	1.00

Surrogate/Tracer recovery	Result	Nominal	Units	Recovery%	Acceptable Limits
Uranium-232 Tracer	20.0	19.0	pCi/g	105	(30%-105%)

Comments:

- U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
 - X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
- TPU and Counting Uncertainty are calculated at the 95% confidence level (1.96-sigma).
The MDC is a sample specific MDC.

**Rad
Certificate of Analysis
Sample Summary**

SDG Number: GEL415150
Lab Sample ID: 415150001

Client: CPRC001
Date Collected: 01/25/2017 13:30
Date Received: 01/27/2017 08:55

Project: CPRC0F16028
Matrix: SOIL
%Moisture: 12.5

Client ID: B36M10
Batch ID: 1635085
Run Date: 02/03/2017 14:39
Data File: S1635085.xls
Prep Batch: 1635085
Prep Date: 02/01/2017 11:16

Method: SRTOT_SEP_PRECIP_GPC
Analyst: BXF1
Aliquot: 0.52 g
Prep Method: EPA 905.0 Modified/DOE RP5

Prep Basis: "Dry Weight Corrected"
SOP Ref: GL-RAD-A-004
Instrument: PIC6A
Count Time: 60 min
Prep SOP Ref: GL-RAD-A-021

CAS No.	Parmname	Qual	Result	Units	Uncert	TPU	MDC	RDL
SR-RAD	Total Strontium		2.69	pCi/g	+/-0.837	1.09	1.06	2.00

Surrogate/Tracer recovery	Result	Nominal	Units	Recovery%	Acceptable Limits
Strontium Carrier	5.50	7.75	mg	71	(40%-110%)

Comments:

- U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
 - X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
- TPU and Counting Uncertainty are calculated at the 95% confidence level (1.96-sigma).
The MDC is a sample specific MDC.

Rad
Certificate of Analysis
Sample Summary

SDG Number: GEL415150
 Lab Sample ID: 415150001

 Client ID: B36M10
 Batch ID: 1635217
 Run Date: 02/20/2017 07:48
 Data File: G415150001.CNF;5
 Prep Batch: 1635217
 Prep Date: 01/30/2017 00:00

Client: CPRC001
 Date Collected: 01/25/2017 13:30
 Date Received: 01/27/2017 08:55

 Method: GAMMA_GS
 Analyst: MXR1
 Aliquot: 136.698 g
 Prep Method: DOE HASL 300, 4.5.2.3/Ga-01

Project: CPRC0F16028
 Matrix: SOIL
 %Moisture: 12.5

 Prep Basis: "Dry Weight Corrected"
 SOP Ref: GL-RAD-A-013
 Instrument: GAM21
 Count Time: 720 min
 Prep SOP Ref: GL-RAD-A-021

CAS No.	Parmname	Qual	Result	Units	Uncert	TPU	MDC	RDL
10045-97-3	Cesium-137	U	8.29E-06	pCi/g	+/-0.0143	0.0143	0.0205	0.100
10198-40-0	Cobalt-60	U	0.00305	pCi/g	+/-0.0127	0.0128	0.0232	0.100
14683-23-9	Europium-152	U	0.00718	pCi/g	+/-0.0218	0.0221	0.0383	0.100
15585-10-1	Europium-154	U	0.0389	pCi/g	+/-0.0417	0.0453	0.0791	0.100
14391-16-3	Europium-155	UX	0.00	pCi/g	+/-0.0297	0.0298	0.0255	0.100
13982-63-3	Radium-226		0.329	pCi/g	+/-0.0536	0.0615	0.0313	1.00
15262-20-1	Radium-228		0.498	pCi/g	+/-0.0976	0.107	0.0729	3.00

Surrogate/Tracer recovery	Result	Nominal	Units	Recovery%	Acceptable Limits
---------------------------	--------	---------	-------	-----------	-------------------

Comments:

U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
 X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
 TPU and Counting Uncertainty are calculated at the 95% confidence level (1.96-sigma).
 The MDC is a sample specific MDC.

**Rad
Certificate of Analysis
Sample Summary**

SDG Number: GEL415150	Client: CPRC001	Project: CPRC0F16028
Lab Sample ID: 415150001	Date Collected: 01/25/2017 13:30	Matrix: SOIL
	Date Received: 01/27/2017 08:55	%Moisture: 12.5
Client ID: B36M10		Prep Basis: "As Received"
Batch ID: 1635249	Method: I129_SEP_LEPS_GS	SOP Ref: GL-RAD-A-006
Run Date: 02/08/2017 14:22	Analyst: MJH1	Instrument: GAM05
Data File: I415150001.CNF;1	Aliquot: 0.998 g	Count Time: 240 min
Prep Batch: 1635249	Prep Method: DOE EML HASL-300,I-01 M	
Prep Date: 02/07/2017 00:00		

CAS No.	Parmname	Qual	Result	Units	Uncert	TPU	MDC	RDL
15046-84-1	Iodine-129	U	-0.232	pCi/g	+/-0.397	0.412	0.606	2.00

Surrogate/Tracer recovery	Result	Nominal	Units	Recovery%	Acceptable Limits
---------------------------	--------	---------	-------	-----------	-------------------

Comments:

U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
 X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
 TPU and Counting Uncertainty are calculated at the 95% confidence level (1.96-sigma).
 The MDC is a sample specific MDC.

**Rad
Certificate of Analysis
Sample Summary**

SDG Number: GEL415150	Client: CPRC001	Project: CPRC0F16028
Lab Sample ID: 415150001	Date Collected: 01/25/2017 13:30	Matrix: SOIL
	Date Received: 01/27/2017 08:55	%Moisture: 12.5
Client ID: B36M10		Prep Basis: "As Received"
Batch ID: 1635477	Method: TC99_EIE_LSC	SOP Ref: GL-RAD-A-059
Run Date: 02/14/2017 06:37	Analyst: LXT2	Instrument: LSCRED
Data File: E1635477.xls	Aliquot: 1.231 g	Count Time: 15 min
Prep Batch: 1635477	Prep Method: DOE EML HASL-300, Tc-02-	
Prep Date: 02/10/2017 10:17		

CAS No.	Parmname	Qual	Result	Units	Uncert	TPU	MDC	RDL
14133-76-7	Technetium-99	U	1.98	pCi/g	+/-2.12	2.14	3.56	5.00

Surrogate/Tracer recovery	Result	Nominal	Units	Recovery%	Acceptable Limits
Technetium-99m Tracer	16400	17800	CPM	92.2	(30%-105%)

Comments:

- U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
 - X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
- TPU and Counting Uncertainty are calculated at the 95% confidence level (1.96-sigma).
The MDC is a sample specific MDC.

Rad
Certificate of Analysis
Sample Summary

SDG Number: GEL415150	Client: CPRC001	Project: CPRC0F16028
Lab Sample ID: 415150001	Date Collected: 01/25/2017 13:30	Matrix: SOIL
	Date Received: 01/27/2017 08:55	%Moisture: 12.5
Client ID: B36M10		Prep Basis: "As Received"
Batch ID: 1636812	Method: TRITIUM_DIST_LSC	SOP Ref: GL-RAD-A-002
Run Date: 02/11/2017 03:47	Analyst: TXP3	Instrument: LSCBLUE
Data File: T1636812.xls	Aliquot: 2.12 g	Count Time: 15 min
Prep Batch: 1636812	Prep Method: EPA 906.0 Modified	
Prep Date: 02/10/2017 09:11		

CAS No.	Parmname	Qual	Result	Units	Uncert	TPU	MDC	RDL
10028-17-8	Tritium	U	-1.1	pCi/g	+/-8.35	8.35	15.3	30.0

Surrogate/Tracer recovery	Result	Nominal	Units	Recovery%	Acceptable Limits

Comments:

- U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
 - X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
- TPU and Counting Uncertainty are calculated at the 95% confidence level (1.96-sigma).
 The MDC is a sample specific MDC.

**Rad
Certificate of Analysis
Sample Summary**

SDG Number: GEL415150	Client: CPRC001	Project: CPRC0F16028
Lab Sample ID: 415150001	Date Collected: 01/25/2017 13:30	Matrix: SOIL
	Date Received: 01/27/2017 08:55	%Moisture: 12.5
Client ID: B36M10		Prep Basis: "As Received"
Batch ID: 1636853	Method: C14_LSC	SOP Ref: GL-RAD-A-003
Run Date: 02/11/2017 00:40	Analyst: TXJ1	Instrument: LSCSILVER
Data File: C1636853.xls	Aliquot: 0.511 g	Count Time: 45 min
Prep Batch: 1636853	Prep Method: EPA EERF C-01 Modified	
Prep Date: 02/10/2017 11:56		

CAS No.	Parmname	Qual	Result	Units	Uncert	TPU	MDC	RDL
14762-75-5	Carbon-14	U	-0.337	pCi/g	+/-2.17	2.17	3.74	5.00

Surrogate/Tracer recovery	Result	Nominal	Units	Recovery%	Acceptable Limits
---------------------------	--------	---------	-------	-----------	-------------------

Comments:

U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
 X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
 TPU and Counting Uncertainty are calculated at the 95% confidence level (1.96-sigma).
 The MDC is a sample specific MDC.

Rad
Certificate of Analysis
Sample Summary

SDG Number: GEL415150	Client: CPRC001	Project: CPRC0F16028
Lab Sample ID: 415150001	Date Collected: 01/25/2017 13:30	Matrix: SOIL
	Date Received: 01/27/2017 08:55	%Moisture: 12.5
Client ID: B36M10		Prep Basis: "Dry Weight Corrected"
Batch ID: 1639089	Method: NI63_LSC	SOP Ref: GL-RAD-A-022
Run Date: 02/15/2017 16:56	Analyst: CXS7	Instrument: LSCBLUE
Data File: N1639089.xls	Aliquot: 0.551 g	Count Time: 30 min
Prep Batch: 1639089	Prep Method: DOE RESL Ni-1, Modified	Prep SOP Ref: GL-RAD-A-021
Prep Date: 02/14/2017 10:35		

CAS No.	Parmname	Qual	Result	Units	Uncert	TPU	MDC	RDL
NI-63	Nickel-63	U	-5.83	pCi/g	+/-4.70	4.70	8.42	10.0

Surrogate/Tracer recovery	Result	Nominal	Units	Recovery%	Acceptable Limits
Nickel Carrier	15.7	24.6	mg	63.9	(40%-110%)

Comments:

- U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
 - X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
- TPU and Counting Uncertainty are calculated at the 95% confidence level (1.96-sigma).
The MDC is a sample specific MDC.

**Rad
Certificate of Analysis
Sample Summary**

SDG Number: GEL415150	Client: CPRC001	Project: CPRC0F16028
Lab Sample ID: 415150001	Date Collected: 01/25/2017 13:30	Matrix: SOIL
	Date Received: 01/27/2017 08:55	%Moisture: 12.5
Client ID: B36M10		Prep Basis: "As Received"
Batch ID: 1634908	Method: ASTM D 2216 (Modified)	SOP Ref: GL-OA-E-020
Run Date: 01/27/2017 15:25	Analyst: MPK1	Instrument: SP-39020004
Data File:		Count Time:
Prep Batch: 1634908		
Prep Date: 01/27/2017 15:25		

CAS No.	Parmname	Qual	Result	Units	MDC
%MOISTURE	Moisture		12.5	percent +/-	

Surrogate/Tracer recovery	Result	Nominal	Units	Recovery%	Acceptable Limits
---------------------------	--------	---------	-------	-----------	-------------------

Comments:

U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
 X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
 TPU and Counting Uncertainty are calculated at the 95% confidence level (1.96-sigma).
 The MDC is a sample specific MDC.

Quality Control Summary

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Report Date: February 22, 2017
Page 1 of 7

Client : CH2MHill Plateau Remediation Company
MSIN R3-50 CHPRC
PO Box 1600
Richland, Washington 99352
Contact: Mr. Scot Fitzgerald
Workorder: 415150

Parmname	NOM	Sample	Qual	QC	Units	QC Criteria	Range	Analyst	Date Time
Rad Alpha Spec									
Batch	1635302								
QC1203718266	MB								
Americium-241			U	-0.0101	pCi/g			KXB2	02/04/1711:47
				Uncert: +/-0.234					
				TPU: +/-0.234					
**Americium-243 Tracer	19.4			16.3	pCi/g	REC: 84	(30%-105%)		
				Uncert: +/-2.17					
				TPU: +/-3.33					
QC1203718267	415150001	DUP							
Americium-241		U	0.00258	U	0.0206	pCi/g			02/04/1711:47
				Uncert: +/-0.192	+/-0.235	RPD: 0	N/A		
				TPU: +/-0.192	+/-0.235	RER: 0.116	(0-2)		
**Americium-243 Tracer	21.2		18.1	21.0	pCi/g	REC: 99	(30%-105%)		
				Uncert: +/-2.19	+/-2.28				
				TPU: +/-3.36	+/-3.53				
QC1203718268	LCS								
Americium-241				17.9	pCi/g	REC: 87	(80%-120%)		02/08/1715:08
				Uncert: +/-2.02					
				TPU: +/-2.90					
**Americium-243 Tracer	19.4			18.4	pCi/g	REC: 95	(30%-105%)		
				Uncert: +/-2.25					
				TPU: +/-3.44					
Batch	1635305								
QC1203718269	MB								
Neptunium-237			U	0.0744	pCi/g			KXB2	02/02/1709:39
				Uncert: +/-0.287					
				TPU: +/-0.287					
**Americium-243 Tracer	2020			2030	pCi/g	REC: 100	(30%-105%)		
QC1203718270	415150001	DUP							
Neptunium-237		U	0.107	U	0.00547	pCi/g			02/02/1709:39
				Uncert: +/-0.301	+/-0.251	RPD: 0	N/A		
				TPU: +/-0.302	+/-0.251	RER: 0.508	(0-2)		
**Americium-243 Tracer	2080		1380	2120	pCi/g	REC: 102	(30%-105%)		
QC1203718271	LCS								
Neptunium-237				42.1	pCi/g	REC: 99	(80%-120%)		02/02/1709:40
				Uncert: +/-3.24					
				TPU: +/-5.61					
**Americium-243 Tracer	2020			2110	pCi/g	REC: 104	(30%-105%)		
Batch	1635307								
QC1203718272	MB								
Plutonium-238			U	-0.0371	pCi/g			KXB2	02/04/1712:10
				Uncert: +/-0.164					
				TPU: +/-0.164					
Plutonium-239/240			U	-0.0185	pCi/g				
				Uncert: +/-0.160					

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 415150

Page 2 of 7

Parname	NOM	Sample	Qual	QC	Units	QC Criteria	Range	Analyst	Date Time
Rad Alpha Spec									
Batch	1635307								
**Plutonium-242 Tracer	17.9	TPU:		+/-0.160					
		Uncert:		14.7	pCi/g	REC:	82 (30%-105%)		
		TPU:		+/-2.32					
				+/-3.44					
QC1203718273 415150001 DUP									
Plutonium-238		U	0.0338	U	0.00503				02/04/1711:47
		Uncert:	+/-0.188		+/-0.231	RPD:	0 N/A		
		TPU:	+/-0.188		+/-0.231	RER:	0.189 (0-2)		
Plutonium-239/240		U	0.0676	U	-0.0453				
		Uncert:	+/-0.231		+/-0.137	RPD:	0 N/A		
		TPU:	+/-0.231		+/-0.137	RER:	0.823 (0-2)		
**Plutonium-242 Tracer	19.5		14.8		17.2	pCi/g	REC:	88 (30%-105%)	
		Uncert:	+/-2.13		+/-2.18				
		TPU:	+/-3.18		+/-3.29				
QC1203718274 LCS									
Plutonium-238				U	0.190	pCi/g			
		Uncert:			+/-0.285				
		TPU:			+/-0.286				
Plutonium-239/240	18.0				17.2	pCi/g	REC:	96 (80%-120%)	
		Uncert:			+/-1.95				
		TPU:			+/-2.90				
**Plutonium-242 Tracer	17.9				17.4	pCi/g	REC:	97 (30%-105%)	
		Uncert:			+/-1.99				
		TPU:			+/-3.00				
Batch	1635308								
QC1203718275 MB									
Thorium-232				U	0.243	pCi/g		KXB2	02/04/1711:41
		Uncert:			+/-0.296				
		TPU:			+/-0.298				
**Thorium-229 Tracer	19.3				16.1	pCi/g	REC:	83 (30%-105%)	
		Uncert:			+/-2.24				
		TPU:			+/-3.60				
QC1203718276 415150001 DUP									
Thorium-232			0.425	U	0.314	pCi/g			
		Uncert:	+/-0.344		+/-0.321	RPD:	10 (0% - 100%)		
		TPU:	+/-0.349		+/-0.324	RER:	0.456 (0-2)		
**Thorium-229 Tracer	20.5		16.9		20.0	pCi/g	REC:	97 (30%-105%)	
		Uncert:	+/-2.16		+/-2.22				
		TPU:	+/-3.50		+/-3.63				
QC1203718277 LCS									
Thorium-232	18.6				17.8	pCi/g	REC:	96 (80%-120%)	02/04/1711:41
		Uncert:			+/-1.83				
		TPU:			+/-2.99				
**Thorium-229 Tracer	19.3				19.5	pCi/g	REC:	101 (30%-105%)	
		Uncert:			+/-1.91				
		TPU:			+/-3.21				
Batch	1635309								
QC1203718278 MB									
Uranium-233/234				U	0.0217	pCi/g		KXB2	02/04/1711:13
		Uncert:			+/-0.248				

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 415150

Page 3 of 7

Parname	NOM	Sample	Qual	QC	Units	QC Criteria	Range	Analyst	Date	Time
Rad Alpha Spec										
Batch	1635309									
Uranium-235/236		TPU:		+/-0.248						
			U	0.0437	pCi/g					
		Uncert:		+/-0.243						
Uranium-238		TPU:		+/-0.243						
			U	-0.0136	pCi/g					
		Uncert:		+/-0.204						
**Uranium-232 Tracer	19.0	TPU:		+/-0.204						
				16.5	pCi/g	REC:	87	(30%-105%)		
		Uncert:		+/-2.27						
		TPU:		+/-3.51						
QC1203718279 415150001 DUP										
Uranium-233/234		U	0.229	U	0.129	pCi/g				02/04/1711:13
		Uncert:	+/-0.319		+/-0.254		RPD:	0	N/A	
		TPU:	+/-0.320		+/-0.255		RER:	0.476	(0-2)	
Uranium-235/236		U	-0.0572	U	0.00	pCi/g				
		Uncert:	+/-0.173		+/-0.183		RPD:	0	N/A	
		TPU:	+/-0.173		+/-0.184		RER:	0.444	(0-2)	
Uranium-238		U	-0.0514	U	0.332	pCi/g				
		Uncert:	+/-0.314		+/-0.358		RPD:	0	N/A	
		TPU:	+/-0.314		+/-0.361		RER:	1.57	(0-2)	
**Uranium-232 Tracer	20.7	20.0		19.8	pCi/g	REC:	96	(30%-105%)		
		Uncert:	+/-2.18		+/-2.43					
		TPU:	+/-3.39		+/-3.78					
QC1203718280 LCS										
Uranium-233/234					22.0	pCi/g				02/04/1711:13
		Uncert:			+/-2.14					
		TPU:			+/-3.55					
Uranium-235/236					1.00	pCi/g				
		Uncert:			+/-0.525					
		TPU:			+/-0.540					
Uranium-238	24.5				24.7	pCi/g	REC:	101	(80%-120%)	
		Uncert:			+/-2.27					
		TPU:			+/-3.91					
**Uranium-232 Tracer	19.0				18.9	pCi/g	REC:	100	(30%-105%)	
		Uncert:			+/-1.99					
		TPU:			+/-3.16					
Rad Gamma Spec										
Batch	1635217									
QC1203718016 MB										
Cesium-137				U	-0.0145	pCi/g			MXR1	02/20/1707:48
		Uncert:			+/-0.0119					
		TPU:			+/-0.0136					
Cobalt-60				U	0.00232	pCi/g				
		Uncert:			+/-0.0109					
		TPU:			+/-0.011					
Europium-152				U	-0.00959	pCi/g				
		Uncert:			+/-0.0275					
		TPU:			+/-0.0278					
Europium-154				U	0.0315	pCi/g				
		Uncert:			+/-0.0737					

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 415150

Page 4 of 7

Parmname	NOM	Sample	Qual	QC	Units	QC Criteria	Range	Analyst	Date Time
Rad Gamma Spec									
Batch	1635217								
Europium-155		TPU:		+/-0.0751					
			U	-0.00585	pCi/g				
		Uncert:		+/-0.0196					
Radium-226		TPU:		+/-0.0197					
			U	-0.00801	pCi/g				
		Uncert:		+/-0.0341					
Radium-228		TPU:		+/-0.0343					
			U	0.0311	pCi/g				
		Uncert:		+/-0.061					
		TPU:		+/-0.0626					
QC1203718017	415150001	DUP							
Cesium-137		U	8.29E-06	U	-0.00226	pCi/g			02/20/1715:10
		Uncert:	+/-0.0143		+/-0.0115		RPD: 0	N/A	
		TPU:	+/-0.0143		+/-0.0115		RER: 0.48	(0-2)	
Cobalt-60		U	0.00305	U	-0.00379	pCi/g			
		Uncert:	+/-0.0127		+/-0.013		RPD: 0	N/A	
		TPU:	+/-0.0128		+/-0.0131		RER: 0.79	(0-2)	
Europium-152		U	0.00718	U	-0.007	pCi/g			
		Uncert:	+/-0.0218		+/-0.0252		RPD: 0	N/A	
		TPU:	+/-0.0221		+/-0.0254		RER: 0.937	(0-2)	
Europium-154		U	0.0389	U	0.0156	pCi/g			
		Uncert:	+/-0.0417		+/-0.0387		RPD: 0	N/A	
		TPU:	+/-0.0453		+/-0.0393		RER: 1.45	(0-2)	
Europium-155		UX	0.00	U	0.0214	pCi/g			
		Uncert:	+/-0.0297		+/-0.0228		RPD: 0	N/A	
		TPU:	+/-0.0298		+/-0.0248		RER: 0.567	(0-2)	
Radium-226			0.329		0.330	pCi/g			
		Uncert:	+/-0.0536		+/-0.0564		RPD: 11	(0% - 20%)	
		TPU:	+/-0.0615		+/-0.0582		RER: 0.607	(0-2)	
Radium-228			0.498		0.538	pCi/g			
		Uncert:	+/-0.0976		+/-0.106		RPD: 6	(0% - 20%)	
		TPU:	+/-0.107		+/-0.110		RER: 0.308	(0-2)	
QC1203718018	LCS								
Americium-241		489			520	pCi/g	REC: 106	(80%-120%)	02/20/1708:52
		Uncert:			+/-7.19				
		TPU:			+/-38.8				
Cesium-137		178			185	pCi/g	REC: 104	(80%-120%)	
		Uncert:			+/-3.13				
		TPU:			+/-8.51				
Cobalt-60		154			149	pCi/g	REC: 97	(80%-120%)	
		Uncert:			+/-3.22				
		TPU:			+/-6.36				
Europium-152				U	0.102	pCi/g			
		Uncert:			+/-1.23				
		TPU:			+/-1.23				
Europium-154				U	-0.339	pCi/g			
		Uncert:			+/-0.784				
		TPU:			+/-0.799				
Europium-155				U	0.225	pCi/g			
		Uncert:			+/-1.03				

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 415150

Page 5 of 7

Parname	NOM	Sample	Qual	QC	Units	QC Criteria	Range	Analyst	Date	Time
Rad Gamma Spec										
Batch	1635217									
		TPU:		+/-1.03						
Radium-226			U	-0.279	pCi/g					
		Uncert:		+/-0.929						
		TPU:		+/-0.937						
Radium-228			U	-3.79	pCi/g					
		Uncert:		+/-2.65						
		TPU:		+/-3.15						
Batch	1635249									
QC1203718103	MB									
Iodine-129			U	0.142	pCi/g			MJH1	02/08/1714:22	
		Uncert:		+/-0.274						
		TPU:		+/-0.282						
QC1203718104	415150001	DUP								
Iodine-129		U	-0.232	U	0.0233	pCi/g			02/08/1714:23	
		Uncert:	+/-0.397		+/-0.235		RPD: 0	N/A		
		TPU:	+/-0.412		+/-0.235		RER: 1.05	(0-2)		
QC1203718105	415150001	MS								
Iodine-129		41.7	U	-0.232	37.8	pCi/g	REC: 91	(75%-125%)	02/08/1714:23	
		Uncert:	+/-0.397		+/-2.21					
		TPU:	+/-0.412		+/-4.37					
QC1203718106	LCS									
Iodine-129		41.7			41.5	pCi/g	REC: 100	(80%-120%)	02/08/1714:24	
		Uncert:			+/-2.24					
		TPU:			+/-4.75					
Rad Gas Flow										
Batch	1635085									
QC1203717596	MB									
Total Strontium			U	0.935	pCi/g			BXF1	02/03/1714:39	
		Uncert:		+/-0.682						
		TPU:		+/-0.724						
**Strontium Carrier		7.75			6.80	mg	REC: 88	(40%-110%)		
QC1203717597	415150001	DUP								
Total Strontium			2.69		2.77	pCi/g			02/03/1714:39	
		Uncert:	+/-0.837		+/-0.929		RPD: 3	(0% - 100%)		
		TPU:	+/-1.09		+/-1.17		RER: 0.103	(0-2)		
**Strontium Carrier		7.75	5.50		4.80	mg	REC: 62	(40%-110%)		
QC1203717598	LCS									
Total Strontium		42.4			44.8	pCi/g	REC: 106	(80%-120%)	02/03/1714:39	
		Uncert:			+/-2.23					
		TPU:			+/-11.7					
**Strontium Carrier		7.75			7.30	mg	REC: 94	(40%-110%)		
Rad Liquid Scintillation										
Batch	1635477									
QC1203718729	MB									
Technetium-99			U	1.22	pCi/g			LXT2	02/14/1707:10	
		Uncert:		+/-1.99						
		TPU:		+/-2.00						
**Technetium-99m Tracer		17800			16800	CPM	REC: 94	(30%-105%)		
QC1203718730	415150001	DUP								

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 415150

Page 6 of 7

Parmname	NOM	Sample	Qual	QC	Units	QC Criteria	Range	Analyst	Date Time
Rad Liquid Scintillation									
Batch	1635477								
Techneium-99		U	1.98	U	2.77	pCi/g			
		Uncert:	+/-2.12		+/-2.06		RPD: 0	N/A	
		TPU:	+/-2.14		+/-2.08		RER: 0.522	(0-2)	
**Techneium-99m Tracer	17800		16400		16900	CPM	REC: 95	(30%-105%)	
QC1203718731	LCS								
Techneium-99		68.3			66.1	pCi/g	REC: 97	(80%-120%)	02/14/1707:43
		Uncert:			+/-3.85				
		TPU:			+/-8.52				
**Techneium-99m Tracer	17800				17100	CPM	REC: 96	(30%-105%)	
Batch	1636812								
QC1203722046	MB								
Tritium				U	-5.48	pCi/g		TXP3	02/11/1704:20
		Uncert:			+/-7.75				
		TPU:			+/-7.75				
QC1203722047	415150001	DUP							
Tritium		U	-1.1	U	3.00	pCi/g			02/11/1704:37
		Uncert:	+/-8.35		+/-8.72		RPD: 0	N/A	
		TPU:	+/-8.35		+/-8.74		RER: 0.666	(0-2)	
QC1203722048	415150001	MS							
Tritium		55.8	U	-1.1	51.1	pCi/g	REC: 92	(75%-125%)	02/11/1704:53
		Uncert:	+/-8.35		+/-12.2				
		TPU:	+/-8.35		+/-16.8				
QC1203722049	LCS								
Tritium		53.8			45.7	pCi/g	REC: 85	(80%-120%)	02/11/1705:09
		Uncert:			+/-11.5				
		TPU:			+/-15.5				
Batch	1636853								
QC1203722176	MB								
Carbon-14				U	-1.53	pCi/g		TXJ1	02/11/1703:30
		Uncert:			+/-2.14				
		TPU:			+/-2.14				
QC1203722177	415150001	DUP							
Carbon-14		U	-0.337	U	1.42	pCi/g			02/11/1704:17
		Uncert:	+/-2.17		+/-2.20		RPD: 0	N/A	
		TPU:	+/-2.17		+/-2.20		RER: 1.11	(0-2)	
QC1203722178	415150001	MS							
Carbon-14		149	U	-0.337	140	pCi/g	REC: 94	(75%-125%)	02/11/1705:04
		Uncert:	+/-2.17		+/-4.42				
		TPU:	+/-2.17		+/-11.2				
QC1203722179	LCS								
Carbon-14		145			140	pCi/g	REC: 97	(80%-120%)	02/11/1705:51
		Uncert:			+/-4.36				
		TPU:			+/-11.2				
Batch	1639089								
QC1203727697	MB								
Nickel-63				U	-4.29	pCi/g		CXS7	02/15/1718:32
		Uncert:			+/-4.01				
		TPU:			+/-4.01				
**Nickel Carrier	24.6				15.3	mg	REC: 62	(40%-110%)	
QC1203727698	415150001	DUP							

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 415150

Page 7 of 7

Parname	NOM	Sample	Qual	QC	Units	QC Criteria	Range	Analyst	Date	Time
Rad Liquid Scintillation										
Batch	1639089									
Nickel-63		U	-5.83	U	3.15	pCi/g				
		Uncert:	+/-4.70		+/-4.72		RPD:	0	N/A	
		TPU:	+/-4.70		+/-4.76		RER:	2.63	(0-2)	
**Nickel Carrier	24.6		15.7		17.2	mg	REC:	70	(40%-110%)	
QC1203727699	LCS									
Nickel-63	204				198	pCi/g	REC:	97	(80%-120%) 02/15/1719:37	
		Uncert:			+/-8.03					
		TPU:			+/-37.4					
**Nickel Carrier	24.6				16.1	mg	REC:	66	(40%-110%)	

Notes:

TPU and Counting Uncertainty are calculated at the 95% confidence level (1.96-sigma).

The Qualifiers in this report are defined as follows:

- * Duplicate analysis not within control limits
- + Correlation coefficient for Method of Standard Additions (MSA) is < 0.995
- < Sample is below the EPA guidance level for Reactive Releasable Cyanide and/or Reactive Releasable Sulfide
- > Result greater than quantifiable range or greater than upper limit of the analysis range
- B The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate).
- B The associated QC sample blank has a result $\geq 2X$ the MDA and, after corrections, result is \geq MDA for this sample
- C Target analyte was detected in the sample and the associated blank. The associated blank concentration is \geq EQL or is $> 5\%$ of the measured concentration and/or decision level for associated samples.
- D Results are reported from a diluted aliquot of sample.
- E Reported value is estimated due to interferences. See comment in narrative.
- M Duplicate precision not met.
- N Spike Sample recovery is outside control limits.
- S Reported value determined by the Method of Standard Additions (MSA)
- U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
- UX Gamma Spectroscopy--Uncertain identification
- W Post-digestion spike recovery for GFAA out of control limit. Sample absorbency $< 50\%$ of spike absorbency.
- X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
- Y Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
- Z Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier

N/A indicates that spike recovery limits do not apply when sample concentration exceeds spike conc. by a factor of 4 or more or %RPD not applicable.

** Indicates analyte is a surrogate compound.

^ The Relative Percent Difference (RPD) obtained from the sample duplicate (DUP) is evaluated against the acceptance criteria when the sample is greater than five times (5X) the contract required detection limit (RL). In cases where either the sample or duplicate value is less than 5X the RL, a control limit of +/- the RL is used to evaluate the DUP result.

For PS, PSD, and SDILT results, the values listed are the measured amounts, not final concentrations.

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless qualified on the QC Summary.