GEL Laboratories LLC
a member of The GEL Group INC

PO Box 30712 Charleston, SC 29417
2040 Savage Road Charleston, SC 29407
P 843.556.8171
F 843.766.1178

gel.com

June 13, 2017

Mr. Scot Fitzgerald
CH2MHill Plateau Remediation Company
MSIN R3-50 CHPRC
PO Box 1600
Richland, Washington 99352

Re: CHPRC SAF I17-006
Work Order: 424397
SDG: GEL424397

Dear Mr. Fitzgerald:

GEL Laboratories, LLC (GEL) appreciates the opportunity to provide the enclosed analytical results for the sample(s) we received on June 01, 2017. This original data report has been prepared and reviewed in accordance with GEL's standard operating procedures.

Our policy is to provide high quality, personalized analytical services to enable you to meet your analytical needs on time every time. We trust that you will find everything in order and to your satisfaction. If you have any questions, please do not hesitate to call me at (843) 556-8171, ext. 4505.

Sincerely,

B Luthman
Brielle Luthman for
Heather Shaffer
Project Manager

Purchase Order: 300071 - 7H
Chain of Custody: 117-006-078, 117-006-129, 117-006-141 and 117-006-143
Enclosures

Table of Contents

Case Narrative

General Narrative
for
CH2MHill Plateau Remediation Company
CHPRC SAF I17-006
SDG: GEL424397

June 13, 2017

Laboratory Identification:

GEL Laboratories LLC
2040 Savage Road
Charleston, South Carolina 29407
(843) 556-8171

Summary

Sample receipt

The sample(s) arrived at GEL Laboratories, LLC, Charleston, South Carolina on June 01, 2017, for analysis. The samples were delivered with proper chain of custody documentation and signatures. All sample containers arrived without any visible signs of tampering or breakage. There are no additional comments concerning sample receipt.

Items of Note All efforts were made by the lab to meet any short hold times. Samples that were analyzed outside of the initial hold time but still within 2X hold time will be noted in the lab case narrative.

Sample Identification
The laboratory received the following samples:

Laboratory Identification	Sample Description
424397001	B39FD4
424397002	B39FB1
424397003	B39FF3
424397004	B39FF7

Case Narrative

Sample analyses were conducted using methodology as outlined in GEL Laboratories, LLC (GEL) Standard Operating Procedures. Any technical or administrative problems during analysis, data review, and reduction are contained in the analytical case narratives in the enclosed data package.

Data Package

The enclosed data package contains the following sections: General Narrative, Chain of Custody and Supporting Documentation, and data from the following fractions: General Chemistry and Radiochemistry.

We certify that this package is in compliance with the SOW, both technically and for completeness, including a full description of, explanation of, and corrective actions for, any and all deviations, from either the analyses requested or the case narrative requested. Release of the data contained in this hard copy data package has been authorized by the Laboratory Analytical Manager (or designee) and the laboratory's client services representative as verified by their signatures on this report.

B Luthman

Brielle Luthman for
Heather Shaffer
Project Manager

Technical Case Narrative
CH2MHill Plateau Remediation Company (CPRC)
SDG #: GEL424397
Work Order #: 424397

General Chemistry

Ion Chromatography

All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable, with the following exceptions.

Technical Information

Sample Dilutions
The following samples 1203801747 (B39FD4DUP), 1203801748 (B39FD4PS) and 424397001 (B39FD4) were diluted because target analyte concentrations exceeded the calibration range.

Analyte	424397 001
Chloride	2X
Sulfate	2X

Radiochemistry

C14_LSC: COMMON

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

Certification Statement

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless otherwise noted in the analytical case narrative.

Chain of Custody and Supporting Documentation

CH2MHill Plateau Remediation Company

CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST

424397

C.O.C. # I17-006-078

Page 1 of 1

Collector	Roy Shepard /CHPRC	**Contact/Requester**	Karen Waters-Husted	**Telephone No.**	509-376-4650
SAF No.	I17-006	**Sampling Origin**	Hanford Site	**Purchase Order/Charge Code**	300071
Project Title	100KR4, May 2017	**Logbook No.**	HNF-N-506 88164	**Ice Chest No.**	GWS 350
Shipped To (Lab)	GEL Laboratories, LLC	**Method of Shipment**	Commercial Carrier	**Bill of Lading/Air Bill No.**	7792 6984 3245
Protocol	CERCLA	**Priority:**	30 Days **PRIORITY**	**Offsite Property No.**	7961

POSSIBLE SAMPLE HAZARDS/REMARKS

** ** Contains Radioactive Material at concentrations that are not regulated for transportation per 49 CFR / IATA Dangerous Goods Regulations but are not releasable per DOE Order 458.1

SPECIAL INSTRUCTIONS **Hold Time**

N/A

Total Activity Exemption: Yes ☑ No ☐

Sample No.	Filter	*	Date	Time	No/Type Container	Sample Analysis	Holding Time	Preservative
B39FD4	N	W	5-31-17	0945	1x125-mL G/P	9056_ANIONS_IC: COMMON	48 Hours	Cool <=6C

June 26, 2017

Relinquished By	Print	Sign	Date/Time	Received By	Print	Sign	Date/Time
	Roy Shepard /CHPRC		MAY 31 2017 1020		Christina Aguilar /CHPRC		MAY 31 2017 1020
Relinquished By	Christina Aguilar /CHPRC		MAY 31 2017 1400	Received By	FEDEX		Date/Time
Relinquished By			FED EX	Received By	STACY BOONE		6-1-17 9:10
Relinquished By			Date/Time	Received By			Date/Time

Matrix *

S	= Soil	DS	= Drum Solids		
SE	= Sediment	DL	= Drum Liquids		
SO	= Solid	T	= Tissue		
SL	= Sludge	WI	= Wipe		
W	= Water	L	= Liquid		
O	= Oil	V	= Vegetation		
A	= Air	X	= Other		

FINAL SAMPLE DISPOSITION Disposal Method (e.g., Return to customer, per lab procedure, used in process) Disposed By Date/Time

PRINTED ON 4/4/2017 FSR ID = FSR28080 A-6004-842 (REV 2)

77105

CH2MHill Plateau Remediation Company

CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST

424397

C.O.C. # **I17-006-129**

Page 1 of 1

Collector	Roy Shepard /CHPRC	**Contact/Requester**	Karen Waters-Husted	**Telephone No.**	509-376-4650
SAF No.	I17-006	**Sampling Origin**	Hanford Site	**Purchase Order/Charge Code**	300071
Project Title	100KR4, May 2017	**Logbook No.**	HNF-N-506 88164	**Ice Chest No.**	Cws-350
Shipped To (Lab)	GEL Laboratories, LLC	**Method of Shipment**	Commercial Carrier	**Bill of Lading/Air Bill No.**	7792 6984 3245
Protocol	CERCLA	**Priority:**	30 Days **PRIORITY**	**Offsite Property No.**	79601

POSSIBLE SAMPLE HAZARDS/REMARKS

** ** Contains Radioactive Material at concentrations that are not regulated for transportation per 49 CFR / IATA Dangerous Goods Regulations but are not releasable per DOE Order 458.1

SPECIAL INSTRUCTIONS **Hold Time**

N/A

Total Activity Exemption: Yes ☑ No ☐

Sample No.	Filter	*	Date	Time	No/Type Container	Sample Analysis	Holding Time	Preservative
B39FB1	N	W	5-31-17	0823	1x500-mL G/P	C14_LSC: COMMON	6 Months	None

June 26, 2017

Relinquished By	Print	Sign	Date/Time	Received By	Print	Sign	Date/Time
	Roy Shepard /CHPRC		MAY 31 2017 1020		Christina Aguilar /CHPRC		MAY 31 2017 1020
Relinquished By	Christina Aguilar /CHPRC		MAY 31 2017 1400	Received By	FEDEX		Date/Time
Relinquished By			FED EX	Received By	STACY BOOUE		6/1/17 9:10
Relinquished By			Date/Time	Received By			Date/Time

Matrix *

S	= Soil	DS	=	Drum Solids	
SE	= Sediment	DL	=	Drum Liquids	
SO	= Solid	T	=	Tissue	
SL	= Sludge	WI	=	Wipe	
W	= Water	L	=	Liquid	
O	= Oil	V	=	Vegetation	
A	= Air	X	=	Other	

FINAL SAMPLE DISPOSITION

Disposal Method (e.g., Return to customer, per lab procedure, used in process)

Disposed By

Date/Time

PRINTED ON 3/27/2017

FSR ID = FSR26415

A-6004-842 (REV 2)

CH2MHill Plateau Remediation Company

CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST

4243997

C.O.C. # **I17-006-141**

Page 1 of 1

Collector	Roy Shepard /CHPRC	**Contact/Requester**	Karen Waters-Husted	**Telephone No.**	509-376-4650
SAF No.	I17-006	**Sampling Origin**	Hanford Site	**Purchase Order/Charge Code**	300071
Project Title	100KR4, May 2017	**Logbook No.**	HNF-N-506 88164	**Ice Chest No.**	GWS-350
Shipped To (Lab)	GEL Laboratories, LLC	**Method of Shipment**	Commercial Carrier	**Bill of Lading/Air Bill No.**	7792 6984 3245
Protocol	CERCLA	**Priority:**	**30 Days** **PRIORITY**	**Offsite Property No.**	79601

POSSIBLE SAMPLE HAZARDS/REMARKS

** ** Contains Radioactive Material at concentrations that are not regulated for transportation per 49 CFR / IATA Dangerous Goods Regulations but are not releasable per DOE Order 458.1

SPECIAL INSTRUCTIONS **Hold Time** Total Activity Exemption: Yes ☑ No ☐

N/A

Sample No.	Filter	*	Date	Time	No/Type Container	Sample Analysis	Holding Time	Preservative
B39FF3	N	W	5-31-17	1130	1x500-mL G/P	C14_LSC: COMMON	6 Months	None

June 26, 2017

Relinquished By	Print	Sign	Date/Time	Received By	Print	Sign	Date/Time
	Roy Shepard /CHPRC		MAY 31 2017 1145		Christina Aguilar /CHPRC		MAY 31 2017 1145
Relinquished By	Christina Aguilar /CHPRC		MAY 31 2017 1400	Received By	FEDEX		Date/Time
Relinquished By			FED EX	Received By	STACY BOONE		6-1-17 9:10
Relinquished By			Date/Time	Received By			Date/Time

Matrix *

S	= Soil	DS	=	Drum Solids
SE	= Sediment	DL	=	Drum Liquids
SO	= Solid	T	=	Tissue
SL	= Sludge	WI	=	Wipe
W	= Water	L	=	Liquid
O	= Oil	V	=	Vegetation
A	= Air	X	=	Other

FINAL SAMPLE DISPOSITION Disposal Method (e.g., Return to customer, per lab procedure, used in process) Disposed By Date/Time

PRINTED ON 3/27/2017 **FSR ID = FSR38923** A-6004-842 (REV 2)

CH2MHill Plateau Remediation Company

CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST

424397

C.O.C. # I17-006-143

Page 1 of 1

Collector	Roy Shepard /CHPRC	**Contact/Requester**	Karen Waters-Husted	**Telephone No.**	509-376-4650
SAF No.	I17-006	**Sampling Origin**	Hanford Site	**Purchase Order/Charge Code**	300071
Project Title	100KR4, May 2017	**Logbook No.**	HNF-N-506 88164	**Ice Chest No.**	GWS-350
Shipped To (Lab)	GEL Laboratories, LLC	**Method of Shipment**	Commercial Carrier	**Bill of Lading/Air Bill No.**	7792 6984 3245 7961
Protocol	CERCLA	**Priority:**	30 Days **PRIORITY**	**Offsite Property No.**	

POSSIBLE SAMPLE HAZARDS/REMARKS

** ** Contains Radioactive Material at concentrations that are not regulated for transportation per 49 CFR / IATA Dangerous Goods Regulations but are not releasable per DOE Order 458.1

SPECIAL INSTRUCTIONS **Hold Time** Total Activity Exemption: Yes ☑ No ☐

N/A

Sample No.	Filter	*	Date	Time	No/Type Container	Sample Analysis	Holding Time	Preservative
B39FF7	N	W	5-31-17	1110	1x500-mL G/P	C14_LSC: COMMON	6 Months	None

June 26, 2017

Relinquished By	Print	Sign	Date/Time	Received By	Print	Sign	Date/Time
Roy Shepard /CHPRC			MAY 31 2017 1145	Christina Aguilar /CHPRC			MAY 31 2017 1145
Christina Aguilar /CHPRC			MAY 31 2017 1400	FEDEX			
			FEDEX	STACY BOONE			6/1/17 9:10

Matrix *

S	= Soil	DS	= Drum Solids		
SE	= Sediment	DL	= Drum Liquids		
SO	= Solid	T	= Tissue		
SL	= Sludge	WI	= Wipe		
W	= Water	L	= Liquid		
O	= Oil	V	= Vegetation		
A	= Air	X	= Other		

FINAL SAMPLE DISPOSITION Disposal Method (e.g., Return to customer, per lab procedure, used in process) Disposed By Date/Time

PRINTED ON 3/27/2017 FSR ID = FSR38924 A-6004-842 (REV 2)

GEL | Laboratories LLC

June 26, 2017

SAMPLE RECEIPT & REVIEW FORM

Client: CPRC	SDG/AR/COC/Work Order: 424397
Received By: Stacy Boone	Date Received: 1-JUNE-17
Carrier and Tracking Number	Circle Applicable: FedEx Express FedEx Ground UPS Field Services Courier Other 7792 6984 3245

Suspected Hazard Information	Yes	No	*If Net Counts > 100cpm on samples not marked "radioactive", contact the Radiation Safety Group for further investigation.
Shipped as a DOT Hazardous?		✓	Hazard Class Shipped: UN#:
COC/Samples marked or classified as radioactive?		✓	Maximum Net Counts Observed* (Observed Counts - Area Background Counts): 8 CPM / mR/Hr Classified as: Rad 1 Rad 2 Rad 3
Is package, COC, and/or Samples marked HAZ?		✓	If yes, select Hazards below, and contact the GEL Safety Group. PCB's Flammable Foreign Soil RCRA Asbestos Beryllium Other:

	Sample Receipt Criteria	Yes	NA	No	Comments/Qualifiers (Required for Non-Conforming Items)
1	Shipping containers received intact and sealed?	✓			Circle Applicable: Seals broken Damaged container Leaking container Other (describe)
2	Chain of custody documents included with shipment?	✓			
3	Samples requiring cold preservation within (0 ≤ 6 deg. C)?*	✓			Preservation Method: Wet Ice Ice Packs Dry ice None Other: *all temperatures are recorded in Celsius TEMP: 1 c
4	Daily check performed and passed on IR temperature gun?	✓			Temperature Device Serial #: IR3-17 Secondary Temperature Device Serial # (If Applicable):
5	Sample containers intact and sealed?	✓			Circle Applicable: Seals broken Damaged container Leaking container Other (describe)
6	Samples requiring chemical preservation at proper pH?	✓			Sample ID's and Containers Affected: If Preservation added, Lot#:
7	Do any samples require Volatile Analysis?			✓	If Yes, Are Encores or Soil Kits present? Yes___ No___ (If yes, take to VOA Freezer) Do VOA vials contain acid preservation? Yes No___ N/A (If unknown, select No) VOA vials free of headspace? Yes No___ N/A Sample ID's and containers affected:
8	Samples received within holding time?	✓			ID's and tests affected:
9	Sample ID's on COC match ID's on bottles?	✓			Sample ID's and containers affected:
10	Date & time on COC match date & time on bottles?	✓			Sample ID's affected:
11	Number of containers received match number indicated on COC?	✓			Sample ID's affected:
12	Are sample containers identifiable as GEL provided?			✓	
13	COC form is properly signed in relinquished/received sections?	✓			

Comments (Use Continuation Form if needed):

PM (or PMA) review: Initials BL Date 6.2.17 Page 1 of ___

GL-CHL-SR-001 Rev 5

Data Review Qualifier Definitions

GEL LABORATORIES LLC
2040 Savage Road Charleston, SC 29407 (843) 556−8171

Report Date: 13−JUN−17

Project Specific Qualifier Definitions for GEL Client Code: CPRC

Qualifier	Qualifier Definition	Department	Fraction
U	Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.		
J	The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate).Value is estimated	Organics	
P	Aroclor target analyte with greater than 25% difference between column analyses.	Organics	
C	Analyte has been confirmed by GC/MS analysis	Organics	Pesticide
B	The analyte was detected in both the associated QC blank and in the sample.	Organics	
E	Concentration exceeds the calibration range of the instrument	Organics	
A	The TIC is a suspected aldol−condensation product	Organics	Semi−Volatile
X	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier		
N	Spike Sample recovery is outside control limits.		
*	Duplicate analysis not within control limits	Inorganics	
>	Result greater than quantifiable range or greater than upper limit of the analysis range	General Chemistry	
Z	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier		
B	The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate).	Inorganics	Metals
D	Results are reported from a diluted aliquot of sample.		
E	Reported value is estimated due to interferences. See comment in narrative.	Inorganics	Metals
M	Duplicate precision not met.	Inorganics	Metals
o	Analyte failed to recover within LCS limits (0rganics only)	Organics	
S	Reported value determined by the Method of Standard Additions (MSA)	Inorganics	
T	Spike and/or spike duplicate sample recovery is outside control limits.	Organics	
W	Post−digestion spike recovery for GFAA out of control limit. Sample absorbency < 50% of spike absorbency.	Inorganics	
B	The associated QC sample blank has a result >= 2X the MDA and, after corrections, result is >= MDA for this sample	Radiological	
Y	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier		
+	Correlation coefficient for Method of Standard Additions (MSA) is < 0.995	Inorganics	
B	The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate).	General Chemistry	
C	Target analyte was detected in the sample and the associated blank. The associated blank concentration is >= EQL or is > 5% of the measured concentration and/or decision level for associated samples.	Inorganics	Metals
C	Target analyte was detected in the sample and the associated blank. The associated blank concentration is >= EQL or is > 5% of the measured concentration and/or decision level for associated samples.	General Chemistry	
<	Sample is below the EPA guidance level for Reactive Releasable Cyanide and/or Reactive Releasable Sulfide	General Chemistry	
UX	Gamma Spectroscopy−−Uncertain identification	Radiological	

Laboratory Certifications

List of current GEL Certifications as of 13 June 2017

State	Certification
Alaska	UST−0110
Arkansas	88−0651
CLIA	42D0904046
California	2940
Colorado	SC00012
Connecticut	PH−0169
Delaware	SC00012
DoD ELAP/ ISO17025 A2LA	2567.01
Florida NELAP	E87156
Foreign Soils Permit	P330−15−00283, P330−15−00253
Georgia	SC00012
Georgia SDWA	967
Hawaii	SC00012
Idaho Chemistry	SC00012
Idaho Radiochemistry	SC00012
Illinois NELAP	200029
Indiana	C−SC−01
Kansas NELAP	E−10332
Kentucky SDWA	90129
Kentucky Wastewater	90129
Louisiana NELAP	03046 (AI33904)
Louisiana SDWA	LA170010
Maryland	270
Massachusetts	M−SC012
Michigan	9976
Mississippi	SC00012
Nebraska	NE−OS−26−13
Nevada	SC000122017−1
New Hampshire NELAP	205415
New Jersey NELAP	SC002
New Mexico	SC00012
New York NELAP	11501
North Carolina	233
North Carolina SDWA	45709
North Dakota	R−158
Oklahoma	9904
Pennsylvania NELAP	68−00485
S.Carolina Radchem	10120002
South Carolina Chemistry	10120001
Tennessee	TN 02934
Texas NELAP	T104704235−17−12
Utah NELAP	SC000122017−22
Vermont	VT87156
Virginia NELAP	460202
Washington	C780
West Virginia	997404

General Chem Analysis

Case Narrative

General Chemistry
Technical Case Narrative
CH2MHill Plateau Remediation Company (CPRC)
SDG #: GEL424397
Work Order #: 424397

Product: **Ion Chromatography**
Analytical Method: 9056_ANIONS_IC
Analytical Procedure: GL-GC-E-086 REV# 25
Analytical Batch: 1670142

The following samples were analyzed using the above methods and analytical procedure(s).

GEL Sample ID#	Client Sample Identification
424397001	B39FD4
1203801745	Method Blank (MB)
1203801746	Laboratory Control Sample (LCS)
1203801747	424397001(B39FD4) Sample Duplicate (DUP)
1203801748	424397001(B39FD4) Post Spike (PS)

The samples in this SDG were analyzed on an "as received" basis.

Data Summary:

All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable, with the following exceptions.

Technical Information

Sample Dilutions
The following samples 1203801747 (B39FD4DUP), 1203801748 (B39FD4PS) and 424397001 (B39FD4) were diluted because target analyte concentrations exceeded the calibration range. Dilutions may be required for many reasons, including to minimize matrix interferences or to bring over range target analyte concentrations into the linear calibration range.

Analyte	424397001
Chloride	2X
Sulfate	2X

Certification Statement

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless otherwise noted in the analytical case narrative.

GEL LABORATORIES LLC
2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Qualifier Definition Report
for

CPRC001 CH2MHill Plateau Remediation Company

Client SDG: GEL424397 GEL Work Order: 424397

The Qualifiers in this report are defined as follows:

B The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate).

D Results are reported from a diluted aliquot of sample.

U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.

Review/Validation

GEL requires all analytical data to be verified by a qualified data reviewer. In addition, all CLP-like deliverables receive a third level review of the fractional data package.

The following data validator verified the information presented in this data report:

Signature:  **Name: Aubrey Kingsbury**

Date: 06 JUN 2017 **Title: Analyst I**

Sample Data Summary

GEL LABORATORIES LLC
2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis

Report Date: June 6, 2017

Company : CH2MHill Plateau Remediation Company
Address : MSIN R3-50 CHPRC
PO Box 1600
Richland, Washington 99352
Contact: Mr. Scot Fitzgerald
Project: CHPRC SAF I17-006

Client Sample ID: B39FD4
Sample ID: 424397001
Matrix: WATER
Collect Date: 31-MAY-17 09:45
Receive Date: 01-JUN-17
Collector: Client

Project: CPRC0I17006
Client ID: CPRC001

Parameter	Qualifier	Result	DL	RL	Units	PF	DF	Analyst	Date	Time	Batch	Method
Ion Chromatography												
9056_ANIONS_IC: COMMON "As Received"												
Fluoride	B	271	33.0	500	ug/L		1	MXL2	06/01/17	1147	1670142	1
Nitrate-N		3970	33.0	250	ug/L		1					
Nitrite-N	U	33.0	33.0	250	ug/L		1					
Chloride	D	9590	134	400	ug/L		2	MXL2	06/01/17	1343	1670142	2
Sulfate	D	35100	266	800	ug/L		2					

The following Analytical Methods were performed:

Method	Description	Analyst Comments
1	9056_ANIONS_IC	
2	9056_ANIONS_IC	

Notes:

Column headers are defined as follows:

DF: Dilution Factor
DL: Detection Limit
MDA: Minimum Detectable Activity
MDC: Minimum Detectable Concentration
Lc/LC: Critical Level
PF: Prep Factor
RL: Reporting Limit
SQL: Sample Quantitation Limit

Quality Control Summary

GEL LABORATORIES LLC
2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Report Date: June 6, 2017

CH2MHill Plateau Remediation Company
MSIN R3-50 CHPRC
PO Box 1600
Richland, Washington

Contact: **Mr. Scot Fitzgerald**

Workorder: **424397**

Parmname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date Time
Ion Chromatography										
Batch 1670142										
QC1203801747 424397001 DUP										
Chloride	D	9590	D	9600	ug/L	0.121		(0%-20%)	MXL2	06/01/17 14:12
Fluoride	B	271	B	275	ug/L	1.24 ^		(+/-500)		06/01/17 12:16
Nitrate-N		3970		3960	ug/L	0.114		(0%-20%)		
Nitrite-N	U	33.0	U	33.0	ug/L	N/A				
Sulfate	D	35100	D	35100	ug/L	0.0222		(0%-20%)		06/01/17 14:12
QC1203801746 LCS										
Chloride	5000			4510	ug/L		90.2	(80%-120%)		06/01/17 10:49
Fluoride	2500			2300	ug/L		92	(80%-120%)		
Nitrate-N	2500			2260	ug/L		90.3	(80%-120%)		
Nitrite-N	2500			2280	ug/L		91.3	(80%-120%)		
Sulfate	10000			9190	ug/L		91.9	(80%-120%)		
QC1203801745 MB										
Chloride			U	67.0	ug/L					06/01/17 10:20
Fluoride			U	33.0	ug/L					
Nitrate-N			U	33.0	ug/L					

GEL LABORATORIES LLC
2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: **424397**

Parmname	NOM		Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date Time
Ion Chromatography											
Batch 1670142											
Nitrite-N				U	33.0	ug/L				MXL2	06/01/17 10:20
Sulfate				U	133	ug/L					
QC1203801748 424397001 PS											
Chloride	5.00	D	4.80	D	10.0	mg/L		105	(75%-125%)		06/01/17 14:41
Fluoride	2.50	B	0.271		2.60	mg/L		93	(75%-125%)		06/01/17 12:45
Nitrate-N	2.50		3.97		6.54	mg/L		103	(75%-125%)		
Nitrite-N	2.50	U	0.00		2.34	mg/L		93.8	(75%-125%)		
Sulfate	10.0	D	17.5	D	28.4	mg/L		109	(75%-125%)		06/01/17 14:41

Notes:

The Qualifiers in this report are defined as follows:

- < Sample is below the EPA guidance level for Reactive Releasable Cyanide and/or Reactive Releasable Sulfide
- > Result greater than quantifiable range or greater than upper limit of the analysis range
- B The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate).
- C Target analyte was detected in the sample and the associated blank. The associated blank concentration is >= EQL or is > 5% of the measured concentration and/or decision level for associated samples.
- D Results are reported from a diluted aliquot of sample.
- N Spike Sample recovery is outside control limits.
- U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
- X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
- Y Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
- Z Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier

GEL LABORATORIES LLC
2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: **424397**

Parmname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date	Time

N/A indicates that spike recovery limits do not apply when sample concentration exceeds spike conc. by a factor of 4 or more or %RPD not applicable.
^ The Relative Percent Difference (RPD) obtained from the sample duplicate (DUP) is evaluated against the acceptance criteria when the sample is greater than five times (5X) the contract required detection limit (RL). In cases where either the sample or duplicate value is less than 5X the RL, a control limit of +/- the RL is used to evaluate the DUP result.
* Indicates that a Quality Control parameter was not within specifications.
For PS, PSD, and SDILT results, the values listed are the measured amounts, not final concentrations.

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless qualified on the QC Summary.

Radiological Analysis

Case Narrative

Radiochemistry
Technical Case Narrative
CH2MHill Plateau Remediation Company (CPRC)
SDG #: GEL424397
Work Order #: 424397

Product: **C14_LSC: COMMON**
Analytical Method: C14_LSC
Analytical Procedure: GL-RAD-A-003 REV# 15
Analytical Batch: 1670749

The following samples were analyzed using the above methods and analytical procedure(s).

GEL Sample ID#	Client Sample Identification
424397002	B39FB1
424397003	B39FF3
424397004	B39FF7
1203803225	Method Blank (MB)
1203803226	424396002(NonSDG) Sample Duplicate (DUP)
1203803227	424396002(NonSDG) Matrix Spike (MS)
1203803228	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

Data Summary:

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

Certification Statement

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless otherwise noted in the analytical case narrative.

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Qualifier Definition Report
for

CPRC001 CH2MHill Plateau Remediation Company

Client SDG: GEL424397 GEL Work Order: 424397

The Qualifiers in this report are defined as follows:

U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.

Review/Validation

GEL requires all analytical data to be verified by a qualified data reviewer. In addition, all CLP-like deliverables receive a third level review of the fractional data package.

The following data validator verified the information presented in this data report:

Signature:  **Name: Theresa Austin**

Date: 13 JUN 2017 **Title: Group Leader**

Sample Data Summary

Rad

Certificate of Analysis
Sample Summary

SDG Number:	**GEL424397**	**Client:**	**CPRC001**	**Project:**	**CPRC0I17006**
Lab Sample ID:	**424397002**	**Date Collected:**	**05/31/2017 08:23**	**Matrix:**	**WATER**
		Date Received:	**06/01/2017 09:10**		
Client ID:	**B39FB1**			**Prep Basis:**	**"As Received"**
Batch ID:	**1670749**	**Method:**	**C14_LSC**	**SOP Ref:**	**GL-RAD-A-003**
Run Date:	**06/09/2017 21:04**	**Analyst:**	**BXM4**	**Instrument:**	**LSCGREEN**
Data File:	**C1670749.xls**	**Aliquot:**	**60.01 mL**	**Count Time:**	**45 min**
Prep Batch:	**1670749**	**Prep Method:**	**EPA EERF C-01 Modified**		
Prep Date:	**06/09/2017 09:17**				

CAS No.	Parmname	Qual	Result	Units	Uncert	TPU	MDC	RDL
14762-75-5	Carbon-14	U	10.9	pCi/L	+/-18.5	18.6	31.4	50.0

Surrogate/Tracer recovery	Result	Nominal	Units	Recovery%	Acceptable Limits

Comments:

U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
TPU and Counting Uncertainty are calculated at the 95% confidence level (1.96-sigma).
The MDC is a sample specific MDC.

Rad

Certificate of Analysis
Sample Summary

SDG Number:	**GEL424397**	**Client:**	**CPRC001**	**Project:**	**CPRC0I17006**
Lab Sample ID:	**424397003**	**Date Collected:**	**05/31/2017 11:30**	**Matrix:**	**WATER**
		Date Received:	**06/01/2017 09:10**		
Client ID:	**B39FF3**			**Prep Basis:**	**"As Received"**
Batch ID:	**1670749**	**Method:**	**C14_LSC**	**SOP Ref:**	**GL-RAD-A-003**
Run Date:	**06/09/2017 21:51**	**Analyst:**	**BXM4**	**Instrument:**	**LSCGREEN**
Data File:	**C1670749.xls**	**Aliquot:**	**60.08 mL**	**Count Time:**	**45 min**
Prep Batch:	**1670749**	**Prep Method:**	**EPA EERF C-01 Modified**		
Prep Date:	**06/09/2017 09:17**				

CAS No.	Parmname	Qual	Result	Units	Uncert	TPU	MDC	RDL
14762-75-5	Carbon-14		498	pCi/L	+/-28.2	96.6	31.3	50.0

Surrogate/Tracer recovery	Result	Nominal	Units	Recovery%	Acceptable Limits

Comments:

TPU and Counting Uncertainty are calculated at the 95% confidence level (1.96-sigma).

The MDC is a sample specific MDC.

Rad

Certificate of Analysis
Sample Summary

SDG Number:	**GEL424397**	**Client:**	**CPRC001**	**Project:**	**CPRC0I17006**
Lab Sample ID:	**424397004**	**Date Collected:**	**05/31/2017 11:10**	**Matrix:**	**WATER**
		Date Received:	**06/01/2017 09:10**		
Client ID:	**B39FF7**			**Prep Basis:**	**"As Received"**
Batch ID:	**1670749**	**Method:**	**C14_LSC**	**SOP Ref:**	**GL-RAD-A-003**
Run Date:	**06/09/2017 22:38**	**Analyst:**	**BXM4**	**Instrument:**	**LSCGREEN**
Data File:	**C1670749.xls**	**Aliquot:**	**60.06 mL**	**Count Time:**	**45 min**
Prep Batch:	**1670749**	**Prep Method:**	**EPA EERF C-01 Modified**		
Prep Date:	**06/09/2017 09:17**				

CAS No.	Parmname	Qual	Result	Units	Uncert	TPU	MDC	RDL
14762-75-5	Carbon-14		232	pCi/L	+/-23.4	49.0	31.3	50.0

Surrogate/Tracer recovery	Result	Nominal	Units	Recovery%	Acceptable Limits

Comments:

TPU and Counting Uncertainty are calculated at the 95% confidence level (1.96-sigma).

The MDC is a sample specific MDC.

Quality Control Summary

GEL LABORATORIES LLC
2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Client : **CH2MHill Plateau Remediation Company**
MSIN R3-50 CHPRC
PO Box 1600
Richland, Washington 99352
Contact: **Mr. Scot Fitzgerald**

Workorder: **424397**

Report Date: June 13, 2017

Parmname	NOM	Sample	Qual	QC	Units	QC Criteria	Range	Analyst	Date Time
Rad Liquid Scintillation									
Batch 1670749									
QC1203803225 MB									
Carbon-14			U	-5.32	pCi/L			BXM4	06/10/1709:05
Uncert:				+/-18.3					
TPU:				+/-18.3					
QC1203803226 424396002 DUP									
Carbon-14		202		193	pCi/L				06/10/1709:52
Uncert:		+/-22.9		+/-22.7		RPD: 5	(0% - 20%)		
TPU:		+/-44.0		+/-42.3		RER: 0.317	(0-2)		
QC1203803227 424396002 MS									
Carbon-14	1250	202		1380	pCi/L	REC: 94	(75%-125%)		06/10/1710:39
Uncert:		+/-22.9		+/-40.2					
TPU:		+/-44.0		+/-259					
QC1203803228 LCS									
Carbon-14	1250			1170	pCi/L	REC: 94	(80%-120%)		06/10/1711:26
Uncert:				+/-37.7					
TPU:				+/-220					

Notes:

TPU and Counting Uncertainty are calculated at the 95% confidence level (1.96-sigma).

The Qualifiers in this report are defined as follows:

<	Sample is below the EPA guidance level for Reactive Releasable Cyanide and/or Reactive Releasable Sulfide
>	Result greater than quantifiable range or greater than upper limit of the analysis range
B	The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate).
B	The associated QC sample blank has a result >= 2X the MDA and, after corrections, result is >= MDA for this sample
C	Target analyte was detected in the sample and the associated blank. The associated blank concentration is >= EQL or is > 5% of the measured concentration and/or decision level for associated samples.
D	Results are reported from a diluted aliquot of sample.
N	Spike Sample recovery is outside control limits.
U	Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
UX	Gamma Spectroscopy--Uncertain identification
X	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
Y	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
Z	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier

GEL LABORATORIES LLC
2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: **424397**

Parmname	NOM	Sample	Qual	QC	Units	QC Criteria	Range	Analyst	Date	Time

N/A indicates that spike recovery limits do not apply when sample concentration exceeds spike conc. by a factor of 4 or more or %RPD not applicable.
** Indicates analyte is a surrogate compound.
^ The Relative Percent Difference (RPD) obtained from the sample duplicate (DUP) is evaluated against the acceptence criteria when the sample is greater than five times (5X) the contract required detection limit (RL). In cases where either the sample or duplicate value is less than 5X the RL, a control limit of +/- the RL is used to evaluate the DUP result.
For PS, PSD, and SDILT results, the values listed are the measured amounts, not final concentrations.

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless qualified on the QC Summary.