

July 10, 2017



PO Box 30712 Charleston, SC 29417
2040 Savage Road Charleston, SC 29407
P 843.556.8171
F 843.766.1178

gel.com

July 10, 2017

Mr. Scot Fitzgerald
CH2MHill Plateau Remediation Company
MSIN R3-50 CHPRC
PO Box 1600
Richland, Washington 99352

Re: CHPRC SAF F17-014
Work Order: 425727
SDG: GEL425727

Dear Mr. Fitzgerald:

GEL Laboratories, LLC (GEL) appreciates the opportunity to provide the enclosed analytical results for the sample(s) we received on June 17, 2017. This original data report has been prepared and reviewed in accordance with GEL's standard operating procedures.

Our policy is to provide high quality, personalized analytical services to enable you to meet your analytical needs on time every time. We trust that you will find everything in order and to your satisfaction. If you have any questions, please do not hesitate to call me at (843) 556-8171, ext. 4505.

Sincerely,

B Luthman
Brielle Luthman for
Heather Shaffer
Project Manager

Purchase Order: 7H
Chain of Custody: F17-014-102
Enclosures



Table of Contents

Case Narrative.....1

Chain of Custody and Supporting Documentation.....5

Data Review Qualifier Definitions.....8

Laboratory Certifications.....10

Radiological Analysis.....12

 Case Narrative.....13

 Sample Data Summary.....17

 Quality Control Summary.....21

Case Narrative

July 10, 2017

General Narrative
for
CH2MHill Plateau Remediation Company
CHPRC SAF F17-014
SDG: GEL425727

July 10, 2017

Laboratory Identification:

GEL Laboratories LLC
2040 Savage Road
Charleston, South Carolina 29407
(843) 556-8171

Summary

Sample receipt

The sample(s) arrived at GEL Laboratories, LLC, Charleston, South Carolina on June 17, 2017, for analysis. The sample was delivered with proper chain of custody documentation and signatures. All sample containers arrived without any visible signs of tampering or breakage. There are no additional comments concerning sample receipt.

Items of Note All efforts were made by the lab to meet any short hold times. Samples that were analyzed outside of the initial hold time but still within 2X hold time will be noted in the lab case narrative.

Sample Identification

The laboratory received the following sample:

Laboratory Identification	Sample Description
425727001	B39901

Case Narrative

Sample analyses were conducted using methodology as outlined in GEL Laboratories, LLC (GEL) Standard Operating Procedures. Any technical or administrative problems during analysis, data review, and reduction are contained in the analytical case narratives in the enclosed data package.

Data Package

The enclosed data package contains the following sections: General Narrative, Chain of Custody and Supporting Documentation, and data from the following fractions: Radiochemistry.

We certify that this package is in compliance with the SOW, both technically and for completeness, including a full description of, explanation of, and corrective actions for, any and all deviations, from either the analyses requested or the case narrative requested. Release of the data contained in this hard copy data package has been authorized by the Laboratory Analytical Manager (or designee) and the laboratory's client services representative as verified by their signatures on this report.

July 10, 2017

B. Luthman
Brielle Luthman for
Heather Shaffer
Project Manager

July 10, 2017

Radiochemistry

Technical Case Narrative

CH2MHill Plateau Remediation Company (CPRC)

SDG #: GEL425727

Work Order #: 425727

I129_SEP_LEPS_GS: COMMON

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

906.0_TRITIUM_LSC:COMMON

All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable, with the following exceptions.

Technical Information

Recounts

Sample 1203815893 (Non SDG 425280008MS) was recounted due to low recovery. The recount is reported.

Miscellaneous Information

Additional Comments

The matrix spike, 1203815893 (Non SDG 425280008MS), aliquot was reduced to conserve sample volume.

TC99_EIE_LSC: COMMON

All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable, with the following exceptions.

Technical Information

Recounts

Sample 1203815897 (LCS) was recounted due to low recovery. The recount is reported.

Certification Statement

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless otherwise noted in the analytical case narrative.

Chain of Custody and Supporting Documentation

July 10, 2017

CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST

CH2MHill Plateau Remediation Company

COLLECTOR: DAVE WIGHT
CHPRC

COMPANY CONTACT: LYNCH, SA
TELEPHONE NO.: 373-5586

PROJECT COORDINATOR: LYNCH, SA

SAMPLING LOCATION: C9563, I-012

PROJECT DESIGNATION: FY2017 200-ZP-1 Remedial Action Wells Sampling and Analysis - Water

PRICE CODE: 7H

AIR QUALITY:

DATA TURNAROUND: 30 Days / 30 Days

ICE CHEST NO.: *SWS 2-6-17 620*

FIELD LOGBOOK NO.: *HNF-N-645 F-61*

ACTUAL SAMPLE DEPTH: *506.63*

COA: 302938

METHOD OF SHIPMENT: FEDERAL EXPRESS

SAF NO.: FI7-014

ORIGINAL

SHIPPED TO: GEL Laboratories, LLC

OFFSITE PROPERTY NO.: *8057*

BILL OF LADING/AIR BILL NO.: *779422470484*

F17-014-102

PAGE 1 OF 1

425727

MATRIX*	POSSIBLE SAMPLE HAZARDS/ REMARKS	None	HNO3 to pH <2	None
A=Air DL=Drum Liquids DS=Drum Solids L=Liquid O=Oil S=Soil SE=Sediment T=Tissue V=Vegetation W=Water WT=Wipe X=Other	*Contains Radioactive Material at concentrations that are not regulated for transportation per 49 CFR/IATA Dangerous Goods Regulations but are not releasable per DOE Order 458.1.	6 Months	6 Months	6 Months
		P	G/P	G/P
		1	1	2
		250mL	500mL	1L
		906.0 TRITIUM LSC; COMMON;	TCS9 EIE LSC; COMMON;	IL29 SEP LEPS CS; COMMON;
SPECIAL HANDLING AND/OR STORAGE	NA	SAMPLE DATE	SAMPLE TIME	
B39901	WATER	<i>6/15/17</i>	<i>1340</i>	<input checked="" type="checkbox"/>

CHAIN OF POSSESSION

SIGN/ PRINT NAMES

SPECIAL INSTRUCTIONS
TRVL-17-090 *COLLECT A WATER SAMPLE 10 FEET BELOW RINGOLD LOWER MUD UNIT

RELINQUISHED BY/REMOVED FROM	DATE/TIME	RECEIVED BY/STORED IN	DATE/TIME
<i>CHPRC</i>	<i>6/15/17</i>	SSU-1	<i>6/15/17</i>
	<i>JUN 16 2017</i>	RECEIVED BY/STORED IN	<i>JUN 16 2017 07:55</i>
SSU-1	<i>JUN 16 2017</i>	Curt Hoffman CHPRC	<i>JUN 16 2017 07:55</i>
RECEIVED BY/REMOVED FROM	<i>JUN 16 2017 14:00</i>	RECEIVED BY/STORED IN	<i>JUN 16 2017 14:00</i>
CHPRC	<i>JUN 16 2017 14:00</i>	RECEIVED BY/STORED IN	<i>JUN 16 2017 14:00</i>
	<i>6/17/17</i>	RECEIVED BY/STORED IN	<i>6/17/17 09:10</i>
RECEIVED BY/REMOVED FROM	<i>6/17/17</i>	RECEIVED BY/STORED IN	<i>6/17/17 09:10</i>
		RECEIVED BY/STORED IN	
		RECEIVED BY/STORED IN	
		RECEIVED BY/STORED IN	

LABORATORY SECTION: RECEIVED BY

FINAL SAMPLE DISPOSITION: DISPOSAL METHOD

PRINTED ON 3/9/2017

FRS ID = FSR42561

TRVL NUM = TRVL-17-090

A-6003-618 (REV 2)

8 of 25

July 10, 2017

SAMPLE RECEIPT & REVIEW FORM

425727

Client: <u>CPRC</u>		SDG/AR/COC/Work Order: <u>425725/425726 HS</u>		
Received By: <u>[Signature]</u>		Date Received: <u>6/17/17</u>		
Carrier and Tracking Number		Circle Applicable: <input checked="" type="checkbox"/> FedEx Express <input type="checkbox"/> FedEx Ground <input type="checkbox"/> UPS <input type="checkbox"/> Field Services <input type="checkbox"/> Courier <input type="checkbox"/> Other <u>779422470484-3</u> <u>779424651557-4</u> <u>779424930270-4</u>		
Suspected Hazard Information		*If Net Counts > 100cpm on samples not marked "radioactive", contact the Radiation Safety Group for further investigation.		
Shipped as a DOT Hazardous?		Hazard Class Shipped: _____ UN#: _____		
COC/Samples marked or classified as radioactive?		Maximum Net Counts Observed* (Observed Counts - Area Background Counts): <u>0</u> CPM/mR/Hr Classified as: Rad 1 Rad 2 Rad 3		
Is package, COC, and/or Samples marked HAZ?		If yes, select Hazards below, and contact the GEL Safety Group. PCB's Flammable Foreign Soil RCRA Asbestos Beryllium Other:		
Sample Receipt Criteria	Yes	NA	No	Comments/Qualifiers (Required for Non-Conforming Items)
1 Shipping containers received intact and sealed?	<input checked="" type="checkbox"/>			Circle Applicable: Seals broken Damaged container Leaking container Other (describe)
2 Chain of custody documents included with shipment?	<input checked="" type="checkbox"/>			
3 Samples requiring cold preservation within (0 ≤ 6 deg. C)?*	<input checked="" type="checkbox"/>			Preservation Method: <u>Wet Ice</u> Ice Packs Dry Ice None Other *all temperatures are recorded in Celsius TEMP: <u>3, 4</u>
4 Daily check performed and passed on IR temperature gun?	<input checked="" type="checkbox"/>			Temperature Device Serial #: <u>103-16</u> Secondary Temperature Device Serial # (if Applicable): _____
5 Sample containers intact and sealed?	<input checked="" type="checkbox"/>			Circle Applicable: Seals broken <u>Damaged container</u> Leaking container Other (describe) <u>2x1L of B 3B3H1 + 440ml vial of B39UNO</u>
6 Samples requiring chemical preservation at proper pH?	<input checked="" type="checkbox"/>			Sample ID's and Containers Affected: If Preservation added, Lot#: _____
7 Do any samples require Volatile Analysis?	<input checked="" type="checkbox"/>			If Yes, Are Encores or Soil Kits present? Yes _____ No <input checked="" type="checkbox"/> (If yes, take to VOA Freezer) Do VOA vials contain acid preservation? Yes <input checked="" type="checkbox"/> No _____ N/A (If unknown, select No) VOA vials free of headspace? Yes <input checked="" type="checkbox"/> No _____ N/A Sample ID's and containers affected:
8 Samples received within holding time?	<input checked="" type="checkbox"/>			ID's and tests affected:
9 Sample ID's on COC match ID's on bottles?	<input checked="" type="checkbox"/>			Sample ID's and containers affected:
10 Date & time on COC match date & time on bottles?	<input checked="" type="checkbox"/>			Sample ID's affected:
11 Number of containers received match number indicated on COC?	<input checked="" type="checkbox"/>			Sample ID's affected:
12 Are sample containers identifiable as GEL provided?	<input checked="" type="checkbox"/>			
13 COC form is properly signed in relinquished/received sections?	<input checked="" type="checkbox"/>			
Comments (Use Continuation Form if needed):				

PM (or PMA) review: Initials [Signature] Date 6/19/17 Page 1 of 1

Data Review Qualifier Definitions

Project Specific Qualifier Definitions for GEL Client Code: CPRC

Qualifier	Qualifier Definition	Department	Fraction
U	Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.		
J	The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate). Value is estimated	Organics	
P	Aroclor target analyte with greater than 25% difference between column analyses.	Organics	
C	Analyte has been confirmed by GC/MS analysis	Organics	Pesticide
B	The analyte was detected in both the associated QC blank and in the sample.	Organics	
E	Concentration exceeds the calibration range of the instrument	Organics	
A	The TIC is a suspected aldol-condensation product	Organics	Semi-Volatile
X	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier		
N	Spike Sample recovery is outside control limits.		
*	Duplicate analysis not within control limits	Inorganics	
>	Result greater than quantifiable range or greater than upper limit of the analysis range	General Chemistry	
Z	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier		
B	The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate).	Inorganics	Metals
D	Results are reported from a diluted aliquot of sample.		
E	Reported value is estimated due to interferences. See comment in narrative.	Inorganics	Metals
M	Duplicate precision not met.	Inorganics	Metals
o	Analyte failed to recover within LCS limits (Organics only)	Organics	
S	Reported value determined by the Method of Standard Additions (MSA)	Inorganics	
T	Spike and/or spike duplicate sample recovery is outside control limits.	Organics	
W	Post-digestion spike recovery for GFAA out of control limit. Sample absorbency < 50% of spike absorbency.	Inorganics	
B	The associated QC sample blank has a result $\geq 2X$ the MDA and, after corrections, result is \geq MDA for this sample	Radiological	
Y	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier		
+	Correlation coefficient for Method of Standard Additions (MSA) is < 0.995	Inorganics	
B	The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate).	General Chemistry	
C	Target analyte was detected in the sample and the associated blank. The associated blank concentration is \geq EQL or is > 5% of the measured concentration and/or decision level for associated samples.	Inorganics	Metals
C	Target analyte was detected in the sample and the associated blank. The associated blank concentration is \geq EQL or is > 5% of the measured concentration and/or decision level for associated samples.	General Chemistry	
<	Sample is below the EPA guidance level for Reactive Releasable Cyanide and/or Reactive Releasable Sulfide	General Chemistry	
UX	Gamma Spectroscopy--Uncertain identification	Radiological	

Laboratory Certifications

List of current GEL Certifications as of 10 July 2017

State	Certification
Alaska	UST-0110
Arkansas	88-0651
CLIA	42D0904046
California	2940
Colorado	SC00012
Connecticut	PH-0169
Delaware	SC00012
DoD ELAP/ ISO17025 A2LA	2567.01
Florida NELAP	E87156
Foreign Soils Permit	P330-15-00283, P330-15-00253
Georgia	SC00012
Georgia SDWA	967
Hawaii	SC00012
Idaho Chemistry	SC00012
Idaho Radiochemistry	SC00012
Illinois NELAP	200029
Indiana	C-SC-01
Kansas NELAP	E-10332
Kentucky SDWA	90129
Kentucky Wastewater	90129
Louisiana NELAP	03046 (AI33904)
Louisiana SDWA	LA170010
Maryland	270
Massachusetts	M-SC012
Michigan	9976
Mississippi	SC00012
Nebraska	NE-OS-26-13
Nevada	SC000122017-1
New Hampshire NELAP	205415
New Jersey NELAP	SC002
New Mexico	SC00012
New York NELAP	11501
North Carolina	233
North Carolina SDWA	45709
North Dakota	R-158
Oklahoma	9904
Pennsylvania NELAP	68-00485
S.Carolina Radchem	10120002
South Carolina Chemistry	10120001
Tennessee	TN 02934
Texas NELAP	T104704235-17-12
Utah NELAP	SC000122017-22
Vermont	VT87156
Virginia NELAP	460202
Washington	C780
West Virginia	997404

Radiological Analysis

Case Narrative

July 10, 2017

Radiochemistry

Technical Case Narrative

CH2MHill Plateau Remediation Company (CPRC)

SDG #: GEL425727

Work Order #: 425727

Product: I129_SEP_LEPS_GS: COMMON

Analytical Method: DOE EML HASL-300,I-01 Modified

Analytical Procedure: GL-RAD-A-006 REV# 21

Analytical Batch: 1674807

The following samples were analyzed using the above methods and analytical procedure(s).

<u>GEL Sample ID#</u>	<u>Client Sample Identification</u>
425727001	B39901
1203813188	Method Blank (MB)
1203813189	425282001(NonSDG) Sample Duplicate (DUP)
1203813190	425282001(NonSDG) Matrix Spike (MS)
1203813191	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

Data Summary:

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

Product: 906.0_TRITIUM_LSC:COMMON

Analytical Method: TRITIUM_DIST_LSC

Analytical Procedure: GL-RAD-A-002 REV# 22

Analytical Batch: 1675960

The following samples were analyzed using the above methods and analytical procedure(s).

<u>GEL Sample ID#</u>	<u>Client Sample Identification</u>
425727001	B39901
1203815891	Method Blank (MB)
1203815892	425280008(NonSDG) Sample Duplicate (DUP)
1203815893	425280008(NonSDG) Matrix Spike (MS)
1203815894	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

Data Summary:

All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable, with the following exceptions.

Technical Information

Recounts

Sample 1203815893 (Non SDG 425280008MS) was recounted due to low recovery. The recount is reported.

Miscellaneous Information

Additional Comments

The matrix spike, 1203815893 (Non SDG 425280008MS), aliquot was reduced to conserve sample volume.

Product: TC99_EIE_LSC: COMMON

Analytical Method: TC99_EIE_LSC

Analytical Procedure: GL-RAD-A-059 REV# 5

Analytical Batch: 1675961

The following samples were analyzed using the above methods and analytical procedure(s).

<u>GEL Sample ID#</u>	<u>Client Sample Identification</u>
425727001	B39901
1203815895	Method Blank (MB)
1203815896	425384006(NonSDG) Sample Duplicate (DUP)
1203815897	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

Data Summary:

All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable, with the following exceptions.

Technical Information

Recounts

Sample 1203815897 (LCS) was recounted due to low recovery. The recount is reported.

Certification Statement

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless otherwise noted in the analytical case narrative.

July 10, 2017

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

**Qualifier Definition Report
for**

CPRC001 CH2MHill Plateau Remediation Company

Client SDG: GEL425727 GEL Work Order: 425727

The Qualifiers in this report are defined as follows:

U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.

Review/Validation

GEL requires all analytical data to be verified by a qualified data reviewer. In addition, all CLP-like deliverables receive a third level review of the fractional data package.

The following data validator verified the information presented in this data report:

Signature:



Name: Theresa Austin

Date: 10 JUL 2017

Title: Group Leader

Sample Data Summary

July 10, 2017

**Certificate of Analysis
Sample Summary**

SDG Number: GEL425727	Client: CPRC001	Project: CPRC0F17014
Lab Sample ID: 425727001	Date Collected: 06/15/2017 13:40	Matrix: WATER
	Date Received: 06/17/2017 09:10	
Client ID: B39901	Method: DOE EML HASL-300,I-01 Mo	Prep Basis: "As Received"
Batch ID: 1674807	Analyst: MJH1	SOP Ref: GL-RAD-A-006
Run Date: 06/22/2017 14:46	Aliquot: 0.4 L	Instrument: XRAY5
Data File: I425727001.CNF;1	Prep Method: DOE EML HASL-300,I-01 M	Count Time: 90 min
Prep Batch: 1674807		
Prep Date: 06/21/2017 00:00		

CAS No.	Parmname	Qual	Result	Units	Uncert	TPU	MDC	RDL
15046-84-1	Iodine-129	U	-1.17	pCi/L	+/-1.25	1.36	2.02	5.00

Surrogate/Tracer recovery	Result	Nominal	Units	Recovery%	Acceptable Limits
---------------------------	--------	---------	-------	-----------	-------------------

Comments:

U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
 TPU and Counting Uncertainty are calculated at the 95% confidence level (1.96-sigma).
 The MDC is a sample specific MDC.

July 10, 2017

**Certificate of Analysis
Sample Summary**

SDG Number: GEL425727	Client: CPRC001	Project: CPRC0F17014
Lab Sample ID: 425727001	Date Collected: 06/15/2017 13:40	Matrix: WATER
	Date Received: 06/17/2017 09:10	
Client ID: B39901	Method: TRITIUM_DIST_LSC	Prep Basis: "As Received"
Batch ID: 1675960	Analyst: BXM4	SOP Ref: GL-RAD-A-002
Run Date: 07/06/2017 08:12	Aliquot: 50 mL	Instrument: LSCGREEN
Data File: T1675960R.xls	Prep Method: EPA 906.0 Modified	Count Time: 45 min
Prep Batch: 1675960		
Prep Date: 07/05/2017 09:31		

CAS No.	Parmname	Qual	Result	Units	Uncert	TPU	MDC	RDL
10028-17-8	Tritium	U	10.3	pCi/L	+/-202	202	354	400

Surrogate/Tracer recovery	Result	Nominal	Units	Recovery%	Acceptable Limits
---------------------------	--------	---------	-------	-----------	-------------------

Comments:

U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
 TPU and Counting Uncertainty are calculated at the 95% confidence level (1.96-sigma).
 The MDC is a sample specific MDC.

July 10, 2017

**Certificate of Analysis
Sample Summary**

SDG Number: GEL425727	Client: CPRC001	Project: CPRC0F17014
Lab Sample ID: 425727001	Date Collected: 06/15/2017 13:40	Matrix: WATER
	Date Received: 06/17/2017 09:10	
Client ID: B39901	Method: TC99_EIE_LSC	Prep Basis: "As Received"
Batch ID: 1675961	Analyst: CXS7	SOP Ref: GL-RAD-A-059
Run Date: 07/04/2017 15:12	Aliquot: 100 mL	Instrument: LSCGREEN
Data File: E1675961R.xls	Prep Method: DOE EML HASL-300, Tc-02-	Count Time: 20 min
Prep Batch: 1675961		
Prep Date: 06/29/2017 13:52		

CAS No.	Parmname	Qual	Result	Units	Uncert	TPU	MDC	RDL
14133-76-7	Technetium-99	U	-6.71	pCi/L	+/-22.6	22.6	39.6	50.0

Surrogate/Tracer recovery	Result	Nominal	Units	Recovery%	Acceptable Limits
Technetium-99m Tracer	29800	33500	CPM	88.9	(30%-105%)

Comments:

U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
 TPU and Counting Uncertainty are calculated at the 95% confidence level (1.96-sigma).
 The MDC is a sample specific MDC.

Quality Control Summary

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Report Date: July 10, 2017
Page 1 of 2

Client : CH2MHill Plateau Remediation Company
MSIN R3-50 CHPRC
PO Box 1600
Richland, Washington 99352
Contact: Mr. Scot Fitzgerald
Workorder: 425727

Parmname	NOM	Sample	Qual	QC	Units	QC Criteria	Range	Analyst	Date Time
Rad Gamma Spec									
Batch	1674807								
QC1203813188	MB								
Iodine-129			U	-0.125	pCi/L			MJH1	06/22/1714:58
				Uncert: +/-0.259					
				TPU: +/-0.265					
QC1203813189	425282001	DUP							
Iodine-129		2.31		2.38	pCi/L				06/22/1715:13
				Uncert: +/-1.14		RPD: 3 (0% - 100%)			
				TPU: +/-1.16		RER: 0.0916 (0-2)			
QC1203813190	425282001	MS							
Iodine-129		32.0	2.31	30.9	pCi/L	REC: 89 (75%-125%)			06/22/1716:03
				Uncert: +/-1.14					
				TPU: +/-1.16					
QC1203813191	LCS								
Iodine-129		26.0		28.9	pCi/L	REC: 111 (80%-120%)			06/22/1716:42
				Uncert: +/-3.92					
				TPU: +/-4.97					
Rad Liquid Scintillation									
Batch	1675960								
QC1203815891	MB								
Tritium			U	66.7	pCi/L			BXM4	07/06/1708:59
				Uncert: +/-205					
				TPU: +/-206					
QC1203815892	425280008	DUP							
Tritium		9520		8980	pCi/L				07/06/1709:46
				Uncert: +/-495		RPD: 6 (0% - 20%)			
				TPU: +/-1910		RER: 0.403 (0-2)			
QC1203815893	425280008	MS							
Tritium		4480	9520	13900	pCi/L	REC: 97 (75%-125%)			07/07/1706:22
				Uncert: +/-495					
				TPU: +/-1910					
QC1203815894	LCS								
Tritium		2230		2110	pCi/L	REC: 94 (80%-120%)			07/06/1710:48
				Uncert: +/-470					
				TPU: +/-622					
Batch	1675961								
QC1203815895	MB								
Technetium-99			U	-1.46	pCi/L			CXS7	07/04/1715:54
				Uncert: +/-23.2					
				TPU: +/-23.2					
**Technetium-99m Tracer		33500		29300	CPM	REC: 88 (30%-105%)			
QC1203815896	425384006	DUP							
Technetium-99		U	17.5	U	35.4	pCi/L			07/04/1716:16
				Uncert: +/-20.8		RPD: 0 N/A			
				TPU: +/-20.9		RER: 1.07 (0-2)			
						REC:			

QC Summary

Workorder: 425727

Page 2 of 2

Parmname	NOM	Sample	Qual	QC	Units	QC Criteria	Range	Analyst	Date Time
Rad Liquid Scintillation									
Batch	1675961								
**Technetium-99m Tracer	33500	33400		28900	CPM		86 (30%-105%)		
QC1203815897 LCS									
Technetium-99	861			713	pCi/L	REC:	83 (80%-120%)		07/05/1707:49
				Uncert:					
				TPU:					
**Technetium-99m Tracer	33500			26600	CPM	REC:	80 (30%-105%)		

Notes:

TPU and Counting Uncertainty are calculated at the 95% confidence level (1.96-sigma).

The Qualifiers in this report are defined as follows:

- B The associated QC sample blank has a result $\geq 2X$ the MDA and, after corrections, result is \geq MDA for this sample
- D Results are reported from a diluted aliquot of sample.
- N Spike Sample recovery is outside control limits.
- U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
- UX Gamma Spectroscopy--Uncertain identification
- X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
- Y Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
- Z Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier

N/A indicates that spike recovery limits do not apply when sample concentration exceeds spike conc. by a factor of 4 or more or %RPD not applicable.

** Indicates analyte is a surrogate compound.

^ The Relative Percent Difference (RPD) obtained from the sample duplicate (DUP) is evaluated against the acceptance criteria when the sample is greater than five times (5X) the contract required detection limit (RL). In cases where either the sample or duplicate value is less than 5X the RL, a control limit of +/- the RL is used to evaluate the DUP result.

For PS, PSD, and SDILT results, the values listed are the measured amounts, not final concentrations.

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless qualified on the QC Summary.