



December 26, 2017

Mr. Scot Fitzgerald
CH2MHill Plateau Remediation Company
MSIN R3-50 CHPRC
PO Box 1600
Richland, Washington 99352

Re: CHPRC SAF W18-012
Work Order: 440130
SDG: GEL440130

Dear Mr. Fitzgerald:

GEL Laboratories, LLC (GEL) appreciates the opportunity to provide the enclosed analytical results for the sample(s) we received on December 15, 2017. This original data report has been prepared and reviewed in accordance with GEL's standard operating procedures.

Our policy is to provide high quality, personalized analytical services to enable you to meet your analytical needs on time every time. We trust that you will find everything in order and to your satisfaction. If you have any questions, please do not hesitate to call me at (843) 556-8171, ext. 4505.

Sincerely,


Kaitlyn Stone for
Heather Shaffer
Project Manager

Purchase Order: 300071 - 7H
Chain of Custody: W18-012-080
Enclosures




Table of Contents

Case Narrative.....	1
Chain of Custody and Supporting Documentation.....	5
Data Review Qualifier Definitions.....	9
Laboratory Certifications.....	11
General Chem Analysis.....	13
Case Narrative.....	14
Sample Data Summary.....	17
Quality Control Summary.....	19

Case Narrative

**General Narrative
for
CH2MHill Plateau Remediation Company
CHPRC SAF W18-012
SDG: GEL440130**

December 26, 2017

Laboratory Identification:

GEL Laboratories LLC
2040 Savage Road
Charleston, South Carolina 29407
(843) 556-8171

Summary

Sample receipt

The sample(s) arrived at GEL Laboratories, LLC, Charleston, South Carolina on December 15, 2017, for analysis. The sample was delivered with proper chain of custody documentation and signatures. All sample containers arrived without any visible signs of tampering or breakage. There are no additional comments concerning sample receipt.

Items of Note All efforts were made by the lab to meet any short hold times. Samples that were analyzed outside of the initial hold time but still within 2X hold time will be noted in the lab case narrative.

Sample Identification

The laboratory received the following sample:

Laboratory Identification	Sample Description
440130001	B3FK60

Case Narrative

Sample analyses were conducted using methodology as outlined in GEL Laboratories, LLC (GEL) Standard Operating Procedures. Any technical or administrative problems during analysis, data review, and reduction are contained in the analytical case narratives in the enclosed data package.

Data Package

The enclosed data package contains the following sections: General Narrative, Chain of Custody and Supporting Documentation, and data from the following fractions: General Chemistry.

We certify that this package is in compliance with the SOW, both technically and for completeness, including a full description of, explanation of, and corrective actions for, any and all deviations, from either the analyses requested or the case narrative requested. Release of the data contained in this hard copy data package has been authorized by the Laboratory Analytical Manager (or designee) and the laboratory's client services representative as verified by their signatures on this report.

12/26/2017

REV.0


Kaitlyn Stone for
Heather Shaffer
Project Manager

**General Chemistry
Technical Case Narrative
CH2MHill Plateau Remediation Company (CPRC)
SDG #: GEL440130
Work Order #: 440130**

Alkalinity

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

Certification Statement

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless otherwise noted in the analytical case narrative.

Chain of Custody and Supporting Documentation

CH2M Hill Plateau Remediation Company

C.O.C. # W18-012-080
Page 1 of 2
TAB 12-17-17

CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST

76
440130

Collector: Juan Aguilar /CHERC
SAF No.: W18-012
Project Title: RCRA, DECEMBER 2017
Shipped To (Lab): GEL Laboratories, LLC
Protocol: RCRA

Contact/Requester: Karen Waters-Husted
Sampling Origin: Hanford Site
Logbook No.: HNF-N-506 -47135
Method of Shipment: Commercial Carrier
Priority: 15 Days **PRIORITY**
Offsite Property No.: 8860

Telephone No.: 509-376-4650
Purchase Order/Charge Code: 300071
Ice Chest No.: GWS-565
Bill of Lading/Air Bill No.: 770994876526

POSSIBLE SAMPLE HAZARDS/REMARK
 ** ** Contains Radioactive Material at concentrations that are not regulated for transportation per 49 CFR / IATA Dangerous Goods Regulations but are not releasable per DOE Order 458.1

SPECIAL INSTRUCTIONS
 N/A

Sample No.	Filter	*	Date	Time	No/Type Container	Sample Analysis	Holding Time	Preservative
B3FK60	N	W	12-13-17	0851	1x250-mL G/P	2320_ALKALINITY: COMMON	14 Days	Cool <=6C

Relinquished By: Juan Aguilar /CHERC Signature	DEC 13 2017 Date/Time	Received By: Troy Backstrom /CHERC Signature	DEC 13 2017 Date/Time
Relinquished By: Juan Aguilar /CHERC Signature	DEC 13 2017 Date/Time	Received By: Troy Backstrom /CHERC Signature	DEC 13 2017 Date/Time
Relinquished By: Troy Backstrom /CHERC Signature	DEC 14 2017 07:00 Date/Time	Received By: Troy Backstrom /CHERC Signature	DEC 14 2017 07:00 Date/Time
Relinquished By: Troy Backstrom /CHERC Signature	DEC 14 2017 1:00 Date/Time	Received By: FEDEX Signature	

Matrix *
 S = Soil DS = Drum Solids
 SE = Sediment DL = Drum Liquid
 SO = Solid T = Tissue
 SL = Sludge WI = Wipe
 W = Water L = Liquid
 O = Oil V = Vegetation
 A = Air X = Other

Disposal Method (e.g., Return to customer, per lab procedure, used in process):

Disposed By: _____ Date/Time: _____



SAMPLE RECEIPT & REVIEW FORM

Client: CPRC		SDG/AR/COC/Work Order: 440130		
Received By: Tracy Boone		Date Received: 15-DEC-2017		
Carrier and Tracking Number		Circle Applicable: <input checked="" type="checkbox"/> FedEx Express <input type="checkbox"/> FedEx Ground <input type="checkbox"/> UPS <input type="checkbox"/> Field Services <input type="checkbox"/> Courier <input type="checkbox"/> Other 7709 9487 6526		
Suspected Hazard Information	Yes <input type="checkbox"/> No <input type="checkbox"/>	*If Net Counts > 100cpm on samples not marked "radioactive", contact the Radiation Safety Group for further investigation.		
Shipped as a DOT Hazardous?	Yes <input type="checkbox"/> No <input checked="" type="checkbox"/>	Hazard Class Shipped: _____ UN#: _____		
COC/Samples marked or classified as radioactive?	Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>	Maximum Net Counts Observed* (Observed Counts - Area Background Counts): <u>0</u> CPM / mR/Hr Classified as: <u>Rad 1</u> Rad 2 Rad 3		
Is package, COC, and/or Samples marked HAZ?	Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>	If yes, select Hazards below, and contact the GEL Safety Group. PCB's Flammable Foreign Soil RCRA Asbestos Beryllium Other:		
Sample Receipt Criteria	Yes	NA	No	Comments/Qualifiers (Required for Non-Conforming Items)
1 Shipping containers received intact and sealed?	<input checked="" type="checkbox"/>			Circle Applicable: Seals broken Damaged container Leaking container Other (describe)
2 Chain of custody documents included with shipment?	<input checked="" type="checkbox"/>			
3 Samples requiring cold preservation within (0 ≤ 6 deg. C)?*	<input checked="" type="checkbox"/>			Preservation Method: <u>Wet Ice</u> Ice Packs Dry Ice None Other: *all temperatures are recorded in Celsius TEMP: 1°C
4 Daily check performed and passed on IR temperature gun?	<input checked="" type="checkbox"/>			Temperature Device Serial #: <u>1R3-17</u> Secondary Temperature Device Serial # (if Applicable):
5 Sample containers intact and sealed?	<input checked="" type="checkbox"/>			Circle Applicable: Seals broken Damaged container Leaking container Other (describe)
6 Samples requiring chemical preservation at proper pH?	<input checked="" type="checkbox"/>			Sample ID's and Containers Affected: If Preservation added, Lot#:
7 Do any samples require Volatile Analysis?	<input checked="" type="checkbox"/>			If Yes, Are Encores or Soil Kits present? Yes ___ No ___ (If yes, take to VOA Freezer) Do VOA vials contain acid preservation? Yes ___ No ___ N/A (If unknown, select No) VOA vials free of headspace? Yes ___ No ___ N/A Sample ID's and containers affected:
8 Samples received within holding time?	<input checked="" type="checkbox"/>			ID's and tests affected:
9 Sample ID's on COC match ID's on bottles?	<input checked="" type="checkbox"/>			Sample ID's and containers affected:
10 Date & time on COC match date & time on bottles?	<input checked="" type="checkbox"/>			Sample ID's affected:
11 Number of containers received match number indicated on COC?	<input checked="" type="checkbox"/>			Sample ID's affected:
12 Are sample containers identifiable as GEL provided?	<input checked="" type="checkbox"/>			
13 COC form is properly signed in relinquished/received sections?	<input checked="" type="checkbox"/>			
Comments (Use Contiquation Form if needed):				

PM (or PMA) review: Initials met Date 12/18/17 Page 1 of 1

Data Review Qualifier Definitions

Project Specific Qualifier Definitions for GEL Client Code: CPRC

Qualifier	Qualifier Definition	Department	Fraction
U	Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.		
J	The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate). Value is estimated	Organics	
P	Aroclor target analyte with greater than 25% difference between column analyses.	Organics	
C	Analyte has been confirmed by GC/MS analysis	Organics	Pesticide
B	The analyte was detected in both the associated QC blank and in the sample.	Organics	
E	Concentration exceeds the calibration range of the instrument	Organics	
A	The TIC is a suspected aldol-condensation product	Organics	Semi-Volatile
X	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier		
N	Spike Sample recovery is outside control limits.		
*	Duplicate analysis not within control limits	Inorganics	
>	Result greater than quantifiable range or greater than upper limit of the analysis range	General Chemistry	
Z	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier		
B	The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate).	Inorganics	Metals
D	Results are reported from a diluted aliquot of sample.		
E	Reported value is estimated due to interferences. See comment in narrative.	Inorganics	Metals
M	Duplicate precision not met.	Inorganics	Metals
o	Analyte failed to recover within LCS limits (Organics only)	Organics	
S	Reported value determined by the Method of Standard Additions (MSA)	Inorganics	
T	Spike and/or spike duplicate sample recovery is outside control limits.	Organics	
W	Post-digestion spike recovery for GFAA out of control limit. Sample absorbency < 50% of spike absorbency.	Inorganics	
B	The associated QC sample blank has a result $\geq 2X$ the MDA and, after corrections, result is \geq MDA for this sample	Radiological	
Y	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier		
+	Correlation coefficient for Method of Standard Additions (MSA) is < 0.995	Inorganics	
B	The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate).	General Chemistry	
C	Target analyte was detected in the sample and the associated blank. The associated blank concentration is \geq EQL or is > 5% of the measured concentration and/or decision level for associated samples.	Inorganics	Metals
C	Target analyte was detected in the sample and the associated blank. The associated blank concentration is \geq EQL or is > 5% of the measured concentration and/or decision level for associated samples.	General Chemistry	
<	Sample is below the EPA guidance level for Reactive Releasable Cyanide and/or Reactive Releasable Sulfide	General Chemistry	
UX	Gamma Spectroscopy--Uncertain identification	Radiological	

Laboratory Certifications

List of current GEL Certifications as of 26 December 2017

State	Certification
Alaska	UST-0110
Arkansas	88-0651
CLIA	42D0904046
California	2940
Colorado	SC00012
Connecticut	PH-0169
Delaware	SC00012
DoD ELAP/ ISO17025 A2LA	2567.01
Florida NELAP	E87156
Foreign Soils Permit	P330-15-00283, P330-15-00253
Georgia	SC00012
Georgia SDWA	967
Hawaii	SC00012
Idaho Chemistry	SC00012
Idaho Radiochemistry	SC00012
Illinois NELAP	200029
Indiana	C-SC-01
Kansas NELAP	E-10332
Kentucky SDWA	90129
Kentucky Wastewater	90129
Louisiana NELAP	03046 (AI33904)
Louisiana SDWA	LA170010
Maryland	270
Massachusetts	M-SC012
Michigan	9976
Mississippi	SC00012
Nebraska	NE-OS-26-13
Nevada	SC000122018-1
New Hampshire NELAP	205415
New Jersey NELAP	SC002
New Mexico	SC00012
New York NELAP	11501
North Carolina	233
North Carolina SDWA	45709
North Dakota	R-158
Oklahoma	9904
Pennsylvania NELAP	68-00485
Puerto Rico	SC00012
S.Carolina Radchem	10120002
South Carolina Chemistry	10120001
Tennessee	TN 02934
Texas NELAP	T104704235-17-12
Utah NELAP	SC000122017-25
Vermont	VT87156
Virginia NELAP	460202
Washington	C780
West Virginia	997404

General Chem Analysis

Case Narrative

**General Chemistry
Technical Case Narrative
CH2MHill Plateau Remediation Company (CPRC)
SDG #: GEL440130
Work Order #: 440130**

Product: Alkalinity

Analytical Method: 2320_ALKALINITY

Analytical Procedure: GL-GC-E-033 REV# 13

Analytical Batch: 1727519

The following samples were analyzed using the above methods and analytical procedure(s).

<u>GEL Sample ID#</u>	<u>Client Sample Identification</u>
440130001	B3FK60
1203941754	Laboratory Control Sample (LCS)
1203941755	440092002(B3FKD4) Sample Duplicate (DUP)

The samples in this SDG were analyzed on an "as received" basis.

Data Summary:

There are no exceptions, anomalies or deviations from the specified methods. All sample data provided in this report met the acceptance criteria specified in the analytical methods and procedures for initial calibration, continuing calibration, instrument controls and process controls where applicable.

Certification Statement

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless otherwise noted in the analytical case narrative.

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

**Qualifier Definition Report
for**

CPRC001 CH2MHill Plateau Remediation Company

Client SDG: GEL440130 GEL Work Order: 440130

The Qualifiers in this report are defined as follows:**Review/Validation**

GEL requires all analytical data to be verified by a qualified data reviewer. In addition, all CLP-like deliverables receive a third level review of the fractional data package.

The following data validator verified the information presented in this data report:

Signature: **Name:** Aubrey Kingsbury**Date:** 26 DEC 2017**Title:** Analyst I

Sample Data Summary

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis

Report Date: December 26, 2017

Company : CH2MHill Plateau Remediation Company
 Address : MSIN R3-50 CHPRC
 PO Box 1600
 Richland, Washington 99352
 Contact: Mr. Scot Fitzgerald
 Project: CHPRC SAF W18-012

Client Sample ID: B3FK60 Project: CPRCOW18012
 Sample ID: 440130001 Client ID: CPRC001
 Matrix: WATER
 Collect Date: 13-DEC-17 08:51
 Receive Date: 15-DEC-17
 Collector: Client

Parameter	Qualifier	Result	DL	RL	Units	PF	DF	Analyst	Date	Time	Batch	Method
Titration and Ion Analysis												
2320_ALKALINITY: COMMON (Alkalinity only) "As Received"												
Alkalinity, Total as CaCO3		97600	1450	4000	ug/L			RXB5	12/21/17	1714	1727519	1

The following Analytical Methods were performed:

Method	Description	Analyst	Comments
1	2320_ALKALINITY		

Notes:

Column headers are defined as follows:

DF: Dilution Factor Lc/LC: Critical Level
 DL: Detection Limit PF: Prep Factor
 MDA: Minimum Detectable Activity RL: Reporting Limit
 MDC: Minimum Detectable Concentration SQL: Sample Quantitation Limit

Quality Control Summary

GEL LABORATORIES LLC

2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Report Date: December 26, 2017

Page 1 of 1

CH2M Hill Plateau Remediation Company
MSIN R3-50 CHPRC
PO Box 1600
Richland, Washington

Contact: Mr. Scot Fitzgerald

Workorder: 440130

Parmname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date	Time
Titration and Ion Analysis											
Batch	1727519										
QC1203941755	440092002	DUP									
Alkalinity, Total as CaCO3		90000		90000	ug/L	0		(0%-20%)	RXB5	12/21/17	16:55
QC1203941754	LCS										
Alkalinity, Total as CaCO3	100000			107000	ug/L		107	(80%-120%)		12/21/17	16:53

Notes:

The Qualifiers in this report are defined as follows:

- < Sample is below the EPA guidance level for Reactive Releasable Cyanide and/or Reactive Releasable Sulfide
- > Result greater than quantifiable range or greater than upper limit of the analysis range
- B The analyte was detected at a value less than the contract required detection limit (RDL), but greater than or equal to the IDL/MDL (as appropriate).
- C Target analyte was detected in the sample and the associated blank. The associated blank concentration is >= EQL or is > 5% of the measured concentration and/or decision level for associated samples.
- D Results are reported from a diluted aliquot of sample.
- N Spike Sample recovery is outside control limits.
- U Analyzed for but not detected above limiting criteria. Includes MDL, MDA, PQL, zero, counting error, and total analytical error.
- X Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
- Y Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
- Z Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier

N/A indicates that spike recovery limits do not apply when sample concentration exceeds spike conc. by a factor of 4 or more or %RPD not applicable.
 ^ The Relative Percent Difference (RPD) obtained from the sample duplicate (DUP) is evaluated against the acceptance criteria when the sample is greater than five times (5X) the contract required detection limit (RL). In cases where either the sample or duplicate value is less than 5X the RL, a control limit of +/- the RL is used to evaluate the DUP result.
 * Indicates that a Quality Control parameter was not within specifications.
 For PS, PSD, and SDILT results, the values listed are the measured amounts, not final concentrations.

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless qualified on the QC Summary.