

RECEIVED NOVEMBER 19, 2010

Revision 1 *etc* 11/22/10



EBERLINE
SERVICES

EBERLINE ANALYTICAL CORPORATION
2838 Wright Avenue
Richmond, California 94804-3849
Phone (510) 235-2623 Fax (510) 235-8438
Toll Free (800) 841-5487
www.eberlineservices.com

November 19, 2010

0096754

Mr. Michael Neely
CH2M Hill Plateau Remediation Company
P.O. Box 1600
Mail Stop - B6-06
Richland, WA 99352

Reference: P.O. #33677
Eberline Analytical S0-04-124-7652, SDG H4220
S0-10-057-7652, SDG H4220

Dear Mr. Neely:

Enclosed is a data report for four water samples designated under SAF No. X10-039 received at Eberline Analytical on April 28, 2010. Results for the originally requested Total Radium analysis were reported on June 2, 2010. This report includes the additionally requested Ra-228 analysis.


Please call if you have any questions concerning this report.

Sincerely,

N. Joseph Verville
Client Services Manager

NJV/ljb

Enclosure: Data Package



1.0 GENERAL

CH2M Hill Plateau Remediation Company (CHPRC) Sample Delivery Group H4220 was composed of four water samples designated under SAF No. X10-039 with a Project Designation of: IU2, IU6DURA, APRIL 2010.

The samples were received as stated on the chain-of-custody documents. Any discrepancies are noted on the Eberline Analytical Sample Receipt Checklist.

2.0 ANALYSIS NOTES

2.1 Total Radium Analysis


The results for both the original and duplicate analyses were less than their respective MDA's, therefore no RPD is calculated, and there is no associated control limit. The activity for the QC LCS was 80%, which is an acceptable recovery. No other problems were encountered during the course of the analyses.

2.2 Radium-228 Analysis

The MDA's for samples B24M49 (3.21 pCi/L), B25044 (4.01 pCi/L), the duplicate of sample B25044 (4.76 pCi/L) and B25050 (3.14 pCi/L) were greater than the RDL of 3.0 pCi/L; Ra-228 activity greater than any sample specific MDA was not observed. The results for both the original and duplicate analyses were less than their respective MDA's, therefore no RPD is calculated, and there is no associated control limit. No other problems were encountered during the course of the analyses.

3.0 Case Narrative Certification Statement

"I certify that this data package is in compliance with the SOW, both technically and for completeness, for other than the conditions detailed above. Release of the data obtained in this hard copy data package has been authorized by the Laboratory Manager or a designee, as verified by the following signature."



N. Joseph Verville
Client Services Manager

11/19/10

Date

EBERLINE ANALYTICAL / RICHMOND
SAMPLE DELIVERY GROUP H4220

SDG 7652
Contact N. Joseph Verville

Client CHPRC
Contract No. 33677
Case no SDG_H4220

S U M M A R Y D A T A S E C T I O N

T A B L E O F C O N T E N T S				
About this section	.	.	.	1
Sample Summaries	.	.	.	3
Prep Batch Summary	.	.	.	5
Work Summary	.	.	.	6
Method Blanks	.	.	.	8
Lab Control Samples	.	.	.	10
Duplicates	.	.	.	12
Data Sheets	.	.	.	14
Method Summaries	.	.	.	18
Report Guides	.	.	.	20
End of Section	.	.	.	34

VB

Prepared by _____

Reviewed by _____

N. Joseph Verville

Lab id EBRLNE
Protocol CHPRC
Version Ver 1.0
Form DVD-TOC
Version 3.06
Report date 11/19/10

EBERLINE ANALYTICAL / RICHMOND

SAMPLE DELIVERY GROUP H4220

SDG 7652
Contact N. Joseph Verville

REPORT GUIDE

Client CHPRC
Contract No. 33677
Case no SDG_H4220

ABOUT THE DATA SUMMARY SECTION

The Data Summary Section of a Data Package has all data, in several useful orders, necessary for first level, routine review of the data package for a Sample Delivery Group (SDG). This section follows the Data Package Narrative, which has an overview of the data package and a discussion of special problems. It is followed by the Raw Data Section, which has full details.

The Data Summary Section has several groups of reports:

SAMPLE SUMMARIES

The Sample and QC Summary Reports show all samples, including QC samples, reported in one SDG. These reports cross-reference client and lab sample identifiers.

PREPARATION BATCH SUMMARY

The Preparation Batch Summary Report shows all preparation batches (lab groupings reflecting how work was organized) relevant to the reported SDG with information necessary to check the completeness and consistency of the SDG.

WORK SUMMARY

The Work Summary Report shows all samples and work done on them relevant to the reported SDG.

METHOD BLANKS

The Method Blank Reports, one for each Method Blank relevant to the SDG, show all results and primary supporting information for the blanks.

LAB CONTROL SAMPLES

The Lab Control Sample Reports, one for each Lab Control Sample relevant to the SDG, show all results, recoveries and primary supporting information for these QC samples.

Lab id EBRLNE
Protocol CHPRC
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 11/19/10

EBERLINE ANALYTICAL/RICHMOND

SAMPLE DELIVERY GROUP H4220

SDG 7652
Contact N. Joseph Verville

GUIDE, cont.

Client CHPRC
Contract No. 33677
Case no SDG H4220

ABOUT THE DATA SUMMARY SECTION

DUPLICATES

The Duplicate Reports, one for each Duplicate and Original sample pair relevant to the SDG, show all results, differences and primary supporting information for these QC samples.

MATRIX SPIKES

The Matrix Spike Reports, one for each Spiked and Original sample pair relevant to the SDG, show all results, recoveries and primary supporting information for these QC samples.

DATA SHEETS

The Data Sheet Reports, one for each client sample in the SDG, show all results and primary supporting information for these samples.

METHOD SUMMARIES

The Method Summary Reports, one for each test used in the SDG, show all results, QC and method performance data for one analyte on one or two pages. (A test is a short code for the method used to do certain work to the client's specification.)

REPORT GUIDES

The Report Guides, one for each of the above groups of reports, have documentation on how to read the associated reports.

REPORT GUIDES
Page 2
SUMMARY DATA SECTION
Page 2

Lab id EBRLNE
Protocol CHPRC
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 11/19/10

EBERLINE ANALYTICAL/RICHMOND

SAMPLE DELIVERY GROUP H4220

LAB SAMPLE SUMMARY

SDG 7652
 Contact N. Joseph Verville

Client CHPRC
 Contract No. 33677
 Case no SDG H4220

LAB						CHAIN OF	
SAMPLE ID	CLIENT SAMPLE ID	LOCATION	MATRIX	LEVEL	SAP NO	CUSTODY	COLLECTED
S004124-01	B24M49	HNF-N-506 28/89	WATER		X10-039	X10-039-290	04/26/10 10:46
S004124-02	B25044	HNF-N-506 31/47	WATER		X10-039	X10-039-314	04/26/10 10:05
S004124-03	B25050	HNF-N-506 31/47	WATER		X10-039	X10-039-320	04/26/10 12:17
S004124-04	B25077	HNF-N-506 28/89	WATER		X10-039	X10-039-342	04/26/10 13:16
S004124-05	Lab Control Sample		WATER		X10-039		
S004124-06	Method Blank		WATER		X10-039		
S004124-07	Duplicate (S004124-01)	HNF-N-506 28/89	WATER		X10-039		04/26/10 10:46
S004124-08	Lab Control Sample		WATER		X10-039		
S004124-09	Method Blank		WATER		X10-039		
S004124-10	Duplicate (S004124-02)	HNF-N-506 31/47	WATER		X10-039		04/26/10 10:05

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-LS
 Version 3.06
 Report date 11/19/10

EBERLINE ANALYTICAL/RICHMOND

SAMPLE DELIVERY GROUP H4220

SDG 7652
 Contact N. Joseph Verville

QC SUMMARY

Client CHPRC
 Contract No. 33677
 Case no SDG H4220

QC BATCH	CHAIN OF CUSTODY	CLIENT SAMPLE ID	MATRIX	SOLIDS	SAMPLE AMOUNT	BASIS AMOUNT	DAYS SINCE RECEIVED	LAB COLL	LAB SAMPLE ID	DEPARTMENT SAMPLE ID
7652	X10-039-290	B24M49	WATER		2.020 L		04/28/10	2	S004124-01	7652-001
	X10-039-314	B25044	WATER		2.020 L		04/28/10	2	S004124-02	7652-002
	X10-039-320	B25050	WATER		2.020 L		04/28/10	2	S004124-03	7652-003
	X10-039-342	B25077	WATER				04/28/10	2	S004124-04	7652-004
		Method Blank	WATER						S004124-06	7652-006
		Method Blank	WATER						S004124-09	7652-009
		Lab Control Sample	WATER						S004124-05	7652-005
		Lab Control Sample	WATER						S004124-08	7652-008
		Duplicate (S004124-01)	WATER		2.020 L		04/28/10	2	S004124-07	7652-007
		Duplicate (S004124-02)	WATER		2.020 L		04/28/10	2	S004124-10	7652-010

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-QS
 Version 3.06
 Report date 11/19/10

EBERLINE ANALYTICAL/RICHMOND

SAMPLE DELIVERY GROUP H4220

SDG 7652
 Contact N. Joseph Verville

PREP BATCH SUMMARY

Client CHPRC
 Contract No. 33677
 Case no SDG H4220

TEST MATRIX	METHOD	PREPARATION ERROR BATCH	PREPARATION ERROR		PLANCHETS ANALYZED			QUALI- FIERS	
			2σ %	CLIENT MORE	RE	BLANK	LCS DUP/ORIG MS/ORIG		
Beta Counting									
AC	WATER	Radium 228 in Water	7276-113	10.4	4		1	1	1/1
Gas Proportional Counting									
RAT	WATER	Total Alpha Radium in Water	7251-071	12.8	4		1	1	1/1

Blank, LCS, Duplicate and Spike planchets are those in the same preparation batch as some Client sample.

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-PBS
 Version 3.06
 Report date 11/19/10

EBERLINE ANALYTICAL/RICHMOND

SAMPLE DELIVERY GROUP H4220

SDG 7652
Contact N. Joseph Verville

LAB WORK SUMMARY

Client CHPRC
Contract No. 33677
Case no SDG H4220

LAB SAMPLE COLLECTED RECEIVED	CLIENT SAMPLE ID LOCATION CUSTODY	SAF No	MATRIX	PLANCHET	TEST	SUF- FIX ANALYZED	REVIEWED	BY	METHOD
S004124-01 04/26/10 04/28/10	B24M49 HNF-N-506 28/89 X10-039-290	X10-039		7652-001 7652-001	AC RAT	11/16/10 05/22/10	11/17/10 05/24/10	BW BW	Radium 228 in Water Total Alpha Radium in Water
S004124-02 04/26/10 04/28/10	B25044 HNF-N-506 31/47 X10-039-314	X10-039		7652-002 7652-002	AC RAT	11/16/10 05/22/10	11/17/10 05/24/10	BW BW	Radium 228 in Water Total Alpha Radium in Water
S004124-03 04/26/10 04/28/10	B25050 HNF-N-506 31/47 X10-039-320	X10-039		7652-003 7652-003	AC RAT	11/16/10 05/22/10	11/17/10 05/24/10	BW BW	Radium 228 in Water Total Alpha Radium in Water
S004124-04 04/26/10 04/28/10	B25077 HNF-N-506 28/89 X10-039-342	X10-039		7652-004 7652-004	AC RAT	11/16/10 05/20/10	11/17/10 05/24/10	BW BW	Radium 228 in Water Total Alpha Radium in Water
S004124-05	Lab Control Sample		WATER	7652-005	RAT	05/20/10	05/24/10	BW	Total Alpha Radium in Water
		X10-039							
S004124-06	Method Blank		WATER	7652-006	RAT	05/22/10	05/24/10	BW	Total Alpha Radium in Water
		X10-039							
S004124-07 04/26/10 04/28/10	Duplicate (S004124-01) HNF-N-506 28/89	X10-039	WATER	7652-007	RAT	05/22/10	05/24/10	BW	Total Alpha Radium in Water
S004124-08	Lab Control Sample		WATER	7652-008	AC	11/16/10	11/17/10	BW	Radium 228 in Water
		X10-039							
S004124-09	Method Blank		WATER	7652-009	AC	11/16/10	11/17/10	BW	Radium 228 in Water
		X10-039							
S004124-10 04/26/10 04/28/10	Duplicate (S004124-02) HNF-N-506 31/47	X10-039	WATER	7652-010	AC	11/16/10	11/17/10	BW	Radium 228 in Water

WORK SUMMARY

Page 1

SUMMARY DATA SECTION

Page 6

Lab id EBRLNE
Protocol CHPRC
Version Ver 1.0
Form DVD-LWS
Version 3.06
Report date 11/19/10

EBERLINE ANALYTICAL/RICHMOND

SAMPLE DELIVERY GROUP H4220

SDG 7652
 Contact N. Joseph Verville

WORK SUMMARY, cont.

Client CHPRC
 Contract No. 33677
 Case no SDG H4220

COUNTS OF TESTS BY SAMPLE TYPE										
TEST	SAP No	METHOD	REFERENCE	CLIENT	MORE	RE	BLANK	LCS	DUP SPIKE	TOTAL
AC	X10-039	Radium 228 in Water	RAISO_SEP_GPC	4			1	1	1	7
RAT	X10-039	Total Alpha Radium in Water	RATOT_GPC	4			1	1	1	7
TOTALS				8			2	2	2	14

WORK SUMMARY

Page 2

SUMMARY DATA SECTION

Page 7

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-LWS
 Version 3.06
 Report date 11/19/10

EBERLINE ANALYTICAL / RICHMOND

SAMPLE DELIVERY GROUP H4220

7652-006

Method Blank

METHOD BLANK

SDG <u>7652</u>	Client/Case no <u>CHPRC</u>	SDG <u>H4220</u>
Contact <u>N. Joseph Verville</u>	Contract No. <u>33677</u>	
Lab sample id <u>S004124-06</u>	Client sample id <u>Method Blank</u>	
Dept sample id <u>7652-006</u>	Material/Matrix <u>WATER</u>	
	SAF No <u>X10-039</u>	

ANALYTE	CAS NO	RESULT pCi/L	2σ ERR (COUNT)	MDA pCi/L	RDL pCi/L	QUALI- FIERS	TEST
Total Radium	ALPHA-RA	0.069	0.081	0.295	1.00	U	RAT

QC-BLANK #73235

METHOD BLANKS

Page 1

SUMMARY DATA SECTION

Page 8

Lab id <u>EBRLNE</u>
Protocol <u>CHPRC</u>
Version <u>Ver 1.0</u>
Form <u>DVD-DS</u>
Version <u>3.06</u>
Report date <u>11/19/10</u>

EBERLINE ANALYTICAL / RICHMOND

SAMPLE DELIVERY GROUP H4220

7652-009

Method Blank

METHOD BLANK

SDG <u>7652</u>	Client/Case no <u>CHPRC</u>	SDG <u>H4220</u>
Contact <u>N. Joseph Verville</u>	Contract <u>No. 33677</u>	
Lab sample id <u>S004124-09</u>	Client sample id <u>Method Blank</u>	
Dept sample id <u>7652-009</u>	Material/Matrix <u>WATER</u>	
	SAF No <u>X10-039</u>	

ANALYTE	CAS NO	RESULT pCi/L	2σ ERR (COUNT)	MDA pCi/L	RDL pCi/L	QUALIFIERS	TEST
Radium 228	15262-20-1	-0.210	1.1	2.78	3.00	U	AC

QC-BLANK #75833

Lab id <u>EBRLNE</u>
Protocol <u>CHPRC</u>
Version <u>Ver 1.0</u>
Form <u>DVD-DS</u>
Version <u>3.06</u>
Report date <u>11/19/10</u>

EBERLINE ANALYTICAL/RICHMOND

SAMPLE DELIVERY GROUP H4220

7652-005

Lab Control Sample

LAB CONTROL SAMPLE

SDG <u>7652</u> Contact <u>N. Joseph Verville</u> Lab sample id <u>S004124-05</u> Dept sample id <u>7652-005</u>	Client/Case no <u>CHPRC</u> SDG <u>H4220</u> Contract No. <u>33577</u> Client sample id <u>Lab Control Sample</u> Material/Matrix _____ <u>WATER</u> SAF No <u>X10-039</u>
---	--

ANALYTE	RESULT	2σ ERR	MDA	RDL	QUALI-	ADDED	2σ ERR	REC	3σ LMIS	PROTOCOL
	pCi/L	(COUNT)	pCi/L	pCi/L	FIERS	TEST	pCi/L	pCi/L	‡ (TOTAL)	LIMITS
Total Radium	49.0	3.8	0.462	1.00		RAT	61.5	2.5	<u>80</u>	81-119 80-120

QC-LCS #73234

LAB CONTROL SAMPLES

Page 1

SUMMARY DATA SECTION

Page 10

Lab id <u>EBRLNE</u>
Protocol <u>CHPRC</u>
Version <u>Ver 1.0</u>
Form <u>DVD-LCS</u>
Version <u>3.06</u>
Report date <u>11/19/10</u>

EBERLINE ANALYTICAL/RICHMOND

SAMPLE DELIVERY GROUP H4220

7652-008

Lab Control Sample

LAB CONTROL SAMPLE

SDG <u>7652</u> Contact <u>N. Joseph Verville</u> Lab sample id <u>S004124-08</u> Dept sample id <u>7652-008</u>	Client/Case no <u>CHPRC</u> SDG <u>H4220</u> Contract <u>No. 33677</u> Client sample id <u>Lab Control Sample</u> Material/Matrix _____ <u>WATER</u> SAF No <u>X10-039</u>
---	--

ANALYTE	RESULT	2σ ERR	MDA	RDL	QUALI-	ADDED	2σ ERR	REC	3σ LMTS	PROTOCOL
	pCi/L	(COUNT)	pCi/L	pCi/L	FIERS	pCi/L	pCi/L	%	(TOTAL)	LIMITS
Radium 228	14.0	1.6	2.97	3.00	AC	14.2	0.57	99	76-124	80-120

QC-LCS #75832

Lab id <u>EBRLNE</u>
Protocol <u>CHPRC</u>
Version <u>Ver 1.0</u>
Form <u>DVD-LCS</u>
Version <u>3.06</u>
Report date <u>11/19/10</u>

EBERLINE ANALYTICAL/RICHMOND

SAMPLE DELIVERY GROUP H4220

7652-007

B24M49

DUPLICATE

<u>SDG 7652</u>	<u>Client/Case no CHPRC</u>	<u>SDG H4220</u>
<u>Contact N. Joseph Verville</u>	<u>Contract No. 33677</u>	
DUPLICATE	ORIGINAL	
Lab sample id <u>S004124-07</u>	Lab sample id <u>S004124-01</u>	Client sample id <u>B24M49</u>
Dept sample id <u>7652-007</u>	Dept sample id <u>7652-001</u>	Location/Matrix <u>HNF-N-506 28/89</u> <u>WATER</u>
	Received <u>04/28/10</u>	Collected/Volume <u>04/26/10 10:46</u> <u>2.020 L</u>
		Custody/SAF No <u>X10-039-290</u> <u>X10-039</u>

ANALYTE	DUPLICATE		MDA pCi/L	RDL pCi/L	QUALI- FIERS	TEST	ORIGINAL		MDA pCi/L	QUALI- FIERS	RPD %	3σ TOT	DER σ
	pCi/L	2σ ERR (COUNT)					pCi/L	2σ ERR (COUNT)					
Total Radium	-0.008	0.093	0.417	1.00	U	RAT	0.123	0.12	0.375	U	-	1.7	

QC-DUP#1 73236

IU2, IU6DURA, APRIL 2010

Lab id <u>EBRLNE</u>
Protocol <u>CHPRC</u>
Version <u>Ver 1.0</u>
Form <u>DVD-DUP</u>
Version <u>3.06</u>
Report date <u>11/19/10</u>

EBERLINE ANALYTICAL/RICHMOND

SAMPLE DELIVERY GROUP H4220

7652-010

B25044

DUPLICATE

SDG <u>7652</u> Contact <u>N. Joseph Verville</u> Duplicates Lab sample id <u>S004124-10</u> Dept sample id <u>7652-010</u>	Client/Case no <u>CHERC</u> SDG <u>H4220</u> Contract No. <u>33677</u> ORIGINAL Lab sample id <u>S004124-02</u> Dept sample id <u>7652-002</u> Received <u>04/28/10</u>
Client sample id <u>B25044</u> Location/Matrix <u>HNF-N-506 31/47</u> <u>WATER</u> Collected/Volume <u>04/26/10 10:05</u> <u>2.020 L</u> Custody/SAF No <u>X10-039-314</u> <u>X10-039</u>	

ANALYTE	DUPLICATE		MDA pCi/L	RDL pCi/L	QUALI- FIERS	TEST	ORIGINAL		MDA pCi/L	QUALI- FIERS	RPD %	3σ TOT	DER σ
	pCi/L	2σ ERR (COUNT)					pCi/L	2σ ERR (COUNT)					
Radium 228	-0.816	1.8	<u>4.76</u>	3.00	U	AC	1.19	2.0	<u>4.01</u>	U	-	1.5	

QC-DUP#2 75834

IU2, IU6DURA, APRIL 2010

Lab id <u>EBRLNE</u>
Protocol <u>CHERC</u>
Version <u>Ver 1.0</u>
Form <u>DVD-DUP</u>
Version <u>3.06</u>
Report date <u>11/19/10</u>

EBERLINE ANALYTICAL / RICHMOND
SAMPLE DELIVERY GROUP H4220

7652-001

B24M49

DATA SHEET

SDG <u>7652</u>	Client/Case no <u>CHPRC</u>	SDG <u>H4220</u>
Contact <u>N. Joseph Verville</u>	Contract No. <u>33677</u>	
Lab sample id <u>S004124-01</u>	Client sample id <u>B24M49</u>	
Dept sample id <u>7652-001</u>	Location/Matrix <u>HNF-N-506 28/89</u>	<u>WATER</u>
Received <u>04/28/10</u>	Collected/Volume <u>04/26/10 10:46</u>	<u>2.020 L</u>
	Custody/SAF No <u>X10-039-290</u>	<u>X10-039</u>

ANALYTE	CAS NO	RESULT pCi/L	2σ ERR (COUNT)	MDA pCi/L	RDL pCi/L	QUALI- FIERS	TEST
Total Radium	ALPHA-RA	0.123	0.12	0.375	1.00	U	RAT
Radium 228	15262-20-1	0.580	1.4	<u>3.21</u>	3.00	U	AC

IU2, IU6DURA, APRIL 2010

Lab id <u>EBRLNE</u>
Protocol <u>CHPRC</u>
Version <u>Ver 1.0</u>
Form <u>DVD-DS</u>
Version <u>3.06</u>
Report date <u>11/19/10</u>

EBERLINE ANALYTICAL / RICHMOND
SAMPLE DELIVERY GROUP H4220

7652-002

B25044

DATA SHEET

SDG <u>7652</u>	Client/Case no <u>CHPRC</u>	SDG <u>H4220</u>
Contact <u>N. Joseph Verville</u>	Contract <u>No. 33677</u>	
Lab sample id <u>S004124-02</u>	Client sample id <u>B25044</u>	
Dept sample id <u>7652-002</u>	Location/Matrix <u>HNF-N-506 31/47</u>	<u>WATER</u>
Received <u>04/28/10</u>	Collected/Volume <u>04/26/10 10:05</u>	<u>2.020 L</u>
	Custody/SAF No <u>X10-039-314</u>	<u>X10-039</u>

ANALYTE	CAS NO	RESULT pCi/L	2σ ERR (COUNT)	MDA pCi/L	RDL pCi/L	QUALI- FIERS	TEST
Total Radium	ALPHA-RA	0.114	0.12	0.354	1.00	U	RAT
Radium 228	15262-20-1	1.19	2.0	<u>4.01</u>	3.00	U	AC

IU2, IU6DURA, APRIL 2010

Lab id <u>EBRLNE</u>
Protocol <u>CHPRC</u>
Version <u>Ver 1.0</u>
Form <u>DVD-DS</u>
Version <u>3.06</u>
Report date <u>11/19/10</u>

EBERLINE ANALYTICAL / RICHMOND
SAMPLE DELIVERY GROUP H4220

7652-003

B25050

DATA SHEET

SDG <u>7652</u>	Client/Case no <u>CHPRC</u>	<u>SDG H4220</u>
Contact <u>N. Joseph Verville</u>	Contract No. <u>33677</u>	
Lab sample id <u>S004124-03</u>	Client sample id <u>B25050</u>	
Dept sample id <u>7652-003</u>	Location/Matrix <u>HNF-N-506 31/47</u>	<u>WATER</u>
Received <u>04/28/10</u>	Collected/Volume <u>04/26/10 12:17</u>	<u>2.020 L</u>
	Custody/SAF No <u>X10-039-320</u>	<u>X10-039</u>

ANALYTE	CAS NO	RESULT	2σ ERR	MDA	RDL	QUALI-	TEST
		pCi/L	(COUNT)	pCi/L	pCi/L	FIERS	
Total Radium	ALPHA-RA	0.088	0.11	0.344	1.00	U	RAT
Radium 228	15262-20-1	-0.813	1.3	<u>3.14</u>	3.00	U	AC

IU2, IU6DURA, APRIL 2010

Lab id <u>EBRLNE</u>
Protocol <u>CHPRC</u>
Version <u>Ver 1.0</u>
Form <u>DVD-DS</u>
Version <u>3.06</u>
Report date <u>11/19/10</u>

EBERLINE ANALYTICAL / RICHMOND
SAMPLE DELIVERY GROUP H4220

7652-004

B25077

DATA SHEET

SDG <u>7652</u>	Client/Case no <u>CHPRC</u>	SDG <u>H4220</u>
Contact <u>N. Joseph Verville</u>	Contract <u>No. 33677</u>	
Lab sample id <u>S004124-04</u>	Client sample id <u>B25077</u>	
Dept sample id <u>7652-004</u>	Location/Matrix <u>HNF-N-506 28/89</u>	<u>WATER</u>
Received <u>04/28/10</u>	Collected <u>04/26/10 13:16</u>	
	Custody/SAF No <u>X10-039-342</u>	<u>X10-039</u>

ANALYTE	CAS NO	RESULT pCi/L	2σ ERR (COUNT)	MDA pCi/L	RDL pCi/L	QUALI- FIERS	TEST
Total Radium	ALPHA-RA	0.156	0.12	0.364	1.00	U	RAT
Radium 228	15262-20-1	-0.689	1.1	2.83	3.00	U	AC

IU2, IU6DURA, APRIL 2010

Lab id <u>EBRLNE</u>
Protocol <u>CHPRC</u>
Version <u>Ver 1.0</u>
Form <u>DVD-DS</u>
Version <u>3.06</u>
Report date <u>11/19/10</u>

EBERLINE ANALYTICAL/RICHMOND

SAMPLE DELIVERY GROUP H4220

Test AC Matrix WATER
 SDG 7652
 Contact N. Joseph Verville

LAB METHOD SUMMARY

RADIUM 228 IN WATER

BETA COUNTING

Client CHPRC
 Contract No. 33677
 Contract SDG H4220

RESULTS

LAB	RAW	SUP-		
SAMPLE ID	TEST FIX	FLANCHET	CLIENT SAMPLE ID	Radium 228
Preparation batch 7276-113				
S004124-01		7652-001	B24M49	U
S004124-02		7652-002	B25044	U
S004124-03		7652-003	B25050	U
S004124-04		7652-004	B25077	U
S004124-08		7652-008	Lab Control Sample	ok
S004124-09		7652-009	Method Blank	U
S004124-10		7652-010	Duplicate (S004124-02)	- U

Nominal values and limits from method RDLs (pCi/L) 3.00

METHOD PERFORMANCE

LAB	RAW	SUP-	MDA	ALIQ	PREP	DILU-	YIELD	EPF	COUNT	FWHM	DRIFT	DAYS	ANAL-		
SAMPLE ID	TEST FIX	CLIENT SAMPLE ID	pCi/L	L	FAC	TION	%	%	min	keV	KeV	HELD	PREPARED	YZED	DETECTOR
Preparation batch 7276-113 2σ prep error 10.4 % Reference Lab Notebook No. 7276 pg.113															
S004124-01		B24M49	<u>3.21</u>	0.500			78	150				<u>204</u>	11/16/10	11/16	GRB-225
S004124-02		B25044	<u>4.01</u>	0.500			75	150				<u>204</u>	11/16/10	11/16	GRB-226
S004124-03		B25050	<u>3.14</u>	0.500			77	150				<u>204</u>	11/16/10	11/16	GRB-227
S004124-04		B25077	2.83	0.500			78	150				<u>204</u>	11/16/10	11/16	GRB-228
S004124-08		Lab Control Sample	2.97	0.500			76	150					11/16/10	11/16	GRB-201
S004124-09		Method Blank	2.78	0.500			74	150					11/16/10	11/16	GRB-202
S004124-10		Duplicate (S004124-02)	<u>4.76</u>	<u>0.300</u>			76	150				<u>204</u>	11/16/10	11/16	GRB-203

Nominal values and limits from method 3.00 0.500 40-110 100 180

PROCEDURES REFERENCE RAISO_SEP_GPC
 SPP-062 Sample Aliquoting, rev 1
 DWP-894 Sequential Separation of Actinium-228 and Radium-226 in Drinking Water (>1 Liter Aliquot), rev 5

AVERAGES ± 2 SD MDA 3.39 ± 1.46
 FOR 7 SAMPLES YIELD 76 ± 3

METHOD SUMMARIES
 Page 1
 SUMMARY DATA SECTION
 Page 18

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-LMS
 Version 3.06
 Report date 11/19/10

EBERLINE ANALYTICAL/RICHMOND

SAMPLE DELIVERY GROUP H4220

Test RAT Matrix WATER
SDG 7652
Contact N. Joseph Verville

LAB METHOD SUMMARY

TOTAL ALPHA RADIUM IN WATER
GAS PROPORTIONAL COUNTING

Client CHPRC
Contract No. 33677
Contract SDG H4220

RESULTS

LAB	RAW	SUF-			
SAMPLE ID	TEST FIX	PLANCHET	CLIENT SAMPLE ID	Total Radium	
Preparation batch 7251-071					
S004124-01		7652-001	B24M49	U	
S004124-02		7652-002	B25044	U	
S004124-03		7652-003	B25050	U	
S004124-04		7652-004	B25077	U	
S004124-05		7652-005	Lab Control Sample	<u>LOW</u>	
S004124-06		7652-006	Method Blank	U	
S004124-07		7652-007	Duplicate (S004124-01)	-	U
Nominal values and limits from method					
			RDLs (pCi/L)	1.00	

METHOD PERFORMANCE

LAB	RAW	SUF-	MDA	ALIQ	PREP	DILU-	YIELD	EFF	COUNT	FWHM	DRIFT	DAYS	ANAL-		
SAMPLE ID	TEST FIX	CLIENT SAMPLE ID	pCi/L	L	FAC	TION	%	%	min	keV	KeV	HELD	PREPARED	YZED	DETECTOR
Preparation batch 7251-071 2σ prep error 12.8 % Reference Lab Notebook No. 7251 pg.071															
S004124-01		B24M49	0.375	0.200			94		100			26	05/18/10	05/22	GAW-111
S004124-02		B25044	0.354	0.200			94		100			26	05/18/10	05/22	GAW-112
S004124-03		B25050	0.344	0.200			94		100			26	05/18/10	05/22	GAW-213
S004124-04		B25077	0.364	0.200			95		100			24	05/18/10	05/20	GAW-216
S004124-05		Lab Control Sample	0.462	0.200			93		100				05/18/10	05/20	GAW-111
S004124-06		Method Blank	0.295	0.200			96		100				05/18/10	05/22	GAW-216
S004124-07		Duplicate (S004124-01)	0.417	0.200			95		100			26	05/18/10	05/22	GAW-109
Nominal values and limits from method			1.00	0.200			40-110		100			180			

PROCEDURES REFERENCE RATOT_GPC
SPP-062 Sample Aliquoting, rev 1
CP-880 Total Radium in Water, rev 2

AVERAGES ± 2 SD MDA 0.373 ± 0.107
FOR 7 SAMPLES YIELD 94 ± 2

Lab id EBRLINE
Protocol CHPRC
Version Ver 1.0
Form DVD-LMS
Version 3.06
Report date 11/19/10

EBERLINE ANALYTICAL / RICHMOND

SAMPLE DELIVERY GROUP H4220

SDG 7652
 Contact N. Joseph Verville

REPORT GUIDE

Client CHPRC
 Contract No. 33677
 Case no SDG H4220

SAMPLE SUMMARY

The Sample and QC Summary Reports show all samples, including QC samples, reported in one Sample Delivery Group (SDG).

The Sample Summary Report fully identifies client samples and gives the corresponding lab sample identification. The QC Summary Report shows at the sample level how the lab organized the samples into batches and generated QC samples. The Preparation Batch and Method Summary Reports show this at the analysis level.

The following notes apply to these reports:

- * LAB SAMPLE ID is the lab's primary identification for a sample.
- * DEPARTMENT SAMPLE ID is an alternate lab id, for example one assigned by a radiochemistry department in a lab.
- * CLIENT SAMPLE ID is the client's primary identification for a sample. It includes any sample preparation done by the client that is necessary to identify the sample.
- * QC BATCH is a lab assigned code that groups samples to be processed and QCed together. These samples should have similar matrices.

QC BATCH is not necessarily the same as SDG, which reflects samples received and reported together.

- * All Lab Control Samples, Method Blanks, Duplicates and Matrix Spikes are shown that QC any of the samples. Due to possible reanalyses, not all results for all these QC samples may be relevant to the SDG. The Lab Control Sample, Method Blank, Duplicate, Matrix Spike and Method Summary Reports detail these relationships.

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-RG
 Version 3.06
 Report date 11/19/10

EBERLINE ANALYTICAL / RICHMOND

SAMPLE DELIVERY GROUP H4220

SDG 7652
 Contact N. Joseph Verville

REPORT GUIDE

Client CHPRC
 Contract No. 33677
 Case no SDG H4220

PREPARATION BATCH SUMMARY

The Preparation Batch Summary Report shows all preparation batches in one Sample Delivery Group (SDG) with information necessary to check the completeness and consistency of the SDG.

The following notes apply to this report:

- * The preparation batches are shown in the same order as the Method Summary Reports are printed.
- * Only analyses of planchets relevant to the SDG are included.
- * Each preparation batch should have at least one Method Blank and LCS in it to validate client sample results.
- * The QUALIFIERS shown are all qualifiers other than U, J, B, L and H that occur on any analysis in the preparation batch. The Method Summary Report has these qualifiers on a per sample basis.

These qualifiers should be reviewed as follows:

- X Some data has been manually entered or modified. Transcription errors are possible.
- P One or more results are 'preliminary'. The data is not ready for final reporting.
- 2 There were two or more results for one analyte on one planchet imported at one time. The results in DVD may not be the same as on the raw data sheets.

Other lab defined qualifiers may occur. In general, these should be addressed in the SDG narrative.

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-RG
 Version 3.06
 Report date 11/19/10

EBERLINE ANALYTICAL / RICHMOND

SAMPLE DELIVERY GROUP H4220

SDG 7652
 Contact N. Joseph Verville

REPORT GUIDE

Client CHPRC
 Contract No. 33677
 Case no SDG H4220

WORK SUMMARY

The Work Summary Report shows all samples, including QC samples, and all relevant analyses in one Sample Delivery Group (SDG). This report is often useful as supporting documentation for an invoice.

The following notes apply to this report:

- * TEST is a code for the method used to measure associated analytes. Results and related information for each analyte are on the Data Sheet Report. In special cases, a test code used in the summary data section is not the same as in associated raw data. In this case, both codes are shown on the Work Summary.
- * SUFFIX is the lab's code to distinguish multiple analyses (recounts, reworks, reanalyses) of a fraction of the sample. The suffix indicates which result is being reported. An empty suffix normally identifies the first attempt to analyze the sample.
- * The LAB SAMPLE ID, TEST and SUFFIX uniquely identify all supporting data for a result. The Method Summary Report for each TEST has method performance data, such as yield, for each lab sample id and suffix and procedures used in the method.
- * PLANCHET is an alternate lab identifier for work done for one test. It, combined with the TEST and SUFFIX, may be the best link to raw data.
- * For QC samples, only analyses that directly QC some regular sample are shown. The Lab Control Sample, Method Blank, Duplicate, Matrix Spike and Method Summary Reports detail these relationships.
- * The SAS (Special Analytical Services) Number is a client or lab assigned code that reflects special processing for samples, such as rapid turn around. Counts of tests done are lists by SAS number since it is likely to affect prices.

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-RG
 Version 3.06
 Report date 11/19/10

REPORT GUIDES

Page 3

SUMMARY DATA SECTION

Page 22

EBERLINE ANALYTICAL / RICHMOND

SAMPLE DELIVERY GROUP H4220

SDG 7652
 Contact N. Joseph Verville

REPORT GUIDE

Client CHPRC
 Contract No. 33677
 Case no SDG H4220

DATA SHEET

The Data Sheet Report shows all results and primary supporting information for one client sample or Method Blank. This report corresponds to both the CLP Inorganics and Organics Data Sheet.

The following notes apply to this report:

- * TEST is a code for the method used to measure an analyte. If the TEST is empty, no data is available; the analyte was not analyzed for.
- * The LAB SAMPLE ID and TEST uniquely identify work within the Summary Data Section of a Data Package. The Work Summary and Method Summary Reports further identify raw data that underlies this work.

The Method Summary Report for each TEST has method performance data, such as yield, for each Lab Sample ID and a list of procedures used in the method.

- * ERRORS can be labeled TOTAL or COUNT. TOTAL implies a preparation (non-counting method) error has been added, as square root of sum of squares, to the counting error denoted by COUNT. The preparation errors, which may vary by preparation batch, are shown on the Method Summary Report.
- * A RESULT can be 'N.R.' (Not Reported). This means the lab did this work but chooses not to report it now, possibly because it was reported at another time.
- * When reporting a Method Blank, a RESULT can be 'N.A.' (Not Applicable). This means there is no reported client sample work in the same preparation batch as the Blank's result. This is likely to occur when the Method Blank is associated with reanalyses of selected work for a few samples in the SDG.

The following qualifiers are defined by the DVD system:

- U The RESULT is less than the MDA (Minimum Detectable Activity).

REPORT GUIDES

Page 4

SUMMARY DATA SECTION

Page 23

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-RG
 Version 3.06
 Report date 11/19/10

EBERLINE ANALYTICAL/RICHMOND

SAMPLE DELIVERY GROUP H4220

SDG 7652
 Contact N. Joseph Verville

GUIDE, cont.

Client CHPRC
 Contract No. 33677
 Case no SDG H4220

DATA SHEET

If the MDA is blank, the ERROR is used as the limit.

J The RESULT is less than the RDL (Required Detection Limit) and no U qualifier is assigned.

B A Method Blank associated with this sample had a result without a U flag and, after correcting for possibly different aliquots, that result is greater than or equal to the MDA for this sample.

Normally, B is not assigned if U is. When method blank subtraction is shown on this report, B flags are assigned based on the unsubtracted values while U's are assigned based on the subtracted ones. Both flags can be assigned in this case.

For each sample result, all Method Blank results in the same preparation batch are compared. The Method Summary Report documents this and other QC relationships.

L Some Lab Control Sample that QC's this sample had a low recovery. The lab can disable assignment of this qualifier.

H Similar to 'L' except the recovery was high.

P The RESULT is 'preliminary'.

X Some data necessary to compute the RESULT, ERROR or MDA was manually entered or modified.

2 There were two or more results available for this analyte. The reported result may not be the same as in the raw data.

Other qualifiers are lab defined. Definitions should be in the SDG narrative.

The following values are underlined to indicate possible problems:

* An MDA is underlined if it is bigger than its RDL.

REPORT GUIDES
 Page 5
 SUMMARY DATA SECTION
 Page 24

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-RG
 Version 3.06
 Report date 11/19/10

EBERLINE ANALYTICAL/RICHMOND

SAMPLE DELIVERY GROUP H4220

SDG 7652
 Contact N. Joseph Verville

GUIDE, cont.

Client CHPRC
 Contract No. 33677
 Case no SDG_H4220

DATA SHEET

- * An ERROR is underlined if the 1.645 sigma counting error is bigger than both the MDA and the RESULT, implying that the MDA may not be a good estimate of the 'real' minimum detectable activity.
- * A negative RESULT is underlined if it is less than the negative of its 2 sigma counting ERROR.
- * When reporting a Method Blank, a RESULT is underlined if greater than its MDA. If the MDA is blank, the 2 sigma counting error is used in the comparison.

REPORT GUIDES

Page 6

SUMMARY DATA SECTION

Page 25

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-RG
 Version 3.06
 Report date 11/19/10

EBERLINE ANALYTICAL/RICHMOND

SAMPLE DELIVERY GROUP H4220

SDG 7652
Contact N. Joseph Verville

REPORT GUIDE

Client CHPRC
Contract No. 33677
Case no SDG H4220

LAB CONTROL SAMPLE

The Lab Control Sample Report shows all results, recoveries and primary supporting information for one Lab Control Sample.

The following notes apply to this report:

- * All fields in common with the Data Sheet Report have similar usage. Refer to its Report Guide for details.
- * An amount ADDED is the lab's value for the actual amount spiked into this sample with its ERROR an estimate of the error of this amount.

An amount added is underlined if its ratio to the corresponding RDL is outside protocol specified limits.

- * REC (Recovery) is RESULT divided by ADDED expressed as a percent.
- * The first, computed limits for the recovery reflect:
 1. The error of RESULT, including that introduced by rounding the result prior to printing.

If the limits are labeled (TOTAL), they include preparation error in the result. If labeled (COUNT), they do not.
 2. The error of ADDED.
 3. A lab specified, per analyte bias. The bias changes the center of the computed limits.
- * The second limits are protocol defined upper and lower QC limits for the recovery.
- * The recovery is underlined if it is outside either of these ranges.

Lab id EBRLE
Protocol CHPRC
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 11/19/10

EBERLINE ANALYTICAL / RICHMOND

SAMPLE DELIVERY GROUP H4220

SDG 7652
 Contact N. Joseph Verville

REPORT GUIDE

Client CHPRC
 Contract No. 33677
 Case no SDG H4220

DUPLICATE

The Duplicate Report shows all results, differences and primary supporting information for one Duplicate and associated Original sample.

The following notes apply to this report:

- * All fields in common with the Data Sheet Report have similar usage. This applies both to the Duplicate and Original sample data. Refer to the Data Sheet Report Guide for details.

If the Duplicate has data for a TEST and the lab did not do this test to the Original, the Original's RESULTS are underlined.

- * The RPD (Relative Percent Difference) is the absolute value of the difference of the RESULTS divided by their average expressed as a percent.

If both RESULTS are less than their MDAs, no RPD is computed and a '-' is printed.

For an analyte, if the lab did work for both samples but has data for only one, the MDA from the sample with data is used as the other's result in the RPD.

- * The first, computed limit is the sum, as square root of sum of squares, of the errors of the results divided by the average result as a percent, hence the relative error of the difference rather than the error of the relative difference. The errors include those introduced by rounding the RESULTS prior to printing.

If this limit is labeled TOT, it includes the preparation error in the RESULTS. If labeled CNT, it does not.

This value reported for this limit is at most 999.

- * The second limit for the RPD is the larger of:
 1. A fixed percentage specified in the protocol.

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-RG
 Version 3.06
 Report date 11/19/10

EBERLINE ANALYTICAL / RICHMOND

SAMPLE DELIVERY GROUP H4220

SDG 7652
 Contact N. Joseph Verville

GUIDE, cont.

Client CHPRC
 Contract No. 33677
 Case no SDG H4220

DUPLICATE

2. A protocol factor (typically 2) times the average MDA as a percent of the average result. This limit applies when the results are close to the MDAs.

- * The RPD is underlined if it is greater than either limit.
- * If specified by the lab, the second limit column is replaced by the Difference Error Ratio (DER), which is the absolute value of the difference of the results divided by the quadratic sum of their one sigma errors, the same errors as used in the first limit.

Except for differences due to rounding, the DER is the same as the RPD divided by the first RPD limit with the limit scaled to 1 sigma.

- * The DER is underlined if it is greater than the sigma factor, typically 2 or 3, shown in the header for the first RPD limit.

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-RG
 Version 3.06
 Report date 11/19/10

SDG 7652
 Contact N. Joseph Verville

REPORT GUIDE

Client CHPRC
 Contract No. 33677
 Case no SDG H4220

MATRIX SPIKE

The Matrix Spike Report shows all results, recoveries and primary supporting information for one Matrix Spike and associated Original sample.

The following notes apply to this report:

- * All fields in common with the Data Sheet Report have similar usage. This applies both to the Spiked and Original sample data. Refer to the Data Sheet Report Guide for details.

If the Spike has data for a TEST and the lab did not do this test to the Original, the Original's RESULTS are underlined.

- * An amount ADDED is the lab's value for the actual amount spiked into the Spike sample with its ERROR an estimate of the error of this amount.

An amount is underlined if its ratio to the corresponding RDL is outside protocol specified limits.

- * REC (Recovery) is the Spike RESULT minus the Original RESULT divided by ADDED expressed as a percent.

- * The first, computed limits for the recovery reflect:

1. The errors of the two RESULTS, including those introduced by rounding them prior to printing.

If the limits are labeled (TOTAL), they include preparation error in the result. If labeled (COUNT), they do not.

2. The error of ADDED.
3. A lab specified, per analyte bias. The bias changes the center of the computed limits.

- * The second limits are protocol defined upper and lower QC limits

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-RG
 Version 3.06
 Report date 11/19/10

EBERLINE ANALYTICAL/RICHMOND

SAMPLE DELIVERY GROUP H4220

SDG 7652
Contact N. Joseph Verville

GUIDE, cont.

Client CHPRC
Contract No. 33677
Case no SDG_H4220

MATRIX SPIKE

for the recovery.

These limits are left blank if the Original RESULT is more than a protocol defined factor (typically 4) times ADDED. This is a way of accounting for that when the spike is small compared to the amount in the original sample, the recovery is unreliable.

* The recovery is underlined (out of spec) if it is outside either of these ranges.

REPORT GUIDES

Page 11

SUMMARY DATA SECTION

Page 30

Lab id EBRLNE
Protocol CHPRC
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 11/19/10

EBERLINE ANALYTICAL / RICHMOND

SAMPLE DELIVERY GROUP H4220

SDG 7652
 Contact N. Joseph Verville

REPORT GUIDE

Client CHPRC
 Contract No. 33677
 Case no SDG H4220

METHOD SUMMARY

The Method Summary Report has two tables. One shows up to five results measured using one method. The other has performance data for the method. There is one report for each TEST, as used on the Data Sheet Report.

The following notes apply to this report:

- * Each table is subdivided into sections, one for each preparation batch. A preparation batch is a group of aliquots prepared at roughly the same time in one work area of the lab using the same method.

There should be Lab Control Sample and Method Blank results in each preparation batch since this close correspondence makes the QC meaningful. Depending on lab policy, Duplicates need not occur in each batch since they QC sample dependencies such as matrix effects.

- * The RAW TEST column shows the test code used in the raw data to identify a particular analysis if it is different than the test code in the header of the report. This occurs in special cases due to method specific details about how the lab labels work.

The Lab Sample or Planchet ID combined with the (Raw) Test Code and Suffix uniquely identify the raw data for each analysis.

- * If a result is less than both its MDA and RDL, it is replaced by just 'U' on this report. If it is greater than or equal to the RDL but less than the MDA, the result is shown with a 'U' flag.

The J and X flags are as on the data sheet.

- * Non-U results for Method Blanks are underlined to indicate possible contamination of other samples in the preparation batch. The Method Blank Report has supporting data.
- * Lab Control Sample and Matrix Spike results are shown as: ok, No data, LOW or HIGH, with the last two underlined. 'No data'

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-RG
 Version 3.06
 Report date 11/19/10

EBERLINE ANALYTICAL / RICHMOND

SAMPLE DELIVERY GROUP H4220

SDG 7652
 Contact N. Joseph Verville

GUIDE, cont.

Client CHPRC
 Contract No. 33677
 Case no SDG H4220

METHOD SUMMARY

means no amount ADDED was specified. 'LOW' and 'HIGH' correspond to when the recovery is underlined on the Lab Control Sample or Matrix Spike Report. See these reports for supporting data.

- * Duplicate sample results are shown as: ok, No data, or OUT, with the last two underlined. 'No data' means there was no original sample data found for this duplicate. 'OUT' corresponds to when the RPD is underlined on the Duplicate Report. See this report for supporting data.
- * If the MDA column is labeled 'MAX MDA', there was more than one result measured by the reported method and the MDA shown is the largest MDA. If not all these results have the same RDL, the MAX MDA reflects only those results with RDL equal to the smallest one.

MDAs are underlined if greater than the printed RDL.

- * Aliquots are underlined if less than the nominal value specified for the method.
- * Preparation factors are underlined if greater than the nominal value specified for the method.
- * Dilution factors are underlined if greater than the nominal value specified for the method.
- * Residues are underlined if outside the range specified for the method. Residues are not printed if yields are.
- * Yields, which may be gravimetric, radiometric or some type of recovery depending on the method, are underlined if outside the range specified for the method.
- * Efficiencies are underlined if outside the range specified for the method. Efficiencies are detector and geometry dependent so this test is only approximate.

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-RG
 Version 3.06
 Report date 11/19/10

REPORT GUIDES
 Page 13
 SUMMARY DATA SECTION
 Page 32

EBERLINE ANALYTICAL / RICHMOND

SAMPLE DELIVERY GROUP H4220

SDG 7652
 Contact N. Joseph Verville

GUIDE, cont.

Client CHPRC
 Contract No. 33677
 Case no SDG H4220

METHOD SUMMARY

- * Count times are underlined if less than the nominal value specified for the method.
- * Resolutions (as FWHM; Full Width at Half Max) are underlined if greater than the method specified limit.
- * Tracer drifts are underlined if their absolute values are greater than the method specified limit. Tracer drifts are not printed if percent moistures are.
- * Days Held are underlined if greater than the holding time specified in the protocol.
- * Analysis dates are underlined if before their planchet's preparation date or, if a limit is specified, too far after it.

For some methods, ratios as percentages and error estimates for them are computed for pairs of results. A ratio column header like '1+3' means the ratio of the first result column and the third result column.

Ratios are not computed for Lab Control Sample, Method Blank or Matrix Spike results since their matrices are not necessarily similar to client samples'.

The error estimate for a ratio of results from one planchet reflects only counting errors since other errors should be correlated. For a ratio involving different planchets, if QC limits are computed based on total errors, the error for the ratio allows for the preparation errors for the planchets.

The ratio is underlined (out of spec) if the absolute value of its difference from the nominal value is greater than its error estimate. If no nominal value is specified, this test is not done.

For Gross Alpha or Gross Beta results, there may be a column showing the sum of other Alpha or Beta emitters. This sum includes all relevant

Lab id EBRLNE
 Protocol CHPRC
 Version Ver 1.0
 Form DVD-RG
 Version 3.06
 Report date 11/19/10

EBERLINE ANALYTICAL/RICHMOND

SAMPLE DELIVERY GROUP H4220

SDG 7652
Contact N. Joseph Verville

GUIDE, cont.

Client CHPRC
Contract No. 33677
Case no SDG H4220

METHOD SUMMARY

results in the DVD database, whether reported or not. Results in the sum are weighted by a particles/decay value specified by the lab for each relevant analyte. Results less than their MDA are not included. No sums are computed for Lab Control, Method Blank or Matrix Spike samples since their various planchets may not be physically related.

If a ratio of total isotopic to Gross Alpha or Beta is shown, the error for the ratio reflects both the error in the Gross result and the sum, as square root of sum of squares, of the errors in the isotopic results.

For total elemental uranium or thorium results, there may be a column showing the total weight computed from associated isotopic results. Ignoring results less than their MDAs, this is a weighted sum of the isotopic results. The weights depend on the molecular weight and half-life of each isotope so as to convert activities (decays) to weight (atoms).

If a ratio of total computed to measured elemental uranium or thorium is shown, the error for the ratio reflects the errors in all the measurements.

REPORT GUIDES
Page 15
SUMMARY DATA SECTION
Page 34

Lab id EBRLNE
Protocol CHPRC
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 11/19/10



RE: Please hold total radium and GAMMALL analyses for all SAF X10-107 samples

Widrig, Dana L to: 'jverville@eberlineservices.com'

10/06/2010 03:04 PM

Cc: "Fitzgerald, Scot L", "Ayres, Doris E", "Caraway, Mary Jo",
"Crisp, Bryan D", "Neely, Michael", "Narquis, Clifford T"

History: This message has been replied to.

Joe,

Per our conversation, we erroneously requested total Radium (RATOT_AEAGEA) for the attached SAFs and associated samples which did not include the desired Ra-228 analysis. Scot Fitzgerald will work with you on getting these samples analyzed for Ra-228 by the proper service list RAISO_SEP_GPC at Eberline. Please ensure all future samples associated with these SAFs are also logged and analyzed accordingly. All COCs should be corrected accordingly (signed and dated) and sent or resent to us for our tracking purposes.

As soon as you obtain the requested MDA information we discussed (preferably 1 pCi/L) and the turnaround times you expect, please let Scot know at (509) 373-7495. Please also let us know in writing so we will be able to track this in our systems.

Of course, let me know if there are questions –

Thanks,

Dana L. Widrig
Project Coordinator
CH2MHill Plateau Remediation Company
SMR/SGRP
Ph: (509) 376-2858
Cell: (509) 392-2297

From: Douglas, James G (Jim)
Sent: Wednesday, October 06, 2010 1:55 PM
To: 'jverville@eberlineservices.com'
Cc: Waters-husted, Karen S; Trent, Stephen J; Narbutovskih, Susan M; Widrig, Dana L; Fitzgerald, Scot L
Subject: RE: Please hold total radium and GAMMALL analyses for all SAF X10-107 samples

Joe:

Thank you for your continued follow-up on this issue.

Please direct future correspondence to our project coordinator Dana Widrig; I'm turning over my participation on this issue to her. Scot Fitzgerald will assist Dana

RED = results received from Eberline

EBERLINE SAMPLES
Total Radium (RATOT_GPC)

APRIL	MAY		JUNE	JULY	AUGUST		SEPTEMBER	
X10-039	X10-053	X10-056	X10-073	X10-081	X10-098	X10-099	X10-107	X10-108
SAMP_NUM	SAMP_NUM							
B24M49	B25923	B259D5	B25NF4	B26535	B26ML6	B26M16	B275D4	B27560
B24M50	B25929		B25NF5	B26541	B26ML7	B26M21	B275D5	
B24M51	B25935			B26547	B26MM7	B26M22	B275F5	
B24M52	B25941			B26548	B26MM8		B275H1	
B24M53	B25947			B26558	B26MN8		B275H7	
B24M54	B25953			B26564	B26MP4		B275J3	
B24M55	B25959			B26570	B26MR4		B275J9	
B24M58	B25965			B26576	B26MR9		B275K5	
B25023	B25971			B265K3	B26MT5		B275L1	
B25025	B25977			B265K9	B26MV1		B275L7	
B25027	B25983			B265L5	B26MV7		B275M3	
B25037	B25989			B265M1	B26MW3		B275M9	
B25038	B25995			B265M7	B26MW9		B275N5	
B25040	B25981			B265N3	B26MX5		B275P1	
B25042	B25987			B265N9	B26MY1		B275P7	
B25044	B259C3			B265P5	B26MY7		B275R3	
B25046	B259C9			B265R1	B26N10		B275R9	
B25047				B265R7	B26N16		B275T5	
B25048				B265T3	B26N22		B27620	
B25050				B265T9				
B25052				B265V5				
B25053				B265W1				
B25054				B265W7				
B25055				B265X3				
B25056				B265X9				
B25062				B265Y5				
B25064				B265Y6				
B25065				B26606				
B25066				B26607				
B25067				B26617				
B25068				B26623				
B25069				B26629				
B25070				B26635				
B25071				B26641				
B25072				B26647				
B25073				B26858				
B25074								
B25075								
B25076								
B25077								

CHPRC	CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST	C.O.C. #	X10-039-314
		Page	1 of 1

Collector JP Herrick CHPRC	Contact/Requester Karen Waters-Husted H4220 (7652)	Telephone No. 376-4650
SAF No. X10-039	Sampling Origin Hanford Site	Purchase Order/Charge Code 300118ES20
Project Title HJ2 H16DJRA APRIL 2010	Logbook No: HNF-N-506 31,47	Ice Chest No. 605-135
Shipped To (Lab) Eberline Services	Method of Shipment Commercial Carrier	Bill of Lading/Air Bill No. 7934 85885718
Protocol CERCLA	Priority: 45 Days	Offsite Property No.

POSSIBLE SAMPLE HAZARDS/REMARKS ** ** Contains Radioactive Material at concentrations that are not regulated for transportation per 49 CFR but are not releasable per DOE Order 5400.5 (1990/1993)	SPECIAL INSTRUCTIONS Do not combine samples from this SAF with other data sets. Need SDG to be stand alone. 100 Area Generator Knowledge Information Form applies. The CACN for all analytical work at WSCF is 300118. TICs are required 8260_VOA method. Do not combine samples	Hold Time Total Activity Exemption: Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>
--	--	---

Sample No.	Filter	*	Date	Time	No/Type Container	Sample Analysis	Holding Time	Preservative
B25044	N	W	4.26.10	1005	1x20-mL P	Activity Scan	6 Months	None
B25044	N	W	↓	↓	1x1000-mL G/P	Total Radium	6 Months	HNO3 to pH <2

Relinquished By JP Herrick CHPRC	Print <i>[Signature]</i>	Sign <i>[Signature]</i>	Date/Time APR 26 2010 15:00	Received By SSU #1	Print <i>[Signature]</i>	Sign <i>[Signature]</i>	Date/Time APR 26 2010 15:00	Matrix * S - Soil DS - Drum Solid SF - Sediment DI - Drum Liquid SO - Solid T - Tissue SL - Sludge WI - Wine W - Water L - Liquid O - Oil V - Vegetation A - Air X - Other
Relinquished By SSU-1	Date/Time APR 27 2010 0800	Received By L.D. Wall CHPRC	Date/Time APR 27 2010 0800					
Relinquished By L.D. Wall CHPRC	Date/Time APR 27 2010 1400	Received By FEDEX	Date/Time					
Relinquished By FEDEX	Date/Time	Received By <i>[Signature]</i>	Date/Time KELCENSON 4/28/10 10:00					
FINAL SAMPLE DISPOSITION	Disposal Method (e.g., Return to customer, per lab procedure, used in process)		Disposed By		Date/Time			

CHPRC

CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST

C.O.C. #

X10-039-342

Page 1 of 1

Collector AR McIntyre	Contact/Requester Karen Waters-Husted H4220 (TUS2)	Telephone No. 376-4650
SAF No. X10-039	Sampling Origin Hanford Site	Purchase Order/Charge Code 300118ES20
Project Title IIJ2 IIJ6DURA APRIL 2010	Logbook No: HNF-N-506 28/89	Ice Chest No. 605-135
Shipped To (Lab) Eberline Services	Method of Shipment Commercial Carrier	Bill of Lading/Air Bill No. 7934 8588 5718
Protocol CERCLA	Priority: 45 Days	Offsite Property No.

<p>POSSIBLE SAMPLE HAZARDS/REMARKS</p> <p>** ** Contains Radioactive Material at concentrations that are not regulated for transportation per 49 CFR but are not releasable per DOE Order 5400.5 (1990/1993)</p>	<p>SPECIAL INSTRUCTIONS Hold Time</p> <p>Do not combine samples from this SAF with other data sets. Need SDG to be stand alone. 100 Area Generator Knowledge Information Form applies. The CACN for all analytical work at WSCF is 300118. TICs are required \$260_VOA method. Do not combine samples</p>	<p>Total Activity Exemption: Yes <input checked="" type="checkbox"/> No <input type="checkbox"/></p>
---	--	--

Sample No.	Filter	*	Date	Time	No/Type Container	Sample Analysis	Holding Time	Preservative
B25077	N	W	4/26/10	1316	1x20-mL P	Activity Scan	6 Months	None
B25077	N	W	↓	↓	1x1000-mL G/P	Total Radium	6 Months	HNO3 to pH <2
REM 4/26/10								

Relinquished By AR McIntyre	Print <i>[Signature]</i>	Sign	Date/Time APR 26 2010 1455	Received By SSU # 1	Print <i>[Signature]</i>	Sign	Date/Time APR 26 2010 1955	Matrix *
Relinquished By SSU-1	Print <i>[Signature]</i>	Sign	Date/Time APR 27 2010 0800	Received By L.D. Wall	Print <i>[Signature]</i>	Sign	Date/Time APR 27 2010 0800	<ul style="list-style-type: none"> S = Soil SE = Sediment SO = Solid SL = Sludge W = Water O = Oil A = Air DS = Drum Solid DM = Drum Liquid T = Tissue WI = Wine L = Liquid V = Vegetation X = Other
Relinquished By L.D. Wall	Print <i>[Signature]</i>	Sign	Date/Time APR 27 2010 1400	Received By FEDEX	Print <i>[Signature]</i>	Sign	Date/Time	
Relinquished By FEDEX	Print <i>[Signature]</i>	Sign	Date/Time	Received By Alex Keelson	Print <i>[Signature]</i>	Sign	Date/Time 4/28/10 10:00	
FINAL SAMPLE DISPOSITION	Disposal Method (e.g., Return to customer, per lab procedure, used in process)						Disposed By	Date/Time



RICHMOND, CA LABORATORY

SAMPLE RECEIPT CHECKLIST

Client: CHPRC City RICHLAND State WA

Date/Time received 4/28/10 10:00 CoC No. X10-089-290, 314, 320, 312

Container I.D. No. GWS 135 Requested TAT (Days) 45 P.O. Received Yes [] No []

INSPECTION

1. Custody seals on shipping container intact? Yes [] No [] N/A []
2. Custody seals on shipping container dated & signed? Yes [] No [] N/A []
3. Custody seals on sample containers intact? Yes [] No [] N/A []
4. Custody seals on sample containers dated & signed? Yes [] No [] N/A []
5. Packing material is: Wet [] Dry [] N/A []
6. Number of samples in shipping container: 4 Sample Matrix WATER
7. Number of containers per sample: 2 (Or see CoC _____)
8. Samples are in correct container Yes [] No []
9. Paperwork agrees with samples? Yes [] No []
10. Samples have: Tape [] Hazard labels [] Rad labels [] Appropriate sample labels []
11. Samples are: In good condition [] Leaking [] Broken Container [] Missing []
12. Samples are: Preserved [] Not preserved [] pH 2 Preservative _____
13. Describe any anomalies:

14. Was P.M. notified of any anomalies? Yes [] No [] Date _____

15. Inspected by JK Date: 4/28/10 Time: 12:30

Customer Sample No.	Beta/Gamma cpm	Ion Chamber mR/hr	Wipe	Customer Sample No.	Beta/Gamma cpm	Ion Chamber mR/hr	wipe
<u>ALL Sample</u>	<u>< 60</u>						

Ion Chamber Ser. No. _____
Alpha Meter Ser. No. _____
Beta/Gamma Meter Ser. No. 100482

Calibration date _____
Calibration date _____
Calibration date 5 Aug 2009