FS-C98028-Field

0052241

DATA DELIVERABLE COVER SHEET

Data Deliverable Cover Sheet



Date: Wednesday, December 30, 1998

Deliverable type (check one): ____ **Preliminary Test Results**

✓ **Final Data Deliverable**

Sample authorization form number: C98-028

Attachments (check all that apply or n/a): ____ **Narrative Summary**

✓ **Test Results**

____ **Chains of Custody**

Total number of pages (including cover sheet): 3 **pages**

Comments:

(Note: If Preliminary results were provided verbally, document information provided below including instrument used and results as Communicated to the project.)

M. A. Baechler
Analyst Signature

12/30/98
Date

Distribution: Sample Management – Doris Ayres H9- 03
Project (specify) None. Distributed electronically

Requires distribution to listed project personnel by Sample Management.

PROJECT

Continued From Page

Hexavalent Chromium Daily Activity Log
Project: 100-HR-3 IAM / SAF C98-028
Date: 3/23/98

1250	Arrived on-site at 200W, HO 68N-1985 mobile laboratory. Set up to analyze samples for hexavalent chromium using a HACH DR 2000 spectrophotometer, S/N 920800020780, Method 950, low range hexavalent chromium.
1255	Prepared 0.10 mg/L standard by adding 600 uL of 5 mg/L HACH hexavalent chromium standard, lot # N5034, no expiration, to 29.4 mL of deionized water.
1425	Analyzed samples.

Sample #	Well #	Concentration (mg/L)	Date Sampled	Time Sampled
Blank	-	<0.005	-	-
0.10 mg/L std.	-	0.098	-	-
B0N4B4	199-D8-68	0.236	3/23/98	11:43
B0N4B5	199-D8-69	0.089	3/23/98	09:05
B0N4B6	199-D8-69	0.090	3/23/98	09:05
B0N4B7	199-D8-70	0.212	3/23/98	09:20
B0N5D2	199-D8-70	0.216	3/23/98	09:30
B0N5D3	199-D8-70	0.242	3/23/98	09:40

1440 Packed up equipment and left site.

M.A. Baechler / M.G. Baechler

Not Used

Continued on Page N/A

M.G. Baechler
Signed

3/23/98
Date

Read and Understood By

Signed

Date

Hexavalent Chromium Daily Activity Log
Project: 100-HR-3 IAM / SAF C98-028
Date: 3/02/98

1215	Arrived on-site at 200W, HO 68N-1985 mobile laboratory. Set up to analyze samples for hexavalent chromium using a HACH DR 2000 spectrophotometer, S/N 920800020780, Method 950, low range hexavalent chromium.
1220	Prepared 0.10 mg/L standard by adding 600 uL of 5 mg/L HACH hexavalent chromium standard, lot # N5034, no expiration, to 29.4 mL of deionized water.
1230	Analyzed samples.

Sample #	Well #	Concentration (mg/L)	Date Sampled	Time Sampled
Blank	-	<0.005	-	-
0.10 mg/L std.	-	0.091	-	-
B0N4B8	199-H4-4	0.123	3/02/98	1005
B0N4B9	199-H4-5	0.150	3/02/98	1121
B0N4C0	199-H4-63	0.072	3/02/98	0911
B0N4C3	199-H4-64	0.045	3/02/98	1035

1300 Packed up equipment and left site.

M.A. Baechler / M.G. Baechler

Not Used

Continued on Page N/A

M.G. Baechler — Signed
3/2/98 — Date
Read and Understood By
Signed
Date

[redacted]