


DATA DELIVERABLE COVER SHEET



Data Deliverable Cover Sheet

Date: Wednesday, December 30, 1998

Deliverable type (check one): Preliminary Test Results
 Final Data Deliverable

Sample authorization form number: C98-055

Attachments (check all that apply or n/a): Narrative Summary
 Test Results
 Chains of Custody

Total number of pages (including cover sheet): 3 pages

Comments:
(Note: If Preliminary results were provided verbally, document information provided below including instrument used and results as Communicated to the project.)

M. G. Baehler
Analyst Signature

12/30/98
Date

Distribution: Sample Management – Doris Ayres H9- 03
Project (specify) None. Distributed electronically

Requires distribution to listed project personnel by Sample Management.

100-KR-4 IAM

LABORATORY No. 22-1240-
N/A

Hexavalent Chromium Daily Activity Log

Project: 100-KR-4 IAM / SAF C98-055

Date: 09/28/98

1410	Arrived on-site at 200W, HO 68N-1985 mobile laboratory. Set up to analyze samples for hexavalent chromium using a HACH DR 2000 spectrophotometer, S/N 980600009425, Method 950, low range hexavalent chromium.
1415	Prepared 0.10 mg/L standard by adding 300 uL of 10 mg/L HACH hexavalent chromium standard, lot # N5272, no expiration, to 29.7 mL of deionized water.
1427	Analyzed samples.

Sample #	Well #	Concentration (mg/L)	Date Sampled	Time Sampled
Blank	-	<0.005	-	-
0.10 mg/L std.	-	0.099	-	-
BOPR05	199-K-18	0.046	09/28/98	10:41
BOPR07	199-K-20	0.076	09/28/98	11:01
BOPR01	199-K-112A	0.095	09/28/98	09:30
BOPR02	199-K-114A	0.128	09/28/98	09:59
BOPR03	199-K-114A	0.128	09/28/98	09:59

1500 Packed up equipment and left site.

M.A. Baechler / M.A. Baechler

Prof

Used

Continued on Page *N/A*

M.A. Baechler *9/29/98*

Read and understood by:

Date

100-HR-3 / KR-4 IAM

EL-1340-1
N/A

Hexavalent Chromium Daily Activity Log

Project: 100-HR-3 IAM / SAF C98-053
100-KR-4 IAM / SAF C98-055
Date: 09/25/98

1337	Arrived on-site at 200W, HO 68N-1985 mobile laboratory. Set up to analyze samples for hexavalent chromium using a HACH DR 2000 spectrophotometer, S/N 980600009425, Method 950, low range hexavalent chromium.
1345	Prepared 0.10 mg/L standard by adding 300 uL of 10 mg/L HACH hexavalent chromium standard, lot # N5272, no expiration, to 29.7 mL of deionized water.
1402	Analyzed samples.

Sample #	Well #	Concentration (mg/L)	Date Sampled	Time Sampled
Blank	-	<0.005	-	-
0.10 mg/L std.	-	0.098	-	-
BOPPP8	199-D8-70	0.238	09/25/98	11:49
BOPPR5	199-D8-70	0.236	09/25/98	11:53
BOPPR6	199-D8-70	0.237	09/25/98	11:58
BOPR04	199-K-117A	0.112	09/25/98	12:40
BOPR08	199-K-117A	0.109	09/25/98	12:43
BOPR09	199-K-117A	0.111	09/25/98	12:49
BOPR10	199-K-117A	0.120	09/25/98	12:53

1430 Packed up equipment and left site.

M.A. Baechler / M.A. Baechler

~~Not Used~~

Continued on Page *1*

M.A. Baechler *9/29/98*

Read and Understood By