


0052268

DATA DELIVERABLE COVER SHEET



Data Deliverable Cover Sheet

Date: Wednesday, December 30, 1998

Deliverable type (check one): Preliminary Test Results
 Final Data Deliverable

Sample authorization form number: C 99-001

Attachments (check all that apply or n/a): Narrative Summary
 Test Results
 Chains of Custody

Total number of pages (including cover sheet): 3 pages

Comments:
(Note: If Preliminary results were provided verbally, document information provided below including instrument used and results as Communicated to the project.)

M.C. Baehler
Analyst Signature

12/30/98
Date

Distribution: Sample Management – Doris Ayres H9- 03
Project (specify) None. Distributed electronically

Requires distribution to listed project personnel by Sample Management.

100-KR-3/KR-4 IAM

EL-1390-1
4/17

Hexavalent Chromium Daily Activity Log
Project: 100-KR-4 IAM / SAF C99-006
100-HR-3 IAM / SAF C99-001
Date: 10/22/98

0815 1330	Arrived on-site at 200W, HO 68N-1985 mobile laboratory. Set up to analyze samples for hexavalent chromium using a HACH DR 2000 spectrophotometer, S/N 980600009425, Method 950, low range hexavalent chromium.
0820	Prepared 0.10 mg/L standard by adding 300 uL of 10 mg/L HACH hexavalent chromium standard, lot # N5272, no expiration, to 29.7 mL of deionized water.
0848 1349	Analyzed samples (B0R4H9, B0R4J1). Analyzed samples.

Sample #	Well #	Concentration (mg/L)	Date Sampled	Time Sampled
Blank	-	<0.005	-	-
0.10 mg/L std.	-	0.105	-	-
B0R4H9	199-K-18	0.046	10/21/98	09:18
B0R4J1	199-K-19	0.085	10/21/98	10:09
B0R4H1	199-K-112A	0.098	10/22/98	10:44
B0PYT9	199-D8-68	0.405	10/22/98	08:37
B0PYV0	199-D8-69	0.143	10/22/98	09:14
B0PYV1	199-D8-69	0.144	10/22/98	09:14
B0PYV2	199-D8-70	0.238	10/22/98	09:39
B0R0B9	199-D8-70	0.236	10/22/98	09:48
B0R0C0	199-D8-70	0.219	10/22/98	09:53

0915 Packed up equipment and left site.
 1405 Packed up equipment and left site.

M. A. Baechler (M. A. Baechler 10/26/98)

Not Used

Continued on Page N

M. A. Baechler 10/26/98

read and understood

100-HR-3 IAM

EL-1340-
N/A

Hexavalent Chromium Daily Activity Log

Project: 100-HR-3 IAM / SAF C99-001

Date: 10/14/98

1300	Arrived on-site at 200W, HO 68N-1985 mobile laboratory. Set up to analyze samples for hexavalent chromium using a HACH DR 2000 spectrophotometer, S/N 980600009425, Method 950, low range hexavalent chromium.
1305	Prepared 0.10 mg/L standard by adding 300 uL of 10 mg/L HACH hexavalent chromium standard, lot # N5272, no expiration, to 29.7 mL of deionized water.
1310	Analyzed samples.

Sample #	Well #	Concentration (mg/L)	Date Sampled	Time Sampled
Blank	-	<0.005 MAD 10/15/98	-	-
0.10 mg/L std.	-	0.0998	-	-
BOPYV3	199-H4-4	0.236	10/14/98	10:03
BOPYV4	199-H4-5	0.132	10/14/98	10:42
BOPYV5	199-H4-63	0.111	10/14/98	12:21
BOPYT6	199-H4-64	0.087	10/14/98	08:07

1340 Packed up equipment and left site.

M. A. Baechler / M. G. Baechler 10/15/98

~~Not Used~~

Continued on Page N/A

Baechler 10/15/98